

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space		냉장고	
	Manufacturer	Giant Star	Storage Period		2022.1	
	Model Name	CBR1000SS	IP	Temp.	IP	Temp.
	Serial Number	GS17090102-10				
Calibration Subject	Description	Resistance Temperature Detector				
	Manufacturer	대진센서제작소				
	Model Name	PT100				
	Serial Number	SS21072202				
	Date of Calibration	2021.7.23				
	Certificate No.	SS-210722-0036	Temperature Range		2~8℃	

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2022.01.01						
2022.01.02						
2022.01.03	10:30			4.3	conf	
2022.01.04	10:00			4.4	conf	
2022.01.05	P:10			4.4	conf	
2022.01.06	P:30			3.9	conf	
2022.01.07	P:50			4.2	conf	
2022.01.08						
2022.01.09						
2022.01.10	P:30			4.2	김승하	
2022.01.11	P:30			3.9	conf	
2022.01.12	P:30			4.4	conf	
2022.01.13	P:30			3.7	conf	
2022.01.14	P:30			3.8	conf	
2022.01.15						
2022.01.16						
2022.01.17	P:00			3.7	conf	
2022.01.18	P:30			4.3	conf	
2022.01.19	P:30			4.0	conf	
2022.01.20	P:30			4.4	conf	
2022.01.21	P:30			4.6	conf	
2022.01.22						
2022.01.23						
2022.01.24	P:30			4.3	conf	
2022.01.25	P:30			4.6	conf	
2022.01.26	P:30			3.8	conf	
2022.01.27	P:30			3.7	conf	
2022.01.28	P:30			4.3	conf	
2022.01.29						
2022.01.30						
2022.01.31						설날연휴

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 3.7 DV: -0.3

3

09:58 SP: 4.0 PV: 3.7 DV: -0.
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-01-02

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 3.8 DV: -0.2

2

07:58 SP: 4.0 PV: 3.8 DV: -0.
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-01-01

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-01-04

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 3.8 DV: -0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-01-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.0 DV: 0.0
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 3.8 DV: -0.2
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 4.0 DV: 0.0
 2022-01-06
 23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.2 DV: 0.2

14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.1 DV: 0.1
 2022-01-05

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.0 DV: 0.0

0
 15:58 SP: 4.0 PV: 4.0 DV: 0.
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.0 DV: 0.0
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 3.9 DV: -0.1

0
 01:58 SP: 4.0 PV: 4.0 DV: 0.
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-08

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 4.0 DV: 0.0
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.9 DV: -0.1

2

02:58 SP: 4.0 PV: 3.8 DV: -0.
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.9 DV: -0.1
 2022-01-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.9 DV: -0.1

1

15:58 SP: 4.0 PV: 4.1 DV: 0.
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.3 DV: 0.3
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 3.7 DV: -0.3

2

03:58 SP: 4.0 PV: 4.2 DV: 0.
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-01-10

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-01-09

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 3.9 DV: -0.1

15:58 SP: 4.0 PV: 4.0 DV: 0.0

14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 3.8 DV: -0.2
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.0 DV: 0.0
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.1 DV: 0.1

2

02:58 SP: 4.0 PV: 3.8 DV: -0.
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-01-12

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3

2

18:58 SP: 4.0 PV: 4.2 DV: 0.

1

17:58 SP: 4.0 PV: 3.9 DV: -0.
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 3.9 DV: -0.1

11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 3.7 DV: -0.3

08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-01-14

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.8 DV: -0.2

1

01:58 SP: 4.0 PV: 4.1 DV: 0.
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-01-13

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1

1

12:58 SP: 4.0 PV: 3.9 DV: -0.
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.0 DV: 0.0
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.0 DV: 0.0

2

01:58 SP: 4.0 PV: 3.8 DV: -0.

1

00:58 SP: 4.0 PV: 3.9 DV: -0.

2022-01-16

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.0 DV: 0.0
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-01-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 3.8 DV: -0.2

3

06:58 SP: 4.0 PV: 3.7 DV: -0.1
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-01-18

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.0 DV: 0.0
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 3.8 DV: -0.2

3

08:58 SP: 4.0 PV: 3.7 DV: -0.1
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-01-17

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.2 DV: 0.2

1

17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.1 DV: 0.1

1

09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.2 DV: 0.2

0

02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-20

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1

2

10:58 SP: 4.0 PV: 4.2 DV: 0.1
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-01-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.1 DV: 0.1
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 3
 15:58 SP: 4.0 PV: 3.7 DV: -0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 2
 11:58 SP: 4.0 PV: 4.2 DV: 0.
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 2
 03:58 SP: 4.0 PV: 4.2 DV: 0.
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 3.8 DV: -0.2
 2022-01-22
 23:58 SP: 4.0 PV: 4.2 DV: 0.2
 0
 22:58 SP: 4.0 PV: 4.0 DV: 0.
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 2
 05:58 SP: 4.0 PV: 3.8 DV: -0.
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 4.0 DV: 0.0
 2022-01-21

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 3.9 DV: -0.1
 2022-01-24
 23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.0 DV: 0.0
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.9 DV: -0.1
 2022-01-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.8 DV: 0.8
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-01-26

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-25

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 4.5 DV: 0.5
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.3 DV: 0.3
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.4 DV: 0.4

2

02:58 SP: 4.0 PV: 4.2 DV: 0.1
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-28

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.7 DV: 0.7
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.1

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 4.0 DV: 0.0
 21:58 SP: 4.0 PV: 4.3 DV: 0.3

3

20:58 SP: 4.0 PV: 3.7 DV: -0.1
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-01-30

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.5 DV: 0.5

1

01:58 SP: 4.0 PV: 3.9 DV: -0.1

5

00:58 SP: 4.0 PV: 4.5 DV: 0.1

2022-01-29

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.1 DV: 0.1

3

04:58 SP: 4.0 PV: 3.7 DV: -0.1
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-01-31

SP Setting Point
 PV Present Value
 DV Difference Value