

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space		냉장고	
	Manufacturer	Giant Star	Storage Period		2022.2	
	Model Name	CBR1000SS	IP	Temp.	IP	Temp.
	Serial Number	GS17090102-10				
Calibration Subject	Description	Resistance Temperature Detector				
	Manufacturer	대진센서제작소				
	Model Name	PT100				
	Serial Number	SS21072202				
	Date of Calibration	2021.7.23				
	Certificate No.	SS-210722-0036	Temperature Range		2~8℃	

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2022.02.01						설날
2022.02.02						설날연휴
2022.02.03	10:15			4.5		
2022.02.04	P:40			4.6		
2022.02.05						
2022.02.06						
2022.02.07	P:30			4.0		
2022.02.08	P:40			4.6		
2022.02.09	P:30			4.3		
2022.02.10	P:30			3.7		
2022.02.11	P:30			4.0		
2022.02.12						
2022.02.13						
2022.02.14	P:30			4.7		
2022.02.15	P:30			4.0		
2022.02.16	P:30			4.5		
2022.02.17	P:30			4.3		
2022.02.18	P:30			3.8		
2022.02.19						
2022.02.20						
2022.02.21	P:30			3.8		
2022.02.22	P:30			3.8		
2022.02.23	P:30			3.9		
2022.02.24	P:30			4.3		
2022.02.25	P:30			3.9		
2022.02.26						
2022.02.27						
2022.02.28	P:30			4.0		

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 3.8 DV: -0.2
3

22:58 SP: 4.0 PV: 4.3 DV: 0.
21:58 SP: 4.0 PV: 3.8 DV: -0.2
20:58 SP: 4.0 PV: 4.2 DV: 0.2
19:58 SP: 4.0 PV: 3.9 DV: -0.1
18:58 SP: 4.0 PV: 3.9 DV: -0.1
17:58 SP: 4.0 PV: 4.3 DV: 0.3
16:58 SP: 4.0 PV: 3.9 DV: -0.1
15:58 SP: 4.0 PV: 3.8 DV: -0.2
14:58 SP: 4.0 PV: 4.4 DV: 0.4
13:58 SP: 4.0 PV: 3.7 DV: -0.3
12:58 SP: 4.0 PV: 4.2 DV: 0.2
11:58 SP: 4.0 PV: 3.7 DV: -0.3
10:58 SP: 4.0 PV: 4.2 DV: 0.2
09:58 SP: 4.0 PV: 3.7 DV: -0.3
08:58 SP: 4.0 PV: 4.2 DV: 0.2
07:58 SP: 4.0 PV: 3.7 DV: -0.3
06:58 SP: 4.0 PV: 4.3 DV: 0.3
05:58 SP: 4.0 PV: 4.1 DV: 0.1
04:58 SP: 4.0 PV: 3.7 DV: -0.3
03:58 SP: 4.0 PV: 4.3 DV: 0.3
02:58 SP: 4.0 PV: 3.7 DV: -0.3
01:58 SP: 4.0 PV: 4.4 DV: 0.4
00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-02-02

5

23:58 SP: 4.0 PV: 4.5 DV: 0.
22:58 SP: 4.0 PV: 3.8 DV: -0.2
21:58 SP: 4.0 PV: 4.3 DV: 0.3
20:58 SP: 4.0 PV: 3.8 DV: -0.2
19:58 SP: 4.0 PV: 3.8 DV: -0.2
18:58 SP: 4.0 PV: 4.3 DV: 0.3
17:58 SP: 4.0 PV: 3.9 DV: -0.1
16:58 SP: 4.0 PV: 4.3 DV: 0.3
15:58 SP: 4.0 PV: 3.7 DV: -0.3
14:58 SP: 4.0 PV: 4.3 DV: 0.3
13:58 SP: 4.0 PV: 4.0 DV: 0.0
12:58 SP: 4.0 PV: 4.2 DV: 0.2
11:58 SP: 4.0 PV: 3.7 DV: -0.3
10:58 SP: 4.0 PV: 4.4 DV: 0.4
09:58 SP: 4.0 PV: 3.7 DV: -0.3
08:58 SP: 4.0 PV: 4.4 DV: 0.4
07:58 SP: 4.0 PV: 3.7 DV: -0.3
06:58 SP: 4.0 PV: 4.4 DV: 0.4
05:58 SP: 4.0 PV: 3.7 DV: -0.3
04:58 SP: 4.0 PV: 4.3 DV: 0.3
03:58 SP: 4.0 PV: 3.9 DV: -0.1
02:58 SP: 4.0 PV: 3.7 DV: -0.3
01:58 SP: 4.0 PV: 4.3 DV: 0.3
00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-01

23:58 SP: 4.0 PV: 3.9 DV: -0.1
22:58 SP: 4.0 PV: 4.3 DV: 0.3
21:58 SP: 4.0 PV: 3.9 DV: -0.1
20:58 SP: 4.0 PV: 4.2 DV: 0.2
19:58 SP: 4.0 PV: 3.7 DV: -0.3
18:58 SP: 4.0 PV: 4.3 DV: 0.3
17:58 SP: 4.0 PV: 3.9 DV: -0.1
16:58 SP: 4.0 PV: 4.2 DV: 0.2
15:58 SP: 4.0 PV: 3.9 DV: -0.1
14:58 SP: 4.0 PV: 4.1 DV: 0.1
13:58 SP: 4.0 PV: 4.0 DV: 0.0
12:58 SP: 4.0 PV: 4.1 DV: 0.1
11:58 SP: 4.0 PV: 3.7 DV: -0.3
10:58 SP: 4.0 PV: 4.3 DV: 0.3
09:58 SP: 4.0 PV: 3.8 DV: -0.2
08:58 SP: 4.0 PV: 4.2 DV: 0.2
0

07:58 SP: 4.0 PV: 4.0 DV: 0.
06:58 SP: 4.0 PV: 4.0 DV: 0.0
05:58 SP: 4.0 PV: 3.7 DV: -0.3
04:58 SP: 4.0 PV: 4.2 DV: 0.2
03:58 SP: 4.0 PV: 3.9 DV: -0.1
02:58 SP: 4.0 PV: 3.7 DV: -0.3
01:58 SP: 4.0 PV: 4.3 DV: 0.3
00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-02-04

23:58 SP: 4.0 PV: 4.2 DV: 0.2
22:58 SP: 4.0 PV: 3.9 DV: -0.1
21:58 SP: 4.0 PV: 4.1 DV: 0.1
20:58 SP: 4.0 PV: 3.9 DV: -0.1
1

19:58 SP: 4.0 PV: 4.1 DV: 0.
18:58 SP: 4.0 PV: 3.8 DV: -0.2
17:58 SP: 4.0 PV: 3.7 DV: -0.3
16:58 SP: 4.0 PV: 3.8 DV: -0.2
15:58 SP: 4.0 PV: 4.3 DV: 0.3
14:58 SP: 4.0 PV: 3.9 DV: -0.1
13:58 SP: 4.0 PV: 4.1 DV: 0.1
12:58 SP: 4.0 PV: 3.7 DV: -0.3
11:58 SP: 4.0 PV: 4.2 DV: 0.2
10:58 SP: 4.0 PV: 3.9 DV: -0.1
09:58 SP: 4.0 PV: 3.9 DV: -0.1
08:58 SP: 4.0 PV: 3.8 DV: -0.2
07:58 SP: 4.0 PV: 3.7 DV: -0.3
06:58 SP: 4.0 PV: 4.1 DV: 0.1
05:58 SP: 4.0 PV: 3.7 DV: -0.3
04:58 SP: 4.0 PV: 4.2 DV: 0.2
2

03:58 SP: 4.0 PV: 3.8 DV: -0.
02:58 SP: 4.0 PV: 3.7 DV: -0.3
01:58 SP: 4.0 PV: 4.3 DV: 0.3
00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-03

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 1
 15:58 SP: 4.0 PV: 4.1 DV: 0.
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.3 DV: 0.3
 2022-02-06
 23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 3
 03:58 SP: 4.0 PV: 4.3 DV: 0.
 2
 02:58 SP: 4.0 PV: 3.8 DV: -0.
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 3.9 DV: -0.1
 2022-02-05

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 6
 16:58 SP: 4.0 PV: 4.6 DV: 0.
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 4
 06:58 SP: 4.0 PV: 4.4 DV: 0.
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 3
 03:58 SP: 4.0 PV: 3.7 DV: -0.
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 4.6 DV: 0.6
 2022-02-08
 23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 0
 15:58 SP: 4.0 PV: 4.0 DV: 0.
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 3
 09:58 SP: 4.0 PV: 4.3 DV: 0.
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 1
 02:58 SP: 4.0 PV: 4.1 DV: 0.
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 3.7 DV: -0.3
 2022-02-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

1
 23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.2 DV: 0.2
 2022-02-10
 23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.1 DV: 0.1
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 3.8 DV: -0.2
 2022-02-09

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 1
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 1
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 2
 12:58 SP: 4.0 PV: 3.8 DV: -0.1
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.2 DV: 0.2
 2022-02-12
 23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 1
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.5 DV: 0.5
 2022-02-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.5 DV: 0.5
 16:58 SP: 4.0 PV: 4.0 DV: 0.0

4

15:58 SP: 4.0 PV: 4.4 DV: 0.
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-02-14

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-02-13

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.1 DV: 0.1
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.0 DV: 0.0

1

09:58 SP: 4.0 PV: 3.9 DV: -0.
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-02-16

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.1 DV: 0.1
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-02-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-18

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.1 DV: 0.1

2022-02-17

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.0 DV: 0.0

1

11:58 SP: 4.0 PV: 4.1 DV: 0.
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.2 DV: 0.2

1

08:58 SP: 4.0 PV: 3.9 DV: -0.
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-02-20

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-02-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.0 DV: 0.0
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 4.0 DV: 0.0
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.2 DV: 0.2

1

09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-22

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.1 DV: 0.1

3

15:58 SP: 4.0 PV: 3.7 DV: -0.1
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.2 DV: 0.2

2

01:58 SP: 4.0 PV: 3.8 DV: -0.1
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-21

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-02-24

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.9 DV: -0.1

1

14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 3.7 DV: -0.3

2

08:58 SP: 4.0 PV: 4.2 DV: 0.1
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 4.0 DV: 0.0

1

04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-02-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.2

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.7 DV: 0.7
 19:58 SP: 4.0 PV: 3.7 DV: -0.3

3

18:58 SP: 4.0 PV: 4.3 DV: 0.
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.5 DV: 0.5

.2

02:58 SP: 4.0 PV: 3.8 DV: -0.
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-02-26

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4

1

14:58 SP: 4.0 PV: 3.9 DV: -0.
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.0 DV: 0.0
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-02-25

23:58 SP: 4.0 PV: 4.8 DV: 0.8
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.6 DV: 0.6

3

04:58 SP: 4.0 PV: 3.7 DV: -0.
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-02-28

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 4.7 DV: 0.7
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 3.8 DV: -0.2

4

02:58 SP: 4.0 PV: 4.4 DV: 0.
 01:58 SP: 4.0 PV: 4.7 DV: 0.7

3

00:58 SP: 4.0 PV: 3.7 DV: -0.
 2022-02-27

SP Setting Point
 PV Present Value
 DV Difference Value