

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space		냉장고	
	Manufacturer	Giant Star	Storage Period		2022.3	
	Model Name	CBR1000SS	IP	Temp.	IP	Temp.
	Serial Number	GS17090102-10				
Calibration Subject	Description	Resistance Temperature Detector				
	Manufacturer	대진센서제작소				
	Model Name	PT100				
	Serial Number	SS21072202				
	Date of Calibration	2021.7.23				
	Certificate No.	SS-210722-0036	Temperature Range		2~8℃	

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2022.03.01						삼일절
2022.03.02	15:30			3.7	anf	
2022.03.03	10:00			4.4	anf	
2022.03.04	15:30			4.3	anf	
2022.03.05						
2022.03.06						
2022.03.07	15:30			4.7	anf	
2022.03.08	15:30			4.6	anf	
2022.03.09						대통령선거일
2022.03.10	15:30			4.4	anf	
2022.03.11	15:30			4.3	anf	
2022.03.12						
2022.03.13						
2022.03.14	15:30			3.7	anf	
2022.03.15	15:30			4.1	anf	
2022.03.16	15:30			4.6	anf	
2022.03.17	15:30			4.9	anf	
2022.03.18	15:30			4.4	anf	
2022.03.19						
2022.03.20						
2022.03.21	15:30			4.4	anf	
2022.03.22	15:30			3.9	anf	
2022.03.23	15:30			3.8	anf	
2022.03.24	15:30			3.7	anf	
2022.03.25	10:00			3.9	anf	
2022.03.26						
2022.03.27						
2022.03.28	15:30			4.4	anf	
2022.03.29	15:30			4.7	anf	
2022.03.30	15:30			3.9	anf	
2022.03.31	15:30			4.6	anf	

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.0 DV: 0.0
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.0 DV: 0.0

6

16:58 SP: 4.0 PV: 4.6 DV: 0.
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.6 DV: -0.4
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-03-02

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.1 DV: 0.1
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-03-01

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 4.1 DV: 0.1
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-03-04

23:58 SP: 4.0 PV: 3.9 DV: -0.1

6

22:58 SP: 4.0 PV: 4.6 DV: 0.
 21:58 SP: 4.0 PV: 4.0 DV: -0.1
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.4 DV: 0.4

2

13:58 SP: 4.0 PV: 4.2 DV: 0.
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.7 DV: 0.7
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4

1

03:58 SP: 4.0 PV: 4.1 DV: 0.
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-03-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 3.8 DV: -0.2

3

22:58 SP: 4.0 PV: 4.3 DV: 0.

21:58 SP: 4.0 PV: 3.7 DV: -0.3

20:58 SP: 4.0 PV: 4.4 DV: 0.4

19:58 SP: 4.0 PV: 3.7 DV: -0.3

18:58 SP: 4.0 PV: 4.4 DV: 0.4

17:58 SP: 4.0 PV: 4.0 DV: 0.0

16:58 SP: 4.0 PV: 4.2 DV: 0.2

15:58 SP: 4.0 PV: 4.2 DV: 0.2

14:58 SP: 4.0 PV: 3.7 DV: -0.3

13:58 SP: 4.0 PV: 4.3 DV: 0.3

12:58 SP: 4.0 PV: 3.7 DV: -0.3

11:58 SP: 4.0 PV: 4.4 DV: 0.4

10:58 SP: 4.0 PV: 4.0 DV: 0.0

09:58 SP: 4.0 PV: 4.6 DV: 0.6

08:58 SP: 4.0 PV: 3.9 DV: -0.1

07:58 SP: 4.0 PV: 4.6 DV: 0.6

06:58 SP: 4.0 PV: 3.9 DV: -0.1

6

05:58 SP: 4.0 PV: 4.6 DV: 0.

04:58 SP: 4.0 PV: 3.9 DV: -0.1

03:58 SP: 4.0 PV: 4.6 DV: 0.6

02:58 SP: 4.0 PV: 3.9 DV: -0.1

01:58 SP: 4.0 PV: 4.6 DV: 0.6

00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-03-06

23:58 SP: 4.0 PV: 4.4 DV: 0.4

22:58 SP: 4.0 PV: 4.1 DV: 0.1

21:58 SP: 4.0 PV: 4.6 DV: 0.6

20:58 SP: 4.0 PV: 3.8 DV: -0.2

19:58 SP: 4.0 PV: 4.5 DV: 0.5

18:58 SP: 4.0 PV: 4.2 DV: 0.2

17:58 SP: 4.0 PV: 4.6 DV: 0.6

16:58 SP: 4.0 PV: 4.2 DV: 0.2

15:58 SP: 4.0 PV: 4.5 DV: 0.5

14:58 SP: 4.0 PV: 4.2 DV: 0.2

13:58 SP: 4.0 PV: 4.6 DV: 0.6

12:58 SP: 4.0 PV: 4.2 DV: 0.2

11:58 SP: 4.0 PV: 4.6 DV: 0.6

10:58 SP: 4.0 PV: 4.7 DV: 0.7

09:58 SP: 4.0 PV: 3.8 DV: -0.2

08:58 SP: 4.0 PV: 4.3 DV: 0.3

07:58 SP: 4.0 PV: 4.3 DV: 0.3

06:58 SP: 4.0 PV: 4.6 DV: 0.6

05:58 SP: 4.0 PV: 4.7 DV: 0.7

04:58 SP: 4.0 PV: 4.0 DV: 0.0

03:58 SP: 4.0 PV: 4.4 DV: 0.4

02:58 SP: 4.0 PV: 4.6 DV: 0.6

01:58 SP: 4.0 PV: 4.7 DV: 0.7

00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-03-05

23:58 SP: 4.0 PV: 4.5 DV: 0.5

22:58 SP: 4.0 PV: 4.0 DV: 0.0

21:58 SP: 4.0 PV: 4.6 DV: 0.6

20:58 SP: 4.0 PV: 3.7 DV: -0.3

19:58 SP: 4.0 PV: 4.4 DV: 0.4

18:58 SP: 4.0 PV: 4.1 DV: 0.1

17:58 SP: 4.0 PV: 4.6 DV: 0.6

16:58 SP: 4.0 PV: 3.8 DV: -0.2

15:58 SP: 4.0 PV: 4.4 DV: 0.4

14:58 SP: 4.0 PV: 4.1 DV: 0.1

6

13:58 SP: 4.0 PV: 4.6 DV: 0.

12:58 SP: 4.0 PV: 3.7 DV: -0.3

11:58 SP: 4.0 PV: 4.4 DV: 0.4

10:58 SP: 4.0 PV: 3.9 DV: -0.1

09:58 SP: 4.0 PV: 4.2 DV: 0.2

08:58 SP: 4.0 PV: 3.7 DV: -0.3

07:58 SP: 4.0 PV: 4.3 DV: 0.3

06:58 SP: 4.0 PV: 3.8 DV: -0.2

05:58 SP: 4.0 PV: 4.2 DV: 0.2

04:58 SP: 4.0 PV: 4.0 DV: 0.0

03:58 SP: 4.0 PV: 4.3 DV: 0.3

02:58 SP: 4.0 PV: 3.9 DV: -0.1

01:58 SP: 4.0 PV: 3.8 DV: -0.2

00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-03-08

23:58 SP: 4.0 PV: 3.9 DV: -0.1

22:58 SP: 4.0 PV: 3.8 DV: -0.2

21:58 SP: 4.0 PV: 4.3 DV: 0.3

20:58 SP: 4.0 PV: 3.7 DV: -0.3

19:58 SP: 4.0 PV: 4.5 DV: 0.5

18:58 SP: 4.0 PV: 3.8 DV: -0.2

17:58 SP: 4.0 PV: 4.3 DV: 0.3

16:58 SP: 4.0 PV: 3.7 DV: -0.3

15:58 SP: 4.0 PV: 4.4 DV: 0.4

14:58 SP: 4.0 PV: 3.7 DV: -0.3

13:58 SP: 4.0 PV: 4.4 DV: 0.4

12:58 SP: 4.0 PV: 3.7 DV: -0.3

11:58 SP: 4.0 PV: 4.3 DV: 0.3

10:58 SP: 4.0 PV: 3.7 DV: -0.3

09:58 SP: 4.0 PV: 4.4 DV: 0.4

1

08:58 SP: 4.0 PV: 3.9 DV: -0.

07:58 SP: 4.0 PV: 3.7 DV: -0.3

06:58 SP: 4.0 PV: 4.4 DV: 0.4

05:58 SP: 4.0 PV: 4.0 DV: 0.0

04:58 SP: 4.0 PV: 4.2 DV: 0.2

03:58 SP: 4.0 PV: 3.7 DV: -0.3

02:58 SP: 4.0 PV: 4.3 DV: 0.3

01:58 SP: 4.0 PV: 3.7 DV: -0.3

00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-03-07

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

```

-----
23:58 SP: 4.0 PV: 3.8 DV: -0.2
22:58 SP: 4.0 PV: 4.6 DV: 0.6
21:58 SP: 4.0 PV: 3.8 DV: -0.2
20:58 SP: 4.0 PV: 4.5 DV: 0.5
19:58 SP: 4.0 PV: 3.7 DV: -0.3
18:58 SP: 4.0 PV: 4.4 DV: 0.4
17:58 SP: 4.0 PV: 3.9 DV: -0.1
16:58 SP: 4.0 PV: 4.2 DV: 0.2
15:58 SP: 4.0 PV: 4.2 DV: 0.2
14:58 SP: 4.0 PV: 4.6 DV: 0.6
13:58 SP: 4.0 PV: 3.9 DV: -0.1
12:58 SP: 4.0 PV: 4.5 DV: 0.5
11:58 SP: 4.0 PV: 3.7 DV: -0.3
10:58 SP: 4.0 PV: 4.3 DV: 0.3
09:58 SP: 4.0 PV: 4.2 DV: 0.2
08:58 SP: 4.0 PV: 4.6 DV: 0.6
07:58 SP: 4.0 PV: 3.7 DV: -0.3
06:58 SP: 4.0 PV: 4.4 DV: 0.4
05:58 SP: 4.0 PV: 4.2 DV: 0.2
04:58 SP: 4.0 PV: 4.6 DV: 0.6
03:58 SP: 4.0 PV: 3.8 DV: -0.2
02:58 SP: 4.0 PV: 4.6 DV: 0.6
01:58 SP: 4.0 PV: 3.7 DV: -0.3
00:58 SP: 4.0 PV: 4.3 DV: 0.3
2022-03-10
23:58 SP: 4.0 PV: 4.2 DV: 0.2
22:58 SP: 4.0 PV: 4.7 DV: 0.7
21:58 SP: 4.0 PV: 3.7 DV: -0.3
20:58 SP: 4.0 PV: 4.4 DV: 0.4
19:58 SP: 4.0 PV: 4.2 DV: 0.2
18:58 SP: 4.0 PV: 4.3 DV: 0.3
17:58 SP: 4.0 PV: 4.6 DV: 0.6
16:58 SP: 4.0 PV: 3.7 DV: -0.3
15:58 SP: 4.0 PV: 4.3 DV: 0.3
14:58 SP: 4.0 PV: 4.2 DV: 0.2
13:58 SP: 4.0 PV: 4.6 DV: 0.6
12:58 SP: 4.0 PV: 3.9 DV: -0.1
11:58 SP: 4.0 PV: 4.6 DV: 0.6
10:58 SP: 4.0 PV: 3.9 DV: -0.1
09:58 SP: 4.0 PV: 4.6 DV: 0.6
08:58 SP: 4.0 PV: 3.9 DV: -0.1
07:58 SP: 4.0 PV: 4.5 DV: 0.5
06:58 SP: 4.0 PV: 3.9 DV: -0.1
05:58 SP: 4.0 PV: 4.6 DV: 0.6
04:58 SP: 4.0 PV: 3.8 DV: -0.2
03:58 SP: 4.0 PV: 4.5 DV: 0.5
02:58 SP: 4.0 PV: 3.9 DV: -0.1
6
01:58 SP: 4.0 PV: 4.6 DV: 0.
00:58 SP: 4.0 PV: 3.9 DV: -0.1
2022-03-09

```

```

23:58 SP: 4.0 PV: 4.8 DV: 0.8
22:58 SP: 4.0 PV: 3.8 DV: -0.2
21:58 SP: 4.0 PV: 4.2 DV: 0.2
20:58 SP: 4.0 PV: 4.3 DV: 0.3
19:58 SP: 4.0 PV: 4.6 DV: 0.6
18:58 SP: 4.0 PV: 4.6 DV: 0.6
17:58 SP: 4.0 PV: 4.0 DV: 0.0
16:58 SP: 4.0 PV: 4.4 DV: 0.4
15:58 SP: 4.0 PV: 4.4 DV: 0.4
14:58 SP: 4.0 PV: 4.7 DV: 0.7
8
13:58 SP: 4.0 PV: 4.8 DV: 0.
12:58 SP: 4.0 PV: 4.8 DV: 0.8
11:58 SP: 4.0 PV: 4.8 DV: 0.8
10:58 SP: 4.0 PV: 4.2 DV: 0.2
09:58 SP: 4.0 PV: 4.3 DV: 0.3
08:58 SP: 4.0 PV: 4.4 DV: 0.4
07:58 SP: 4.0 PV: 4.3 DV: 0.3
06:58 SP: 4.0 PV: 4.6 DV: 0.6
05:58 SP: 4.0 PV: 4.7 DV: 0.7
04:58 SP: 4.0 PV: 4.7 DV: 0.7
03:58 SP: 4.0 PV: 4.0 DV: 0.0
02:58 SP: 4.0 PV: 4.4 DV: 0.4
01:58 SP: 4.0 PV: 4.5 DV: 0.5
00:58 SP: 4.0 PV: 4.3 DV: 0.3
2022-03-12
23:58 SP: 4.0 PV: 4.5 DV: 0.5
22:58 SP: 4.0 PV: 4.6 DV: 0.6
3
21:58 SP: 4.0 PV: 4.3 DV: 0.
20:58 SP: 4.0 PV: 4.4 DV: 0.4
19:58 SP: 4.0 PV: 4.6 DV: 0.6
18:58 SP: 4.0 PV: 4.7 DV: 0.7
17:58 SP: 4.0 PV: 4.8 DV: 0.8
16:58 SP: 4.0 PV: 4.2 DV: 0.2
15:58 SP: 4.0 PV: 4.5 DV: 0.5
14:58 SP: 4.0 PV: 4.6 DV: 0.6
13:58 SP: 4.0 PV: 4.7 DV: 0.7
7
12:58 SP: 4.0 PV: 4.7 DV: 0.
11:58 SP: 4.0 PV: 4.3 DV: 0.3
10:58 SP: 4.0 PV: 4.5 DV: 0.5
09:58 SP: 4.0 PV: 4.2 DV: 0.2
08:58 SP: 4.0 PV: 4.6 DV: 0.6
07:58 SP: 4.0 PV: 4.1 DV: 0.1
06:58 SP: 4.0 PV: 4.6 DV: 0.6
05:58 SP: 4.0 PV: 4.0 DV: 0.0
04:58 SP: 4.0 PV: 4.6 DV: 0.6
03:58 SP: 4.0 PV: 4.1 DV: 0.1
02:58 SP: 4.0 PV: 4.6 DV: 0.6
01:58 SP: 4.0 PV: 3.7 DV: -0.3
00:58 SP: 4.0 PV: 4.5 DV: 0.5
2022-03-11

```

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.7 DV: 0.7
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.3 DV: 0.3

6

08:58 SP: 4.0 PV: 4.6 DV: 0.
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.6 DV: 0.6

6

03:58 SP: 4.0 PV: 4.6 DV: 0.
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-03-14

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.1 DV: 0.1
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.8 DV: 0.8
 16:58 SP: 4.0 PV: 4.7 DV: 0.7
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.8 DV: 0.8
 00:58 SP: 4.0 PV: 4.8 DV: 0.8

2022-03-13

3

*23:58 SP: 4.0 PV: 3.7 DV: -0.
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 3.9 DV: -0.1

4

17:58 SP: 4.0 PV: 4.4 DV: 0.
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4

7

06:58 SP: 4.0 PV: 4.7 DV: 0.
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 3.7 DV: -0.3

3

01:58 SP: 4.0 PV: 4.3 DV: 0.
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-03-16

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.7 DV: 0.7
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.1 DV: 0.1

4

13:58 SP: 4.0 PV: 4.4 DV: 0.
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 4.7 DV: 0.7
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.5 DV: 0.5
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-03-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 5
 20:58 SP: 4.0 PV: 4.5 DV: 0.
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 4
 17:58 SP: 4.0 PV: 4.4 DV: 0.
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.7 DV: 0.7
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.8 DV: 0.8
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.8 DV: 0.8
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 3.6 DV: -0.4
 01:58 SP: 4.0 PV: 4.4 DV: 0.4

2
 00:58 SP: 4.0 PV: 4.2 DV: 0.

2022-03-18

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.2 DV: 0.2

6

14:58 SP: 4.0 PV: 4.6 DV: 0.
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 3.9 DV: -0.1

5

00:58 SP: 4.0 PV: 4.5 DV: 0.
 2022-03-17

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 4.7 DV: 0.7

2

21:58 SP: 4.0 PV: 3.8 DV: -0.
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.5 DV: 0.5
 16:58 SP: 4.0 PV: 4.7 DV: 0.7
 15:58 SP: 4.0 PV: 4.0 DV: 0.0
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.7 DV: 0.7
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.3 DV: 0.3

7

*02:58 SP: 4.0 PV: 4.7 DV: 0.
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-03-20

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.6 DV: 0.6

8

21:58 SP: 4.0 PV: 4.8 DV: 0.
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.7 DV: 0.7
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.8 DV: 0.8
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 3.9 DV: -0.1

4

05:58 SP: 4.0 PV: 4.4 DV: 0.
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.8 DV: 0.8
 00:58 SP: 4.0 PV: 4.8 DV: 0.8

2022-03-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 4.8 DV: 0.8
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.7 DV: 0.7
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.5 DV: 0.5

B

12:58 SP: 4.0 PV: 4.8 DV: 0.
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-03-22

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.7 DV: 0.7
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.7 DV: 0.7
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-03-21

23:58 SP: 4.0 PV: 4.8 DV: 0.8
 22:58 SP: 4.0 PV: 4.8 DV: 0.8
 21:58 SP: 4.0 PV: 4.8 DV: 0.8
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.8 DV: 0.8
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.7 DV: 0.7
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.7 DV: 0.7

2022-03-24

23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.7 DV: 0.7

8

15:58 SP: 4.0 PV: 4.8 DV: 0.
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.3 DV: 0.3

3

08:58 SP: 4.0 PV: 4.3 DV: 0.
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.6 DV: 0.6

3

04:58 SP: 4.0 PV: 3.7 DV: -0.
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-03-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP:	4.0 PV:	4.6 DV:	0.6	23:58 SP:	4.0 PV:	4.8 DV:	0.8
22:58 SP:	4.0 PV:	4.6 DV:	0.6	22:58 SP:	4.0 PV:	4.8 DV:	0.8
21:58 SP:	4.0 PV:	4.0 DV:	0.0	21:58 SP:	4.0 PV:	3.7 DV:	-0.3
20:58 SP:	4.0 PV:	4.4 DV:	0.4	20:58 SP:	4.0 PV:	4.5 DV:	0.5
19:58 SP:	4.0 PV:	4.6 DV:	0.6	19:58 SP:	4.0 PV:	4.7 DV:	0.7
18:58 SP:	4.0 PV:	4.7 DV:	0.7	18:58 SP:	4.0 PV:	4.3 DV:	0.3
17:58 SP:	4.0 PV:	4.6 DV:	0.6	4			
16:58 SP:	4.0 PV:	4.5 DV:	0.5	17:58 SP:	4.0 PV:	4.4 DV:	0.
15:58 SP:	4.0 PV:	4.5 DV:	0.5	16:58 SP:	4.0 PV:	4.6 DV:	0.6
14:58 SP:	4.0 PV:	4.6 DV:	0.6	15:58 SP:	4.0 PV:	4.7 DV:	0.7
13:58 SP:	4.0 PV:	4.3 DV:	0.3	14:58 SP:	4.0 PV:	4.2 DV:	0.2
12:58 SP:	4.0 PV:	4.6 DV:	0.6	13:58 SP:	4.0 PV:	4.6 DV:	0.6
11:58 SP:	4.0 PV:	4.7 DV:	0.7	12:58 SP:	4.0 PV:	4.8 DV:	0.8
10:58 SP:	4.0 PV:	4.2 DV:	0.2	11:58 SP:	4.0 PV:	4.1 DV:	0.1
09:58 SP:	4.0 PV:	4.5 DV:	0.5	10:58 SP:	4.0 PV:	4.6 DV:	0.6
08:58 SP:	4.0 PV:	4.7 DV:	0.7	09:58 SP:	4.0 PV:	4.6 DV:	0.6
07:58 SP:	4.0 PV:	4.7 DV:	0.7	08:58 SP:	4.0 PV:	4.2 DV:	0.2
06:58 SP:	4.0 PV:	4.6 DV:	0.6	07:58 SP:	4.0 PV:	4.3 DV:	0.3
05:58 SP:	4.0 PV:	4.7 DV:	0.7	06:58 SP:	4.0 PV:	4.6 DV:	0.6
04:58 SP:	4.0 PV:	4.8 DV:	0.8	05:58 SP:	4.0 PV:	4.7 DV:	0.7
03:58 SP:	4.0 PV:	4.8 DV:	0.8	04:58 SP:	4.0 PV:	3.7 DV:	-0.3
02:58 SP:	4.0 PV:	4.8 DV:	0.8	03:58 SP:	4.0 PV:	4.4 DV:	0.4
01:58 SP:	4.0 PV:	4.9 DV:	0.9	02:58 SP:	4.0 PV:	4.7 DV:	0.7
00:58 SP:	4.0 PV:	4.8 DV:	0.8	01:58 SP:	4.0 PV:	4.7 DV:	0.7
2022-03-26				00:58 SP:	4.0 PV:	3.9 DV:	-0.1
23:58 SP:	4.0 PV:	4.3 DV:	0.3	2022-03-28			
22:58 SP:	4.0 PV:	4.5 DV:	0.5	23:58 SP:	4.0 PV:	4.4 DV:	0.4
21:58 SP:	4.0 PV:	4.7 DV:	0.7	22:58 SP:	4.0 PV:	4.3 DV:	0.3
20:58 SP:	4.0 PV:	4.7 DV:	0.7	21:58 SP:	4.0 PV:	4.6 DV:	0.6
19:58 SP:	4.0 PV:	4.6 DV:	0.6	20:58 SP:	4.0 PV:	4.7 DV:	0.7
18:58 SP:	4.0 PV:	4.6 DV:	0.6	19:58 SP:	4.0 PV:	4.7 DV:	0.7
17:58 SP:	4.0 PV:	4.6 DV:	0.6	1			
16:58 SP:	4.0 PV:	4.7 DV:	0.7	18:58 SP:	4.0 PV:	3.9 DV:	-0.
15:58 SP:	4.0 PV:	4.8 DV:	0.8	4			
14:58 SP:	4.0 PV:	4.8 DV:	0.8	17:58 SP:	4.0 PV:	4.4 DV:	0.
13:58 SP:	4.0 PV:	4.8 DV:	0.8	16:58 SP:	4.0 PV:	4.6 DV:	0.6
12:58 SP:	4.0 PV:	4.8 DV:	0.8	15:58 SP:	4.0 PV:	4.8 DV:	0.8
11:58 SP:	4.0 PV:	4.8 DV:	0.8	14:58 SP:	4.0 PV:	4.2 DV:	0.2
10:58 SP:	4.0 PV:	4.8 DV:	0.8	13:58 SP:	4.0 PV:	4.4 DV:	0.4
09:58 SP:	4.0 PV:	4.1 DV:	0.1	12:58 SP:	4.0 PV:	4.6 DV:	0.6
08:58 SP:	4.0 PV:	4.3 DV:	0.3	11:58 SP:	4.0 PV:	4.7 DV:	0.7
07:58 SP:	4.0 PV:	4.4 DV:	0.4	10:58 SP:	4.0 PV:	4.8 DV:	0.8
06:58 SP:	4.0 PV:	4.5 DV:	0.5	09:58 SP:	4.0 PV:	4.2 DV:	0.2
05:58 SP:	4.0 PV:	4.7 DV:	0.7	08:58 SP:	4.0 PV:	4.5 DV:	0.5
04:58 SP:	4.0 PV:	4.7 DV:	0.7	07:58 SP:	4.0 PV:	4.7 DV:	0.7
03:58 SP:	4.0 PV:	4.2 DV:	0.2	06:58 SP:	4.0 PV:	4.8 DV:	0.8
02:58 SP:	4.0 PV:	4.4 DV:	0.4	05:58 SP:	4.0 PV:	4.0 DV:	0.0
01:58 SP:	4.0 PV:	4.6 DV:	0.6	04:58 SP:	4.0 PV:	4.4 DV:	0.4
00:58 SP:	4.0 PV:	4.7 DV:	0.7	03:58 SP:	4.0 PV:	4.7 DV:	0.7
2022-03-25				02:58 SP:	4.0 PV:	4.7 DV:	0.7
--				01:58 SP:	4.0 PV:	3.6 DV:	-0.4
				00:58 SP:	4.0 PV:	4.4 DV:	0.4
				2022-03-27			

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.3

23:58 SP: 4.0 PV: 3.7 DV: -0.3	23:58 SP: 4.0 PV: 4.6 DV: 0.6
22:58 SP: 4.0 PV: 4.5 DV: 0.5	22:58 SP: 4.0 PV: 4.3 DV: 0.3
21:58 SP: 4.0 PV: 3.7 DV: -0.3	21:58 SP: 4.0 PV: 4.7 DV: 0.7
20:58 SP: 4.0 PV: 4.4 DV: 0.4	20:58 SP: 4.0 PV: 4.2 DV: 0.2
19:58 SP: 4.0 PV: 4.2 DV: 0.2	19:58 SP: 4.0 PV: 4.7 DV: 0.7
18:58 SP: 4.0 PV: 4.7 DV: 0.7	18:58 SP: 4.0 PV: 4.0 DV: 0.0
17:58 SP: 4.0 PV: 3.9 DV: -0.1	17:58 SP: 4.0 PV: 4.6 DV: 0.6
6	16:58 SP: 4.0 PV: 4.3 DV: -0.4
16:58 SP: 4.0 PV: 4.6 DV: 0.	15:58 SP: 4.0 PV: 4.3 DV: 0.3
15:58 SP: 4.0 PV: 4.1 DV: 0.1	6
14:58 SP: 4.0 PV: 4.6 DV: 0.6	14:58 SP: 4.0 PV: 4.6 DV: 0.
13:58 SP: 4.0 PV: 4.0 DV: 0.0	13:58 SP: 4.0 PV: 4.8 DV: 0.8
12:58 SP: 4.0 PV: 4.6 DV: 0.6	12:58 SP: 4.0 PV: 4.8 DV: 0.8
11:58 SP: 4.0 PV: 3.9 DV: -0.1	11:58 SP: 4.0 PV: 4.2 DV: 0.2
3	10:58 SP: 4.0 PV: 4.4 DV: 0.4
10:58 SP: 4.0 PV: 4.3 DV: 0.	09:58 SP: 4.0 PV: 4.7 DV: 0.7
09:58 SP: 4.0 PV: 4.7 DV: 0.7	08:58 SP: 4.0 PV: 4.2 DV: 0.2
08:58 SP: 4.0 PV: 4.1 DV: 0.1	07:58 SP: 4.0 PV: 4.6 DV: 0.6
07:58 SP: 4.0 PV: 4.4 DV: 0.4	06:58 SP: 4.0 PV: 3.6 DV: -0.4
06:58 SP: 4.0 PV: 4.7 DV: 0.7	05:58 SP: 4.0 PV: 4.4 DV: 0.4
05:58 SP: 4.0 PV: 4.8 DV: 0.8	04:58 SP: 4.0 PV: 4.7 DV: 0.7
04:58 SP: 4.0 PV: 3.9 DV: -0.1	03:58 SP: 4.0 PV: 3.8 DV: -0.2
03:58 SP: 4.0 PV: 4.4 DV: 0.4	02:58 SP: 4.0 PV: 4.5 DV: 0.5
02:58 SP: 4.0 PV: 4.7 DV: 0.7	01:58 SP: 4.0 PV: 4.2 DV: 0.2
01:58 SP: 4.0 PV: 4.2 DV: 0.2	00:58 SP: 4.0 PV: 4.3 DV: 0.3
00:58 SP: 4.0 PV: 4.6 DV: 0.6	2022-03-31
2022-03-29	23:58 SP: 4.0 PV: 4.3 DV: 0.3
	22:58 SP: 4.0 PV: 4.7 DV: 0.7
	21:58 SP: 4.0 PV: 3.7 DV: -0.3
	20:58 SP: 4.0 PV: 4.4 DV: 0.4
	19:58 SP: 4.0 PV: 4.7 DV: 0.7
	18:58 SP: 4.0 PV: 4.0 DV: 0.0
	17:58 SP: 4.0 PV: 4.5 DV: 0.5
	16:58 SP: 4.0 PV: 4.7 DV: 0.7
	1
	15:58 SP: 4.0 PV: 3.9 DV: -0.
	14:58 SP: 4.0 PV: 4.4 DV: 0.4
	13:58 SP: 4.0 PV: 4.7 DV: 0.7
	12:58 SP: 4.0 PV: 3.9 DV: -0.1
	11:58 SP: 4.0 PV: 4.6 DV: 0.6
	10:58 SP: 4.0 PV: 4.2 DV: 0.2
	09:58 SP: 4.0 PV: 4.7 DV: 0.7
	08:58 SP: 4.0 PV: 3.7 DV: -0.3
	07:58 SP: 4.0 PV: 4.4 DV: 0.4
	06:58 SP: 4.0 PV: 4.7 DV: 0.7
	05:58 SP: 4.0 PV: 3.7 DV: -0.3
	04:58 SP: 4.0 PV: 4.5 DV: 0.5
	03:58 SP: 4.0 PV: 4.2 DV: 0.2
	02:58 SP: 4.0 PV: 4.7 DV: 0.7
	01:58 SP: 4.0 PV: 4.1 DV: 0.1
	00:58 SP: 4.0 PV: 4.7 DV: 0.7
	2022-03-30

SP Setting Point
PV Present Value
DV Difference Value