

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space		함온고1	
	Manufacturer	SANYO	Storage Period		2022.6	
	Model Name	MPR-720R	IP	Temp.	IP	Temp.
	Serial Number	07120004		15~25°C 25°C 1-25°C		
Calibration Subject	Description					
	Manufacturer					
	Model Name					
	Serial Number					
	Date of Calibration	2021.8.10				
	Certificate No.	SP2108-720R-06	Temperature Range		16~22°C(19±3°C)	

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2022.06.01						지방선거일
2022.06.02	P:30			18		
2022.06.03	P:30			18		
2022.06.04						
2022.06.05						
2022.06.06						현충일
2022.06.07	P:30			18		
2022.06.08	P:30			19		
2022.06.09	P:40			19		
2022.06.10	P:30			18		
2022.06.11						
2022.06.12						
2022.06.13	P:30			18		
2022.06.14	P:30			18		
2022.06.15	P:30			18		
2022.06.16	P:30			18		
2022.06.17	P:30			18		
2022.06.18						
2022.06.19						
2022.06.20	P:10			18		
2022.06.21	P:30			19		
2022.06.22	P:30			19		
2022.06.23	P:30			18		
2022.06.24	P:30			18		
2022.06.25						
2022.06.26						
2022.06.27	P:30			18		
2022.06.28	P:30			18		
2022.06.29	P:30			19		HD-AI IP o/dc
2022.06.30	P:30			19		


date of copying

sign

부록7. 서식II-01-07, 온도기록지(IP 보관)-자동온도 기록용지 부착형

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space		항온고1	
	Manufacturer	SANYO	Storage Period		2022.5.31 ~ 2022.6.30	
	Model Name	MPR-720R	IP	Temp.	IP	Temp.
	Serial Number	07120004	15~25°C 25°C 1~25°C			
Calibration Subject	Description					
	Manufacturer					
	Model Name					
	Serial Number					
	Date of Calibration					
Certificate No.		SP2108-720R-06	Temperature Range		16~22°C(19±3°C)	



date of copying

sign