

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.09.01	p:40	Left	2	6	5	[Signature]	
		Right	3	4	3		
2022.09.02	p:30	Left	2	5	3	[Signature]	
		Right	3	4	3		
2022.09.03		Left					
		Right					
2022.09.04		Left					
		Right					
2022.09.05	p:30	Left	2	5	2	[Signature]	
		Right	3	4	3		
2022.09.06	p:30	Left	2	5	3	[Signature]	
		Right	3	4	4		
2022.09.07	p:30	Left	2	6	2	[Signature]	
		Right	3	4	3		
2022.09.08	p:30	Left	2	6	4	[Signature]	
		Right	3	4	4		
2022.09.09		Left					추석연휴
		Right					
2022.09.10		Left					추석
		Right					
2022.09.11		Left					추석연휴
		Right					
2022.09.12		Left					대체휴일
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.09.13	P:30	Left	2	6	4	[Signature]	
		Right	2	4	3		
2022.09.14	P:30	Left	2	6	5	[Signature]	
		Right	3	4	4		
2022.09.15	P:30	Left	2	6	4	[Signature]	
		Right	3	4	4		
2022.09.16	P:30	Left	2	6	5	[Signature]	
		Right	3	4	4		
2022.09.17		Left					
		Right					
2022.09.18		Left					
		Right					
2022.09.19	P:30	Left	2	6	3	[Signature]	
		Right	2	4	3		
2022.09.20	P:30	Left	2	5	5	[Signature]	
		Right	3	4	4		
2022.09.21	P:30	Left	2	6	5	[Signature]	
		Right	2	5	4		
2022.09.22	P:30	Left	2	6	5	[Signature]	
		Right	3	5	4		
2022.09.23	P:30	Left	2	6	5	[Signature]	
		Right	3	5	4		
2022.09.24		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.09.25		Left					
		Right					
2022.09.26	p:30	Left	2	6	4	ml	
		Right	2	5	4		
2022.09.27	p:10	Left	2	6	2	ml	
		Right	3	5	3		
2022.09.28	p:30	Left	2	6	5	ml	
		Right	3	5	4		
2022.09.29	p:45	Left	2	6	4	ml	
		Right	2	5	4		
2022.09.30	p:30	Left	2	6	5	ml	
		Right	3	5	4		
		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.5 DV: 0.5
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-09-02

2

23:58 SP: 4.0 PV: 4.2 DV: 0.
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.5 DV: 0.5
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-09-01

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 3.8 DV: -0.2
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.5 DV: 0.5
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-04

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.8 DV: -0.2

3

09:58 SP: 4.0 PV: 4.3 DV: 0.
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-06

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-09-05

23:58 SP: 4.0 PV: 4.8 DV: 0.8
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.8 DV: -0.3
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.5 DV: 0.5

0

07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-09-08

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4

3

12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 3.7 DV: -0.3

5

06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-09-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.2 DV: 0.3
 19:58 SP: 4.0 PV: 4.2 DV: 0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.2 DV: 0.7
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.2 DV: 0.3
 14:58 SP: 4.0 PV: 3.8 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.0 DV: 0.0
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.2 DV: 0.2
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-10

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.2 DV: 0.1
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.7 DV: -0.2
 09:58 SP: 4.0 PV: 4.2 DV: 0.3
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.9 DV: 0.9

2022-09-09

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.8 DV: -0.1
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.2 DV: 0.2

3

17:58 SP: 4.0 PV: 3.7 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.3
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.0
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 3.8 DV: -0.2
 07:58 SP: 4.0 PV: 4.3 DV: 0.3
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.1 DV: 0.1

2

03:58 SP: 4.0 PV: 4.2 DV: 0.1
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-12

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3

5

19:58 SP: 4.0 PV: 4.5 DV: 0.1
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.8 DV: -0.1
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-09-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 3.9 DV: -0.1

5

18:58 SP: 4.0 PV: 4.5 DV: 0.
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 3.9 DV: -0.1

6

14:58 SP: 4.0 PV: 4.6 DV: 0.
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.7 DV: -0.3

4

08:58 SP: 4.0 PV: 4.4 DV: 0.
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-09-14

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.0 DV: 0.0
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.3 DV: 0.3

2

11:58 SP: 4.0 PV: 4.2 DV: 0.
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-09-13

SP Setting Point
 PV Present Value
 DV Difference Value

23:54 SP: 4.0 PV: 3.8 DV: -0.2
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.5 DV: 0.5
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 4.2 DV: 0.2
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.3 DV: 0.3

4

09:54 SP: 4.0 PV: 4.4 DV: 0.
 Condensor#1 Temp#1. OK...
 Compressor#1 Output OK...
 System Checking Process...

08:54:58

2022-09-15

주소창에 한글로 혈액냉장고

www.gsmedical.co.kr

CBR-1000SS
 GIANT STAR Ltd.

07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-09-15

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:54 SP: 4.0 PV: 3.6 DV: -0.4
 22:54 SP: 4.0 PV: 4.0 DV: 0.0
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 3.6 DV: -0.4
 13:54 SP: 4.0 PV: 4.7 DV: 0.7
 12:54 SP: 4.0 PV: 4.9 DV: 0.9
 11:54 SP: 4.0 PV: 4.7 DV: 0.7
 10:54 SP: 4.0 PV: 4.7 DV: 0.7
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 3.6 DV: -0.4
 06:54 SP: 4.0 PV: 3.6 DV: -0.4
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.9 DV: 0.9
 03:54 SP: 4.0 PV: 4.8 DV: 0.8

6

02:54 SP: 4.0 PV: 4.6 DV: 0.
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.6 DV: 0.6

2022-09-17

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 3.7 DV: -0.3
 04:54 SP: 4.0 PV: 4.5 DV: 0.5
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 4.5 DV: 0.5

2022-09-16

SP Setting Point
 PV Present Value
 DV Difference Value

23:18 SP: 4.0 PV: 4.2 DV: 0.2
 22:18 SP: 4.0 PV: 3.7 DV: -0.3
 21:18 SP: 4.0 PV: 4.4 DV: 0.4
 20:18 SP: 4.0 PV: 3.7 DV: -0.3
 19:18 SP: 4.0 PV: 4.4 DV: 0.4
 18:18 SP: 4.0 PV: 3.7 DV: -0.3
 17:18 SP: 4.0 PV: 4.5 DV: 0.5
 16:18 SP: 4.0 PV: 3.7 DV: -0.3
 15:18 SP: 4.0 PV: 4.3 DV: 0.3
 14:18 SP: 4.0 PV: 3.7 DV: -0.3
 13:18 SP: 4.0 PV: 4.4 DV: 0.4
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 4.5 DV: 0.5
 10:18 SP: 4.0 PV: 4.2 DV: 0.2

Condensor1 Temp#1. OK...
 Compressor#1 Output OK...
 System Checking Process...

09:18:47

2022-09-19

주소창에 한글로 혈액냉장고

www.gsmedical.co.kr

CBR-1000SS

GIANT STAR Ltd.

08:54 SP: 4.0 PV: 4.2 DV: 0.2

07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.1 DV: 0.1
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-19

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 3.7 DV: -0.3

5

19:54 SP: 4.0 PV: 4.5 DV: 0.
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.0 DV: 0.0
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 3.7 DV: -0.3
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.8 DV: 0.8
 05:54 SP: 4.0 PV: 4.7 DV: 0.7
 04:54 SP: 4.0 PV: 4.8 DV: 0.8
 03:54 SP: 4.0 PV: 4.8 DV: 0.8
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2022-09-18

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:18 SP: 4.0 PV: 4.6 DV: 0.6
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 4.4 DV: 0.4
 19:18 SP: 4.0 PV: 4.4 DV: 0.4
 18:18 SP: 4.0 PV: 4.6 DV: 0.6
 17:18 SP: 4.0 PV: 4.6 DV: 0.6
 16:18 SP: 4.0 PV: 4.5 DV: 0.5
 15:18 SP: 4.0 PV: 4.5 DV: 0.5
 14:18 SP: 4.0 PV: 4.1 DV: 0.1
 13:18 SP: 4.0 PV: 4.6 DV: 0.6
 12:18 SP: 4.0 PV: 4.2 DV: 0.2
 11:18 SP: 4.0 PV: 4.6 DV: 0.6
 10:18 SP: 4.0 PV: 4.8 DV: 0.8
 09:18 SP: 4.0 PV: 3.8 DV: -0.2
 08:18 SP: 4.0 PV: 4.5 DV: 0.5
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 4.2 DV: 0.2
 05:18 SP: 4.0 PV: 3.8 DV: -0.2
 04:18 SP: 4.0 PV: 3.9 DV: -0.1
 03:18 SP: 4.0 PV: 3.7 DV: -0.3
 02:18 SP: 4.0 PV: 3.6 DV: -0.4
 01:18 SP: 4.0 PV: 4.8 DV: 0.8
 00:18 SP: 4.0 PV: 4.8 DV: 0.8

2022-09-21

23:18 SP: 4.0 PV: 4.6 DV: 0.6
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 4.6 DV: 0.6
 20:18 SP: 4.0 PV: 4.5 DV: 0.5

6

19:18 SP: 4.0 PV: 4.6 DV: 0.
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 4.6 DV: 0.6
 16:18 SP: 4.0 PV: 4.6 DV: 0.6
 15:18 SP: 4.0 PV: 4.0 DV: 0.0
 14:18 SP: 4.0 PV: 4.6 DV: 0.6
 13:18 SP: 4.0 PV: 4.6 DV: 0.6
 12:18 SP: 4.0 PV: 4.8 DV: 0.8

1

11:18 SP: 4.0 PV: 3.9 DV: -0.
 10:18 SP: 4.0 PV: 4.2 DV: 0.2
 09:18 SP: 4.0 PV: 4.6 DV: 0.6
 08:18 SP: 4.0 PV: 3.7 DV: -0.3

4

07:18 SP: 4.0 PV: 4.4 DV: 0.
 06:18 SP: 4.0 PV: 4.6 DV: 0.6
 05:18 SP: 4.0 PV: 3.8 DV: -0.2
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.6 DV: 0.6
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 4.2 DV: 0.2

2022-09-20

23:18 SP: 4.0 PV: 4.3 DV: 0.3
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 4.4 DV: 0.4
 20:18 SP: 4.0 PV: 4.6 DV: 0.6
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 3.7 DV: -0.3
 17:18 SP: 4.0 PV: 4.3 DV: 0.3

6

16:18 SP: 4.0 PV: 4.6 DV: 0.
 15:18 SP: 4.0 PV: 4.8 DV: 0.8
 14:18 SP: 4.0 PV: 4.7 DV: 0.7
 13:18 SP: 4.0 PV: 4.3 DV: 0.3
 12:18 SP: 4.0 PV: 4.2 DV: 0.2
 11:18 SP: 4.0 PV: 4.0 DV: 0.0
 10:18 SP: 4.0 PV: 4.3 DV: 0.3
 09:18 SP: 4.0 PV: 4.0 DV: 0.0
 08:18 SP: 4.0 PV: 4.3 DV: 0.3
 07:18 SP: 4.0 PV: 4.6 DV: 0.6
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 4.2 DV: 0.2
 04:18 SP: 4.0 PV: 4.5 DV: 0.5
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.4 DV: 0.4
 01:18 SP: 4.0 PV: 4.7 DV: 0.7
 00:18 SP: 4.0 PV: 4.2 DV: 0.2

2022-09-23

23:18 SP: 4.0 PV: 4.5 DV: 0.5
 22:18 SP: 4.0 PV: 4.7 DV: 0.7
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 4.3 DV: 0.3
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 4.6 DV: 0.6
 17:18 SP: 4.0 PV: 4.7 DV: 0.7
 16:18 SP: 4.0 PV: 4.1 DV: 0.1
 15:18 SP: 4.0 PV: 4.4 DV: 0.4
 14:18 SP: 4.0 PV: 4.5 DV: 0.5
 13:18 SP: 4.0 PV: 4.6 DV: 0.6
 12:18 SP: 4.0 PV: 4.7 DV: 0.7
 11:18 SP: 4.0 PV: 4.7 DV: 0.7
 10:18 SP: 4.0 PV: 4.8 DV: 0.8
 09:18 SP: 4.0 PV: 4.6 DV: 0.6
 08:18 SP: 4.0 PV: 4.4 DV: 0.4
 07:18 SP: 4.0 PV: 4.6 DV: 0.6
 06:18 SP: 4.0 PV: 4.4 DV: 0.4
 05:18 SP: 4.0 PV: 4.0 DV: 0.0
 04:18 SP: 4.0 PV: 4.4 DV: 0.4
 03:18 SP: 4.0 PV: 4.1 DV: 0.1
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 4.0 DV: 0.0
 00:18 SP: 4.0 PV: 4.1 DV: 0.1

2022-09-22

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:18 SP: 4.0 PV: 3.7 DV: -0.3
 22:18 SP: 4.0 PV: 3.6 DV: -0.4
 21:18 SP: 4.0 PV: 4.3 DV: 0.3
 20:18 SP: 4.0 PV: 3.8 DV: -0.2
 19:18 SP: 4.0 PV: 3.9 DV: -0.1
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.7 DV: -0.3
 15:18 SP: 4.0 PV: 4.0 DV: 0.0
 14:18 SP: 4.0 PV: 3.8 DV: -0.2
 13:18 SP: 4.0 PV: 4.0 DV: 0.0
 12:18 SP: 4.0 PV: 3.8 DV: -0.2
 11:18 SP: 4.0 PV: 3.6 DV: -0.4
 10:18 SP: 4.0 PV: 4.0 DV: 0.0
 09:18 SP: 4.0 PV: 3.8 DV: -0.2
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 3.6 DV: -0.4
 06:18 SP: 4.0 PV: 4.2 DV: 0.2
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 4.0 DV: 0.0
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 3.6 DV: -0.4
 01:18 SP: 4.0 PV: 4.3 DV: 0.3
 00:18 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-25

23:18 SP: 4.0 PV: 4.2 DV: 0.2
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 4.2 DV: 0.2
 20:18 SP: 4.0 PV: 4.4 DV: 0.4

0

19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 4.3 DV: 0.3
 17:18 SP: 4.0 PV: 4.1 DV: 0.1
 16:18 SP: 4.0 PV: 4.3 DV: 0.3
 15:18 SP: 4.0 PV: 4.3 DV: 0.3
 14:18 SP: 4.0 PV: 4.3 DV: 0.3
 13:18 SP: 4.0 PV: 4.3 DV: 0.3
 12:18 SP: 4.0 PV: 4.6 DV: 0.6
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 4.3 DV: 0.3
 09:18 SP: 4.0 PV: 4.0 DV: 0.0
 08:18 SP: 4.0 PV: 4.4 DV: 0.4
 07:18 SP: 4.0 PV: 4.1 DV: 0.1
 06:18 SP: 4.0 PV: 4.3 DV: 0.3
 05:18 SP: 4.0 PV: 3.8 DV: -0.2
 04:18 SP: 4.0 PV: 4.2 DV: 0.2
 03:18 SP: 4.0 PV: 3.7 DV: -0.3
 02:18 SP: 4.0 PV: 4.4 DV: 0.4
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 4.3 DV: 0.3

2022-09-24

23:18 SP: 4.0 PV: 4.3 DV: 0.3
 22:18 SP: 4.0 PV: 4.1 DV: 0.1
 21:18 SP: 4.0 PV: 4.4 DV: 0.4
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 4.2 DV: 0.2
 18:18 SP: 4.0 PV: 4.6 DV: 0.6
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 4.2 DV: 0.2
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 4.4 DV: 0.4
 13:18 SP: 4.0 PV: 4.1 DV: 0.1
 12:18 SP: 4.0 PV: 4.2 DV: 0.2
 11:18 SP: 4.0 PV: 4.6 DV: 0.6
 10:18 SP: 4.0 PV: 4.7 DV: 0.7
 09:18 SP: 4.0 PV: 3.8 DV: -0.2
 08:18 SP: 4.0 PV: 4.4 DV: 0.4
 07:18 SP: 4.0 PV: 4.2 DV: 0.2
 06:18 SP: 4.0 PV: 4.3 DV: 0.3
 05:18 SP: 4.0 PV: 4.3 DV: 0.3
 04:18 SP: 4.0 PV: 4.5 DV: 0.5
 03:18 SP: 4.0 PV: 4.1 DV: 0.1
 02:18 SP: 4.0 PV: 4.2 DV: 0.2
 01:18 SP: 4.0 PV: 4.3 DV: 0.3
 00:18 SP: 4.0 PV: 4.4 DV: 0.4

2022-09-27

23:18 SP: 4.0 PV: 4.2 DV: 0.2
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 4.4 DV: 0.4
 20:18 SP: 4.0 PV: 4.3 DV: 0.3
 19:18 SP: 4.0 PV: 4.4 DV: 0.4
 18:18 SP: 4.0 PV: 4.4 DV: 0.4
 17:18 SP: 4.0 PV: 4.4 DV: 0.4
 16:18 SP: 4.0 PV: 4.5 DV: 0.5
 15:18 SP: 4.0 PV: 4.7 DV: 0.7
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.3 DV: 0.3
 12:18 SP: 4.0 PV: 4.3 DV: 0.3
 11:18 SP: 4.0 PV: 4.6 DV: 0.6
 10:18 SP: 4.0 PV: 4.7 DV: 0.7
 09:18 SP: 4.0 PV: 3.8 DV: -0.2
 08:18 SP: 4.0 PV: 4.2 DV: 0.2
 07:18 SP: 4.0 PV: 4.5 DV: 0.5
 06:18 SP: 4.0 PV: 3.6 DV: -0.4
 05:18 SP: 4.0 PV: 3.6 DV: -0.4
 04:18 SP: 4.0 PV: 4.3 DV: 0.3
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 3.7 DV: -0.3

2022-09-26

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.9

23:18 SP: 4.0 PV: 4.3 DV: 0.3
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 4.5 DV: 0.5
 20:18 SP: 4.0 PV: 4.8 DV: 0.8
 19:18 SP: 4.0 PV: 3.8 DV: -0.2
 18:18 SP: 4.0 PV: 4.3 DV: 0.3
 17:18 SP: 4.0 PV: 4.0 DV: 0.0
 16:18 SP: 4.0 PV: 4.2 DV: 0.2
 15:18 SP: 4.0 PV: 4.4 DV: 0.4
 14:18 SP: 4.0 PV: 3.8 DV: -0.2
 13:18 SP: 4.0 PV: 4.2 DV: 0.2
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 4.3 DV: 0.3
 10:18 SP: 4.0 PV: 3.8 DV: -0.2
 09:18 SP: 4.0 PV: 4.1 DV: 0.1
 08:18 SP: 4.0 PV: 3.8 DV: -0.2
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 4.0 DV: 0.0
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 3.6 DV: -0.4
 03:18 SP: 4.0 PV: 4.4 DV: 0.4
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 4.1 DV: 0.1
 00:18 SP: 4.0 PV: 3.9 DV: -0.1

2022-09-29

23:18 SP: 4.0 PV: 4.6 DV: 0.6
 22:18 SP: 4.0 PV: 4.7 DV: 0.7
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 4.3 DV: 0.3
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 4.3 DV: 0.3
 17:18 SP: 4.0 PV: 4.4 DV: 0.4
 16:18 SP: 4.0 PV: 4.6 DV: 0.6
 15:18 SP: 4.0 PV: 3.8 DV: -0.2
 14:18 SP: 4.0 PV: 4.2 DV: 0.2
 13:18 SP: 4.0 PV: 4.2 DV: 0.2
 12:18 SP: 4.0 PV: 4.5 DV: 0.5
 11:18 SP: 4.0 PV: 4.7 DV: 0.7
 10:18 SP: 4.0 PV: 4.8 DV: 0.8
 09:18 SP: 4.0 PV: 3.8 DV: -0.2
 08:18 SP: 4.0 PV: 4.3 DV: 0.3
 07:18 SP: 4.0 PV: 4.7 DV: 0.7
 06:18 SP: 4.0 PV: 3.8 DV: -0.2
 05:18 SP: 4.0 PV: 4.2 DV: 0.2
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 4.5 DV: 0.5
 00:18 SP: 4.0 PV: 3.9 DV: -0.1

2022-09-28

23:18 SP: 4.0 PV: 4.5 DV: 0.5
 22:18 SP: 4.0 PV: 4.1 DV: 0.1
 21:18 SP: 4.0 PV: 4.2 DV: 0.2
 20:18 SP: 4.0 PV: 4.4 DV: 0.4
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 4.6 DV: 0.6
 16:18 SP: 4.0 PV: 4.7 DV: 0.7
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 4.3 DV: 0.3
 13:18 SP: 4.0 PV: 4.5 DV: 0.5
 12:18 SP: 4.0 PV: 4.6 DV: 0.6
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 4.3 DV: 0.3
 09:18 SP: 4.0 PV: 4.0 DV: 0.0
 08:18 SP: 4.0 PV: 4.0 DV: 0.0
 07:18 SP: 4.0 PV: 4.4 DV: 0.4
 06:18 SP: 4.0 PV: 4.0 DV: 0.0
 05:18 SP: 4.0 PV: 4.4 DV: 0.4
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 4.4 DV: 0.4
 00:18 SP: 4.0 PV: 4.1 DV: 0.1

2022-09-30

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

date of copying

sign