

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.08.01	9:30	Left	2	6	3	[Signature]	
		Right	3	5	4		
2022.08.02	9:30	Left	2	6	4	[Signature]	
		Right	3	4	3		
2022.08.03	10:00	Left	2	6	4	[Signature]	
		Right	3	4	3		
2022.08.04	9:30	Left	2	6	2	[Signature]	
		Right	2	4	3		
2022.08.05	9:30	Left	2	6	5	[Signature]	
		Right	3	4	4		
2022.08.06		Left					
		Right					
2022.08.07		Left					
		Right					
2022.08.08	9:30	Left	2	6	3	[Signature]	
		Right	2	4	3		
2022.08.09	9:30	Left	2	6	3	[Signature]	
		Right	3	4	3		
2022.08.10	9:30	Left	2	6	4	[Signature]	
		Right	3	4	4		
2022.08.11	9:45	Left	2	6	6	[Signature]	
		Right	3	4	4		
2022.08.12	9:30	Left	2	6	3	[Signature]	
		Right	2	4	3		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.08.13		Left					
		Right					
2022.08.14		Left					
		Right					
2022.08.15		Left					광복절
		Right					
2022.08.16	9:30	Left	2	6	4	김옥순	
		Right	3	5	4		
2022.08.17	9:30	Left	2	6	3	ml	
		Right	3	4	3		
2022.08.18	9:45	Left	2	6	4	ml	
		Right	3	4	4		
2022.08.19	9:30	Left	3	6	5	ml	
		Right	3	4	3		
2022.08.20		Left					
		Right					
2022.08.21		Left					
		Right					
2022.08.22	9:30	Left	2	6	6	ml	
		Right	2	4	4		
2022.08.23	9:30	Left	2	6	3	ml	
		Right	3	4	3		
2022.08.24	9:30	Left	2	6	3	ml	
		Right	2	4	3		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.08.25	P:30	Left	3	6	3	[Signature]	
		Right	2	4	3		
2022.08.26	P:40	Left	3	6	5	[Signature]	
		Right	3	4	3		
2022.08.27		Left					
		Right					
2022.08.28		Left					
		Right					
2022.08.29	P:30	Left	2	6	4	[Signature]	
		Right	2	5	4		
2022.08.30	P:30	Left	2	6	4	[Signature]	
		Right	2	5	4		
2022.08.31	P:30	Left	2	6	5	[Signature]	Gravity IP old
		Right	2	4	4		

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3

4

05:58 SP: 4.0 PV: 4.4 DV: 0.
 04:58 SP: 4.0 PV: 4.0 DV: 0.0
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-08-02

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.0 DV: 0.0

2

16:58 SP: 4.0 PV: 3.8 DV: -0.
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-08-01

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 4.7 DV: 0.7

7

00:58 SP: 4.0 PV: 4.7 DV: 0.

2022-08-04

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.3 DV: 0.3

3

13:58 SP: 4.0 PV: 3.7 DV: -0.
 12:58 SP: 4.0 PV: 4.3 DV: 0.4
 11:58 SP: 4.0 PV: 4.3 DV: 0.2
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 3.7 DV: -0.2
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-08-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-08-06

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-08-05

SP Setting Point
 PV Present Value
 DV Difference Value

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.5 DV: 0.6
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.1 DV: 0.1
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.8 DV: 0.8

7

01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.7 DV: 0.7

2022-08-08

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.7 DV: 0.7

6

19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.7 DV: 0.7
 13:58 SP: 4.0 PV: 4.2 DV: 0.2

4

12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 4.5 DV: 0.5

2

08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 3.8 DV: -0.2
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 3.8 DV: -0.2

4

00:58 SP: 4.0 PV: 4.4 DV: 0.4
 2022-08-07

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-08-10

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.0 DV: 0.0
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-08-09

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.5 DV: 0.5
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.7 DV: 0.7

2022-08-12

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.1 DV: 0.1

5

19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-08-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 3.8 DV: -0.2
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.0 DV: 0.0
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 .5
 16:58 SP: 4.0 PV: 4.5 DV: 0
 15:58 SP: 4.0 PV: 4.1 DV: 0.1
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 4.0 DV: 0.0
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.9 DV: -0.1
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.5 DV: 0.5
 2022-08-14
 23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 1
 21:58 SP: 4.0 PV: 3.9 DV: -0.
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 3
 07:58 SP: 4.0 PV: 3.7 DV: -0.
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.4 DV: 0.4
 2022-08-13

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.1 DV: 0.1
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 4.3 DV: 0.3
 2022-08-16
 23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 4
 06:58 SP: 4.0 PV: 4.4 DV: 0.
 05:58 SP: 4.0 PV: 3.6 DV: -0.4
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 4.5 DV: 0.5
 2022-08-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.1 DV: 0.1
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3

3

01:58 SP: 4.0 PV: 4.3 DV: 0.
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-08-18

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.0 DV: 0.0
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-08-17

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.0 DV: 0.0
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.5 DV: 0.5

3

09:58 SP: 4.0 PV: 3.7 DV: -0.
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-08-20

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.8 DV: -0.2
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 4.1 DV: 0.1
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.1 DV: 0.1
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 3.9 DV: -0.1
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-08-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.9 DV: -0.1
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-08-22

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 4.0 DV: 0.0
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 3.9 DV: -0.1
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.8 DV: -0.2
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 4.6 DV: 0.6

2

01:58 SP: 4.0 PV: 3.8 DV: -0.1
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-08-21

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-08-24

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 3.8 DV: -0.2

6

06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-08-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.6 DV: -0.4
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.0 DV: 0.0
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.2 DV: 0.2
 2022-08-26

3

23:58 SP: 4.0 PV: 3.7 DV: -0.
 22:58 SP: 4.0 PV: 4.4 DV: 0.4

1

21:58 SP: 4.0 PV: 4.1 DV: 0.
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.2 DV: 0.2

6

15:58 SP: 4.0 PV: 4.6 DV: 0.
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 3.8 DV: -0.2
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.7 DV: -0.3

4

05:58 SP: 4.0 PV: 4.4 DV: 0.
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 3.8 DV: -0.2
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 3.9 DV: -0.1
 2022-08-25

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 3.7 DV: -0.3

5

18:58 SP: 4.0 PV: 4.5 DV: 0.
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.1 DV: 0.1
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3
 2022-08-28

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.7 DV: 0.7
 16:58 SP: 4.0 PV: 4.7 DV: 0.7
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 3.7 DV: -0.3
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 3.6 DV: -0.4
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.0 DV: 0.0
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.0 DV: 0.0
 2022-08-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.8

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 4.0 DV: 0.0
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 4.2 DV: 0.2
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.2 DV: 0.2

3

09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.6 DV: -0.4
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.7 DV: -0.3
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-08-30

23:58 SP: 4.0 PV: 3.9 DV: -0.1
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.1 DV: 0.1
 15:58 SP: 4.0 PV: 3.8 DV: -0.2
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 3.7 DV: -0.3
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 3.9 DV: -0.1
 10:58 SP: 4.0 PV: 4.3 DV: 0.3
 09:58 SP: 4.0 PV: 3.7 DV: -0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 4.1 DV: 0.1
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.1 DV: 0.1
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 3.6 DV: -0.4
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-08-29

23:58 SP: 4.0 PV: 4.1 DV: 0.1
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.8 DV: -0.2
 18:58 SP: 4.0 PV: 4.5 DV: 0.5
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 3.8 DV: -0.2
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 3.7 DV: -0.3

4

08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 3.6 DV: -0.4
 02:58 SP: 4.0 PV: 4.5 DV: 0.5

0

01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-08-31

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

date of copying

sign