

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.7
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.07.01	P:30	Left	2	7	4	[Signature]	7월 1일 2회
		Right	2	6	3		
2022.07.02		Left					
		Right					
2022.07.03		Left					
		Right					
2022.07.04	P:30	Left	2	8	2	[Signature]	
		Right	2	7	2		
2022.07.05	P:30	Left	2	7	3	[Signature]	
		Right	2	6	2		
2022.07.06	P:30	Left	2	7	5	[Signature]	
		Right	2	6	4		
2022.07.07	P:40	Left	2	8	5	[Signature]	
		Right	2	6	4		
2022.07.08	P:30	Left	2	7	4	[Signature]	
		Right	2	6	2		
2022.07.09		Left					
		Right					
2022.07.10		Left					
		Right					
2022.07.11	P:30	Left	2	7	3	[Signature]	CTC118 IP 1회
		Right	2	7	3		
2022.07.12	P:30	Left	2	6	4	[Signature]	
		Right	2	6	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.7
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.07.13	p:30	Left	2	6	3	ml	
		Right	2	6	2		
2022.07.14	p:30	Left	2	6	4	ml	
		Right	2	6	2		
2022.07.15	p:30	Left	2	7	4	ml	
		Right	2	7	5		
2022.07.16		Left					
		Right					
2022.07.17		Left					
		Right					
2022.07.18	p:30	Left	2	6	2	ml	
		Right	3	5	3		
2022.07.19	p:40	Left	2	7	2	ml	
		Right	2	6	3		
2022.07.20	p:30	Left	2	6	3	ml	
		Right	3	4	3		
2022.07.21	p:45	Left	2	6	3	ml	
		Right	3	4	3		
2022.07.22	p:30	Left	2	6	4	ml	
		Right	3	4	3		
2022.07.23		Left					
		Right					
2022.07.24		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.7
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.07.25	9:30	Left	2	6	4	김종민	
		Right	2	6	5		
2022.07.26	9:30	Left	2	6	3	안	
		Right	3	4	4		
2022.07.27	9:20	Left	2	6	5	안	
		Right	3	4	4		
2022.07.28	9:30	Left	2	6	3	안	
		Right	3	6	3		
2022.07.29	9:30	Left	2	6	3	안	
		Right	3	4	3		
2022.07.30		Left					
		Right					
2022.07.31		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.9 DV: 0.9
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 4.7 DV: 0.7
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.7 DV: 0.7
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.7 DV: 0.7
 02:58 SP: 4.0 PV: 4.9 DV: 0.9
 01:58 SP: 4.0 PV: 4.9 DV: 0.9
 00:58 SP: 4.0 PV: 4.9 DV: 0.9

2022-07-02

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.3 DV: 0.3
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.7 DV: 0.7
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.1 DV: 0.4
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.7 DV: 0.7

7

07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.5 DV: 0.5
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-07-01

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 2

21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 3.8 DV: -0.2
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.8 DV: 0.8
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 4.0 DV: 0.0
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-07-04

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 4.7 DV: 0.7
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.8 DV: 0.8
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 3.7 DV: -0.3
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 3.9 DV: -0.1
 00:58 SP: 4.0 PV: 4.0 DV: 0.0

2022-07-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 4.7 DV: 0.7
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.4 DV: 0.4
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.8 DV: 0.8

8

10:58 SP: 4.0 PV: 4.8 DV: 0.
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.1 DV: 0.1

2022-07-06

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 3.9 DV: -0.1
 16:58 SP: 4.0 PV: 4.4 DV: 0.4

6

15:58 SP: 4.0 PV: 4.6 DV: 0.
 14:58 SP: 4.0 PV: 4.7 DV: 0.7
 13:58 SP: 4.0 PV: 4.7 DV: 0.2
 12:58 SP: 4.0 PV: 4.5 DV: 0.5
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 4.7 DV: 0.7
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.5 DV: 0.5
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-07-05

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 3.7 DV: -0.1
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 4.6 DV: 0.6
 17:58 SP: 4.0 PV: 4.7 DV: 0.7
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.6 DV: 0.6
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.3 DV: 0.8
 09:58 SP: 4.0 PV: 4.2 DV: 0.2

6

08:58 SP: 4.0 PV: 4.6 DV: 0.
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.3 DV: 0.3
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.8 DV: 0.8

8

03:58 SP: 4.0 PV: 4.8 DV: 0.
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 3.8 DV: -0.2
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-07-08

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.8 DV: 0.8
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 4.0 DV: 0.0
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 4.8 DV: 0.8
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.9 DV: 0.9
 13:58 SP: 4.0 PV: 4.8 DV: 0.8
 12:58 SP: 4.0 PV: 4.8 DV: 0.8
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.2 DV: 0.2
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.8 DV: 0.8
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 4.4 DV: 0.4

2022-07-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 3.9 DV: -0.1
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.3 DV: 0.3
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.1 DV: 0.1
 16:58 SP: 4.0 PV: 4.0 DV: -0.4
 15:58 SP: 4.0 PV: 4.0 DV: 0.6
 14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.0 DV: 0.0
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 4.9 DV: 0.9
 10:58 SP: 4.0 PV: 4.9 DV: 0.9
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.7 DV: 0.7
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 4.5 DV: 0.5

2022-07-10

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.8 DV: 0.8
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.8 DV: 0.8
 18:58 SP: 4.0 PV: 4.8 DV: 0.8
 17:58 SP: 4.0 PV: 4.8 DV: 0.8
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.7 DV: 0.7
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.1 DV: 0.6
 08:58 SP: 4.0 PV: 4.2 DV: 0.7
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.2 DV: 0.3
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.8 DV: 0.8
 00:58 SP: 4.0 PV: 4.8 DV: 0.8

2022-07-09

23:58 SP: 4.0 PV: 4.8 DV: 0.8
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 4.8 DV: 0.8
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.4 DV: 0.4
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.7 DV: 0.7
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.3 DV: 0.3
 02:58 SP: 4.0 PV: 4.6 DV: 0.6
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 4.2 DV: 0.2

2022-07-12

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 4.5 DV: 0.5
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 4.8 DV: 0.8
 17:58 SP: 4.0 PV: 4.8 DV: 0.8
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.8 DV: 0.8
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.8 DV: 0.8
 12:58 SP: 4.0 PV: 4.0 DV: -0.1
 11:58 SP: 4.0 PV: 4.1 DV: 0.6
 10:58 SP: 4.0 PV: 4.0 DV: 0.8
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.3 DV: 0.3
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 3.8 DV: -0.2
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.3 DV: 0.3
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-07-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 4.1 DV: 0.1
 21:58 SP: 4.0 PV: 4.8 DV: 0.8
 20:58 SP: 4.0 PV: 4.1 DV: 0.1
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.7 DV: 0.7
 17:58 SP: 4.0 PV: 4.8 DV: 0.8
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.9 DV: 0.9

14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.1 DV: 0.1
 12:58 SP: 4.0 PV: 4.4 DV: 0.4
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 4.8 DV: 0.8
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.5 DV: 0.5
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 4.8 DV: 0.8
 04:58 SP: 4.0 PV: 4.8 DV: 0.8
 03:58 SP: 4.0 PV: 4.1 DV: 0.1
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.3 DV: 0.3

2022-07-14

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.8 DV: 0.8
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.8 DV: 0.8
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.8 DV: 0.8

12:58 SP: 4.0 PV: 4.8 DV: 0.8
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.4 DV: 0.4
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 4.1 DV: 0.1
 04:58 SP: 4.0 PV: 4.4 DV: 0.4
 03:58 SP: 4.0 PV: 4.6 DV: 0.6
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 4.8 DV: 0.8
 00:58 SP: 4.0 PV: 4.8 DV: 0.8

2022-07-13

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.2 DV: 0.3
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.7 DV: 0.7
 07:58 SP: 4.0 PV: 3.8 DV: -0.2
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.6 DV: 0.6
 04:58 SP: 4.0 PV: 4.8 DV: 0.8
 03:58 SP: 4.0 PV: 4.8 DV: 0.8

9

02:58 SP: 4.0 PV: 4.9 DV: 0.9
 01:58 SP: 4.0 PV: 4.9 DV: 0.9
 00:58 SP: 4.0 PV: 4.9 DV: 0.9

2022-07-16

23:58 SP: 4.0 PV: 4.8 DV: 0.8
 22:58 SP: 4.0 PV: 4.8 DV: 0.8
 21:58 SP: 4.0 PV: 4.0 DV: 0.0
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.6 DV: 0.6
 18:58 SP: 4.0 PV: 4.8 DV: 0.8
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.7 DV: 0.7
 13:58 SP: 4.0 PV: 4.8 DV: 0.8
 12:58 SP: 4.0 PV: 4.8 DV: 0.8
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.8 DV: 0.8
 09:58 SP: 4.0 PV: 3.6 DV: -0.4
 08:58 SP: 4.0 PV: 3.6 DV: -0.2
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 3.8 DV: -0.2
 02:58 SP: 4.0 PV: 3.6 DV: -0.4
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 3.6 DV: -0.4

2022-07-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 4.9 DV: 0.9
 22:58 SP: 4.0 PV: 4.2 DV: 0.3
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.7 DV: 0.7
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.2 DV: 0.2
 15:58 SP: 4.0 PV: 4.7 DV: 0.7
 14:58 SP: 4.0 PV: 4.5 DV: 0.5
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 4.2 DV: 0.2
 11:58 SP: 4.0 PV: 4.1 DV: 0.1

10:58 SP: 4.0 PV: 4.7 DV: 0.7
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.8 DV: 0.8

07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.2 DV: 0.2
 05:58 SP: 4.0 PV: 4.8 DV: 0.8
 2022-07-17

23:58 SP: 4.0 PV: 4.7 DV: 0.7
 22:58 SP: 4.0 PV: 4.2 DV: 0.2
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.8 DV: 0.8
 19:58 SP: 4.0 PV: 4.8 DV: 0.8
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 4.7 DV: 0.7
 12:58 SP: 4.0 PV: 4.1 DV: 0.1
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 4.6 DV: 0.6
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.1 DV: 0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 4.8 DV: 0.8
 05:58 SP: 4.0 PV: 4.7 DV: 0.7
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.7 DV: 0.7
 01:58 SP: 4.0 PV: 4.7 DV: 0.7
 00:58 SP: 4.0 PV: 3.9 DV: -0.1

2022-07-20

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 4.7 DV: 0.7
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 4.8 DV: 0.8
 18:58 SP: 4.0 PV: 4.2 DV: 0.2
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.8 DV: 0.8
 15:58 SP: 4.0 PV: 4.8 DV: 0.8

1

14:58 SP: 4.0 PV: 3.9 DV: -0.1
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.8 DV: 0.8
 09:58 SP: 4.0 PV: 4.8 DV: 0.8
 08:58 SP: 4.0 PV: 4.8 DV: 0.8
 07:58 SP: 4.0 PV: 4.8 DV: 0.8
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 05:58 SP: 4.0 PV: 4.8 DV: 0.8
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 4.9 DV: 0.9

2022-07-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.3 DV: 0.3
 21:58 SP: 4.0 PV: 4.1 DV: 0.1
 20:58 SP: 4.0 PV: 4.2 DV: 0.2
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 4.0 DV: 0.0
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.7 DV: 0.7
 11:58 SP: 4.0 PV: 4.2 DV: 0.2
 10:58 SP: 4.0 PV: 4.5 DV: 0.5
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.6 DV: 0.6
 07:58 SP: 4.0 PV: 4.7 DV: 0.7
 06:58 SP: 4.0 PV: 4.0 DV: 0.0
 5
 05:58 SP: 4.0 PV: 4.5 DV: 0.
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 2
 03:58 SP: 4.0 PV: 3.8 DV: -0.
 02:58 SP: 4.0 PV: 4.4 DV: 0.4
 7
 01:58 SP: 4.0 PV: 4.7 DV: 0.
 00:58 SP: 4.0 PV: 4.7 DV: 0.7
 2022-07-22
 23:58 SP: 4.0 PV: 4.0 DV: 0.0
 22:58 SP: 4.0 PV: 4.5 DV: 0.5
 21:58 SP: 4.0 PV: 4.7 DV: 0.7
 20:58 SP: 4.0 PV: 4.7 DV: 0.7
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 3
 18:58 SP: 4.0 PV: 4.3 DV: 0.
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.3 DV: 0.3
 15:58 SP: 4.0 PV: 4.6 DV: 0.6
 14:58 SP: 4.0 PV: 4.8 DV: 0.8
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 4.6 DV: 0.6
 11:58 SP: 4.0 PV: 4.8 DV: 0.8
 10:58 SP: 4.0 PV: 4.8 DV: 0.8
 09:58 SP: 4.0 PV: 4.7 DV: 0.7
 08:58 SP: 4.0 PV: 4.4 DV: 0.4
 07:58 SP: 4.0 PV: 3.9 DV: -0.1
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 4.6 DV: 0.6
 03:58 SP: 4.0 PV: 4.8 DV: 0.8
 02:58 SP: 4.0 PV: 4.8 DV: 0.8
 01:58 SP: 4.0 PV: 4.1 DV: 0.1
 00:58 SP: 4.0 PV: 4.4 DV: 0.4
 2022-07-21

23:58 SP: 4.0 PV: 4.4 DV: 0.4
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.5 DV: 0.5
 20:58 SP: 4.0 PV: 3.9 DV: -0.1
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.6 DV: 0.6
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.2 DV: 0.2
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 0
 00:58 SP: 4.0 PV: 4.0 DV: 0.
 2022-07-24
 23:58 SP: 4.0 PV: 4.2 DV: 0.2
 3
 22:58 SP: 4.0 PV: 3.7 DV: -0.
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.4 DV: 0.4
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.5 DV: 0.5
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.9 DV: -0.1
 09:58 SP: 4.0 PV: 4.2 DV: 0.2
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 4.6 DV: 0.6
 05:58 SP: 4.0 PV: 4.8 DV: 0.8
 04:58 SP: 4.0 PV: 4.7 DV: 0.7
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 4.7 DV: 0.7
 2022-07-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 4.2 DV: 0.2
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.6 DV: 0.6
 19:58 SP: 4.0 PV: 3.9 DV: -0.1
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 3.9 DV: -0.1
 15:58 SP: 4.0 PV: 4.2 DV: 0.2
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.3 DV: 0.3
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.6 DV: 0.6
 06:58 SP: 4.0 PV: 3.8 DV: -0.2
 05:58 SP: 4.0 PV: 4.5 DV: 0.5
 04:58 SP: 4.0 PV: 3.9 DV: -0.1
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 4.2 DV: 0.2
 01:58 SP: 4.0 PV: 4.2 DV: 0.2
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-07-26

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 3.7 DV: -0.3
 21:58 SP: 4.0 PV: 4.4 DV: 0.4
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.5 DV: 0.5
 18:58 SP: 4.0 PV: 3.8 DV: -0.2
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.3 DV: 0.3
 10:58 SP: 4.0 PV: 3.7 DV: -0.3

3

09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4
 00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-07-25

23:58 SP: 4.0 PV: 3.7 DV: -0.3
 22:58 SP: 4.0 PV: 4.4 DV: 0.4
 21:58 SP: 4.0 PV: 3.7 DV: -0.3
 20:58 SP: 4.0 PV: 4.4 DV: 0.4
 19:58 SP: 4.0 PV: 4.1 DV: 0.1
 18:58 SP: 4.0 PV: 3.9 DV: -0.1
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 3.7 DV: -0.3
 07:58 SP: 4.0 PV: 4.4 DV: 0.4
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.0 DV: 0.0
 03:58 SP: 4.0 PV: 4.2 DV: 0.2
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.4 DV: 0.4

3

00:58 SP: 4.0 PV: 3.7 DV: -0.3

2022-07-26

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.9 DV: -0.1
 21:58 SP: 4.0 PV: 4.2 DV: 0.2
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 3.7 DV: -0.3
 17:58 SP: 4.0 PV: 4.3 DV: 0.3
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.4 DV: 0.4
 14:58 SP: 4.0 PV: 3.7 DV: -0.3
 13:58 SP: 4.0 PV: 4.5 DV: 0.5
 12:58 SP: 4.0 PV: 3.9 DV: -0.1
 11:58 SP: 4.0 PV: 4.6 DV: 0.6
 10:58 SP: 4.0 PV: 3.8 DV: -0.2
 09:58 SP: 4.0 PV: 4.5 DV: 0.5
 08:58 SP: 4.0 PV: 3.9 DV: -0.1
 07:58 SP: 4.0 PV: 4.5 DV: 0.5
 06:58 SP: 4.0 PV: 3.9 DV: -0.1
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 4.3 DV: 0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 4.6 DV: 0.6

2022-07-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.7

23:58 SP: 4.0 PV: 4.6 DV: 0.6
 22:58 SP: 4.0 PV: 4.6 DV: 0.6
 21:58 SP: 4.0 PV: 4.6 DV: 0.6
 20:58 SP: 4.0 PV: 4.7 DV: 0.7
 19:58 SP: 4.0 PV: 4.2 DV: 0.2
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 4.6 DV: 0.6
 16:58 SP: 4.0 PV: 4.6 DV: 0.6
 15:58 SP: 4.0 PV: 4.7 DV: 0.7
 14:58 SP: 4.0 PV: 3.8 DV: -0.2
 13:58 SP: 4.0 PV: 4.4 DV: 0.4
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.2 DV: 0.2

3

10:58 SP: 4.0 PV: 3.7 DV: -0.1
 09:58 SP: 4.0 PV: 4.3 DV: 0.3
 08:58 SP: 4.0 PV: 4.0 DV: 0.0
 07:58 SP: 4.0 PV: 4.2 DV: 0.2
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.4 DV: 0.4
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.6 DV: 0.6
 00:58 SP: 4.0 PV: 3.8 DV: -0.2

2022-07-30

23:58 SP: 4.0 PV: 4.5 DV: 0.5
 22:58 SP: 4.0 PV: 3.8 DV: -0.2
 21:58 SP: 4.0 PV: 4.3 DV: 0.3
 20:58 SP: 4.0 PV: 3.7 DV: -0.3
 19:58 SP: 4.0 PV: 4.4 DV: 0.4
 18:58 SP: 4.0 PV: 4.0 DV: 0.0
 17:58 SP: 4.0 PV: 4.2 DV: 0.2
 16:58 SP: 4.0 PV: 3.7 DV: -0.3
 15:58 SP: 4.0 PV: 4.3 DV: 0.3
 14:58 SP: 4.0 PV: 4.6 DV: 0.6
 13:58 SP: 4.0 PV: 3.9 DV: -0.1
 12:58 SP: 4.0 PV: 4.3 DV: 0.3
 11:58 SP: 4.0 PV: 4.5 DV: 0.5
 10:58 SP: 4.0 PV: 3.7 DV: -0.3
 09:58 SP: 4.0 PV: 4.4 DV: 0.4
 08:58 SP: 4.0 PV: 4.2 DV: 0.2
 07:58 SP: 4.0 PV: 3.7 DV: -0.3
 06:58 SP: 4.0 PV: 4.4 DV: 0.4
 05:58 SP: 4.0 PV: 3.7 DV: -0.3
 04:58 SP: 4.0 PV: 4.5 DV: 0.5
 03:58 SP: 4.0 PV: 3.9 DV: -0.1
 02:58 SP: 4.0 PV: 4.3 DV: 0.3
 01:58 SP: 4.0 PV: 3.7 DV: -0.3

4

00:58 SP: 4.0 PV: 4.4 DV: 0.1
 2022-07-29

23:58 SP: 4.0 PV: 4.3 DV: 0.3
 22:58 SP: 4.0 PV: 4.1 DV: 0.1
 21:58 SP: 4.0 PV: 3.8 DV: -0.2
 20:58 SP: 4.0 PV: 4.3 DV: 0.3
 19:58 SP: 4.0 PV: 3.7 DV: -0.3
 18:58 SP: 4.0 PV: 4.4 DV: 0.4
 17:58 SP: 4.0 PV: 3.7 DV: -0.3
 16:58 SP: 4.0 PV: 4.5 DV: 0.5
 15:58 SP: 4.0 PV: 3.7 DV: -0.3
 14:58 SP: 4.0 PV: 4.3 DV: 0.3
 13:58 SP: 4.0 PV: 4.2 DV: 0.2
 12:58 SP: 4.0 PV: 3.7 DV: -0.3
 11:58 SP: 4.0 PV: 4.4 DV: 0.4
 10:58 SP: 4.0 PV: 4.0 DV: 0.0
 09:58 SP: 4.0 PV: 3.8 DV: -0.2
 08:58 SP: 4.0 PV: 4.3 DV: 0.3
 07:58 SP: 4.0 PV: 4.0 DV: 0.0
 06:58 SP: 4.0 PV: 3.7 DV: -0.3
 05:58 SP: 4.0 PV: 4.3 DV: 0.3
 04:58 SP: 4.0 PV: 3.7 DV: -0.3
 03:58 SP: 4.0 PV: 4.4 DV: 0.4
 02:58 SP: 4.0 PV: 3.7 DV: -0.3
 01:58 SP: 4.0 PV: 4.5 DV: 0.5
 00:58 SP: 4.0 PV: 3.8 DV: -0.2
 2022-07-31

SP Setting Point
 PV Present Value
 DV Difference Value