

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.12
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.12.01	p:40	Left	2	6	3	ml	
		Right	2	4	3		
2022.12.02	p:35	Left	3	7	5	ml	
		Right	3	4	4		
2022.12.03		Left					
		Right					
2022.12.04		Left					
		Right					
2022.12.05	p:30	Left	2	7	5	ml	
		Right	3	4	4		
2022.12.06	p:30	Left	2	7	5	ml	
		Right	3	4	4		
2022.12.07	p:30	Left	2	6	5	ml	
		Right	3	4	4		
2022.12.08	p:30	Left	3	6	5	ml	
		Right	3	4	4		
2022.12.09	p:30	Left	3	6	5	ml	
		Right	3	4	4		
2022.12.10		Left					
		Right					
2022.12.11		Left					
		Right					
2022.12.12	p:30	Left	2	7	5	ml	
		Right	3	4	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.12
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8		
			Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2022.12.13	9:30	Left	3	6	5	[Signature]	
		Right	3	4	4		
2022.12.14	9:30	Left	2	6	5	[Signature]	
		Right	3	4	3		
2022.12.15	9:30	Left	3	6	5	[Signature]	
		Right	3	4	3		
2022.12.16	10:50	Left	3	6	6	[Signature]	
		Right	3	4	4		
2022.12.17		Left					
		Right					
2022.12.18		Left					
		Right					
2022.12.19	9:40	Left	2	7	4	[Signature]	
		Right	3	4	4		
2022.12.20	9:20	Left	2	7	6	[Signature]	
		Right	3	4	3		
2022.12.21	9:30	Left	2	5	4	[Signature]	
		Right	3	5	5		
2022.12.22	9:30	Left	4	5	4	[Signature]	
		Right	3	5	4		
2022.12.23	9:35	Left	3	4	4	[Signature]	
		Right	3	4	4		
2022.12.24		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2022.12
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8°C

Date	Time	Probe Position	Temperature(°C)			Verified by	Comment
			Min	Max	Current		
2022.12.25		Left					
		Right					
2022.12.26	P:30	Left	3	4	4	anf	
		Right	3	5	4		
2022.12.27	P:30	Left	3	4	4	anf	
		Right	3	5	4		
2022.12.28	P:30	Left	3	4	4	anf	
		Right	3	5	4		
2022.12.29	P:30	Left	3	4	4	anf	
		Right	3	5	4		
2022.12.30	P:30	Left	3	4	3	anf	
		Right	3	5	4		
2022.12.31		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 4.0 DV: 0.0
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 3.7 DV: -0.3
 17:18 SP: 4.0 PV: 4.2 DV: 0.2
 3
 16:18 SP: 4.0 PV: 3.7 DV: -0.1
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 4.3 DV: 0.3
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 2
 12:18 SP: 4.0 PV: 4.2 DV: 0.0
 11:18 SP: 4.0 PV: 3.9 DV: -0.1
 10:18 SP: 4.0 PV: 4.1 DV: 0.1
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 4.2 DV: 0.2
 07:18 SP: 4.0 PV: 3.9 DV: -0.1
 06:18 SP: 4.0 PV: 4.0 DV: 0.0
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 4.2 DV: 0.2
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.2 DV: 0.2
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 4.0 DV: 0.0
 2022-12-02
 23:18 SP: 4.0 PV: 4.0 DV: 0.0
 22:18 SP: 4.0 PV: 4.0 DV: 0.0
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 4.1 DV: 0.1
 19:18 SP: 4.0 PV: 4.3 DV: 0.3
 18:18 SP: 4.0 PV: 3.8 DV: -0.2
 17:18 SP: 4.0 PV: 4.2 DV: 0.2
 16:18 SP: 4.0 PV: 4.0 DV: 0.0
 15:18 SP: 4.0 PV: 3.7 DV: -0.3
 14:18 SP: 4.0 PV: 4.3 DV: -0.3
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 12:18 SP: 4.0 PV: 4.2 DV: 0.2
 11:18 SP: 4.0 PV: 3.7 DV: -0.3
 10:18 SP: 4.0 PV: 3.8 DV: -0.2
 09:18 SP: 4.0 PV: 4.4 DV: 0.4
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 4.3 DV: 0.3
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 3.9 DV: -0.1
 03:18 SP: 4.0 PV: 4.3 DV: 0.3
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 4.4 DV: 0.4
 2022-12-01

23:18 SP: 4.0 PV: 4.0 DV: 0.0
 22:18 SP: 4.0 PV: 4.0 DV: 0.0
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 4.0 DV: 0.0
 1
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 4.1 DV: 0.1
 15:18 SP: 4.0 PV: 3.8 DV: -0.2
 14:18 SP: 4.0 PV: 4.1 DV: 0.1
 13:18 SP: 4.0 PV: 3.9 DV: -0.1
 12:18 SP: 4.0 PV: 3.9 DV: -0.1
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 3.7 DV: -0.3
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 3.8 DV: -0.2
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 3.7 DV: -0.3
 03:18 SP: 4.0 PV: 4.0 DV: 0.0
 1
 02:18 SP: 4.0 PV: 3.9 DV: -0.1
 01:18 SP: 4.0 PV: 3.7 DV: -0.3
 00:18 SP: 4.0 PV: 4.0 DV: 0.0
 2022-12-04
 23:18 SP: 4.0 PV: 3.8 DV: -0.2
 22:18 SP: 4.0 PV: 3.7 DV: -0.3
 21:18 SP: 4.0 PV: 4.2 DV: 0.2
 20:18 SP: 4.0 PV: 3.8 DV: -0.2
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 3.7 DV: -0.3
 17:18 SP: 4.0 PV: 4.1 DV: 0.1
 16:18 SP: 4.0 PV: 3.9 DV: -0.1
 15:18 SP: 4.0 PV: 4.0 DV: 0.0
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.0 DV: 0.0
 12:18 SP: 4.0 PV: 3.9 DV: -0.1
 11:18 SP: 4.0 PV: 3.9 DV: -0.1
 10:18 SP: 4.0 PV: 3.9 DV: -0.1
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 4.1 DV: 0.1
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 4.1 DV: 0.1
 04:18 SP: 4.0 PV: 3.8 DV: -0.2
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 4.0 DV: 0.0
 00:18 SP: 4.0 PV: 3.8 DV: -0.2
 2022-12-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 3.9 DV: -0.1
 22:18 SP: 4.0 PV: 3.8 DV: -0.2
 21:18 SP: 4.0 PV: 3.8 DV: -0.2
 20:18 SP: 4.0 PV: 3.8 DV: -0.2
 19:18 SP: 4.0 PV: 3.9 DV: -0.1
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.8 DV: -0.2
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 3.8 DV: -0.2
 13:18 SP: 4.0 PV: 3.7 DV: -0.3
 12:18 SP: 4.0 PV: 4.1 DV: 0.1
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 3.7 DV: -0.3
 09:18 SP: 4.0 PV: 4.2 DV: 0.2
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 4.4 DV: 0.4
 06:18 SP: 4.0 PV: 3.8 DV: -0.2
 05:18 SP: 4.0 PV: 4.2 DV: 0.2
 04:18 SP: 4.0 PV: 3.7 DV: -0.3
 03:18 SP: 4.0 PV: 4.2 DV: 0.2
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 4.3 DV: 0.3
 00:18 SP: 4.0 PV: 3.9 DV: -0.1
 2022-12-06
 23:18 SP: 4.0 PV: 4.1 DV: 0.1
 22:18 SP: 4.0 PV: 4.0 DV: 0.0
 21:18 SP: 4.0 PV: 4.1 DV: 0.1
 20:18 SP: 4.0 PV: 4.0 DV: 0.0
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 3.7 DV: -0.3
 17:18 SP: 4.0 PV: 4.3 DV: 0.3
 16:18 SP: 4.0 PV: 3.8 DV: -0.2
 0
 15:18 SP: 4.0 PV: 4.0 DV: 0.
 14:18 SP: 4.0 PV: 3.7 DV: -0.3
 13:18 SP: 4.0 PV: 4.3 DV: 0.3
 12:18 SP: 4.0 PV: 3.8 DV: -0.2
 4
 11:18 SP: 4.0 PV: 4.4 DV: 0.
 10:18 SP: 4.0 PV: 3.9 DV: -0.1
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 4.2 DV: 0.2
 3
 07:18 SP: 4.0 PV: 3.7 DV: -0.
 06:18 SP: 4.0 PV: 4.2 DV: 0.2
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 2
 04:18 SP: 4.0 PV: 4.2 DV: 0.
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 4.1 DV: 0.1
 2022-12-05

23:18 SP: 4.0 PV: 3.7 DV: -0.3
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 3.7 DV: -0.3
 20:18 SP: 4.0 PV: 4.2 DV: 0.2
 19:18 SP: 4.0 PV: 3.7 DV: -0.3
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 4.0 DV: 0.0
 16:18 SP: 4.0 PV: 3.8 DV: -0.2
 3
 15:18 SP: 4.0 PV: 4.3 DV: 0.
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.2 DV: 0.2
 12:18 SP: 4.0 PV: 4.1 DV: 0.1
 11:18 SP: 4.0 PV: 3.9 DV: -0.1
 10:18 SP: 4.0 PV: 4.3 DV: 0.3
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 3.8 DV: -0.2
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 4.2 DV: 0.2
 05:18 SP: 4.0 PV: 3.8 DV: -0.2
 04:18 SP: 4.0 PV: 4.2 DV: 0.2
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 4.0 DV: 0.0
 00:18 SP: 4.0 PV: 3.7 DV: -0.3
 2022-12-08
 23:18 SP: 4.0 PV: 4.1 DV: 0.1
 22:18 SP: 4.0 PV: 3.8 DV: -0.2
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 3.7 DV: -0.3
 19:18 SP: 4.0 PV: 3.8 DV: -0.2
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.7 DV: -0.3
 15:18 SP: 4.0 PV: 4.2 DV: 0.2
 14:18 SP: 4.0 PV: 3.8 DV: -0.2
 13:18 SP: 4.0 PV: 4.1 DV: 0.1
 12:18 SP: 4.0 PV: 3.8 DV: -0.2
 11:18 SP: 4.0 PV: 4.2 DV: 0.2
 10:18 SP: 4.0 PV: 3.8 DV: -0.2
 1
 09:18 SP: 4.0 PV: 3.9 DV: -0.
 08:18 SP: 4.0 PV: 3.8 DV: -0.2
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 3.8 DV: -0.2
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 3.7 DV: -0.3
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 3.8 DV: -0.2
 2022-12-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 3.9 DV: -0.1
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 3.8 DV: -0.2
 20:18 SP: 4.0 PV: 4.0 DV: 0.0
 19:18 SP: 4.0 PV: 3.8 DV: -0.2
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.8 DV: -0.2
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 3.9 DV: -0.1
 12:18 SP: 4.0 PV: 3.9 DV: -0.1
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 3.7 DV: -0.3
 09:18 SP: 4.0 PV: 4.2 DV: 0.2
 08:18 SP: 4.0 PV: 3.8 DV: -0.2
 07:18 SP: 4.0 PV: 4.0 DV: 0.0
 06:18 SP: 4.0 PV: 3.9 DV: -0.1
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 4.2 DV: 0.2
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 4.1 DV: 0.1
 01:18 SP: 4.0 PV: 3.7 DV: -0.3
 00:18 SP: 4.0 PV: 3.9 DV: -0.1
 2022-12-12
 23:18 SP: 4.0 PV: 4.2 DV: 0.2
 22:18 SP: 4.0 PV: 3.8 DV: -0.2
 21:18 SP: 4.0 PV: 4.1 DV: 0.1
 20:18 SP: 4.0 PV: 3.7 DV: -0.3
 19:18 SP: 4.0 PV: 4.2 DV: 0.2
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 3.8 DV: -0.2
 16:18 SP: 4.0 PV: 4.3 DV: 0.3
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 3.7 DV: -0.3
 13:18 SP: 4.0 PV: 4.2 DV: 0.2
 12:18 SP: 4.0 PV: 4.1 DV: 0.1
 11:18 SP: 4.0 PV: 4.3 DV: 0.3
 10:18 SP: 4.0 PV: 3.9 DV: -0.1
 09:18 SP: 4.0 PV: 4.1 DV: 0.1
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 4.6 DV: 0.6
 06:18 SP: 4.0 PV: 3.8 DV: -0.2
 05:18 SP: 4.0 PV: 4.4 DV: 0.4
 04:18 SP: 4.0 PV: 3.7 DV: -0.3
 03:18 SP: 4.0 PV: 4.4 DV: 0.4
 02:18 SP: 4.0 PV: 3.7 DV: -0.3
 01:18 SP: 4.0 PV: 4.3 DV: 0.3
 00:18 SP: 4.0 PV: 3.7 DV: -0.3
 2022-12-11

23:18 SP: 4.0 PV: 4.3 DV: 0.3
 22:18 SP: 4.0 PV: 3.7 DV: -0.3
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 4.3 DV: 0.3
 19:18 SP: 4.0 PV: 3.9 DV: -0.1
 18:18 SP: 4.0 PV: 3.8 DV: -0.2
 17:18 SP: 4.0 PV: 4.4 DV: 0.4
 16:18 SP: 4.0 PV: 3.9 DV: -0.1
 15:18 SP: 4.0 PV: 4.2 DV: 0.2
 14:18 SP: 4.0 PV: 4.0 DV: 0.0
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 12:18 SP: 4.0 PV: 4.3 DV: 0.3
 11:18 SP: 4.0 PV: 4.0 DV: 0.0
 2
 10:18 SP: 4.0 PV: 4.2 DV: 0.
 09:18 SP: 4.0 PV: 4.1 DV: 0.1
 08:18 SP: 4.0 PV: 4.6 DV: 0.6
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 4.4 DV: 0.4
 05:18 SP: 4.0 PV: 4.2 DV: 0.2
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.4 DV: 0.4
 01:18 SP: 4.0 PV: 3.7 DV: -0.3
 00:18 SP: 4.0 PV: 4.3 DV: 0.3
 2022-12-10
 23:18 SP: 4.0 PV: 4.1 DV: 0.1
 22:18 SP: 4.0 PV: 4.2 DV: 0.2
 21:18 SP: 4.0 PV: 4.1 DV: 0.1
 20:18 SP: 4.0 PV: 4.2 DV: 0.2
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 4.2 DV: 0.2
 16:18 SP: 4.0 PV: 3.8 DV: -0.2
 15:18 SP: 4.0 PV: 4.4 DV: 0.4
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.2 DV: 0.2
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 4.3 DV: 0.3
 10:18 SP: 4.0 PV: 3.9 DV: -0.1
 09:18 SP: 4.0 PV: 4.2 DV: 0.2
 08:18 SP: 4.0 PV: 4.0 DV: 0.0
 07:18 SP: 4.0 PV: 4.0 DV: 0.0
 06:18 SP: 4.0 PV: 3.8 DV: -0.2
 05:18 SP: 4.0 PV: 4.4 DV: 0.4
 1
 04:18 SP: 4.0 PV: 3.9 DV: -0.
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 4.2 DV: 0.2
 2022-12-09

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 3.8 DV: -0.2
 22:18 SP: 4.0 PV: 4.0 DV: 0.0
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 4.0 DV: 0.0
 19:18 SP: 4.0 PV: 3.7 DV: -0.3
 18:18 SP: 4.0 PV: 4.1 DV: 0.1
 17:18 SP: 4.0 PV: 3.7 DV: -0.3
 16:18 SP: 4.0 PV: 4.1 DV: 0.1
 15:18 SP: 4.0 PV: 3.7 DV: -0.3
 14:18 SP: 4.0 PV: 4.2 DV: 0.2

.2

13:18 SP: 4.0 PV: 3.8 DV: -0.2
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 4.2 DV: 0.2
 09:18 SP: 4.0 PV: 3.9 DV: -0.1
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 4.2 DV: 0.2
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 4.5 DV: 0.5
 04:18 SP: 4.0 PV: 3.8 DV: -0.2
 03:18 SP: 4.0 PV: 4.6 DV: 0.6
 02:18 SP: 4.0 PV: 3.9 DV: -0.1
 01:18 SP: 4.0 PV: 4.3 DV: 0.3
 00:18 SP: 4.0 PV: 3.7 DV: -0.3

2022-12-14

23:18 SP: 4.0 PV: 4.3 DV: 0.3
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 3.8 DV: -0.2
 20:18 SP: 4.0 PV: 4.3 DV: 0.3
 19:18 SP: 4.0 PV: 3.7 DV: -0.3
 18:18 SP: 4.0 PV: 4.2 DV: 0.2
 17:18 SP: 4.0 PV: 3.7 DV: -0.3
 16:18 SP: 4.0 PV: 4.4 DV: 0.4
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.3 DV: 0.3
 12:18 SP: 4.0 PV: 3.9 DV: -0.1
 11:18 SP: 4.0 PV: 4.1 DV: 0.1
 10:18 SP: 4.0 PV: 3.7 DV: -0.3
 09:18 SP: 4.0 PV: 4.2 DV: 0.2
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 4.2 DV: 0.2
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 4.1 DV: 0.1
 03:18 SP: 4.0 PV: 3.7 DV: -0.3
 02:18 SP: 4.0 PV: 4.2 DV: 0.2
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 4.1 DV: 0.1

2022-12-13

23:18 SP: 4.0 PV: 3.9 DV: -0.1
 22:18 SP: 4.0 PV: 3.8 DV: -0.2
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 3.8 DV: -0.2
 19:18 SP: 4.0 PV: 3.8 DV: -0.2
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 3.8 DV: -0.2
 16:18 SP: 4.0 PV: 3.9 DV: -0.1
 15:18 SP: 4.0 PV: 3.8 DV: -0.2
 14:18 SP: 4.0 PV: 3.7 DV: -0.3
 13:18 SP: 4.0 PV: 3.7 DV: -0.3

1

12:18 SP: 4.0 PV: 4.1 DV: 0.
 11:18 SP: 4.0 PV: 3.9 DV: -0.1
 10:18 SP: 4.0 PV: 3.8 DV: -0.2
 09:18 SP: 4.0 PV: 3.8 DV: -0.2

3

08:18 SP: 4.0 PV: 3.7 DV: -0.
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 3.9 DV: -0.1
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 3.9 DV: -0.1
 03:18 SP: 4.0 PV: 3.9 DV: -0.1
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 3.8 DV: -0.2

2022-12-16

23:18 SP: 4.0 PV: 3.7 DV: -0.3
 22:18 SP: 4.0 PV: 4.1 DV: 0.1
 21:18 SP: 4.0 PV: 3.8 DV: -0.2
 20:18 SP: 4.0 PV: 4.0 DV: 0.0
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 4.0 DV: 0.0
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.9 DV: -0.1
 15:18 SP: 4.0 PV: 4.0 DV: 0.0
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 3.7 DV: -0.3
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 3.8 DV: -0.2
 04:18 SP: 4.0 PV: 3.9 DV: -0.1
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 3.9 DV: -0.1
 01:18 SP: 4.0 PV: 3.9 DV: -0.1
 00:18 SP: 4.0 PV: 3.9 DV: -0.1

2022-12-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 3.8 DV: -0.2
 22:18 SP: 4.0 PV: 4.0 DV: 0.0
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 3.8 DV: -0.2
 18:18 SP: 4.0 PV: 4.1 DV: 0.1
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 3.7 DV: -0.3
 15:18 SP: 4.0 PV: 3.7 DV: -0.3
 14:18 SP: 4.0 PV: 4.2 DV: 0.2
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 12:18 SP: 4.0 PV: 3.7 DV: -0.3
 11:18 SP: 4.0 PV: 4.3 DV: 0.3
 10:18 SP: 4.0 PV: 3.9 DV: -0.1
 09:18 SP: 4.0 PV: 4.1 DV: 0.1
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 3.7 DV: -0.3
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 3.7 DV: -0.3
 04:18 SP: 4.0 PV: 3.7 DV: -0.3
 03:18 SP: 4.0 PV: 3.7 DV: -0.3
 02:18 SP: 4.0 PV: 4.0 DV: 0.0
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 4.1 DV: 0.1
 2022-12-18
 23:18 SP: 4.0 PV: 3.8 DV: -0.2
 22:18 SP: 4.0 PV: 3.9 DV: -0.1
 21:18 SP: 4.0 PV: 3.9 DV: -0.1
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 4.0 DV: 0.0
 18:18 SP: 4.0 PV: 3.8 DV: -0.2
 17:18 SP: 4.0 PV: 4.1 DV: 0.1
 16:18 SP: 4.0 PV: 3.9 DV: -0.1
 15:18 SP: 4.0 PV: 3.9 DV: -0.1
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 3.8 DV: -0.2
 1
 12:18 SP: 4.0 PV: 3.9 DV: -0.1
 11:18 SP: 4.0 PV: 3.8 DV: -0.2
 10:18 SP: 4.0 PV: 4.0 DV: 0.0
 1
 09:18 SP: 4.0 PV: 3.9 DV: -0.1
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 3.8 DV: -0.2
 06:18 SP: 4.0 PV: 3.7 DV: -0.3
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 3.9 DV: -0.1
 03:18 SP: 4.0 PV: 3.8 DV: -0.2
 02:18 SP: 4.0 PV: 3.7 DV: -0.3
 01:18 SP: 4.0 PV: 3.7 DV: -0.3
 00:18 SP: 4.0 PV: 3.9 DV: -0.1
 2022-12-17

23:18 SP: 4.0 PV: 4.4 DV: 0.4
 22:18 SP: 4.0 PV: 3.8 DV: -0.2
 21:18 SP: 4.0 PV: 4.3 DV: 0.3
 20:18 SP: 4.0 PV: 4.6 DV: 0.6
 19:18 SP: 4.0 PV: 4.7 DV: 0.7
 18:18 SP: 4.0 PV: 4.3 DV: 0.3
 5
 17:18 SP: 4.0 PV: 4.5 DV: 0.1
 16:18 SP: 4.0 PV: 4.7 DV: 0.7
 15:18 SP: 4.0 PV: 4.7 DV: 0.7
 14:18 SP: 4.0 PV: 4.7 DV: 0.7
 13:18 SP: 4.0 PV: 4.6 DV: 0.6
 12:18 SP: 4.0 PV: 4.7 DV: 0.7
 11:18 SP: 4.0 PV: 4.6 DV: 0.6
 10:18 SP: 4.0 PV: 4.7 DV: 0.7
 09:18 SP: 4.0 PV: 3.9 DV: -0.1
 08:18 SP: 4.0 PV: 4.5 DV: 0.5
 07:18 SP: 4.0 PV: 4.6 DV: 0.6
 06:18 SP: 4.0 PV: 4.7 DV: 0.7
 05:18 SP: 4.0 PV: 4.8 DV: 0.8
 04:18 SP: 4.0 PV: 4.8 DV: 0.8
 03:18 SP: 4.0 PV: 4.8 DV: 0.8
 02:18 SP: 4.0 PV: 4.8 DV: 0.8
 01:18 SP: 4.0 PV: 4.8 DV: 0.8
 00:18 SP: 4.0 PV: 4.6 DV: 0.6
 2022-12-20
 23:18 SP: 4.0 PV: 3.9 DV: -0.1
 22:18 SP: 4.0 PV: 4.3 DV: 0.3
 21:18 SP: 4.0 PV: 4.5 DV: 0.5
 20:18 SP: 4.0 PV: 4.7 DV: 0.7
 19:18 SP: 4.0 PV: 4.2 DV: 0.2
 18:18 SP: 4.0 PV: 4.5 DV: 0.5
 17:18 SP: 4.0 PV: 4.7 DV: 0.7
 16:18 SP: 4.0 PV: 4.7 DV: 0.7
 15:18 SP: 4.0 PV: 4.6 DV: 0.6
 14:18 SP: 4.0 PV: 3.9 DV: -0.1
 13:18 SP: 4.0 PV: 4.4 DV: 0.4
 12:18 SP: 4.0 PV: 4.7 DV: 0.7
 11:18 SP: 4.0 PV: 4.6 DV: 0.6
 10:18 SP: 4.0 PV: 3.8 DV: -0.2
 09:18 SP: 4.0 PV: 3.7 DV: -0.3
 08:18 SP: 4.0 PV: 3.8 DV: -0.2
 07:18 SP: 4.0 PV: 4.2 DV: 0.2
 06:18 SP: 4.0 PV: 3.9 DV: -0.1
 05:18 SP: 4.0 PV: 4.0 DV: 0.0
 2
 04:18 SP: 4.0 PV: 3.8 DV: -0.1
 03:18 SP: 4.0 PV: 3.7 DV: -0.3
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 3.8 DV: -0.2
 00:18 SP: 4.0 PV: 3.9 DV: -0.1
 2022-12-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 4.4 DV: 0.4
 22:18 SP: 4.0 PV: 4.5 DV: 0.5
 21:18 SP: 4.0 PV: 4.6 DV: 0.6
 20:18 SP: 4.0 PV: 4.7 DV: 0.7
 19:18 SP: 4.0 PV: 4.2 DV: 0.2
 18:18 SP: 4.0 PV: 4.4 DV: 0.4
 17:18 SP: 4.0 PV: 4.7 DV: 0.7
 16:18 SP: 4.0 PV: 4.8 DV: 0.8
 15:18 SP: 4.0 PV: 4.3 DV: 0.3
 14:18 SP: 4.0 PV: 4.5 DV: 0.5
 13:18 SP: 4.0 PV: 4.4 DV: 0.4
 12:18 SP: 4.0 PV: 4.3 DV: 0.3
 11:18 SP: 4.0 PV: 4.4 DV: 0.4
 10:18 SP: 4.0 PV: 4.6 DV: 0.6
 09:18 SP: 4.0 PV: 4.3 DV: 0.3
 08:18 SP: 4.0 PV: 4.6 DV: 0.6
 07:18 SP: 4.0 PV: 4.2 DV: 0.2
 06:18 SP: 4.0 PV: 4.6 DV: 0.6
 05:18 SP: 4.0 PV: 4.7 DV: 0.7
 04:18 SP: 4.0 PV: 4.7 DV: 0.7
 03:18 SP: 4.0 PV: 4.4 DV: 0.4
 02:18 SP: 4.0 PV: 3.8 DV: -0.2
 01:18 SP: 4.0 PV: 4.3 DV: 0.3

6

00:18 SP: 4.0 PV: 4.6 DV: 0.

2022-12-22

3

23:18 SP: 4.0 PV: 4.3 DV: 0.
 22:18 SP: 4.0 PV: 4.5 DV: 0.5
 21:18 SP: 4.0 PV: 4.7 DV: 0.7
 20:18 SP: 4.0 PV: 4.1 DV: 0.1
 19:18 SP: 4.0 PV: 4.5 DV: 0.5
 18:18 SP: 4.0 PV: 4.7 DV: 0.7
 17:18 SP: 4.0 PV: 4.1 DV: 0.1
 16:18 SP: 4.0 PV: 4.5 DV: 0.5
 15:18 SP: 4.0 PV: 4.6 DV: 0.6
 14:18 SP: 4.0 PV: 4.7 DV: 0.7
 13:18 SP: 4.0 PV: 4.8 DV: 0.8
 12:18 SP: 4.0 PV: 4.8 DV: 0.8
 11:18 SP: 4.0 PV: 4.7 DV: 0.7
 10:18 SP: 4.0 PV: 3.8 DV: -0.2

5

09:18 SP: 4.0 PV: 4.5 DV: 0.
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 4.3 DV: 0.3
 06:18 SP: 4.0 PV: 4.6 DV: 0.6
 05:18 SP: 4.0 PV: 4.3 DV: 0.3
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 4.7 DV: 0.7
 02:18 SP: 4.0 PV: 4.7 DV: 0.7
 01:18 SP: 4.0 PV: 4.7 DV: 0.7
 00:18 SP: 4.0 PV: 4.6 DV: 0.6

2022-12-21

23:18 SP: 4.0 PV: 4.7 DV: 0.7
 22:18 SP: 4.0 PV: 4.6 DV: 0.6
 21:18 SP: 4.0 PV: 4.0 DV: 0.0
 20:18 SP: 4.0 PV: 4.5 DV: 0.5
 19:18 SP: 4.0 PV: 4.7 DV: 0.7
 18:18 SP: 4.0 PV: 4.6 DV: 0.6
 17:18 SP: 4.0 PV: 3.9 DV: -0.1
 16:18 SP: 4.0 PV: 4.5 DV: 0.5
 15:18 SP: 4.0 PV: 4.6 DV: 0.6
 14:18 SP: 4.0 PV: 4.5 DV: 0.5
 13:18 SP: 4.0 PV: 3.9 DV: -0.1
 12:18 SP: 4.0 PV: 4.4 DV: 0.4
 11:18 SP: 4.0 PV: 4.7 DV: 0.7
 10:18 SP: 4.0 PV: 4.6 DV: 0.6
 09:18 SP: 4.0 PV: 4.0 DV: 0.0
 08:18 SP: 4.0 PV: 4.5 DV: 0.5
 07:18 SP: 4.0 PV: 4.7 DV: 0.7
 06:18 SP: 4.0 PV: 4.6 DV: 0.6
 05:18 SP: 4.0 PV: 3.9 DV: -0.1
 04:18 SP: 4.0 PV: 4.4 DV: 0.4
 03:18 SP: 4.0 PV: 4.7 DV: 0.7
 02:18 SP: 4.0 PV: 4.8 DV: 0.8

8

01:18 SP: 4.0 PV: 4.8 DV: 0.

00:18 SP: 4.0 PV: 4.8 DV: 0.8

2022-12-24

23:18 SP: 4.0 PV: 4.8 DV: 0.8
 22:18 SP: 4.0 PV: 4.8 DV: 0.8
 21:18 SP: 4.0 PV: 4.8 DV: 0.8
 20:18 SP: 4.0 PV: 4.8 DV: 0.8

8

19:18 SP: 4.0 PV: 4.8 DV: 0.
 18:18 SP: 4.0 PV: 4.7 DV: 0.7
 17:18 SP: 4.0 PV: 4.7 DV: 0.7
 16:18 SP: 4.0 PV: 4.7 DV: 0.7
 15:18 SP: 4.0 PV: 4.7 DV: 0.7
 14:18 SP: 4.0 PV: 4.6 DV: 0.6
 13:18 SP: 4.0 PV: 4.0 DV: 0.0
 12:18 SP: 4.0 PV: 4.4 DV: 0.4
 11:18 SP: 4.0 PV: 4.5 DV: 0.5
 10:18 SP: 4.0 PV: 4.7 DV: 0.7
 09:18 SP: 4.0 PV: 4.3 DV: 0.3
 08:18 SP: 4.0 PV: 4.4 DV: 0.4
 07:18 SP: 4.0 PV: 4.5 DV: 0.5
 06:18 SP: 4.0 PV: 4.7 DV: 0.7
 05:18 SP: 4.0 PV: 4.7 DV: 0.7
 04:18 SP: 4.0 PV: 4.7 DV: 0.7
 03:18 SP: 4.0 PV: 4.6 DV: 0.6
 02:18 SP: 4.0 PV: 4.7 DV: 0.7
 01:18 SP: 4.0 PV: 4.7 DV: 0.7
 00:18 SP: 4.0 PV: 4.2 DV: 0.2

2022-12-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

23:18 SP: 4.0 PV: 4.7 DV: 0.7
 22:18 SP: 4.0 PV: 4.8 DV: 0.8
 21:18 SP: 4.0 PV: 4.2 DV: 0.2
 20:18 SP: 4.0 PV: 4.4 DV: 0.4
 19:18 SP: 4.0 PV: 4.6 DV: 0.6
 18:18 SP: 4.0 PV: 4.7 DV: 0.7
 17:18 SP: 4.0 PV: 4.8 DV: 0.8

8

16:18 SP: 4.0 PV: 4.8 DV: 0.
 15:18 SP: 4.0 PV: 4.8 DV: 0.8
 14:18 SP: 4.0 PV: 4.7 DV: 0.7
 13:18 SP: 4.0 PV: 4.6 DV: 0.6
 12:18 SP: 4.0 PV: 4.5 DV: 0.5
 11:18 SP: 4.0 PV: 4.4 DV: 0.4
 10:18 SP: 4.0 PV: 4.6 DV: 0.6
 09:18 SP: 4.0 PV: 4.2 DV: 0.2
 08:18 SP: 4.0 PV: 3.7 DV: -0.3
 07:18 SP: 4.0 PV: 4.4 DV: 0.4
 06:18 SP: 4.0 PV: 4.3 DV: 0.3
 05:18 SP: 4.0 PV: 4.3 DV: 0.3
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 4.2 DV: 0.2
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 4.6 DV: 0.6
 00:18 SP: 4.0 PV: 4.2 DV: 0.2

2022-12-26

23:18 SP: 4.0 PV: 4.6 DV: 0.6
 22:18 SP: 4.0 PV: 4.7 DV: 0.7
 21:18 SP: 4.0 PV: 4.5 DV: 0.5
 20:18 SP: 4.0 PV: 3.9 DV: -0.1
 19:18 SP: 4.0 PV: 4.4 DV: 0.4
 18:18 SP: 4.0 PV: 4.7 DV: 0.7
 17:18 SP: 4.0 PV: 4.6 DV: 0.6
 16:18 SP: 4.0 PV: 4.0 DV: 0.0
 15:18 SP: 4.0 PV: 4.6 DV: 0.6
 14:18 SP: 4.0 PV: 4.7 DV: 0.7
 13:18 SP: 4.0 PV: 4.5 DV: 0.5
 12:18 SP: 4.0 PV: 3.8 DV: -0.2
 11:18 SP: 4.0 PV: 4.4 DV: 0.4
 10:18 SP: 4.0 PV: 4.7 DV: 0.7
 09:18 SP: 4.0 PV: 4.6 DV: 0.6
 08:18 SP: 4.0 PV: 3.9 DV: -0.1
 07:18 SP: 4.0 PV: 4.4 DV: 0.4
 06:18 SP: 4.0 PV: 4.7 DV: 0.7
 05:18 SP: 4.0 PV: 4.2 DV: 0.2
 04:18 SP: 4.0 PV: 4.6 DV: 0.6
 03:18 SP: 4.0 PV: 4.7 DV: 0.7
 02:18 SP: 4.0 PV: 4.6 DV: 0.6
 01:18 SP: 4.0 PV: 3.7 DV: -0.3
 00:18 SP: 4.0 PV: 4.4 DV: 0.4

2022-12-25

23:18 SP: 4.0 PV: 3.9 DV: -0.1
 22:18 SP: 4.0 PV: 4.4 DV: 0.4
 21:18 SP: 4.0 PV: 4.1 DV: 0.1
 20:18 SP: 4.0 PV: 4.4 DV: 0.4
 19:18 SP: 4.0 PV: 4.6 DV: 0.6
 18:18 SP: 4.0 PV: 3.9 DV: -0.1
 17:18 SP: 4.0 PV: 4.4 DV: 0.4
 16:18 SP: 4.0 PV: 4.1 DV: 0.1
 15:18 SP: 4.0 PV: 4.4 DV: 0.4
 14:18 SP: 4.0 PV: 4.0 DV: 0.0

13:18 SP: 4.0 PV: 4.4 DV: 0.4
 12:18 SP: 4.0 PV: 4.7 DV: 0.7
 11:18 SP: 4.0 PV: 4.8 DV: 0.8
 10:18 SP: 4.0 PV: 4.8 DV: 0.8
 09:18 SP: 4.0 PV: 4.8 DV: 0.8
 08:18 SP: 4.0 PV: 4.8 DV: 0.8
 07:18 SP: 4.0 PV: 4.8 DV: 0.8
 06:18 SP: 4.0 PV: 4.8 DV: 0.8
 05:18 SP: 4.0 PV: 4.7 DV: 0.7
 04:18 SP: 4.0 PV: 4.2 DV: 0.2
 03:18 SP: 4.0 PV: 4.5 DV: 0.5
 02:18 SP: 4.0 PV: 4.7 DV: 0.7
 01:18 SP: 4.0 PV: 4.8 DV: 0.8
 00:18 SP: 4.0 PV: 4.8 DV: 0.8

2022-12-28

23:18 SP: 4.0 PV: 4.7 DV: 0.7
 22:18 SP: 4.0 PV: 4.6 DV: 0.6
 21:18 SP: 4.0 PV: 4.5 DV: 0.5
 20:18 SP: 4.0 PV: 4.5 DV: 0.5
 19:18 SP: 4.0 PV: 4.1 DV: 0.1
 18:18 SP: 4.0 PV: 4.3 DV: 0.3
 17:18 SP: 4.0 PV: 4.3 DV: 0.3
 16:18 SP: 4.0 PV: 4.4 DV: 0.4
 15:18 SP: 4.0 PV: 4.2 DV: 0.2
 14:18 SP: 4.0 PV: 4.4 DV: 0.4
 13:18 SP: 4.0 PV: 4.4 DV: 0.4
 12:18 SP: 4.0 PV: 4.6 DV: 0.6
 11:18 SP: 4.0 PV: 4.7 DV: 0.7
 10:18 SP: 4.0 PV: 4.6 DV: 0.6
 09:18 SP: 4.0 PV: 4.7 DV: 0.7
 08:18 SP: 4.0 PV: 4.8 DV: 0.8
 07:18 SP: 4.0 PV: 4.8 DV: 0.8
 06:18 SP: 4.0 PV: 4.7 DV: 0.7
 05:18 SP: 4.0 PV: 4.1 DV: 0.1
 04:18 SP: 4.0 PV: 4.4 DV: 0.4
 03:18 SP: 4.0 PV: 4.6 DV: 0.6
 02:18 SP: 4.0 PV: 4.3 DV: 0.3
 01:18 SP: 4.0 PV: 4.5 DV: 0.5
 00:18 SP: 4.0 PV: 4.6 DV: 0.6

2022-12-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2022.12

10:18 SP: 4.0 PV: 3.8 DV: -0.2
3

09:18 SP: 4.0 PV: 4.3 DV: 0.
08:18 SP: 4.0 PV: 4.4 DV: 0.4
07:18 SP: 4.0 PV: 4.6 DV: 0.6
06:18 SP: 4.0 PV: 4.8 DV: 0.8
05:18 SP: 4.0 PV: 4.8 DV: 0.8
04:18 SP: 4.0 PV: 4.8 DV: 0.8
03:18 SP: 4.0 PV: 4.8 DV: 0.8
02:18 SP: 4.0 PV: 4.7 DV: 0.7
01:18 SP: 4.0 PV: 4.7 DV: 0.7
00:18 SP: 4.0 PV: 4.6 DV: 0.6

2022-12-30

23:18 SP: 4.0 PV: 4.7 DV: 0.7
22:18 SP: 4.0 PV: 4.0 DV: 0.0
21:18 SP: 4.0 PV: 4.3 DV: 0.3
20:18 SP: 4.0 PV: 4.4 DV: 0.4
19:18 SP: 4.0 PV: 4.6 DV: 0.6
18:18 SP: 4.0 PV: 4.8 DV: 0.8
17:18 SP: 4.0 PV: 3.9 DV: -0.1
16:18 SP: 4.0 PV: 4.2 DV: 0.2
15:18 SP: 4.0 PV: 4.6 DV: 0.6
14:18 SP: 4.0 PV: 4.8 DV: 0.8
13:18 SP: 4.0 PV: 3.8 DV: -0.2
12:18 SP: 4.0 PV: 4.2 DV: 0.2
11:18 SP: 4.0 PV: 4.6 DV: 0.6
10:18 SP: 4.0 PV: 4.8 DV: 0.8
09:18 SP: 4.0 PV: 3.8 DV: -0.2
08:18 SP: 4.0 PV: 4.3 DV: 0.3
07:18 SP: 4.0 PV: 4.6 DV: 0.6
06:18 SP: 4.0 PV: 4.7 DV: 0.7
05:18 SP: 4.0 PV: 3.9 DV: -0.1
04:18 SP: 4.0 PV: 4.2 DV: 0.2
03:18 SP: 4.0 PV: 4.5 DV: 0.5
02:18 SP: 4.0 PV: 4.0 DV: 0.0
01:18 SP: 4.0 PV: 4.4 DV: 0.4
00:18 SP: 4.0 PV: 4.6 DV: 0.6

2022-12-29

23:18 SP: 4.0 PV: 4.6 DV: 0.6
22:18 SP: 4.0 PV: 4.8 DV: 0.8
21:18 SP: 4.0 PV: 4.8 DV: 0.8
20:18 SP: 4.0 PV: 4.2 DV: 0.2
19:18 SP: 4.0 PV: 4.5 DV: 0.5
18:18 SP: 4.0 PV: 4.7 DV: 0.7
17:18 SP: 4.0 PV: 4.8 DV: 0.8
16:18 SP: 4.0 PV: 4.8 DV: 0.8
15:18 SP: 4.0 PV: 4.8 DV: 0.8
14:18 SP: 4.0 PV: 4.8 DV: 0.8
13:18 SP: 4.0 PV: 4.7 DV: 0.7
12:18 SP: 4.0 PV: 4.1 DV: 0.1
11:18 SP: 4.0 PV: 4.4 DV: 0.4
10:18 SP: 4.0 PV: 4.7 DV: 0.7
09:18 SP: 4.0 PV: 4.8 DV: 0.8
08:18 SP: 4.0 PV: 4.8 DV: 0.8
07:18 SP: 4.0 PV: 4.8 DV: 0.8

6

06:18 SP: 4.0 PV: 4.6 DV: 0.
05:18 SP: 4.0 PV: 4.6 DV: 0.6
04:18 SP: 4.0 PV: 4.4 DV: 0.4
03:18 SP: 4.0 PV: 4.3 DV: 0.3
02:18 SP: 4.0 PV: 3.9 DV: -0.1
01:18 SP: 4.0 PV: 4.3 DV: 0.3
00:18 SP: 4.0 PV: 4.6 DV: 0.6

2022-12-31

23:18 SP: 4.0 PV: 4.7 DV: 0.7
22:18 SP: 4.0 PV: 4.8 DV: 0.8
21:18 SP: 4.0 PV: 4.8 DV: 0.8
20:18 SP: 4.0 PV: 4.8 DV: 0.8
19:18 SP: 4.0 PV: 4.7 DV: 0.7
18:18 SP: 4.0 PV: 4.6 DV: 0.6
17:18 SP: 4.0 PV: 4.6 DV: 0.6
16:18 SP: 4.0 PV: 4.7 DV: 0.7

7

15:18 SP: 4.0 PV: 4.7 DV: 0.
14:18 SP: 4.0 PV: 4.8 DV: 0.8
13:18 SP: 4.0 PV: 4.8 DV: 0.8
12:18 SP: 4.0 PV: 3.9 DV: -0.1
11:18 SP: 4.0 PV: 4.2 DV: 0.2

SP Setting Point
PV Present Value
DV Difference Value