

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2022.10
	Model Name	MPR-720R	Investigational Product	보관조건 15~25℃ 25℃ 1~25℃
	Serial Number	07120004		
Calibration Subject	Description			
	Manufacturer			
	Model Name			
	Serial Number			
	Date of Calibration	2022.8.5		
	Certificate No.	SP2208-720R-07	Temperature Range	16~22℃(19±3℃)

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2022.10.01						
2022.10.02						
2022.10.03						개천절
2022.10.04	P:30			18		
2022.10.05	P:30			18		
2022.10.06	P:30			18		
2022.10.07	P:50			19		
2022.10.08						
2022.10.09						
2022.10.10						(한글날)대체휴일
2022.10.11	P:30			19		
2022.10.12	P:50			18		
2022.10.13	P:40			18		
2022.10.14	P:30			19		
2022.10.15						
2022.10.16						
2022.10.17	P:30			18		
2022.10.18	P:30			19		
2022.10.19	P:30			18		
2022.10.20	P:30			18		
2022.10.21	P:30			18		
2022.10.22						
2022.10.23						
2022.10.24	P:30			18		
2022.10.25	P:30			19		
2022.10.26	P:30			19		
2022.10.27	P:20			19		
2022.10.28	P:30			18		
2022.10.29						
2022.10.30						
2022.10.31	10:00			19		


date of copying

sign

부록7. 서식II-01-07, 온도기록지(IP 보관)-자동온도 기록용지 부착형

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2022.9.27~2022.10.10
	Model Name	MPR-720R	Investigational Product	보관조건 15~25℃ 25℃ 1~25℃
	Serial Number	07120004		
Calibration Subject	Description			
	Manufacturer			
	Model Name			
	Serial Number			
	Date of Calibration	2022.8.5		
	Certificate No.	SP2208-720R-07	Temperature Range	16~22℃(19±3℃)



date of copying

sign