

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.4
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-4		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2022.4.4		
	Certificate No.	22008908-8	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.04.01		Left					
		Right					
2023.04.02		Left					
		Right					
2023.04.03	p:30	Left	3	4	3	ml	
		Right	3	5	4		
2023.04.04	p:30	Left	2	4	4	ml	온도계 교체
		Right	3	5	4		
2023.04.05		Left					
		Right					
2023.04.06		Left					
		Right					
2023.04.07		Left					
		Right					
2023.04.08		Left					
		Right					
2023.04.09		Left					
		Right					
2023.04.10		Left					
		Right					
2023.04.11		Left					
		Right					
2023.04.12		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.4
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.04.01		Left					
		Right					
2023.04.02		Left					
		Right					
2023.04.03		Left					
		Right					
2023.04.04	13:00	Left	2	4	3	[Signature]	온도계 교체
		Right	3	5	4		
2023.04.05	11:30	Left	2	4	3	[Signature]	
		Right	3	5	3		
2023.04.06	11:30	Left	3	4	4	[Signature]	
		Right	3	5	4		
2023.04.07	9:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.08		Left					
		Right					
2023.04.09		Left					
		Right					
2023.04.10	11:30	Left	3	4	3	[Signature]	
		Right	3	5	4		
2023.04.11	11:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.12	11:30	Left	3	4	4	[Signature]	
		Right	3	5	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.4
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8°C

Date	Time	Probe Position	Temperature(°C)			Verified by	Comment
			Min	Max	Current		
2023.04.13	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.14	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.15		Left					
		Right					
2023.04.16		Left					
		Right					
2023.04.17	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.18	1:30	Left	2	4	3	[Signature]	
		Right	3	5	5		
2023.04.19	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.20	1:40	Left	2	4	2	[Signature]	
		Right	3	5	3		
2023.04.21	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.22		Left					
		Right					
2023.04.23		Left					
		Right					
2023.04.24	1:30	Left	2	4	3	[Signature]	
		Right	3	5	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.4
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.04.25	9:30	Left	2	4	3	[Signature]	4/25 6C1118 IP [Signature]
		Right	3	5	4		
2023.04.26	9:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.27	9:30	Left	2	4	3	[Signature]	
		Right	3	5	4		
2023.04.28	9:30	Left	2	4	3	[Signature]	
		Right	3	5	5		
2023.04.29		Left					
		Right					
2023.04.30		Left					
		Right					
1900-01-00		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.8 DV: 0.8
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.7 DV: 0.7
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.5 DV: 0.5
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.3 DV: 0.3

1

02:08 SP: 4.0 PV: 4.1 DV: 0.

.4

01:08 SP: 4.0 PV: 4.4 DV: 0

00:08 SP: 4.0 PV: 4.6 DV: 0.6

2023-04-02

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2023-04-01

23:08 SP: 4.0 PV: 4.6 DV: 0.6

22:08 SP: 4.0 PV: 3.8 DV: -0.2

3

21:08 SP: 4.0 PV: 4.3 DV: 0.

20:08 SP: 4.0 PV: 4.6 DV: 0.6

19:08 SP: 4.0 PV: 4.5 DV: 0.5

18:08 SP: 4.0 PV: 4.3 DV: 0.3

17:08 SP: 4.0 PV: 4.2 DV: 0.2

16:08 SP: 4.0 PV: 4.4 DV: 0.4

15:08 SP: 4.0 PV: 4.3 DV: 0.3

14:08 SP: 4.0 PV: 4.5 DV: 0.5

13:08 SP: 4.0 PV: 3.9 DV: -0.1

12:08 SP: 4.0 PV: 4.2 DV: 0.2

11:08 SP: 4.0 PV: 4.4 DV: 0.4

10:08 SP: 4.0 PV: 4.6 DV: 0.6

09:08 SP: 4.0 PV: 4.4 DV: 0.4

08:08 SP: 4.0 PV: 4.6 DV: 0.6

07:08 SP: 4.0 PV: 4.4 DV: 0.4

06:08 SP: 4.0 PV: 4.1 DV: 0.1

05:08 SP: 4.0 PV: 4.2 DV: 0.2

2

04:08 SP: 4.0 PV: 4.2 DV: 0.

03:08 SP: 4.0 PV: 4.3 DV: 0.3

02:08 SP: 4.0 PV: 4.4 DV: 0.4

01:08 SP: 4.0 PV: 4.2 DV: 0.2

00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-04-04

23:08 SP: 4.0 PV: 3.9 DV: -0.1

22:08 SP: 4.0 PV: 4.2 DV: 0.2

21:08 SP: 4.0 PV: 4.6 DV: 0.6

20:08 SP: 4.0 PV: 3.9 DV: -0.1

19:08 SP: 4.0 PV: 4.2 DV: 0.2

18:08 SP: 4.0 PV: 4.5 DV: 0.5

17:08 SP: 4.0 PV: 4.0 DV: 0.0

16:08 SP: 4.0 PV: 4.3 DV: 0.3

15:08 SP: 4.0 PV: 4.1 DV: 0.1

14:08 SP: 4.0 PV: 4.6 DV: 0.6

13:08 SP: 4.0 PV: 4.7 DV: 0.7

12:08 SP: 4.0 PV: 4.8 DV: 0.8

11:08 SP: 4.0 PV: 4.7 DV: 0.7

10:08 SP: 4.0 PV: 4.7 DV: 0.7

09:08 SP: 4.0 PV: 4.7 DV: 0.7

08:08 SP: 4.0 PV: 4.7 DV: 0.7

07:08 SP: 4.0 PV: 4.6 DV: 0.6

06:08 SP: 4.0 PV: 4.2 DV: 0.2

05:08 SP: 4.0 PV: 4.6 DV: 0.6

04:08 SP: 4.0 PV: 4.7 DV: 0.7

03:08 SP: 4.0 PV: 4.7 DV: 0.7

02:08 SP: 4.0 PV: 4.7 DV: 0.7

01:08 SP: 4.0 PV: 4.7 DV: 0.7

00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-04-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 4
 21:08 SP: 4.0 PV: 4.4 DV: 0.
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.5 DV: 0.5
 2023-04-06
 23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.8 DV: 0.8
 07:08 SP: 4.0 PV: 4.8 DV: 0.8
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 4.8 DV: 0.8
 04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.8 DV: 0.8
 02:08 SP: 4.0 PV: 4.7 DV: 0.7
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.7 DV: 0.7
 2023-04-05

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 4
 05:08 SP: 4.0 PV: 4.4 DV: 0.
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.4 DV: 0.4
 2023-04-08
 23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3
 2023-04-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 4.8 DV: 0.8
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-04-10

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3

2

20:08 SP: 4.0 PV: 4.2 DV: 0.
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.6 DV: 0.6

5

13:08 SP: 4.0 PV: 4.5 DV: 0.
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4

6

06:08 SP: 4.0 PV: 4.6 DV: 0.
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.4 DV: 0.4

6

02:08 SP: 4.0 PV: 4.6 DV: 0.
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-04-09

23:08 SP: 4.0 PV: 3.6 DV: -0.4
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.4 DV: -0.6

2

16:08 SP: 4.0 PV: 4.2 DV: 0.
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.8 DV: 0.8
 12:08 SP: 4.0 PV: 4.8 DV: 0.8
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.4 DV: 0.4

4

04:08 SP: 4.0 PV: 4.4 DV: 0.
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2023-04-12

23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4

5

18:08 SP: 4.0 PV: 4.5 DV: 0.
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.8 DV: 0.8
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.9 DV: 0.9
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 3.7 DV: -0.3
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.5 DV: 0.5
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.8 DV: 0.8

2023-04-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.0 DV: 0.0

3

14:08 SP: 4.0 PV: 3.7 DV: -0.1
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.6 DV: -0.4

3

10:08 SP: 4.0 PV: 4.3 DV: 0.1
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2023-04-14

6

23:08 SP: 4.0 PV: 4.6 DV: 0.1
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 3.8 DV: -0.2

0

16:08 SP: 4.0 PV: 4.0 DV: 0.1
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.7 DV: -0.3
 03:08 SP: 4.0 PV: 3.6 DV: -0.4
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2023-04-13

18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.0 DV: 0.0
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2023-04-16

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 3.6 DV: -0.4
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2023-04-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.2 DV: 0.2
6
22:08 SP: 4.0 PV: 4.6 DV: 0.
21:08 SP: 4.0 PV: 3.9 DV: -0.1
20:08 SP: 4.0 PV: 4.2 DV: 0.2
19:08 SP: 4.0 PV: 4.4 DV: 0.4
18:08 SP: 4.0 PV: 4.7 DV: 0.7
17:08 SP: 4.0 PV: 3.9 DV: -0.1
16:08 SP: 4.0 PV: 4.4 DV: 0.4
15:08 SP: 4.0 PV: 4.8 DV: 0.8
14:08 SP: 4.0 PV: 3.9 DV: -0.1
13:08 SP: 4.0 PV: 4.1 DV: 0.1
12:08 SP: 4.0 PV: 4.5 DV: 0.5
11:08 SP: 4.0 PV: 3.8 DV: -0.2
10:08 SP: 4.0 PV: 4.3 DV: 0.3
09:08 SP: 4.0 PV: 4.3 DV: 0.3
08:08 SP: 4.0 PV: 3.9 DV: -0.1
07:08 SP: 4.0 PV: 4.5 DV: 0.5
06:08 SP: 4.0 PV: 3.8 DV: -0.2
05:08 SP: 4.0 PV: 4.4 DV: 0.4
04:08 SP: 4.0 PV: 4.0 DV: 0.0
03:08 SP: 4.0 PV: 4.1 DV: 0.1
02:08 SP: 4.0 PV: 4.0 DV: 0.0
01:08 SP: 4.0 PV: 4.5 DV: 0.5
00:08 SP: 4.0 PV: 3.8 DV: -0.2
2023-04-18
23:08 SP: 4.0 PV: 4.2 DV: 0.2
22:08 SP: 4.0 PV: 4.1 DV: 0.1
21:08 SP: 4.0 PV: 4.1 DV: 0.1
20:08 SP: 4.0 PV: 4.6 DV: 0.6
19:08 SP: 4.0 PV: 3.9 DV: -0.1
18:08 SP: 4.0 PV: 4.6 DV: 0.6
17:08 SP: 4.0 PV: 3.7 DV: -0.3
16:08 SP: 4.0 PV: 3.4 DV: -0.6
15:08 SP: 4.0 PV: 4.4 DV: 0.4
14:08 SP: 4.0 PV: 4.6 DV: 0.6
13:08 SP: 4.0 PV: 3.7 DV: -0.3
12:08 SP: 4.0 PV: 4.2 DV: 0.2
11:08 SP: 4.0 PV: 4.3 DV: 0.3
10:08 SP: 4.0 PV: 4.8 DV: 0.8
09:08 SP: 4.0 PV: 3.8 DV: -0.2
08:08 SP: 4.0 PV: 4.2 DV: 0.2
07:08 SP: 4.0 PV: 4.0 DV: 0.0
06:08 SP: 4.0 PV: 4.4 DV: 0.4
05:08 SP: 4.0 PV: 4.0 DV: 0.0
04:08 SP: 4.0 PV: 4.6 DV: 0.6
03:08 SP: 4.0 PV: 4.0 DV: 0.0
02:08 SP: 4.0 PV: 4.2 DV: 0.2
01:08 SP: 4.0 PV: 4.2 DV: 0.2
00:08 SP: 4.0 PV: 4.3 DV: 0.3
2023-04-17

23:08 SP: 4.0 PV: 4.8 DV: 0.8
22:08 SP: 4.0 PV: 4.9 DV: 0.9
21:08 SP: 4.0 PV: 4.8 DV: 0.8
20:08 SP: 4.0 PV: 4.8 DV: 0.8
19:08 SP: 4.0 PV: 4.6 DV: 0.6
18:08 SP: 4.0 PV: 4.6 DV: 0.6
17:08 SP: 4.0 PV: 4.5 DV: 0.5
16:08 SP: 4.0 PV: 4.3 DV: 0.3
15:08 SP: 4.0 PV: 4.4 DV: 0.4
14:08 SP: 4.0 PV: 4.8 DV: 0.8
13:08 SP: 4.0 PV: 4.7 DV: 0.7
12:08 SP: 4.0 PV: 4.8 DV: 0.8
11:08 SP: 4.0 PV: 4.4 DV: 0.4
10:08 SP: 4.0 PV: 4.6 DV: 0.6
09:08 SP: 4.0 PV: 4.6 DV: 0.6
08:08 SP: 4.0 PV: 4.8 DV: 0.8
8
07:08 SP: 4.0 PV: 4.8 DV: 0.
06:08 SP: 4.0 PV: 4.8 DV: 0.8
05:08 SP: 4.0 PV: 4.8 DV: 0.8
04:08 SP: 4.0 PV: 4.8 DV: 0.8
8
03:08 SP: 4.0 PV: 4.8 DV: 0.
02:08 SP: 4.0 PV: 4.8 DV: 0.8
01:08 SP: 4.0 PV: 4.8 DV: 0.8
00:08 SP: 4.0 PV: 4.8 DV: 0.8
2023-04-20
23:08 SP: 4.0 PV: 4.8 DV: 0.8
22:08 SP: 4.0 PV: 4.8 DV: 0.8
21:08 SP: 4.0 PV: 4.7 DV: 0.7
20:08 SP: 4.0 PV: 4.6 DV: 0.6
19:08 SP: 4.0 PV: 4.8 DV: 0.8
18:08 SP: 4.0 PV: 3.9 DV: -0.1
17:08 SP: 4.0 PV: 4.1 DV: 0.1
16:08 SP: 4.0 PV: 4.1 DV: 0.1
15:08 SP: 4.0 PV: 4.0 DV: 0.0
14:08 SP: 4.0 PV: 4.0 DV: 0.0
13:08 SP: 4.0 PV: 4.3 DV: 0.3
12:08 SP: 4.0 PV: 4.5 DV: 0.5
11:08 SP: 4.0 PV: 3.9 DV: -0.1
10:08 SP: 4.0 PV: 4.6 DV: 0.6
09:08 SP: 4.0 PV: 3.8 DV: -0.2
08:08 SP: 4.0 PV: 4.1 DV: 0.1
07:08 SP: 4.0 PV: 4.6 DV: 0.6
06:08 SP: 4.0 PV: 3.8 DV: -0.2
05:08 SP: 4.0 PV: 4.2 DV: 0.2
04:08 SP: 4.0 PV: 4.6 DV: 0.6
03:08 SP: 4.0 PV: 3.8 DV: -0.2
3
02:08 SP: 4.0 PV: 4.3 DV: 0.
01:08 SP: 4.0 PV: 4.0 DV: 0.0
00:08 SP: 4.0 PV: 4.1 DV: 0.1
2023-04-19

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.8 DV: 0.8
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.8 DV: 0.8
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 5
 03:08 SP: 4.0 PV: 4.5 DV: 0.
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.6 DV: 0.6
 2023-04-22
 23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.8 DV: 0.8
 20:08 SP: 4.0 PV: 4.8 DV: 0.8
 19:08 SP: 4.0 PV: 4.8 DV: 0.8
 18:08 SP: 4.0 PV: 4.8 DV: 0.8
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 8
 16:08 SP: 4.0 PV: 4.8 DV: 0.
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.8 DV: 0.8
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.8 DV: 0.8
 00:08 SP: 4.0 PV: 4.8 DV: 0.8
 2023-04-21

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.1 DV: 0.1
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.4 DV: 0.4
 2023-04-24
 23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 0
 19:08 SP: 4.0 PV: 4.0 DV: 0.
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 1
 16:08 SP: 4.0 PV: 3.9 DV: -0.
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.5 DV: 0.5
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2023-04-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.3 DV: 0.3
 2023-04-25

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.0 DV: 0.0
 2023-04-26

23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 4.8 DV: 0.8
 21:08 SP: 4.0 PV: 4.8 DV: 0.8
 20:08 SP: 4.0 PV: 4.8 DV: 0.8
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.7 DV: 0.7
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.8 DV: 0.8
 15:08 SP: 4.0 PV: 4.8 DV: 0.8
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.8 DV: 0.8
 05:08 SP: 4.0 PV: 4.8 DV: 0.8
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-04-28

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.8 DV: 0.8
 18:08 SP: 4.0 PV: 4.8 DV: 0.8
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.5 DV: 0.5
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2023-04-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.4

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.8 DV: 0.8
 21:08 SP: 4.0 PV: 4.8 DV: 0.8
 20:08 SP: 4.0 PV: 4.8 DV: 0.8
 19:08 SP: 4.0 PV: 4.8 DV: 0.8
 18:08 SP: 4.0 PV: 4.8 DV: 0.8
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2023-04-30

23:08 SP: 4.0 PV: 4.1 DV: 0.1

4

22:08 SP: 4.0 PV: 4.4 DV: 0.
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.6 DV: 0.6

1

16:08 SP: 4.0 PV: 4.1 DV: 0.
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.8 DV: 0.8

2023-04-29

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

date of copying

sign