

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6		
			Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.09.01	9:30	Left	2	4	3	ml	
		Right	3	5	4		
2023.09.02		Left					
		Right					
2023.09.03		Left					
		Right					
2023.09.04	9:30	Left	2	4	3	ml	
		Right	3	5	3		
2023.09.05	9:30	Left	3	5	4	ml	
		Right	3	5	3		
2023.09.06	9:30	Left	3	4	3	ml	
		Right	3	4	3		
2023.09.07	9:30	Left	3	4	4	ml	
		Right	3	4	4		
2023.09.08	9:30	Left	3	4	4	ml	
		Right	3	4	3		
2023.09.09		Left					
		Right					
2023.09.10		Left					
		Right					
2023.09.11	9:30	Left	3	4	3	ml	
		Right	3	4	3		
2023.09.12	9:30	Left	2	4	3	ml	
		Right	3	4	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6		
			Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.09.13	9:30	Left	2	4	3	ml	
		Right	2	5	3		
2023.09.14	9:30	Left	2	4	4	ml	
		Right	3	5	4		
2023.09.15	9:30	Left	2	4	4	ml	
		Right	3	5	4		
2023.09.16		Left					
		Right					
2023.09.17		Left					
		Right					
2023.09.18	9:30	Left	3	4	4	ml	
		Right	2	5	4		
2023.09.19	9:30	Left	3	4	4	ml	
		Right	3	4	4		
2023.09.20	9:30	Left	3	4	4	ml	
		Right	3	5	4		
2023.09.21	9:30	Left	3	4	3	ml	
		Right	3	5	3		
2023.09.22	9:30	Left	3	4	4	ml	
		Right	3	5	4		
2023.09.23		Left					
		Right					
2023.09.24		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2023.9
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Certificate No.	23007746-6		
			Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2023.09.25	9:30	Left	3	4	4	ml	
		Right	2	5	4		
2023.09.26	9:30	Left	3	5	3	ml	
		Right	3	5	4		
2023.09.27	9:30	Left	2	4	4	ml	
		Right	3	5	4		
2023.09.28		Left					추석연휴
		Right					
2023.09.29		Left					추석
		Right					
2023.09.30		Left					
		Right					
1900-01-00		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.7 DV: 0.7
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.8 DV: 0.8
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2023-09-02

23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.4 DV: 0.4

4

13:08 SP: 4.0 PV: 4.4 DV: 0.
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.9 DV: -0.1

4

01:08 SP: 4.0 PV: 4.4 DV: 0.
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2023-09-01

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.1 DV: 0.1
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 5.4 DV: 1.4

4

09:08 SP: 4.0 PV: 4.4 DV: 0.
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.8 DV: 0.8
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 4.8 DV: 0.8
 04:08 SP: 4.0 PV: 4.8 DV: 0.8
 03:08 SP: 4.0 PV: 4.8 DV: 0.8
 02:08 SP: 4.0 PV: 4.8 DV: 0.8
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 4.8 DV: 0.8

2023-09-04

23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 4.8 DV: 0.8
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.2 DV: 0.2

4

03:08 SP: 4.0 PV: 4.4 DV: 0.
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-09-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.6 DV: -0.4
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 3.6 DV: -0.4

4

12:08 SP: 4.0 PV: 4.4 DV: 0.
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 3.6 DV: -0.4
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 3.6 DV: -0.4
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-09-06

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 3.6 DV: -0.4
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 3.6 DV: -0.4
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.7 DV: -0.3
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2023-09-05

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.6 DV: -0.4
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.6 DV: -0.4

11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2023-09-08

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.5 DV: 0.5
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2023-09-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2023-09-10

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.7 DV: 0.7
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.7 DV: 0.7
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2023-09-09

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.6 DV: 0.6

6

07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.7 DV: 0.7
 04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2023-09-12

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2023-09-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6

6

06:08 SP: 4.0 PV: 4.6 DV: 0.
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-09-14

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 4.7 DV: 0.7
 04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.7 DV: 0.7
 02:08 SP: 4.0 PV: 4.7 DV: 0.7
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-09-13

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 4.0 DV: 0.0

5

09:08 SP: 4.0 PV: 4.5 DV: 0.
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-09-16

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-09-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.5 DV: 0.5
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.8 DV: 0.8
 11:08 SP: 4.0 PV: 4.8 DV: 0.8
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-09-18

4

23:08 SP: 4.0 PV: 4.4 DV: 0.
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.3 DV: 0.3

4

17:08 SP: 4.0 PV: 4.4 DV: 0.
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.5 DV: 0.5
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-09-17

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.8 DV: 0.8

8

17:08 SP: 4.0 PV: 4.8 DV: 0.
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.9 DV: 0.9
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.3 DV: 0.3

5

00:08 SP: 4.0 PV: 4.5 DV: 0.

2023-09-20

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.5 DV: 0.5

4

10:08 SP: 4.0 PV: 4.4 DV: 0.
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-09-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.8 DV: 0.8
 12:08 SP: 4.0 PV: 4.8 DV: 0.8
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2023-09-22

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4

1

02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2023-09-21

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.3 DV: 0.3

6

20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2023-09-24

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.9 DV: -0.1

2

13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-09-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2023-09-26

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 5.0 DV: 1.0
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-09-25

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 3.7 DV: -0.3

2

18:08 SP: 4.0 PV: 4.2 DV: 0.0
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2023-09-28

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2023-09-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.9

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.1 DV: 0.1
 12:08 SP: 4.0 PV: 4.0 DV: 0.0
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 3.4 DV: -0.6
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 3.6 DV: -0.4
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-09-30

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2023-09-29

SP Setting Point
 PV Present Value
 DV Difference Value