

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2023.8
	Model Name	MPR-720R	Investigational Product	25°C이하 15~25°C 1~25°C
	Serial Number	07120004		
Calibration Subject	Description	Thermometer		
	Manufacturer	DeltaTrak	Temperature Range	16~22°C(19±3°C)
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)		
	Certificate No.	100397082022103		

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2023.08.01	P:30	18.6	18.8	18.7	ml	
2023.08.02	P:30	18.6	18.9	18.7	ml	
2023.08.03	P:30	18.6	18.8	18.7	ml	
2023.08.04	P:30	18.6	18.9	18.7	ml	
2023.08.05						
2023.08.06						
2023.08.07	P:30	18.6	18.8	18.7	ml	
2023.08.08	P:30	18.6	18.9	18.7	ml	
2023.08.09	P:30	18.6	18.8	18.7	ml	
2023.08.10	P:30	18.6	18.8	18.7	ml	
2023.08.11	P:30	18.6	18.8	18.7	ml	
2023.08.12						
2023.08.13						
2023.08.14	P:30	18.6	18.8	18.7	ml	
2023.08.15						광복절
2023.08.16	P:30	18.6	18.8	18.7	ml	
2023.08.17	P:30	18.6	18.9	18.7	ml	
2023.08.18	P:30	18.6	18.8	18.8	ml	
2023.08.19						
2023.08.20						
2023.08.21	P:30	18.6	18.8	18.7	ml	
2023.08.22	P:30	18.6	18.8	18.7	ml	
2023.08.23	P:30	18.6	18.8	18.7	ml	
2023.08.24	P:30	18.6	18.9	18.7	ml	
2023.08.25	P:30	18.6	18.8	18.7	ml	
2023.08.26						
2023.08.27						
2023.08.28	P:30	18.6	18.9	18.7	ml	
2023.08.29	P:50	18.6	18.8	18.7	ml	
2023.08.30	P:30	18.6	18.9	18.8	ml	
2023.08.31	P:30	18.6	18.9	18.7	ml	

date of copying

sign