

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2023.9
	Model Name	MPR-720R	Investigational Product	25°C이하 15~25°C 1~25°C
	Serial Number	07120004		
Calibration Subject	Description	Thermometer		
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)		
	Certificate No.	100397082022103	Temperature Range	16~22°C(19±3°C)

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2023.09.01	9:30	18.6	18.8	18.7		
2023.09.02						
2023.09.03						
2023.09.04	9:30	18.6	18.9	18.7		
2023.09.05	9:30	18.6	18.9	18.7		
2023.09.06	9:30	18.6	18.9	18.7		
2023.09.07	9:30	18.6	18.8	18.8		
2023.09.08	9:30	18.6	18.8	18.7		
2023.09.09						
2023.09.10						
2023.09.11	9:30	18.6	18.9	18.8		
2023.09.12	9:30	18.6	18.8	18.7		
2023.09.13	9:30	18.6	18.9	18.7		
2023.09.14	9:30	18.6	18.8	18.6		
2023.09.15	9:30	18.6	18.8	18.7		
2023.09.16						
2023.09.17						
2023.09.18	9:30	18.6	18.9	18.7		
2023.09.19	9:30	18.6	18.8	18.7		
2023.09.20	9:30	18.6	18.8	18.7		
2023.09.21	9:30	18.6	18.8	18.7		
2023.09.22	9:30	18.6	18.9	18.8		
2023.09.23						
2023.09.24						
2023.09.25	9:30	18.6	18.9	18.7		
2023.09.26	9:30	18.6	18.8	18.7		
2023.09.27	9:30	18.6	18.8	18.7		
2023.09.28						추석연휴
2023.09.29						추석
2023.09.30						

date of copying

sign