

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2023.10
	Model Name	MPR-720R	Investigational Product	25℃이하 15~25℃ 1~25℃
	Serial Number	07120004		
Calibration Subject	Description	Thermometer		
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)		
	Certificate No.	100397082022103	Temperature Range	16~22℃(19±3℃)

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2023.10.01						
2023.10.02						임시공휴일
2023.10.03						개천절
2023.10.04	P:30	18.6	18.9	18.8		
2023.10.05	P:30	18.6	18.8	18.7		
2023.10.06	P:30	18.6	18.8	18.7		
2023.10.07						
2023.10.08						
2023.10.09						한글날
2023.10.10	P:30	18.6	18.8	18.7		
2023.10.11	P:30	18.6	18.8	18.7		
2023.10.12	P:40	18.6	18.8	18.7		
2023.10.13	P:30	18.6	18.8	18.7		
2023.10.14						
2023.10.15						
2023.10.16	P:30	18.5	18.8	18.7		
2023.10.17	P:30	18.5	18.8	18.6		
2023.10.18	P:30	18.6	18.8	18.7		
2023.10.19	P:45	18.6	18.8	18.6		
2023.10.20	P:30	18.5	18.8	18.7		
2023.10.21						
2023.10.22						
2023.10.23	P:30	18.5	18.8	18.7		
2023.10.24	P:30	18.6	18.8	18.7		
2023.10.25	P:30	18.5	18.8	18.6		
2023.10.26	P:30	18.5	18.8	18.6		
2023.10.27	10:30	18.5	18.8	18.6		
2023.10.28						
2023.10.29						
2023.10.30	P:30	18.5	18.8	18.6		
2023.10.31	P:30	18.6	18.8	18.7		

date of copying

sign