

## 부록6. 서식II-01-06, 온도기록지(IP 보관)

## IP STORAGE TEMPERATURE LOG

|                     |                         |                     |                         |           |
|---------------------|-------------------------|---------------------|-------------------------|-----------|
| Storage Subject     | Description             | 약품냉장고               | Storage Space           | 냉장고       |
|                     | Manufacturer            | Giant Star          | Storage Period          | 2023.12   |
|                     | Model Name              | CBR1000SS           | Investigational Product | 냉장보관대상 IP |
|                     | Serial Number           | GS17090102-10       |                         |           |
| Calibration Subject | Probe Position          | Left                |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 848811              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-5          |                         |           |
|                     | Probe Position          | Right               |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 855630              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-6          | Temperature Range       | 2~8℃      |

| Date       | Time | Probe Position | Temperature(℃) |     |         | Verified by | Comment |
|------------|------|----------------|----------------|-----|---------|-------------|---------|
|            |      |                | Min            | Max | Current |             |         |
| 2023.12.01 | 9:30 | Left           | 2              | 4   | 3       | ml          |         |
|            |      | Right          | 3              | 5   | 4       |             |         |
| 2023.12.02 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |
| 2023.12.03 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |
| 2023.12.04 | 9:30 | Left           | 2              | 4   | 4       | 김복현         |         |
|            |      | Right          | 2              | 4   | 4       |             |         |
| 2023.12.05 | 9:30 | Left           | 2              | 4   | 3       | ml          |         |
|            |      | Right          | 2              | 5   | 4       |             |         |
| 2023.12.06 | 9:30 | Left           | 2              | 4   | 3       | ml          |         |
|            |      | Right          | 3              | 5   | 4       |             |         |
| 2023.12.07 | 9:30 | Left           | 2              | 4   | 3       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |
| 2023.12.08 | 9:30 | Left           | 2              | 4   | 3       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |
| 2023.12.09 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |
| 2023.12.10 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |
| 2023.12.11 | 9:30 | Left           | 2              | 4   | 4       | ml          |         |
|            |      | Right          | 3              | 5   | 4       |             |         |
| 2023.12.12 | 9:30 | Left           | 2              | 4   | 4       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |
| 2023.12.13 | 9:30 | Left           | 2              | 4   | 4       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |

부록6. 서식II-01-06, 온도기록지(IP 보관)

## IP STORAGE TEMPERATURE LOG

|                     |                         |                     |                         |           |
|---------------------|-------------------------|---------------------|-------------------------|-----------|
| Storage Subject     | Description             | 약품냉장고               | Storage Space           | 냉장고       |
|                     | Manufacturer            | Giant Star          | Storage Period          | 2023.12   |
|                     | Model Name              | CBR1000SS           | Investigational Product | 냉장보관대상 IP |
|                     | Serial Number           | GS17090102-10       |                         |           |
| Calibration Subject | Probe Position          | Left                |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 848811              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-5          |                         |           |
|                     | Probe Position          | Right               |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 855630              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-6          | Temperature Range       | 2~8℃      |

| Date       | Time  | Probe Position | Temperature(℃) |     |         | Verified by | Comment |
|------------|-------|----------------|----------------|-----|---------|-------------|---------|
|            |       |                | Min            | Max | Current |             |         |
| 2023.12.14 | 9:30  | Left           | 2              | 4   | 3       | [Signature] |         |
|            |       | Right          | 3              | 5   | 4       |             |         |
| 2023.12.15 | 9:30  | Left           | 2              | 4   | 4       | [Signature] |         |
|            |       | Right          | 3              | 4   | 4       |             |         |
| 2023.12.16 |       | Left           |                |     |         |             |         |
|            |       | Right          |                |     |         |             |         |
| 2023.12.17 |       | Left           |                |     |         |             |         |
|            |       | Right          |                |     |         |             |         |
| 2023.12.18 | 9:30  | Left           | 2              | 4   | 3       | [Signature] |         |
|            |       | Right          | 3              | 4   | 4       |             |         |
| 2023.12.19 | 9:30  | Left           | 2              | 4   | 3       | [Signature] |         |
|            |       | Right          | 3              | 4   | 4       |             |         |
| 2023.12.20 | 9:30  | Left           | 2              | 4   | 4       | [Signature] |         |
|            |       | Right          | 2              | 4   | 4       |             |         |
| 2023.12.21 | 10:00 | Left           | 2              | 4   | 3       | [Signature] |         |
|            |       | Right          | 2              | 4   | 3       |             |         |
| 2023.12.22 | 9:30  | Left           | 2              | 4   | 3       | [Signature] |         |
|            |       | Right          | 2              | 4   | 3       |             |         |
| 2023.12.23 |       | Left           |                |     |         |             |         |
|            |       | Right          |                |     |         |             |         |
| 2023.12.24 |       | Left           |                |     |         |             |         |
|            |       | Right          |                |     |         |             |         |
| 2023.12.25 |       | Left           |                |     |         |             | 성탄절     |
|            |       | Right          |                |     |         |             |         |
| 2023.12.26 | 9:30  | Left           | 3              | 4   | 3       | [Signature] |         |
|            |       | Right          | 2              | 4   | 3       |             |         |

## 부록6. 서식II-01-06, 온도기록지(IP 보관)

## IP STORAGE TEMPERATURE LOG

|                     |                         |                     |                         |           |
|---------------------|-------------------------|---------------------|-------------------------|-----------|
| Storage Subject     | Description             | 약품냉장고               | Storage Space           | 냉장고       |
|                     | Manufacturer            | Giant Star          | Storage Period          | 2023.12   |
|                     | Model Name              | CBR1000SS           | Investigational Product | 냉장보관대상 IP |
|                     | Serial Number           | GS17090102-10       |                         |           |
| Calibration Subject | Probe Position          | Left                |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 848811              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-5          |                         |           |
|                     | Probe Position          | Right               |                         |           |
|                     | Description             | Digital Thermometer |                         |           |
|                     | Manufacturer            | SATO                |                         |           |
|                     | Model Name              | PC-3500             |                         |           |
|                     | Serial Number           | 855630              |                         |           |
|                     | Date of Calibration     | 2023.3.27           |                         |           |
|                     | Date of Calibration Due | 2024.3.27           |                         |           |
|                     | Certificate No.         | 23007746-6          | Temperature Range       | 2~8℃      |

| Date       | Time | Probe Position | Temperature(℃) |     |         | Verified by | Comment |
|------------|------|----------------|----------------|-----|---------|-------------|---------|
|            |      |                | Min            | Max | Current |             |         |
| 2023.12.27 | P:30 | Left           | 3              | 4   | 4       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |
| 2023.12.28 | P:30 | Left           | 3              | 4   | 3       | ml          |         |
|            |      | Right          | 3              | 4   | 3       |             |         |
| 2023.12.29 | P:30 | Left           | 3              | 4   | 4       | ml          |         |
|            |      | Right          | 3              | 4   | 4       |             |         |
| 2023.12.30 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |
| 2023.12.31 |      | Left           |                |     |         |             |         |
|            |      | Right          |                |     |         |             |         |

<Min/Max>  
 ① PC-3500, 12238-02, 온습도계 : 온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨.  
 ② 냉동고 : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.  
 ③ testo : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행. <testo 온도이탈 알람 설정값> 실온 : <15.0℃, >25.0℃, 즉시 / 냉장 : <2.0℃, >8.0℃, 즉시

date of copying

sign

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 3.6 DV: -0.4  
 22:08 SP: 4.0 PV: 4.2 DV: 0.2  
 21:08 SP: 4.0 PV: 3.8 DV: -0.2  
 20:08 SP: 4.0 PV: 4.1 DV: 0.1  
 19:08 SP: 4.0 PV: 3.9 DV: -0.1  
 18:08 SP: 4.0 PV: 3.7 DV: -0.3  
 17:08 SP: 4.0 PV: 4.3 DV: 0.3  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 3.6 DV: -0.4  
 14:08 SP: 4.0 PV: 4.2 DV: 0.2  
 13:08 SP: 4.0 PV: 3.9 DV: -0.1  
 12:08 SP: 4.0 PV: 3.6 DV: -0.4  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 3.9 DV: -0.1  
 09:08 SP: 4.0 PV: 3.7 DV: -0.3  
 08:08 SP: 4.0 PV: 4.2 DV: 0.2  
 07:08 SP: 4.0 PV: 3.8 DV: -0.2  
 06:08 SP: 4.0 PV: 3.6 DV: -0.4  
 05:08 SP: 4.0 PV: 4.2 DV: 0.2  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 4.1 DV: 0.1  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 3.7 DV: -0.3  
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-12-02

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 3.6 DV: -0.4  
 21:08 SP: 4.0 PV: 4.2 DV: 0.2  
 20:08 SP: 4.0 PV: 3.9 DV: -0.1  
 19:08 SP: 4.0 PV: 4.0 DV: 0.0  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 3.9 DV: -0.1  
 14:08 SP: 4.0 PV: 3.7 DV: -0.3  
 13:08 SP: 4.0 PV: 3.6 DV: -0.4  
 12:08 SP: 4.0 PV: 3.6 DV: -0.4  
 11:08 SP: 4.0 PV: 3.4 DV: -0.6  
 10:08 SP: 4.0 PV: 4.2 DV: 0.2  
 09:08 SP: 4.0 PV: 3.8 DV: -0.2  
 08:08 SP: 4.0 PV: 3.7 DV: -0.3  
 07:08 SP: 4.0 PV: 4.2 DV: 0.2  
 06:08 SP: 4.0 PV: 3.8 DV: -0.2  
 05:08 SP: 4.0 PV: 4.0 DV: 0.0  
 04:08 SP: 4.0 PV: 3.8 DV: -0.2  
 03:08 SP: 4.0 PV: 3.6 DV: -0.4  
 02:08 SP: 4.0 PV: 4.2 DV: 0.2  
 01:08 SP: 4.0 PV: 3.8 DV: -0.2  
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2023-12-01

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 4.1 DV: 0.1  
 21:08 SP: 4.0 PV: 3.8 DV: -0.2  
 20:08 SP: 4.0 PV: 4.1 DV: 0.1  
 19:08 SP: 4.0 PV: 3.9 DV: -0.1  
 18:08 SP: 4.0 PV: 4.0 DV: 0.0  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 3.7 DV: -0.3  
 15:08 SP: 4.0 PV: 4.3 DV: 0.3  
 14:08 SP: 4.0 PV: 3.8 DV: -0.2  
 13:08 SP: 4.0 PV: 4.1 DV: 0.1  
 12:08 SP: 4.0 PV: 3.9 DV: -0.1  
 11:08 SP: 4.0 PV: 3.7 DV: -0.3  
 10:08 SP: 4.0 PV: 4.2 DV: 0.2  
 09:08 SP: 4.0 PV: 3.8 DV: -0.2  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.0 DV: 0.0  
 05:08 SP: 4.0 PV: 3.7 DV: -0.3  
 04:08 SP: 4.0 PV: 4.3 DV: 0.3  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.1 DV: 0.1  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2023-12-04

23:08 SP: 4.0 PV: 3.9 DV: -0.1  
 22:08 SP: 4.0 PV: 4.0 DV: 0.0  
 21:08 SP: 4.0 PV: 4.0 DV: 0.0  
 20:08 SP: 4.0 PV: 4.0 DV: 0.0  
 19:08 SP: 4.0 PV: 3.9 DV: -0.1  
 18:08 SP: 4.0 PV: 4.0 DV: 0.0  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 4.0 DV: 0.0  
 13:08 SP: 4.0 PV: 3.9 DV: -0.1  
 12:08 SP: 4.0 PV: 4.0 DV: 0.0  
 11:08 SP: 4.0 PV: 4.0 DV: 0.0  
 10:08 SP: 4.0 PV: 3.9 DV: -0.1  
 09:08 SP: 4.0 PV: 4.0 DV: 0.0  
 08:08 SP: 4.0 PV: 3.7 DV: -0.3

3

07:08 SP: 4.0 PV: 4.3 DV: 0.3  
 06:08 SP: 4.0 PV: 3.8 DV: -0.2  
 05:08 SP: 4.0 PV: 3.6 DV: -0.4  
 04:08 SP: 4.0 PV: 4.2 DV: 0.2  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 3.6 DV: -0.4  
 01:08 SP: 4.0 PV: 4.2 DV: 0.2  
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2023-12-03

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 4.3 DV: 0.3  
 21:08 SP: 4.0 PV: 3.9 DV: -0.1  
 20:08 SP: 4.0 PV: 4.0 DV: 0.0  
 19:08 SP: 4.0 PV: 4.0 DV: 0.0  
 18:08 SP: 4.0 PV: 3.7 DV: -0.3  
 17:08 SP: 4.0 PV: 4.3 DV: 0.3  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 4.1 DV: 0.1  
 14:08 SP: 4.0 PV: 3.8 DV: -0.2  
 13:08 SP: 4.0 PV: 4.1 DV: 0.1  
 12:08 SP: 4.0 PV: 3.8 DV: -0.2  
 11:08 SP: 4.0 PV: 3.9 DV: -0.1  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 3.9 DV: -0.1

1

07:08 SP: 4.0 PV: 3.9 DV: -0.

2

06:08 SP: 4.0 PV: 3.8 DV: -0.  
 05:08 SP: 4.0 PV: 4.1 DV: 0.1  
 04:08 SP: 4.0 PV: 3.8 DV: -0.2  
 03:08 SP: 4.0 PV: 4.0 DV: 0.0  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 3.7 DV: -0.3  
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2023-12-06

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 3.6 DV: -0.4  
 21:08 SP: 4.0 PV: 4.2 DV: 0.2  
 20:08 SP: 4.0 PV: 3.8 DV: -0.2  
 19:08 SP: 4.0 PV: 4.0 DV: 0.0  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 3.7 DV: -0.3  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 3.9 DV: -0.1  
 13:08 SP: 4.0 PV: 3.7 DV: -0.3  
 12:08 SP: 4.0 PV: 3.6 DV: -0.4  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 3.7 DV: -0.3  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.0 DV: 0.0  
 05:08 SP: 4.0 PV: 3.8 DV: -0.2  
 04:08 SP: 4.0 PV: 4.1 DV: 0.1

2

03:08 SP: 4.0 PV: 3.8 DV: -0.  
 02:08 SP: 4.0 PV: 4.1 DV: 0.1  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2023-12-05

23:08 SP: 4.0 PV: 3.9 DV: -0.1  
 22:08 SP: 4.0 PV: 3.8 DV: -0.2  
 21:08 SP: 4.0 PV: 4.0 DV: 0.0  
 20:08 SP: 4.0 PV: 3.8 DV: -0.2  
 19:08 SP: 4.0 PV: 4.0 DV: 0.0  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 3.6 DV: -0.4  
 14:08 SP: 4.0 PV: 4.0 DV: 0.0

1

13:08 SP: 4.0 PV: 3.9 DV: -0.  
 12:08 SP: 4.0 PV: 3.9 DV: -0.1  
 11:08 SP: 4.0 PV: 3.7 DV: -0.3  
 10:08 SP: 4.0 PV: 4.2 DV: 0.2  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 4.0 DV: 0.0  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.1 DV: 0.1  
 05:08 SP: 4.0 PV: 4.0 DV: 0.0  
 04:08 SP: 4.0 PV: 3.7 DV: -0.3  
 03:08 SP: 4.0 PV: 4.2 DV: 0.2  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 4.1 DV: 0.1  
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2023-12-08

23:08 SP: 4.0 PV: 4.3 DV: 0.3

1

22:08 SP: 4.0 PV: 3.9 DV: -0.  
 21:08 SP: 4.0 PV: 3.9 DV: -0.1  
 20:08 SP: 4.0 PV: 3.7 DV: -0.3  
 19:08 SP: 4.0 PV: 3.6 DV: -0.4  
 18:08 SP: 4.0 PV: 3.6 DV: -0.4  
 17:08 SP: 4.0 PV: 3.6 DV: -0.4  
 16:08 SP: 4.0 PV: 3.6 DV: -0.4  
 15:08 SP: 4.0 PV: 3.7 DV: -0.3

4

14:08 SP: 4.0 PV: 3.6 DV: -0.  
 13:08 SP: 4.0 PV: 3.4 DV: -0.6  
 12:08 SP: 4.0 PV: 3.6 DV: -0.4

1

11:08 SP: 4.0 PV: 3.9 DV: -0.  
 10:08 SP: 4.0 PV: 4.1 DV: 0.1  
 09:08 SP: 4.0 PV: 3.7 DV: -0.3  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 3.8 DV: -0.2  
 06:08 SP: 4.0 PV: 3.6 DV: -0.4  
 05:08 SP: 4.0 PV: 4.2 DV: 0.2  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 3.6 DV: -0.4

2

02:08 SP: 4.0 PV: 4.2 DV: 0.  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2023-12-07

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.2 DV: 0.2  
 22:08 SP: 4.0 PV: 4.4 DV: 0.4  
 21:08 SP: 4.0 PV: 4.6 DV: 0.6  
 20:08 SP: 4.0 PV: 4.4 DV: 0.4  
 19:08 SP: 4.0 PV: 3.7 DV: -0.3  
 18:08 SP: 4.0 PV: 4.2 DV: 0.2  
 17:08 SP: 4.0 PV: 4.1 DV: 0.1  
 16:08 SP: 4.0 PV: 4.1 DV: 0.1  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 3.7 DV: -0.3  
 13:08 SP: 4.0 PV: 4.2 DV: 0.2  
 12:08 SP: 4.0 PV: 3.9 DV: -0.1  
 11:08 SP: 4.0 PV: 3.7 DV: -0.3  
 10:08 SP: 4.0 PV: 4.3 DV: 0.3  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 3.6 DV: -0.4  
 05:08 SP: 4.0 PV: 4.3 DV: 0.3  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 4.4 DV: 0.4  
 02:08 SP: 4.0 PV: 3.8 DV: -0.2  
 01:08 SP: 4.0 PV: 4.2 DV: 0.2  
 00:08 SP: 4.0 PV: 4.0 DV: 0.0  
 2023-12-10  
 23:08 SP: 4.0 PV: 4.4 DV: 0.4  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.3 DV: 0.3  
 20:08 SP: 4.0 PV: 3.8 DV: -0.2  
 19:08 SP: 4.0 PV: 4.3 DV: 0.3  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 4.1 DV: 0.1  
 16:08 SP: 4.0 PV: 4.0 DV: 0.0  
 15:08 SP: 4.0 PV: 3.7 DV: -0.3  
 14:08 SP: 4.0 PV: 4.2 DV: 0.2  
 13:08 SP: 4.0 PV: 3.9 DV: -0.1  
 12:08 SP: 4.0 PV: 4.0 DV: 0.0  
 11:08 SP: 4.0 PV: 3.7 DV: -0.3  
 10:08 SP: 4.0 PV: 4.3 DV: 0.3  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 4.4 DV: 0.4  
 07:08 SP: 4.0 PV: 3.8 DV: -0.2  
 06:08 SP: 4.0 PV: 4.3 DV: 0.3  
 05:08 SP: 4.0 PV: 3.8 DV: -0.2  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 3.8 DV: -0.2  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 3.8 DV: -0.2  
 00:08 SP: 4.0 PV: 3.9 DV: -0.1  
 2023-12-09

23:08 SP: 4.0 PV: 4.6 DV: 0.6  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.3 DV: 0.3  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 4.5 DV: 0.5  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 4.5 DV: 0.5  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 4.4 DV: 0.4  
 14:08 SP: 4.0 PV: 3.9 DV: -0.1  
 13:08 SP: 4.0 PV: 4.2 DV: 0.2  
 12:08 SP: 4.0 PV: 3.8 DV: -0.2  
 11:08 SP: 4.0 PV: 4.3 DV: 0.3  
 10:08 SP: 4.0 PV: 3.9 DV: -0.1  
 09:08 SP: 4.0 PV: 4.1 DV: 0.1  
 08:08 SP: 4.0 PV: 4.3 DV: 0.3  
 07:08 SP: 4.0 PV: 3.8 DV: -0.2  
 06:08 SP: 4.0 PV: 4.4 DV: 0.4  
 05:08 SP: 4.0 PV: 3.9 DV: -0.1  
 04:08 SP: 4.0 PV: 4.4 DV: 0.4  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.2 DV: 0.2  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 4.0 DV: 0.0  
 2023-12-12  
 23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 3.8 DV: -0.2  
 20:08 SP: 4.0 PV: 3.9 DV: -0.1  
 19:08 SP: 4.0 PV: 3.8 DV: -0.2  
 18:08 SP: 4.0 PV: 3.8 DV: -0.2  
 17:08 SP: 4.0 PV: 4.0 DV: 0.0  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 3.7 DV: -0.3  
 14:08 SP: 4.0 PV: 3.7 DV: -0.3  
 13:08 SP: 4.0 PV: 3.6 DV: -0.4  
 12:08 SP: 4.0 PV: 3.6 DV: -0.4  
 11:08 SP: 4.0 PV: 4.1 DV: 0.1  
 10:08 SP: 4.0 PV: 3.9 DV: -0.1  
 09:08 SP: 4.0 PV: 4.1 DV: 0.1  
 08:08 SP: 4.0 PV: 3.8 DV: -0.2  
 07:08 SP: 4.0 PV: 4.1 DV: 0.1  
 06:08 SP: 4.0 PV: 3.9 DV: -0.1  
 05:08 SP: 4.0 PV: 3.8 DV: -0.2  
 04:08 SP: 4.0 PV: 4.2 DV: 0.2  
 03:08 SP: 4.0 PV: 4.1 DV: 0.1  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 4.3 DV: 0.3  
 00:08 SP: 4.0 PV: 3.8 DV: -0.2  
 2023-12-11

SP        Setting Point  
 PV        Present Value  
 DV        Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.1 DV: 0.1  
 22:08 SP: 4.0 PV: 4.4 DV: 0.4  
 21:08 SP: 4.0 PV: 4.0 DV: 0.0  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 4.1 DV: 0.1  
 18:08 SP: 4.0 PV: 4.1 DV: 0.1  
 17:08 SP: 4.0 PV: 4.5 DV: 0.5  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 3.9 DV: -0.1  
 14:08 SP: 4.0 PV: 3.9 DV: -0.1  
 13:08 SP: 4.0 PV: 3.4 DV: -0.6  
 12:08 SP: 4.0 PV: 4.2 DV: 0.2  
 11:08 SP: 4.0 PV: 3.7 DV: -0.3  
 10:08 SP: 4.0 PV: 4.3 DV: 0.3  
 09:08 SP: 4.0 PV: 3.8 DV: -0.2  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.1 DV: 0.1  
 05:08 SP: 4.0 PV: 4.0 DV: 0.0  
 04:08 SP: 4.0 PV: 4.0 DV: 0.0  
 03:08 SP: 4.0 PV: 3.7 DV: -0.3  
 02:08 SP: 4.0 PV: 4.2 DV: 0.2  
 01:08 SP: 4.0 PV: 3.8 DV: -0.2  
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-12-14

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 4.2 DV: 0.2  
 21:08 SP: 4.0 PV: 3.8 DV: -0.2  
 20:08 SP: 4.0 PV: 4.2 DV: 0.2  
 19:08 SP: 4.0 PV: 3.9 DV: -0.1

3

18:08 SP: 4.0 PV: 4.3 DV: 0.1  
 17:08 SP: 4.0 PV: 3.9 DV: -0.1  
 16:08 SP: 4.0 PV: 4.4 DV: 0.4  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 4.4 DV: 0.4  
 13:08 SP: 4.0 PV: 3.8 DV: -0.2  
 12:08 SP: 4.0 PV: 4.4 DV: 0.4  
 11:08 SP: 4.0 PV: 3.9 DV: -0.1  
 10:08 SP: 4.0 PV: 4.2 DV: 0.2  
 09:08 SP: 4.0 PV: 4.5 DV: 0.5  
 08:08 SP: 4.0 PV: 4.0 DV: 0.0  
 07:08 SP: 4.0 PV: 4.4 DV: 0.4  
 06:08 SP: 4.0 PV: 3.9 DV: -0.1  
 05:08 SP: 4.0 PV: 4.3 DV: 0.3  
 04:08 SP: 4.0 PV: 4.6 DV: 0.6  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.2 DV: 0.2  
 01:08 SP: 4.0 PV: 4.0 DV: 0.0  
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2023-12-13

23:08 SP: 4.0 PV: 4.3 DV: 0.3  
 22:08 SP: 4.0 PV: 4.1 DV: 0.1  
 21:08 SP: 4.0 PV: 4.2 DV: 0.2  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 4.0 DV: 0.0  
 18:08 SP: 4.0 PV: 4.4 DV: 0.4  
 17:08 SP: 4.0 PV: 4.5 DV: 0.5  
 16:08 SP: 4.0 PV: 4.4 DV: 0.4  
 15:08 SP: 4.0 PV: 3.8 DV: -0.2  
 14:08 SP: 4.0 PV: 4.2 DV: 0.2  
 13:08 SP: 4.0 PV: 4.0 DV: 0.0  
 12:08 SP: 4.0 PV: 4.3 DV: 0.3  
 11:08 SP: 4.0 PV: 4.5 DV: 0.5  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 3.7 DV: -0.3  
 08:08 SP: 4.0 PV: 4.2 DV: 0.2  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.0 DV: 0.0  
 05:08 SP: 4.0 PV: 3.9 DV: -0.1  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.1 DV: 0.1  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2023-12-16

23:08 SP: 4.0 PV: 4.2 DV: 0.2  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.3 DV: 0.3  
 20:08 SP: 4.0 PV: 4.0 DV: 0.0  
 19:08 SP: 4.0 PV: 4.3 DV: 0.3  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 4.0 DV: 0.0  
 16:08 SP: 4.0 PV: 4.1 DV: 0.1  
 15:08 SP: 4.0 PV: 4.4 DV: 0.4  
 14:08 SP: 4.0 PV: 3.8 DV: -0.2  
 13:08 SP: 4.0 PV: 4.3 DV: 0.3  
 12:08 SP: 4.0 PV: 4.0 DV: 0.0  
 11:08 SP: 4.0 PV: 4.4 DV: 0.4  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 4.2 DV: 0.2  
 08:08 SP: 4.0 PV: 4.1 DV: 0.1  
 07:08 SP: 4.0 PV: 4.4 DV: 0.4  
 06:08 SP: 4.0 PV: 3.9 DV: -0.1  
 05:08 SP: 4.0 PV: 4.2 DV: 0.2  
 04:08 SP: 4.0 PV: 4.0 DV: 0.0  
 03:08 SP: 4.0 PV: 4.1 DV: 0.1  
 02:08 SP: 4.0 PV: 4.4 DV: 0.4  
 01:08 SP: 4.0 PV: 3.8 DV: -0.2  
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2023-12-15

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.0 DV: 0.0  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.0 DV: 0.0  
 20:08 SP: 4.0 PV: 3.7 DV: -0.3  
 19:08 SP: 4.0 PV: 4.2 DV: 0.2  
 18:08 SP: 4.0 PV: 3.7 DV: -0.3  
 17:08 SP: 4.0 PV: 4.3 DV: 0.3  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 4.3 DV: 0.3  
 14:08 SP: 4.0 PV: 3.8 DV: -0.2  
 13:08 SP: 4.0 PV: 4.0 DV: 0.0  
 12:08 SP: 4.0 PV: 4.1 DV: 0.1  
 11:08 SP: 4.0 PV: 3.9 DV: -0.1  
 10:08 SP: 4.0 PV: 4.3 DV: 0.3  
 09:08 SP: 4.0 PV: 3.7 DV: -0.3  
 08:08 SP: 4.0 PV: 4.2 DV: 0.2  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.4 DV: 0.4  
 05:08 SP: 4.0 PV: 3.9 DV: -0.1  
 04:08 SP: 4.0 PV: 4.2 DV: 0.2  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.1 DV: 0.1  
 01:08 SP: 4.0 PV: 3.7 DV: -0.3  
 00:08 SP: 4.0 PV: 4.2 DV: 0.2  
 2023-12-18  
 23:08 SP: 4.0 PV: 4.0 DV: 0.0  
 22:08 SP: 4.0 PV: 4.1 DV: 0.1  
 21:08 SP: 4.0 PV: 4.1 DV: 0.1  
 20:08 SP: 4.0 PV: 4.4 DV: 0.4  
 19:08 SP: 4.0 PV: 3.8 DV: -0.2  
 18:08 SP: 4.0 PV: 4.2 DV: 0.2  
 17:08 SP: 4.0 PV: 4.0 DV: 0.0  
 4  
 16:08 SP: 4.0 PV: 4.4 DV: 0.  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 4.2 DV: 0.2  
 13:08 SP: 4.0 PV: 4.1 DV: 0.1  
 12:08 SP: 4.0 PV: 4.2 DV: 0.2  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 4.3 DV: 0.3  
 09:08 SP: 4.0 PV: 4.3 DV: 0.3  
 08:08 SP: 4.0 PV: 4.4 DV: 0.4  
 07:08 SP: 4.0 PV: 4.5 DV: 0.5  
 06:08 SP: 4.0 PV: 4.0 DV: 0.0  
 05:08 SP: 4.0 PV: 4.2 DV: 0.2  
 04:08 SP: 4.0 PV: 4.3 DV: 0.3  
 03:08 SP: 4.0 PV: 4.3 DV: 0.3  
 02:08 SP: 4.0 PV: 4.3 DV: 0.3  
 01:08 SP: 4.0 PV: 4.2 DV: 0.2  
 00:08 SP: 4.0 PV: 4.3 DV: 0.3  
 2023-12-17

23:08 SP: 4.0 PV: 3.8 DV: -0.2  
 22:08 SP: 4.0 PV: 4.2 DV: 0.2  
 21:08 SP: 4.0 PV: 3.8 DV: -0.2  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 3.8 DV: -0.2  
 18:08 SP: 4.0 PV: 4.3 DV: 0.3  
 17:08 SP: 4.0 PV: 3.8 DV: -0.2  
 16:08 SP: 4.0 PV: 4.4 DV: 0.4  
 15:08 SP: 4.0 PV: 3.9 DV: -0.1  
 14:08 SP: 4.0 PV: 4.1 DV: 0.1  
 13:08 SP: 4.0 PV: 4.0 DV: 0.0  
 12:08 SP: 4.0 PV: 4.0 DV: 0.0  
 11:08 SP: 4.0 PV: 4.5 DV: 0.5  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 4.3 DV: 0.3  
 08:08 SP: 4.0 PV: 3.8 DV: -0.2  
 07:08 SP: 4.0 PV: 4.3 DV: 0.3  
 06:08 SP: 4.0 PV: 3.9 DV: -0.1  
 5  
 05:08 SP: 4.0 PV: 4.5 DV: 0.  
 04:08 SP: 4.0 PV: 3.9 DV: -0.1  
 03:08 SP: 4.0 PV: 4.3 DV: 0.3  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 4.5 DV: 0.5  
 00:08 SP: 4.0 PV: 3.9 DV: -0.1  
 2023-12-20  
 23:08 SP: 4.0 PV: 4.4 DV: 0.4  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.4 DV: 0.4  
 20:08 SP: 4.0 PV: 3.9 DV: -0.1  
 19:08 SP: 4.0 PV: 4.4 DV: 0.4  
 18:08 SP: 4.0 PV: 3.9 DV: -0.1  
 17:08 SP: 4.0 PV: 4.5 DV: 0.5  
 16:08 SP: 4.0 PV: 3.8 DV: -0.2  
 15:08 SP: 4.0 PV: 4.3 DV: 0.3  
 14:08 SP: 4.0 PV: 3.9 DV: -0.1  
 13:08 SP: 4.0 PV: 4.4 DV: 0.4  
 12:08 SP: 4.0 PV: 3.7 DV: -0.3  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 3.9 DV: -0.1  
 09:08 SP: 4.0 PV: 4.1 DV: 0.1  
 08:08 SP: 4.0 PV: 3.9 DV: -0.1  
 07:08 SP: 4.0 PV: 3.7 DV: -0.3  
 06:08 SP: 4.0 PV: 4.2 DV: 0.2  
 05:08 SP: 4.0 PV: 3.9 DV: -0.1  
 04:08 SP: 4.0 PV: 4.1 DV: 0.1  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.1 DV: 0.1  
 01:08 SP: 4.0 PV: 4.0 DV: 0.0  
 00:08 SP: 4.0 PV: 4.0 DV: 0.0  
 2023-12-19

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.6 DV: 0.6  
 22:08 SP: 4.0 PV: 4.7 DV: 0.7  
 21:08 SP: 4.0 PV: 4.7 DV: 0.7  
 20:08 SP: 4.0 PV: 4.7 DV: 0.7  
 19:08 SP: 4.0 PV: 4.6 DV: 0.6  
 18:08 SP: 4.0 PV: 4.5 DV: 0.5  
 17:08 SP: 4.0 PV: 4.4 DV: 0.4  
 16:08 SP: 4.0 PV: 4.3 DV: 0.3  
 15:08 SP: 4.0 PV: 4.2 DV: 0.2  
 14:08 SP: 4.0 PV: 4.0 DV: 0.0  
 13:08 SP: 4.0 PV: 4.4 DV: 0.4  
 12:08 SP: 4.0 PV: 4.6 DV: 0.6  
 11:08 SP: 4.0 PV: 4.7 DV: 0.7  
 10:08 SP: 4.0 PV: 4.7 DV: 0.7  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 3.9 DV: -0.1  
 07:08 SP: 4.0 PV: 4.4 DV: 0.4  
 06:08 SP: 4.0 PV: 4.6 DV: 0.6  
 05:08 SP: 4.0 PV: 4.7 DV: 0.7  
 04:08 SP: 4.0 PV: 4.7 DV: 0.7  
 03:08 SP: 4.0 PV: 4.6 DV: 0.6  
 02:08 SP: 4.0 PV: 4.6 DV: 0.6  
 01:08 SP: 4.0 PV: 4.4 DV: 0.4  
 00:08 SP: 4.0 PV: 4.2 DV: 0.2  
 2023-12-22  
 23:08 SP: 4.0 PV: 4.1 DV: 0.1  
 22:08 SP: 4.0 PV: 4.4 DV: 0.4  
 21:08 SP: 4.0 PV: 4.6 DV: 0.6  
 20:08 SP: 4.0 PV: 4.7 DV: 0.7  
 19:08 SP: 4.0 PV: 4.7 DV: 0.7  
 18:08 SP: 4.0 PV: 4.6 DV: 0.6  
 17:08 SP: 4.0 PV: 4.6 DV: 0.6  
 16:08 SP: 4.0 PV: 4.5 DV: 0.5  
 5  
 15:08 SP: 4.0 PV: 4.5 DV: 0.  
 14:08 SP: 4.0 PV: 4.4 DV: 0.4  
 13:08 SP: 4.0 PV: 4.3 DV: 0.3  
 12:08 SP: 4.0 PV: 4.2 DV: 0.2  
 5  
 11:08 SP: 4.0 PV: 4.5 DV: 0.  
 10:08 SP: 4.0 PV: 4.7 DV: 0.7  
 09:08 SP: 4.0 PV: 3.8 DV: -0.2  
 08:08 SP: 4.0 PV: 4.3 DV: 0.3  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.4 DV: 0.4  
 05:08 SP: 4.0 PV: 3.9 DV: -0.1  
 04:08 SP: 4.0 PV: 4.4 DV: 0.4  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.4 DV: 0.4  
 01:08 SP: 4.0 PV: 3.9 DV: -0.1  
 00:08 SP: 4.0 PV: 4.4 DV: 0.4  
 2023-12-21

23:08 SP: 4.0 PV: 4.6 DV: 0.6  
 22:08 SP: 4.0 PV: 4.2 DV: 0.2  
 21:08 SP: 4.0 PV: 4.6 DV: 0.6  
 20:08 SP: 4.0 PV: 4.7 DV: 0.7  
 19:08 SP: 4.0 PV: 4.3 DV: 0.3  
 18:08 SP: 4.0 PV: 4.6 DV: 0.6  
 17:08 SP: 4.0 PV: 4.7 DV: 0.7  
 16:08 SP: 4.0 PV: 4.7 DV: 0.7  
 15:08 SP: 4.0 PV: 4.6 DV: 0.6  
 14:08 SP: 4.0 PV: 4.2 DV: 0.2  
 13:08 SP: 4.0 PV: 4.0 DV: 0.0  
 12:08 SP: 4.0 PV: 4.5 DV: 0.5  
 11:08 SP: 4.0 PV: 4.7 DV: 0.7  
 10:08 SP: 4.0 PV: 4.6 DV: 0.6  
 09:08 SP: 4.0 PV: 4.2 DV: 0.2  
 08:08 SP: 4.0 PV: 4.6 DV: 0.6  
 07:08 SP: 4.0 PV: 4.6 DV: 0.6  
 06:08 SP: 4.0 PV: 4.3 DV: 0.3  
 05:08 SP: 4.0 PV: 4.6 DV: 0.6  
 04:08 SP: 4.0 PV: 4.7 DV: 0.7  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 3.7 DV: -0.3  
 01:08 SP: 4.0 PV: 4.4 DV: 0.4  
 00:08 SP: 4.0 PV: 4.6 DV: 0.6  
 2023-12-24  
 23:08 SP: 4.0 PV: 4.7 DV: 0.7  
 22:08 SP: 4.0 PV: 3.8 DV: -0.2  
 1  
 21:08 SP: 4.0 PV: 3.9 DV: -0.  
 20:08 SP: 4.0 PV: 4.4 DV: 0.4  
 19:08 SP: 4.0 PV: 4.6 DV: 0.6  
 18:08 SP: 4.0 PV: 4.7 DV: 0.7  
 17:08 SP: 4.0 PV: 3.8 DV: -0.2  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 4.4 DV: 0.4  
 14:08 SP: 4.0 PV: 4.6 DV: 0.6  
 13:08 SP: 4.0 PV: 4.7 DV: 0.7  
 12:08 SP: 4.0 PV: 4.6 DV: 0.6  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 4.6 DV: 0.6  
 09:08 SP: 4.0 PV: 4.7 DV: 0.7  
 08:08 SP: 4.0 PV: 4.7 DV: 0.7  
 07:08 SP: 4.0 PV: 4.6 DV: 0.6  
 06:08 SP: 4.0 PV: 4.3 DV: 0.3  
 05:08 SP: 4.0 PV: 4.1 DV: 0.1  
 04:08 SP: 4.0 PV: 4.6 DV: 0.6  
 03:08 SP: 4.0 PV: 4.7 DV: 0.7  
 02:08 SP: 4.0 PV: 4.7 DV: 0.7  
 01:08 SP: 4.0 PV: 4.6 DV: 0.6  
 00:08 SP: 4.0 PV: 4.2 DV: 0.2  
 2023-12-23

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.6 DV: 0.6  
 22:08 SP: 4.0 PV: 4.6 DV: 0.6  
 21:08 SP: 4.0 PV: 4.6 DV: 0.6  
 20:08 SP: 4.0 PV: 4.6 DV: 0.6  
 19:08 SP: 4.0 PV: 4.6 DV: 0.6  
 18:08 SP: 4.0 PV: 4.6 DV: 0.6  
 17:08 SP: 4.0 PV: 4.6 DV: 0.6  
 16:08 SP: 4.0 PV: 3.9 DV: -0.1  
 15:08 SP: 4.0 PV: 4.1 DV: 0.1  
 14:08 SP: 4.0 PV: 4.3 DV: 0.3  
 13:08 SP: 4.0 PV: 4.5 DV: 0.5  
 12:08 SP: 4.0 PV: 4.6 DV: 0.6  
 11:08 SP: 4.0 PV: 4.8 DV: 0.8  
 10:08 SP: 4.0 PV: 4.7 DV: 0.7  
 09:08 SP: 4.0 PV: 4.7 DV: 0.7  
 08:08 SP: 4.0 PV: 4.7 DV: 0.7  
 07:08 SP: 4.0 PV: 4.3 DV: 0.3  
 06:08 SP: 4.0 PV: 4.6 DV: 0.6  
 05:08 SP: 4.0 PV: 4.7 DV: 0.7  
 04:08 SP: 4.0 PV: 4.7 DV: 0.7  
 03:08 SP: 4.0 PV: 4.6 DV: 0.6  
 02:08 SP: 4.0 PV: 4.3 DV: 0.3  
 01:08 SP: 4.0 PV: 4.6 DV: 0.6  
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-12-26

23:08 SP: 4.0 PV: 4.7 DV: 0.7  
 22:08 SP: 4.0 PV: 4.6 DV: 0.6  
 21:08 SP: 4.0 PV: 4.6 DV: 0.6  
 20:08 SP: 4.0 PV: 4.4 DV: 0.4  
 19:08 SP: 4.0 PV: 4.2 DV: 0.2  
 18:08 SP: 4.0 PV: 4.5 DV: 0.5  
 17:08 SP: 4.0 PV: 4.7 DV: 0.7  
 16:08 SP: 4.0 PV: 4.7 DV: 0.7  
 15:08 SP: 4.0 PV: 4.7 DV: 0.7  
 14:08 SP: 4.0 PV: 4.7 DV: 0.7  
 13:08 SP: 4.0 PV: 4.6 DV: 0.6  
 12:08 SP: 4.0 PV: 4.4 DV: 0.4  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2

.5

10:08 SP: 4.0 PV: 4.5 DV: 0  
 09:08 SP: 4.0 PV: 4.7 DV: 0.7  
 08:08 SP: 4.0 PV: 4.7 DV: 0.7  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 3.7 DV: -0.3  
 05:08 SP: 4.0 PV: 4.4 DV: 0.4  
 04:08 SP: 4.0 PV: 4.6 DV: 0.6  
 03:08 SP: 4.0 PV: 4.4 DV: 0.4  
 02:08 SP: 4.0 PV: 4.6 DV: 0.6  
 01:08 SP: 4.0 PV: 4.7 DV: 0.7  
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-12-25

23:08 SP: 4.0 PV: 4.4 DV: 0.4  
 22:08 SP: 4.0 PV: 4.0 DV: 0.0  
 21:08 SP: 4.0 PV: 4.4 DV: 0.4  
 20:08 SP: 4.0 PV: 3.9 DV: -0.1  
 19:08 SP: 4.0 PV: 4.2 DV: 0.2  
 18:08 SP: 4.0 PV: 4.4 DV: 0.4  
 17:08 SP: 4.0 PV: 4.0 DV: 0.0  
 16:08 SP: 4.0 PV: 4.3 DV: 0.3

4

15:08 SP: 4.0 PV: 4.4 DV: 0.  
 14:08 SP: 4.0 PV: 4.4 DV: 0.4  
 13:08 SP: 4.0 PV: 4.6 DV: 0.6  
 12:08 SP: 4.0 PV: 4.6 DV: 0.6  
 11:08 SP: 4.0 PV: 4.8 DV: 0.8  
 10:08 SP: 4.0 PV: 4.8 DV: 0.8  
 09:08 SP: 4.0 PV: 4.8 DV: 0.8  
 08:08 SP: 4.0 PV: 4.8 DV: 0.8  
 07:08 SP: 4.0 PV: 4.7 DV: 0.7  
 06:08 SP: 4.0 PV: 4.8 DV: 0.8  
 05:08 SP: 4.0 PV: 4.7 DV: 0.7  
 04:08 SP: 4.0 PV: 4.8 DV: 0.8  
 03:08 SP: 4.0 PV: 4.7 DV: 0.7  
 02:08 SP: 4.0 PV: 4.8 DV: 0.8  
 01:08 SP: 4.0 PV: 4.8 DV: 0.8  
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-12-28

23:08 SP: 4.0 PV: 4.7 DV: 0.7  
 22:08 SP: 4.0 PV: 4.7 DV: 0.7  
 21:08 SP: 4.0 PV: 4.7 DV: 0.7  
 20:08 SP: 4.0 PV: 4.8 DV: 0.8  
 19:08 SP: 4.0 PV: 4.7 DV: 0.7  
 18:08 SP: 4.0 PV: 4.7 DV: 0.7  
 17:08 SP: 4.0 PV: 4.7 DV: 0.7  
 16:08 SP: 4.0 PV: 4.7 DV: 0.7  
 15:08 SP: 4.0 PV: 4.7 DV: 0.7  
 14:08 SP: 4.0 PV: 4.7 DV: 0.7  
 13:08 SP: 4.0 PV: 4.7 DV: 0.7  
 12:08 SP: 4.0 PV: 4.6 DV: 0.6  
 11:08 SP: 4.0 PV: 4.6 DV: 0.6  
 10:08 SP: 4.0 PV: 3.8 DV: -0.2  
 09:08 SP: 4.0 PV: 4.0 DV: 0.0  
 08:08 SP: 4.0 PV: 4.3 DV: 0.3  
 07:08 SP: 4.0 PV: 4.5 DV: 0.5  
 06:08 SP: 4.0 PV: 4.7 DV: 0.7  
 05:08 SP: 4.0 PV: 4.8 DV: 0.8  
 04:08 SP: 4.0 PV: 4.7 DV: 0.7  
 03:08 SP: 4.0 PV: 4.7 DV: 0.7  
 02:08 SP: 4.0 PV: 4.7 DV: 0.7  
 01:08 SP: 4.0 PV: 4.7 DV: 0.7  
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2023-12-27

SP Setting Point  
 PV Present Value  
 DV Difference Value

## 부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2023.12

23:08 SP: 4.0 PV: 4.0 DV: 0.0  
 22:08 SP: 4.0 PV: 4.3 DV: 0.3  
 21:08 SP: 4.0 PV: 4.1 DV: 0.1  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 4.1 DV: 0.1  
 18:08 SP: 4.0 PV: 4.3 DV: 0.3  
 17:08 SP: 4.0 PV: 4.0 DV: 0.0  
 16:08 SP: 4.0 PV: 4.4 DV: 0.4  
 15:08 SP: 4.0 PV: 4.0 DV: 0.0  
 14:08 SP: 4.0 PV: 4.3 DV: 0.3  
 13:08 SP: 4.0 PV: 4.5 DV: 0.5  
 12:08 SP: 4.0 PV: 3.9 DV: -0.1  
 11:08 SP: 4.0 PV: 4.2 DV: 0.2  
 10:08 SP: 4.0 PV: 4.2 DV: 0.2  
 09:08 SP: 4.0 PV: 4.4 DV: 0.4  
 08:08 SP: 4.0 PV: 4.6 DV: 0.6  
 07:08 SP: 4.0 PV: 3.9 DV: -0.1  
 06:08 SP: 4.0 PV: 4.2 DV: 0.2  
 05:08 SP: 4.0 PV: 4.4 DV: 0.4  
 04:08 SP: 4.0 PV: 4.6 DV: 0.6  
 6  
 03:08 SP: 4.0 PV: 4.6 DV: 0.  
 02:08 SP: 4.0 PV: 3.9 DV: -0.1  
 01:08 SP: 4.0 PV: 4.2 DV: 0.2  
 00:08 SP: 4.0 PV: 4.1 DV: 0.1  
 2023-12-29

23:08 SP: 4.0 PV: 4.4 DV: 0.4  
 22:08 SP: 4.0 PV: 3.9 DV: -0.1  
 21:08 SP: 4.0 PV: 4.2 DV: 0.2  
 20:08 SP: 4.0 PV: 4.3 DV: 0.3  
 19:08 SP: 4.0 PV: 4.4 DV: 0.4  
 18:08 SP: 4.0 PV: 4.1 DV: 0.1  
 17:08 SP: 4.0 PV: 4.3 DV: 0.3  
 16:08 SP: 4.0 PV: 4.3 DV: 0.3  
 15:08 SP: 4.0 PV: 4.3 DV: 0.3  
 14:08 SP: 4.0 PV: 4.4 DV: 0.4  
 13:08 SP: 4.0 PV: 4.4 DV: 0.4  
 12:08 SP: 4.0 PV: 4.4 DV: 0.4  
 11:08 SP: 4.0 PV: 4.6 DV: 0.6  
 10:08 SP: 4.0 PV: 4.6 DV: 0.6  
 09:08 SP: 4.0 PV: 4.6 DV: 0.6  
 08:08 SP: 4.0 PV: 4.6 DV: 0.6  
 07:08 SP: 4.0 PV: 4.6 DV: 0.6  
 06:08 SP: 4.0 PV: 4.6 DV: 0.6  
 05:08 SP: 4.0 PV: 4.0 DV: 0.0  
 04:08 SP: 4.0 PV: 4.2 DV: 0.2  
 03:08 SP: 4.0 PV: 4.1 DV: 0.1  
 02:08 SP: 4.0 PV: 4.3 DV: 0.3  
 01:08 SP: 4.0 PV: 4.4 DV: 0.4  
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2023-12-31

23:08 SP: 4.0 PV: 4.0 DV: 0.0  
 22:08 SP: 4.0 PV: 4.3 DV: 0.3  
 21:08 SP: 4.0 PV: 4.4 DV: 0.4  
 20:08 SP: 4.0 PV: 4.6 DV: 0.6  
 19:08 SP: 4.0 PV: 3.9 DV: -0.1  
 18:08 SP: 4.0 PV: 4.2 DV: 0.2  
 17:08 SP: 4.0 PV: 4.1 DV: 0.1  
 16:08 SP: 4.0 PV: 4.4 DV: 0.4  
 15:08 SP: 4.0 PV: 4.5 DV: 0.5  
 14:08 SP: 4.0 PV: 3.9 DV: -0.1  
 13:08 SP: 4.0 PV: 4.2 DV: 0.2  
 12:08 SP: 4.0 PV: 4.4 DV: 0.4  
 11:08 SP: 4.0 PV: 4.6 DV: 0.6  
 10:08 SP: 4.0 PV: 4.6 DV: 0.6  
 09:08 SP: 4.0 PV: 3.9 DV: -0.1  
 08:08 SP: 4.0 PV: 4.2 DV: 0.2  
 07:08 SP: 4.0 PV: 4.4 DV: 0.4  
 06:08 SP: 4.0 PV: 4.0 DV: 0.0  
 05:08 SP: 4.0 PV: 4.3 DV: 0.3  
 04:08 SP: 4.0 PV: 4.5 DV: 0.5  
 03:08 SP: 4.0 PV: 3.9 DV: -0.1  
 02:08 SP: 4.0 PV: 4.2 DV: 0.2  
 01:08 SP: 4.0 PV: 4.1 DV: 0.1  
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2023-12-30

SP Setting Point  
 PV Present Value  
 DV Difference Value

8 of 8

date of copying

sign