

## 부록6. 서식II-01-06, 온도기록지(IP 보관)

## IP STORAGE TEMPERATURE LOG

|                     |                         |                         |                         |                             |
|---------------------|-------------------------|-------------------------|-------------------------|-----------------------------|
| Storage Subject     | Description             | 전기냉장냉동기                 | Storage Space           | 항온고1                        |
|                     | Manufacturer            | SANYO                   | Storage Period          | 2023.12                     |
|                     | Model Name              | MPR-720R                | Investigational Product | 25°C이하<br>15~25°C<br>1~25°C |
|                     | Serial Number           | 07120004                |                         |                             |
| Calibration Subject | Description             | Thermometer             |                         |                             |
|                     | Manufacturer            | DeltaTrak               |                         |                             |
|                     | Model Name              | 12238-02                |                         |                             |
|                     | Serial Number           | 10039708                |                         |                             |
|                     | Date of Calibration     | 2022.10.3(valid for 2y) |                         |                             |
|                     | Date of Calibration Due | 2024.10.3               |                         |                             |
|                     | Certificate No.         | 100397082022103         | Temperature Range       | 16~22°C(19±3°C)             |

| Date       | Time  | Temperature(°C) |      |         | Verified by | Comment |
|------------|-------|-----------------|------|---------|-------------|---------|
|            |       | Min             | Max  | Current |             |         |
| 2023.12.01 | 9:30  | 18.6            | 18.9 | 18.7    | mf          |         |
| 2023.12.02 |       |                 |      |         |             |         |
| 2023.12.03 |       |                 |      |         |             |         |
| 2023.12.04 | 9:30  | 18.5            | 18.9 | 18.7    | 김종민         |         |
| 2023.12.05 | 9:30  | 18.6            | 18.9 | 18.6    | mf          |         |
| 2023.12.06 | 9:30  | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.07 | 9:30  | 18.6            | 18.9 | 18.8    | mf          |         |
| 2023.12.08 | 9:30  | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.09 |       |                 |      |         |             |         |
| 2023.12.10 |       |                 |      |         |             |         |
| 2023.12.11 | 9:30  | 18.6            | 18.9 | 18.8    | mf          |         |
| 2023.12.12 | 9:30  | 18.6            | 18.9 | 18.7    | mf          |         |
| 2023.12.13 | 9:30  | 18.6            | 18.9 | 18.7    | mf          |         |
| 2023.12.14 | 9:30  | 18.6            | 18.9 | 18.8    | mf          |         |
| 2023.12.15 | 9:30  | 18.6            | 18.9 | 18.8    | mf          |         |
| 2023.12.16 |       |                 |      |         |             |         |
| 2023.12.17 |       |                 |      |         |             |         |
| 2023.12.18 | 9:30  | 18.6            | 18.9 | 18.7    | mf          |         |
| 2023.12.19 | 9:30  | 18.6            | 18.9 | 18.7    | mf          |         |
| 2023.12.20 | 9:30  | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.21 | 10:00 | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.22 | 9:30  | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.23 |       |                 |      |         |             |         |
| 2023.12.24 |       |                 |      |         |             |         |
| 2023.12.25 |       |                 |      |         |             | 성탄절     |
| 2023.12.26 | 9:30  | 18.6            | 18.9 | 18.8    | mf          |         |
| 2023.12.27 | 9:30  | 18.7            | 18.9 | 18.7    | mf          |         |
| 2023.12.28 | 9:30  | 18.7            | 18.9 | 18.8    | mf          |         |
| 2023.12.29 | 9:30  | 18.7            | 18.9 | 18.7    | mf          |         |
| 2023.12.30 |       |                 |      |         |             |         |
| 2023.12.31 |       |                 |      |         |             |         |

&lt;Min/Max&gt;

- ① PC-3500, 12238-02, 온습도계 : 온도기록 후 reset하여 그 시점이후 Min/Max 새로 기록됨.  
 ② 냉동고 : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.  
 ③ testo : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행.  
 <testo 온도이탈 알람 설정값> 실온 : <15.0°C, >25.0°C, 즉시 / 냉장 : <2.0°C, >8.0°C, 즉시

date of copying

sign