

STATEMENT OF ACCURACY

iMiniPlus (Dry Ice) Temperature Data Loggers



Product: **MP-OE-N-8-L**

S. No.: **MH-DD-0291-0066**

Date: **29 Nov 2023**

Checked by: *luamp*



Authorized Quality Document for Cryopak Verification Technologies.
Reference: ISO 9001:2015. ACC-039 Revision 6 Page Form Page 1/1.

iMiniPlus (Dry Ice) data loggers are calibrated for temperature in CVT's Calibration Chamber. The reference equipment used is traceable to National Institute of Standards and Technology (NIST). Temperature Reference Meter: Fluke -Hart Scientific 1521/1522 Digital Thermometer Readout w/Burns Engineering Platinum Resistance Thermometer Probe. This certificate is valid for 2 years from the date of purchase or as dated on this certificate (1 trip for ST logger).

Temperature range: -100°C to +40 °C (N range)

Accuracy: $\pm 1.0^{\circ}\text{C}$ (-100°C to -80°C)

$\pm 0.7^{\circ}\text{C}$ (-80°C to -10°C)

$\pm 0.6^{\circ}\text{C}$ (-10°C to +40°C)

Resolution: 0.1°C

www.cryopakdigital.com

ISO 9001:2015
Certified Company