

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.1
Calibration Subject	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8°C

Date	Time	Probe Position	Temperature(°C)			Verified by	Comment
			Min	Max	Current		
2024.01.01		Left					신정
		Right					
2024.01.02	p:30	Left	3	4	3	ml	
		Right	3	4	3		
2024.01.03	p:30	Left	3	4	4	ml	
		Right	2	4	4		
2024.01.04	p:30	Left	3	4	4	ml	
		Right	2	4	3		
2024.01.05	p:30	Left	3	4	4	ml	
		Right	2	4	4		
2024.01.06		Left					
		Right					
2024.01.07		Left					
		Right					
2024.01.08	p:30	Left	3	4	4	ml	
		Right	2	5	4		
2024.01.09	p:30	Left	3	4	3	ml	
		Right	3	4	3		
2024.01.10	p:30	Left	3	4	4	ml	
		Right	3	4	4		
2024.01.11	p:30	Left	3	4	4	ml	
		Right	3	4	4		
2024.01.12	p:30	Left	3	4	3	ml	
		Right	3	4	4		
2024.01.13		Left					
		Right					

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.1
Calibration Subject	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2024.01.14		Left					
		Right					
2024.01.15	9:30	Left	3	4	4	ml	
		Right	3	4	4		
2024.01.16	9:30	Left	3	4	3	ml	
		Right	2	4	3		
2024.01.17	9:30	Left	3	4	3	ml	
		Right	2	4	3		
2024.01.18	9:30	Left	2	4	3	ml	
		Right	3	5	4		
2024.01.19	9:30	Left	2	4	3	ml	
		Right	3	5	4		
2024.01.20		Left					
		Right					
2024.01.21		Left					
		Right					
2024.01.22	9:40	Left	2	4	3	ml	1404-0038 IP 보관
		Right	3	4	3		
2024.01.23	9:30	Left	2	4	3	ml	
		Right	3	4	4		
2024.01.24	9:30	Left	2	4	3	ml	
		Right	2	4	4		
2024.01.25	9:30	Left	2	4	4	김지현	
		Right	2	4	4		
2024.01.26	9:30	Left	2	4	4	김지현	
		Right	2	4	4		

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.1
Calibration Subject	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-5		
	Probe Position	Right		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855630		
	Date of Calibration	2023.3.27		
	Date of Calibration Due	2024.3.27		
	Certificate No.	23007746-6	Temperature Range	2~8℃

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2024.01.27		Left					
		Right					
2024.01.28		Left					
		Right					
2024.01.29	9:30	Left	2	4	3.4		1404-0041 IP 보관
		Right	3	5	3.4		
2024.01.30	9:30	Left	2	4	3		
		Right	3	4	3		
2024.01.31	9:30	Left	2	4	3		
		Right	3	4	3.4		

<Min/Max>
 ① PC-3500, 12238-02, 온습도계 : 온도기록 후 .reset하여 그 시점 이후 Min/Max 새로 기록됨.
 ② 냉동고 : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.
 ③ testo : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행. <testo 온도이탈 알람 설정값> 실온 : <15.5℃, >24.5℃, 즉시 / 냉장 : <2.5℃, >7.5℃, 즉시

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2024-01-02

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.0 DV: 0.0

2

11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-01-01

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.6 DV: 0.6

7

04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.7 DV: 0.7
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2024-01-04

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.7 DV: 0.7
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3

3

03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-01-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.8 DV: 0.8
 16:08 SP: 4.0 PV: 4.8 DV: 0.8
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.8 DV: 0.8
 13:08 SP: 4.0 PV: 4.8 DV: 0.8

7

12:08 SP: 4.0 PV: 4.7 DV: 0.
 11:08 SP: 4.0 PV: 4.7 DV: 0.7
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.7 DV: 0.7
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-01-06

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.4 DV: 0.4

6

21:08 SP: 4.0 PV: 4.6 DV: 0.
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.8 DV: 0.8
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 5.4 DV: 1.4
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.7 DV: 0.7
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-01-05

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.7 DV: 0.7
 11:08 SP: 4.0 PV: 4.8 DV: 0.8
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.5 DV: 0.5
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2024-01-08

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.7 DV: 0.7
 12:08 SP: 4.0 PV: 4.7 DV: 0.7
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2024-01-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 4.1 DV: 0.1
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.0 DV: 0.0
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.5 DV: 0.5
 2024-01-10
 23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.5 DV: 0.6
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.4 DV: 0.4
 2024-01-09

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.5 DV: 0.6
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.6 DV: 0.6
 2024-01-12
 23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.5 DV: 0.5
 2024-01-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3
 2024-01-14
 23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.4 DV: 0.4
 2024-01-13

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 4.8 DV: 0.8
 20:08 SP: 4.0 PV: 4.8 DV: 0.8
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 7
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 4.8 DV: 0.8
 05:08 SP: 4.0 PV: 4.7 DV: 0.7
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 4
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.8 DV: 0.8
 2024-01-16
 23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 6
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 2
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-01-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 3.9 DV: -0.1	23:08 SP: 4.0 PV: 3.6 DV: -0.4
22:08 SP: 4.0 PV: 4.1 DV: 0.1	22:08 SP: 4.0 PV: 4.2 DV: 0.2
21:08 SP: 4.0 PV: 3.9 DV: -0.1	21:08 SP: 4.0 PV: 3.9 DV: -0.1
20:08 SP: 4.0 PV: 3.7 DV: -0.3	20:08 SP: 4.0 PV: 3.9 DV: -0.1
19:08 SP: 4.0 PV: 4.2 DV: 0.2	19:08 SP: 4.0 PV: 4.0 DV: 0.0
18:08 SP: 4.0 PV: 3.8 DV: -0.2	18:08 SP: 4.0 PV: 3.9 DV: -0.1
17:08 SP: 4.0 PV: 4.3 DV: 0.3	17:08 SP: 4.0 PV: 3.9 DV: -0.1
16:08 SP: 4.0 PV: 3.9 DV: -0.1	0
15:08 SP: 4.0 PV: 3.8 DV: -0.2	16:08 SP: 4.0 PV: 4.0 DV: 0.
14:08 SP: 4.0 PV: 3.7 DV: -0.3	15:08 SP: 4.0 PV: 3.7 DV: -0.3
13:08 SP: 4.0 PV: 4.3 DV: 0.3	14:08 SP: 4.0 PV: 4.2 DV: 0.2
12:08 SP: 4.0 PV: 3.8 DV: -0.2	13:08 SP: 4.0 PV: 3.9 DV: -0.1
11:08 SP: 4.0 PV: 3.6 DV: -0.4	12:08 SP: 4.0 PV: 4.1 DV: 0.1
10:08 SP: 4.0 PV: 4.2 DV: 0.2	11:08 SP: 4.0 PV: 3.9 DV: -0.1
09:08 SP: 4.0 PV: 3.9 DV: -0.1	10:08 SP: 4.0 PV: 3.9 DV: -0.1
08:08 SP: 4.0 PV: 3.9 DV: -0.1	09:08 SP: 4.0 PV: 3.8 DV: -0.2
07:08 SP: 4.0 PV: 3.8 DV: -0.2	08:08 SP: 4.0 PV: 3.7 DV: -0.3
06:08 SP: 4.0 PV: 4.1 DV: 0.1	07:08 SP: 4.0 PV: 4.2 DV: 0.2
05:08 SP: 4.0 PV: 3.8 DV: -0.2	06:08 SP: 4.0 PV: 3.9 DV: -0.1
04:08 SP: 4.0 PV: 3.9 DV: -0.1	05:08 SP: 4.0 PV: 4.2 DV: 0.2
03:08 SP: 4.0 PV: 3.8 DV: -0.2	04:08 SP: 4.0 PV: 3.9 DV: -0.1
02:08 SP: 4.0 PV: 3.9 DV: -0.1	03:08 SP: 4.0 PV: 4.1 DV: 0.1
01:08 SP: 4.0 PV: 3.8 DV: -0.2	02:08 SP: 4.0 PV: 3.8 DV: -0.2
00:08 SP: 4.0 PV: 3.9 DV: -0.1	01:08 SP: 4.0 PV: 3.8 DV: -0.2
2024-01-18	00:08 SP: 4.0 PV: 3.8 DV: -0.2
23:08 SP: 4.0 PV: 3.9 DV: -0.1	2024-01-20
22:08 SP: 4.0 PV: 3.9 DV: -0.1	23:08 SP: 4.0 PV: 4.0 DV: 0.0
21:08 SP: 4.0 PV: 3.9 DV: -0.1	22:08 SP: 4.0 PV: 3.8 DV: -0.2
20:08 SP: 4.0 PV: 3.9 DV: -0.1	21:08 SP: 4.0 PV: 3.6 DV: -0.4
19:08 SP: 4.0 PV: 3.8 DV: -0.2	20:08 SP: 4.0 PV: 4.2 DV: 0.2
18:08 SP: 4.0 PV: 3.8 DV: -0.2	19:08 SP: 4.0 PV: 3.8 DV: -0.2
17:08 SP: 4.0 PV: 3.7 DV: -0.3	18:08 SP: 4.0 PV: 3.7 DV: -0.3
16:08 SP: 4.0 PV: 3.7 DV: -0.3	17:08 SP: 4.0 PV: 4.1 DV: 0.1
15:08 SP: 4.0 PV: 3.6 DV: -0.4	16:08 SP: 4.0 PV: 3.8 DV: -0.2
14:08 SP: 4.0 PV: 3.6 DV: -0.4	15:08 SP: 4.0 PV: 4.0 DV: 0.0
13:08 SP: 4.0 PV: 3.6 DV: -0.4	14:08 SP: 4.0 PV: 3.9 DV: -0.1
12:08 SP: 4.0 PV: 3.4 DV: -0.6	13:08 SP: 4.0 PV: 3.7 DV: -0.3
11:08 SP: 4.0 PV: 4.4 DV: 0.4	12:08 SP: 4.0 PV: 3.6 DV: -0.4
10:08 SP: 4.0 PV: 5.1 DV: 1.1	11:08 SP: 4.0 PV: 3.7 DV: -0.3
09:08 SP: 4.0 PV: 4.8 DV: 0.8	10:08 SP: 4.0 PV: 4.2 DV: 0.2
08:08 SP: 4.0 PV: 4.7 DV: 0.7	09:08 SP: 4.0 PV: 3.8 DV: -0.2
07:08 SP: 4.0 PV: 3.8 DV: -0.2	08:08 SP: 4.0 PV: 3.7 DV: -0.3
06:08 SP: 4.0 PV: 3.9 DV: -0.1	07:08 SP: 4.0 PV: 4.1 DV: 0.1
05:08 SP: 4.0 PV: 4.4 DV: 0.4	06:08 SP: 4.0 PV: 3.9 DV: -0.1
04:08 SP: 4.0 PV: 4.6 DV: 0.6	05:08 SP: 4.0 PV: 3.9 DV: -0.1
03:08 SP: 4.0 PV: 4.8 DV: 0.8	04:08 SP: 4.0 PV: 3.9 DV: -0.1
02:08 SP: 4.0 PV: 4.7 DV: 0.7	03:08 SP: 4.0 PV: 3.8 DV: -0.2
01:08 SP: 4.0 PV: 4.6 DV: 0.6	02:08 SP: 4.0 PV: 4.0 DV: 0.0
00:08 SP: 4.0 PV: 4.4 DV: 0.4	01:08 SP: 4.0 PV: 3.9 DV: -0.1
2024-01-17	00:08 SP: 4.0 PV: 4.1 DV: 0.1
	2024-01-19

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 3.8 DV: -0.2

22:08 SP: 4.0 PV: 4.1 DV: 0.1

1

21:08 SP: 4.0 PV: 3.9 DV: -0.1

20:08 SP: 4.0 PV: 4.0 DV: 0.0

19:08 SP: 4.0 PV: 3.9 DV: -0.1

18:08 SP: 4.0 PV: 3.9 DV: -0.1

17:08 SP: 4.0 PV: 3.9 DV: -0.1

16:08 SP: 4.0 PV: 3.8 DV: -0.2

15:08 SP: 4.0 PV: 3.8 DV: -0.2

14:08 SP: 4.0 PV: 4.0 DV: 0.0

13:08 SP: 4.0 PV: 3.9 DV: -0.1

12:08 SP: 4.0 PV: 3.7 DV: -0.3

11:08 SP: 4.0 PV: 3.7 DV: -0.3

10:08 SP: 4.0 PV: 3.6 DV: -0.4

09:08 SP: 4.0 PV: 4.2 DV: 0.2

08:08 SP: 4.0 PV: 3.9 DV: -0.1

07:08 SP: 4.0 PV: 3.9 DV: -0.1

06:08 SP: 4.0 PV: 3.9 DV: -0.1

05:08 SP: 4.0 PV: 3.8 DV: -0.2

04:08 SP: 4.0 PV: 3.9 DV: -0.1

03:08 SP: 4.0 PV: 3.8 DV: -0.2

02:08 SP: 4.0 PV: 3.9 DV: -0.1

01:08 SP: 4.0 PV: 3.9 DV: -0.1

00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-01-22

23:08 SP: 4.0 PV: 3.9 DV: -0.1

22:08 SP: 4.0 PV: 3.8 DV: -0.2

21:08 SP: 4.0 PV: 3.8 DV: -0.2

20:08 SP: 4.0 PV: 3.9 DV: -0.1

19:08 SP: 4.0 PV: 3.9 DV: -0.1

18:08 SP: 4.0 PV: 3.8 DV: -0.2

17:08 SP: 4.0 PV: 4.0 DV: 0.0

16:08 SP: 4.0 PV: 3.8 DV: -0.2

15:08 SP: 4.0 PV: 4.0 DV: 0.0

14:08 SP: 4.0 PV: 3.9 DV: -0.1

13:08 SP: 4.0 PV: 4.5 DV: 0.5

12:08 SP: 4.0 PV: 3.7 DV: -0.3

11:08 SP: 4.0 PV: 3.9 DV: -0.1

10:08 SP: 4.0 PV: 3.8 DV: -0.2

09:08 SP: 4.0 PV: 3.9 DV: -0.1

08:08 SP: 4.0 PV: 3.8 DV: -0.2

07:08 SP: 4.0 PV: 4.0 DV: 0.0

06:08 SP: 4.0 PV: 3.9 DV: -0.1

05:08 SP: 4.0 PV: 3.8 DV: -0.2

04:08 SP: 4.0 PV: 4.1 DV: 0.1

03:08 SP: 4.0 PV: 3.8 DV: -0.2

02:08 SP: 4.0 PV: 4.0 DV: 0.0

01:08 SP: 4.0 PV: 3.8 DV: -0.2

00:08 SP: 4.0 PV: 3.7 DV: -0.3

2024-01-21

23:08 SP: 4.0 PV: 3.9 DV: -0.1

22:08 SP: 4.0 PV: 3.9 DV: -0.1

21:08 SP: 4.0 PV: 3.9 DV: -0.1

20:08 SP: 4.0 PV: 3.9 DV: -0.1

19:08 SP: 4.0 PV: 3.9 DV: -0.1

18:08 SP: 4.0 PV: 3.8 DV: -0.2

17:08 SP: 4.0 PV: 3.9 DV: -0.1

2

16:08 SP: 4.0 PV: 3.8 DV: -0.1

15:08 SP: 4.0 PV: 4.0 DV: 0.0

14:08 SP: 4.0 PV: 3.8 DV: -0.2

13:08 SP: 4.0 PV: 3.7 DV: -0.3

12:08 SP: 4.0 PV: 4.1 DV: 0.1

11:08 SP: 4.0 PV: 3.9 DV: -0.1

10:08 SP: 4.0 PV: 3.8 DV: -0.2

09:08 SP: 4.0 PV: 4.0 DV: 0.0

08:08 SP: 4.0 PV: 3.9 DV: -0.1

07:08 SP: 4.0 PV: 3.8 DV: -0.2

06:08 SP: 4.0 PV: 3.9 DV: -0.1

05:08 SP: 4.0 PV: 3.9 DV: -0.1

04:08 SP: 4.0 PV: 3.9 DV: -0.1

03:08 SP: 4.0 PV: 3.9 DV: -0.1

02:08 SP: 4.0 PV: 3.9 DV: -0.1

01:08 SP: 4.0 PV: 3.9 DV: -0.1

00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-01-24

23:08 SP: 4.0 PV: 3.9 DV: -0.1

22:08 SP: 4.0 PV: 3.8 DV: -0.2

21:08 SP: 4.0 PV: 3.9 DV: -0.1

20:08 SP: 4.0 PV: 3.8 DV: -0.2

19:08 SP: 4.0 PV: 3.8 DV: -0.2

18:08 SP: 4.0 PV: 3.9 DV: -0.1

17:08 SP: 4.0 PV: 3.8 DV: -0.2

16:08 SP: 4.0 PV: 3.8 DV: -0.2

15:08 SP: 4.0 PV: 3.7 DV: -0.3

14:08 SP: 4.0 PV: 3.7 DV: -0.3

13:08 SP: 4.0 PV: 3.6 DV: -0.4

12:08 SP: 4.0 PV: 4.1 DV: 0.1

11:08 SP: 4.0 PV: 3.8 DV: -0.2

10:08 SP: 4.0 PV: 3.7 DV: -0.3

09:08 SP: 4.0 PV: 3.6 DV: -0.4

08:08 SP: 4.0 PV: 3.8 DV: -0.2

07:08 SP: 4.0 PV: 3.8 DV: -0.2

06:08 SP: 4.0 PV: 3.8 DV: -0.2

05:08 SP: 4.0 PV: 3.9 DV: -0.1

04:08 SP: 4.0 PV: 3.9 DV: -0.1

03:08 SP: 4.0 PV: 3.9 DV: -0.1

02:08 SP: 4.0 PV: 3.8 DV: -0.2

01:08 SP: 4.0 PV: 3.8 DV: -0.2

00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-01-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-01-26

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.0 DV: 0.0
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.6 DV: -0.4
 04:08 SP: 4.0 PV: 3.6 DV: -0.4
 03:08 SP: 4.0 PV: 3.4 DV: -0.6
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-01-25

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-01-28

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.9 DV: -0.1
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-01-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.1

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.7 DV: -0.3
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.4 DV: -0.6
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 3.8 DV: -0.2
 2024-01-29

23:08 SP: 4.0 PV: 3.6 DV: -0.2
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.9 DV: -0.1

2

17:08 SP: 4.0 PV: 3.8 DV: -0.1
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-01-31

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-01-30

SP Setting Point
 PV Present Value
 DV Difference Value