

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.1
	Model Name	MPR-720R		
	Serial Number	07120004		
Calibration Subject	Description	Thermometer	Investigational Product	25°C이하
	Manufacturer	DeltaTrak		15~25°C
	Model Name	12238-02		1~25°C
	Serial Number	10039708	Temperature Range	
	Date of Calibration	2022.10.3(valid for 2y)		
	Date of Calibration Due	2024.10.3		
	Certificate No.	100397082022103		

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2024.01.01						신정
2024.01.02	P:30	18.6	18.9	18.7		
2024.01.03	P:30	18.6	18.9	18.8		
2024.01.04	P:30	18.7	18.9	18.8		
2024.01.05	P:30	18.6	18.9	18.8		
2024.01.06						
2024.01.07						
2024.01.08	P:30	18.6	18.9	18.7		
2024.01.09	P:30	18.6	18.9	18.8		
2024.01.10	P:30	18.6	18.9	18.8		
2024.01.11	P:30	18.7	18.9	18.8		
2024.01.12	P:30	18.6	18.9	18.7		
2024.01.13						
2024.01.14						
2024.01.15	P:30	18.6	18.9	18.7		
2024.01.16	P:30	18.7	18.9	18.8		
2024.01.17	P:30	18.7	18.9	18.7		
2024.01.18	P:30	18.7	18.9	18.8		
2024.01.19	P:30	18.7	18.9	18.7		
2024.01.20						
2024.01.21						
2024.01.22	P:40	18.6	18.9	18.8		
2024.01.23	P:30	18.6	18.9	18.7		
2024.01.24	P:30	18.6	18.8	18.7		
2024.01.25	P:30	18.6	18.8	18.7		
2024.01.26	P:30	18.6	18.8	18.7		
2024.01.27						
2024.01.28						
2024.01.29	P:30	18.6	18.8	18.7		
2024.01.30	P:30	18.6	18.8	18.8		
2024.01.31	P:30	18.6	18.8	18.8		

<Min/Max>

- ① PC-3500, 12238-02, 온습도계 : 온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨.
 ② 냉동고 : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.
 ③ testo : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행.
 <testo 온도이탈 알람 설정값> 실온 : <15.5°C, >24.5°C, 즉시 / 냉장 : <2.5°C, >7.5°C, 즉시

date of copying

sign