

부록6. 서식II-01-06, 온도기록지(IP 보관)

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.2
Calibration Subject	Model Name	MPR-720R	Investigational Product	25℃이하 15~25℃ 1~25℃
	Serial Number	07120004		
	Description	Thermometer		
	Manufacturer	DeltaTrak		
	Model Name	12238-02	Temperature Range	16~22℃(19±3℃)
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)		
	Date of Calibration Due	2024.10.3		
	Certificate No.	100397082022103		

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.02.01	P:30	18.6	18.8	18.7	mf	
2024.02.02	P:30	18.6	18.8	18.8	mf	
2024.02.03						
2024.02.04						
2024.02.05	P:30	18.6	18.9	18.7	mf	
2024.02.06	P:30	18.6	18.8	18.8	mf	
2024.02.07	P:30	18.6	18.8	18.7	mf	
2024.02.08	P:30	18.6	18.8	18.7	mf	
2024.02.09						설날연휴
2024.02.10						설날
2024.02.11						
2024.02.12						설날 대체공휴일
2024.02.13	P:30	18.6	18.9	18.8	mf	
2024.02.14	P:30	18.6	18.9	18.7	mf	
2024.02.15	P:30	18.6	18.8	18.7	mf	
2024.02.16	P:30	18.6	18.8	18.7	mf	
2024.02.17						
2024.02.18						
2024.02.19	P:30	18.6	18.8	18.7	mf	
2024.02.20	P:30	18.6	18.8	18.7	mf	
2024.02.21	P:30	18.6	18.8	18.7	mf	
2024.02.22	P:30	18.6	18.8	18.7	mf	
2024.02.23	P:30	18.6	18.8	18.8	mf	
2024.02.24						
2024.02.25						
2024.02.26	P:30	18.6	18.8	18.8	mf	
2024.02.27	P:30	18.6	18.8	18.7	mf	
2024.02.28	P:30	18.6	18.8	18.7	mf	
2024.02.29	P:30	18.6	18.8	18.7	mf	

<Min/Max>

① PC-3500, 12238-02, 온습도계 : 온도기록 후 reset하여 그 시점이후 Min/Max 새로 기록됨.

② MP-OE-N-8-L : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.

③ Saveris2-T3 : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행. <온도이탈 알람 설정값> 실온 : <15.5℃, >24.5℃, 즉시 / 냉동 : <-85.0℃, >-30.0℃, 즉시

④ TTM-004 : 냉장고2 내장 온도기록계로 현재 온도만 측정. 온도이탈시 알람음. <온도이탈 알람 설정값> <2.5℃, >7.5℃

⑤ MS-Pro2 : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 전화감, 지난 온도기록 이후 알람 없음을 확인 후 진행. <온도이탈 알람 설정값> 냉장 : <2.5℃ >7.5℃ 10초 후

date of copying

sign