

## IP STORAGE TEMPERATURE LOG

|                     |                         |                         |                         |                             |
|---------------------|-------------------------|-------------------------|-------------------------|-----------------------------|
| Storage Subject     | Description             | 전기냉장냉동기                 | Storage Space           | 항온고1                        |
|                     | Manufacturer            | SANYO                   | Storage Period          | 2024.3                      |
|                     | Model Name              | MPR-720R                | Investigational Product | 25°C이하<br>15~25°C<br>1~25°C |
|                     | Serial Number           | 07120004                |                         |                             |
| Calibration Subject | Description             | Thermometer             |                         |                             |
|                     | Manufacturer            | DeltaTrak               |                         |                             |
|                     | Model Name              | 12238-02                |                         |                             |
|                     | Serial Number           | 10039708                |                         |                             |
|                     | Date of Calibration     | 2022.10.3(valid for 2y) |                         |                             |
|                     | Date of Calibration Due | 2024.10.3               | Temperature Range       | 16~22°C(19±3°C)             |
|                     | Certificate No.         | 100397082022103         |                         |                             |

| Date       | Time | Temperature(°C) |      |         | Verified by | Comment |
|------------|------|-----------------|------|---------|-------------|---------|
|            |      | Min             | Max  | Current |             |         |
| 2024.03.01 |      |                 |      |         |             | 삼일절     |
| 2024.03.02 |      |                 |      |         |             |         |
| 2024.03.03 |      |                 |      |         |             |         |
| 2024.03.04 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.05 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.06 | P:30 | 18.6            | 18.9 | 18.8    | ml          |         |
| 2024.03.07 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.08 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.09 |      |                 |      |         |             |         |
| 2024.03.10 |      |                 |      |         |             |         |
| 2024.03.11 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.12 | P:30 | 18.6            | 18.8 | 18.8    | ml          |         |
| 2024.03.13 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.14 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.15 | P:30 | 18.6            | 18.8 | 18.8    | ml          |         |
| 2024.03.16 |      |                 |      |         |             |         |
| 2024.03.17 |      |                 |      |         |             |         |
| 2024.03.18 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.19 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.20 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.21 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.22 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.23 |      |                 |      |         |             |         |
| 2024.03.24 |      |                 |      |         |             |         |
| 2024.03.25 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.26 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.27 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.28 | P:30 | 18.6            | 18.8 | 18.7    | ml          |         |
| 2024.03.29 | P:30 | 18.6            | 18.8 | 18.8    | ml          |         |
| 2024.03.30 |      |                 |      |         |             |         |
| 2024.03.31 |      |                 |      |         |             |         |

&lt;Min/Max&gt;

- ① PC-3500, 12238-02, 온습도계 : 온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨.
- ② MP-OE-N-8-L : PC연결 후 reset해야 하므로, Min/Max 누적기록이며 월1회 reset 후 새로 기록됨.
- ③ Saveris2-T3 : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 문자와 이메일 전송됨, 지난 온도기록 이후 알람 없음을 확인 후 진행. <온도이탈 알람 설정값> 실온 : <15.5°C, >24.5°C, 즉시 / 냉동 : <-85.0°C, >-30.0°C, 즉시
- ④ TTM-004 : 냉장고2 내장 온도기록계로 현재 온도만 측정. 온도이탈시 알람음. <온도이탈 알람 설정값> <2.5°C, >7.5°C
- ⑤ MS-Pro2 : 기기로 Min/Max 확인방법 없으며, 온도이탈시 관리약사에게 전화감, 지난 온도기록 이후 알람 없음을 확인 후 진행. <온도이탈 알람 설정값> 냉장 : <2.5°C >7.5°C 10초 후

date of copying

sign