

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.4
	Model Name	MPR-720R	Temperature Range	15~25°C
	Serial Number	07120004	Setting Point	19.0°C
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15°C, >25°C, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2024.04.01	P:30	18.6	18.8	18.8	mf	
2024.04.02	P:30	18.6	18.8	18.7	mf	
2024.04.03	P:30	18.6	18.8	18.7	mf	
2024.04.04	P:30	18.6	18.8	18.7	mf	
2024.04.05	P:30	18.6	18.8	18.7	mf	
2024.04.06						
2024.04.07						
2024.04.08	P:30	18.6	18.8	18.7	mf	
2024.04.09	P:30	18.6	18.8	18.7	mf	
2024.04.10						국회의원 선거
2024.04.11	P:30	18.6	18.9	18.7	mf	
2024.04.12	P:30	18.6	18.8	18.8	mf	
2024.04.13						
2024.04.14						
2024.04.15	P:30	18.6	18.8	18.7	mf	
2024.04.16	P:40	18.6	18.8	18.6	mf	
2024.04.17	P:30	18.6	18.8	18.7	mf	
2024.04.18	P:30	18.6	18.8	18.8	mf	
2024.04.19	P:30	18.6	18.8	18.7	mf	
2024.04.20						
2024.04.21						
2024.04.22	P:30	18.6	18.8	18.7	mf	
2024.04.23	P:30	18.6	18.8	18.7	mf	
2024.04.24	P:30	18.6	18.8	18.7	mf	
2024.04.25	P:30	18.6	18.8	18.7	mf	
2024.04.26	P:30	18.6	18.8	18.8	mf	
2024.04.27						
2024.04.28						
2024.04.29	P:30	18.6	18.8	18.7	mf	
2024.04.30	P:30	18.6	18.8	18.7	mf	

date of copying

sign