

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.5
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.05.01						근로자의 날
2024.05.02	P:30	18.6	18.8	18.7		
2024.05.03	P:30	18.6	18.8	18.7		
2024.05.04						
2024.05.05						
2024.05.06						어린이날 대체공휴일
2024.05.07	P:30	18.6	18.9	18.7		
2024.05.08	P:30	18.6	18.8	18.7		
2024.05.09	P:30	18.6	18.8	18.7		
2024.05.10	P:30	18.6	18.8	18.7		
2024.05.11						
2024.05.12						
2024.05.13	P:30	18.6	18.9	18.7		
2024.05.14	P:40	18.6	18.8	18.8		
2024.05.15						부처님오신날
2024.05.16	P:30	18.6	18.8	18.7		
2024.05.17	P:30	18.6	18.9	18.7		
2024.05.18						
2024.05.19						
2024.05.20	P:30	18.6	18.9	18.7		
2024.05.21	P:30	18.6	18.8	18.8		
2024.05.22	P:30	18.6	18.8	18.7		
2024.05.23	P:30	18.6	18.8	18.7		
2024.05.24	P:30	18.6	18.8	18.7		
2024.05.25						
2024.05.26						
2024.05.27	P:30	18.6	18.8	18.7		
2024.05.28	P:30	18.6	18.8	18.7		
2024.05.29	P:10	18.6	18.8	18.7		
2024.05.30	P:40	18.6	18.8	18.7		
2024.05.31	P:30	18.6	18.8	18.7		

date of copying

sign