







IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.6
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.06.01						
2024.06.02						
2024.06.03	P:30	18.6	18.8	18.8		
2024.06.04	P:30	18.6	18.8	18.7		
2024.06.05	P:30	18.6	18.8	18.7		
2024.06.06						현충일
2024.06.07	P:30	18.6	18.8	18.7		
2024.06.08						
2024.06.09						
2024.06.10	P:30	18.6	18.8	18.7		
2024.06.11	P:30	18.6	18.9	18.8		
2024.06.12	P:40	18.6	18.8	18.8		
2024.06.13	P:30	18.6	18.8	18.6		
2024.06.14	P:30	18.6	18.9	18.7		
2024.06.15						
2024.06.16						
2024.06.17	P:30	18.6	18.9	18.7		
2024.06.18	P:30	18.6	18.8	18.7		
2024.06.19	P:30	18.6	18.8	18.8		
2024.06.20	P:30	18.6	18.9	18.7		
2024.06.21	P:30	18.6	18.9	18.7		
2024.06.22						
2024.06.23						
2024.06.24	P:30	18.6	18.9	18.7		
2024.06.25	P:30	18.6	18.9	18.7		
2024.06.26	P:30	18.6	18.9	18.7		
2024.06.27	P:30	18.6	18.8	18.7		
2024.06.28	P:30	18.6	18.9	18.8		
2024.06.29						
2024.06.30						

date of copying

sign