

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.7
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.07.01	P:30	18.6	18.9	18.7		
2024.07.02	P:30	18.6	18.9	18.7		
2024.07.03	P:30	18.6	18.9	18.8		
2024.07.04	P:30	18.6	18.9	18.7		
2024.07.05	P:30	18.6	18.9	18.8		
2024.07.06						
2024.07.07						
2024.07.08	P:30	18.6	18.9	18.7		
2024.07.09	P:30	18.7	18.9	18.8		
2024.07.10	P:30	18.7	18.9	18.8		
2024.07.11	P:30	18.6	18.9	18.8		
2024.07.12	P:30	18.6	18.9	18.7		
2024.07.13						
2024.07.14						
2024.07.15	P:30	18.6	18.9	18.8		
2024.07.16	P:30	18.6	18.9	18.8		
2024.07.17	P:30	18.6	18.9	18.8		
2024.07.18	P:30	18.6	18.9	18.8		
2024.07.19	P:30	18.6	18.9	18.8		
2024.07.20						
2024.07.21						
2024.07.22	P:30	18.6	18.9	18.8		
2024.07.23	P:30	18.6	18.9	18.8		
2024.07.24	P:30	18.6	18.9	18.9		
2024.07.25	P:30	18.7	18.9	18.8		
2024.07.26	P:30	18.7	18.9	18.8		
2024.07.27						
2024.07.28						
2024.07.29	P:30	18.7	18.9	18.7		
2024.07.30	P:30	18.7	18.9	18.8		
2024.07.31	P:30	18.7	18.9	18.8		

date of copying

sign