

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.11
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2316576		
	Date of Calibration	2024.3.7	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2025.3.7	Alarm Method	Call to Pharmacist
	Certificate No.	24005617-4	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.11.01	P:30	3	5	4	aml	
2024.11.02						
2024.11.03						
2024.11.04	P:30	3	5	4	aml	
2024.11.05	P:00	3	5	4	aml	
2024.11.06	P:30	3	5	4	aml	
2024.11.07	P:30	3	5	4	aml	
2024.11.08	P:30	3	5	4	aml	
2024.11.09						
2024.11.10						
2024.11.11	P:30	3	5	4	aml	
2024.11.12	P:30	3	5	4	aml	
2024.11.13	P:30	3	5	4	aml	
2024.11.14	P:30	3	5	4	aml	
2024.11.15	P:30	3	5	4	aml	
2024.11.16						
2024.11.17						
2024.11.18	P:30	3	5	4	aml	
2024.11.19	P:30	3	5	4	aml	ESR-18-14354 EP 02/26
2024.11.20	P:30	3	5	4	aml	
2024.11.21	P:30	3	5	4	aml	
2024.11.22	P:30	3	5	4	aml	
2024.11.23						
2024.11.24						
2024.11.25	P:40	3	5	4	aml	
2024.11.26	P:30	3	5	4	aml	
2024.11.27	P:30	3	5	4	aml	
2024.11.28	P:45	3	5	4	aml	
2024.11.29	P:30	3	5	4	aml	
2024.11.30						

date of copying

sign

부록8. 서식I-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

4
 23:08 SP: 4.0 PV: 3.6 DV: -0.4
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.4 DV: 0.4

2
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.7 DV: -0.3
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2024-11-02

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.5 DV: 0.5

2

21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.5 DV: 0.5
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2024-11-01

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.6 DV: -0.4
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 3.6 DV: -0.4

2024-11-04

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 3.7 DV: -0.3
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.6 DV: -0.4
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-11-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.1 DV: 0.1
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.6 DV: 0.6
 2024-11-06
 23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-11-05

23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.8 DV: 0.8
 12:08 SP: 4.0 PV: 4.7 DV: 0.7
 11:08 SP: 4.0 PV: 4.8 DV: 0.8
 8
 10:08 SP: 4.0 PV: 4.8 DV: 0.
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.8 DV: -0.2
 2024-11-08
 23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.6 DV: 0.6
 2024-11-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.2 DV: 0.2

1

09:08 SP: 4.0 PV: 4.1 DV: 0.
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2024-11-10

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 3.9 DV: -0.1

2

17:08 SP: 4.0 PV: 4.2 DV: 0.
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.8 DV: -0.2

.2

04:08 SP: 4.0 PV: 4.2 DV: 0
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2024-11-09

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 3.6 DV: -0.4
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 4.4 DV: 0.4

6

07:08 SP: 4.0 PV: 4.6 DV: 0.
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2024-11-12

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.5 DV: 0.5
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2024-11-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.0 DV: 0.0
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 2
 06:08 SP: 4.0 PV: 4.2 DV: 0.
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-11-14
 23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-11-13

SP Setting Point
 PV Present Value
 DV Difference Value

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.8 DV: 0.8
 00:08 SP: 4.0 PV: 4.8 DV: 0.8
 2024-11-16
 23:08 SP: 4.0 PV: 4.8 DV: 0.8
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 3
 14:08 SP: 4.0 PV: 4.3 DV: 0.
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 .7
 10:08 SP: 4.0 PV: 4.7 DV: 0
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 5
 02:08 SP: 4.0 PV: 4.5 DV: 0.
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.2 DV: 0.2
 2024-11-15

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.1 DV: 0.1

6

18:08 SP: 4.0 PV: 4.6 DV: 0.
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.8 DV: 0.8
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2024-11-18

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 3.9 DV: -0.1

3

05:08 SP: 4.0 PV: 4.3 DV: 0.
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-11-17

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.5 DV: 0.5
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-11-20

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

-- -- --

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.4 DV: 0.4

5

09:08 SP: 4.0 PV: 4.5 DV: 0.
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2024-11-22

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.8 DV: 0.8
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2

6

07:08 SP: 4.0 PV: 4.6 DV: 0.
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.8 DV: 0.8
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2024-11-21

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2024-11-24

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.1 DV: 0.1

5

10:08 SP: 4.0 PV: 4.5 DV: 0.
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-11-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.8 DV: 0.8
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 4.5 DV: 0.5
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.7 DV: 0.7
 06:08 SP: 4.0 PV: 3.8 DV: -0.2

4

05:08 SP: 4.0 PV: 4.4 DV: 0.
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.7 DV: 0.7
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2024-11-26

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.8 DV: 0.8
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 3.8 DV: -0.2

2

07:08 SP: 4.0 PV: 4.2 DV: 0.
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2024-11-25

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 3.6 DV: -0.4
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 3.6 DV: -0.4
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.0 DV: 0.0
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 3.6 DV: -0.4
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-11-28

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 3.6 DV: -0.4
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 4.8 DV: 0.8
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2024-11-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.11

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 3.9 DV: -0.1
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.5 DV: 0.5

2024-11-30

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.2 DV: 0.2
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.0 DV: 0.0
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.6 DV: -0.4
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-11-29

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

date of copying

sign