

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.11
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2025.3.7	Alarm Method	Call to Pharmacist
	Certificate No.	24005617-1	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.11.01	P:30	19	20	20	aml	
2024.11.02						
2024.11.03						
2024.11.04	P:30	19	20	20	aml	
2024.11.05	P:00	19	20	19	aml	
2024.11.06	P:30	17	20	19	aml	
2024.11.07	P:30	17	20	19	aml	
2024.11.08	P:30	17	20	19	aml	
2024.11.09						
2024.11.10						
2024.11.11	P:30	17	20	20	aml	
2024.11.12	P:30	19	20	20	aml	
2024.11.13	P:30	19	20	20	aml	
2024.11.14	P:30	17	20	19	aml	
2024.11.15	P:30	19	20	20	aml	
2024.11.16						
2024.11.17						
2024.11.18	P:30	17	20	19	aml	
2024.11.19	P:30	17	20	19	aml	
2024.11.20	P:30	17	20	19	aml	
2024.11.21	P:30	17	20	20	aml	
2024.11.22	P:30	17	20	20	aml	
2024.11.23						
2024.11.24						
2024.11.25	P:40	17	20	19	aml	
2024.11.26	P:30	17	20	19	aml	
2024.11.27	P:30	17	20	19	aml	
2024.11.28	P:45	17	20	19	aml	
2024.11.29	P:30	17	20	19	aml	
2024.11.30						

date of copying

sign