

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7		
	Date of Calibration Due	2025.3.7		
	Certificate No.	24005617-1		
	Probe Position	Right	Temperature Range	2~8℃
	Description	Digital Thermometer	Setting Point	4.0℃
	Manufacturer	SATO	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Model Name	PC-3500		
	Serial Number	2316576	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration	2024.3.7	Alarm Method	Call to Pharmacist
	Date of Calibration Due	2025.3.7	Alarm Clear	000
	Certificate No.	24005617-4		

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2024.08.01	9:30	Left	2	4	4	[Signature]	
		Right	3	5	4		
2024.08.02	10:00	Left	2	4	3	[Signature]	
		Right	3	5	3		
2024.08.03		Left					
		Right					
2024.08.04		Left					
		Right					
2024.08.05	9:30	Left	2	4	4	[Signature]	
		Right	3	5	4		
2024.08.06	9:30	Left	2	4	4	[Signature]	
		Right	3	5	4		
2024.08.07	9:30	Left	2	4	4	[Signature]	
		Right	3	5	4		
2024.08.08	9:30	Left	2	4	4	[Signature]	
		Right	3	5	4		
2024.08.09	9:30	Left	3	4	4	[Signature]	
		Right	3	5	4		
2024.08.10		Left					
		Right					
2024.08.11		Left					
		Right					

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7		
	Date of Calibration Due	2025.3.7		
	Certificate No.	24005617-1		
	Probe Position	Right	Temperature Range	2~8℃
	Description	Digital Thermometer	Setting Point	4.0℃
	Manufacturer	SATO	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Model Name	PC-3500		
	Serial Number	2316576	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration	2024.3.7	Alarm Method	Call to Pharmacist
	Date of Calibration Due	2025.3.7	Alarm Clear	000
	Certificate No.	24005617-4		

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2024.08.12	9:30	Left	2	4	4	anf	
		Right	3	5	4		
2024.08.13	9:30	Left	2	4	3	anf	
		Right	3	5	4		
2024.08.14	9:30	Left	2	4	3	anf	
		Right	3	5	4		
2024.08.15		Left					광복절
		Right					
2024.08.16	9:30	Left	2	4	4	anf	
		Right	3	5	4		
2024.08.17		Left					
		Right					
2024.08.18		Left					
		Right					
2024.08.19	9:30	Left	2	4	3	anf	
		Right	3	5	4		
2024.08.20	9:30	Left	2	4	3	anf	
		Right	3	4	4		
2024.08.21	9:30	Left	2	3	3	anf	
		Right	3	4	4		
2024.08.22	9:30	Left	2	4	4	anf	
		Right	3	5	4		

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2024.8
	Model Name	CBR1000SS	Investigational Product	냉장보관대상 IP
	Serial Number	GS17090102-10		
Calibration Subject	Probe Position	Left		
	Description	Digital Thermometer		
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7		
	Date of Calibration Due	2025.3.7		
	Certificate No.	24005617-1		
	Probe Position	Right	Temperature Range	2~8℃
	Description	Digital Thermometer	Setting Point	4.0℃
	Manufacturer	SATO	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Model Name	PC-3500		
	Serial Number	2316576	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration	2024.3.7	Alarm Method	Call to Pharmacist
	Date of Calibration Due	2025.3.7	Alarm Clear	000
	Certificate No.	24005617-4		

Date	Time	Probe Position	Temperature(℃)			Verified by	Comment
			Min	Max	Current		
2024.08.23	9:30	Left	2	4	4	ml	
		Right	3	5	5		
2024.08.24		Left					
		Right					
2024.08.25		Left					
		Right					
2024.08.26	9:30	Left	2	4	4	김지현	
		Right	3	5	4		
2024.08.27	9:30	Left	2	4	3	ml	
		Right	3	5	4		
2024.08.28	9:30	Left	2	4	3	ml	
		Right	3	4	4		1404-0004 IP
2024.08.29	9:30	Left	2	4	3	ml	
		Right	3	4	3		
2024.08.30	9:30	Left	2	4	3	ml	
		Right	3	5	4		
2024.08.31		Left					
		Right					

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-08-02
 23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.5 DV: 0.5
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.5 DV: 0.5
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.0 DV: 0.0
 2024-08-01

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.5 DV: 0.5
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.3 DV: 0.3
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 3
 16:08 SP: 4.0 PV: 4.3 DV: 0.4
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.5 DV: 0.5
 2024-08-04
 23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 3
 05:08 SP: 4.0 PV: 4.3 DV: 0.4
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.7 DV: 0.7
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 4.2 DV: 0.2
 2024-08-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.3 DV: 0.3
 2024-08-06
 23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.6 DV: 0.6
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.4 DV: 0.4
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.3 DV: 0.3
 2024-08-05

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.8 DV: 0.8
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 3
 03:08 SP: 4.0 PV: 4.3 DV: 0.
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-08-08
 23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 4.5 DV: 0.5
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.3 DV: 0.3
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 5
 16:08 SP: 4.0 PV: 4.5 DV: 0.
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.1 DV: 0.1
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 4.1 DV: 0.1
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.1 DV: 0.1
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-08-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.2 DV: 0.2
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.4 DV: 0.4

2024-08-10

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.1 DV: 0.1
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-08-09

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.2 DV: 0.2
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2024-08-12

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 4.5 DV: 0.5
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2024-08-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.4 DV: 0.4
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 4.1 DV: 0.1
 11:08 SP: 4.0 PV: 3.9 DV: -0.1
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-08-14
 23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.7 DV: -0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 3.6 DV: -0.4
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.6 DV: -0.4
 04:08 SP: 4.0 PV: 4.2 DV: 0.2
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 4.0 DV: 0.0
 2024-08-13

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 4.1 DV: 0.1
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.4 DV: -0.6
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.0 DV: 0.0
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 4
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.5 DV: 0.5
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2024-08-16
 23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.3 DV: 0.3
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-08-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.1 DV: 0.1
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 3.4 DV: -0.6
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.9 DV: -0.1
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 4.0 DV: 0.0
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.0 DV: 0.0
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2024-08-18

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.1 DV: 0.1
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2024-08-17

23:08 SP: 4.0 PV: 3.6 DV: -0.4
 22:08 SP: 4.0 PV: 3.4 DV: -0.6
 21:08 SP: 4.0 PV: 3.6 DV: -0.4
 20:08 SP: 4.0 PV: 3.4 DV: -0.6
 19:08 SP: 4.0 PV: 3.4 DV: -0.6
 18:08 SP: 4.0 PV: 3.4 DV: -0.6
 17:08 SP: 4.0 PV: 3.4 DV: -0.6
 16:08 SP: 4.0 PV: 3.4 DV: -0.6
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-08-20

1

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.6 DV: -0.4
 20:08 SP: 4.0 PV: 3.6 DV: -0.4
 19:08 SP: 4.0 PV: 3.6 DV: -0.4
 18:08 SP: 4.0 PV: 3.6 DV: -0.4
 17:08 SP: 4.0 PV: 3.6 DV: -0.4
 16:08 SP: 4.0 PV: 3.6 DV: -0.4
 15:08 SP: 4.0 PV: 3.4 DV: -0.6
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2024-08-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 3.6 DV: -0.4
 18:08 SP: 4.0 PV: 3.4 DV: -0.6
 17:08 SP: 4.0 PV: 3.6 DV: -0.4
 16:08 SP: 4.0 PV: 3.4 DV: -0.6
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 3.4 DV: -0.6
 13:08 SP: 4.0 PV: 3.4 DV: -0.6
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 3.6 DV: -0.4
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.7 DV: -0.3
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-08-22
 23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.5 DV: -0.4
 15:08 SP: 4.0 PV: 3.4 DV: -0.6
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 3.6 DV: -0.4
 09:08 SP: 4.0 PV: 3.6 DV: -0.4
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.0 DV: 0.0
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.7 DV: -0.3
 03:08 SP: 4.0 PV: 3.6 DV: -0.4
 02:08 SP: 4.0 PV: 3.4 DV: -0.6
 01:08 SP: 4.0 PV: 3.4 DV: -0.6
 00:08 SP: 4.0 PV: 3.6 DV: -0.4
 2024-08-21

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 1
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.7 DV: -0.3
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-08-24
 23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.6 DV: -0.4
 15:08 SP: 4.0 PV: 3.4 DV: -0.6
 14:08 SP: 4.0 PV: 3.4 DV: -0.6
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2024-08-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 3.7 DV: -0.3
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 3.7 DV: -0.3
 04:08 SP: 4.0 PV: 3.8 DV: -0.2

2

03:08 SP: 4.0 PV: 3.8 DV: -0.1
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-08-26

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 3.9 DV: -0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 3.8 DV: -0.2
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 3.7 DV: -0.3
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 4.0 DV: 0.0
 11:08 SP: 4.0 PV: 3.8 DV: -0.2
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.8 DV: -0.2
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2024-08-25

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 3.7 DV: -0.3
 20:08 SP: 4.0 PV: 3.7 DV: -0.3
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.7 DV: -0.3
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 3.4 DV: -0.6
 13:08 SP: 4.0 PV: 3.4 DV: -0.6
 12:08 SP: 4.0 PV: 3.4 DV: -0.6
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2024-08-28

23:08 SP: 4.0 PV: 3.6 DV: -0.4
 22:08 SP: 4.0 PV: 3.6 DV: -0.4
 21:08 SP: 4.0 PV: 3.6 DV: -0.4
 20:08 SP: 4.0 PV: 3.6 DV: -0.4
 19:08 SP: 4.0 PV: 3.6 DV: -0.4
 18:08 SP: 4.0 PV: 3.4 DV: -0.6

6

17:08 SP: 4.0 PV: 3.4 DV: -0.1
 16:08 SP: 4.0 PV: 3.4 DV: -0.6
 15:08 SP: 4.0 PV: 3.4 DV: -0.6
 14:08 SP: 4.0 PV: 3.6 DV: -0.4
 13:08 SP: 4.0 PV: 3.4 DV: -0.6
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2024-08-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2024.8

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 3.6 DV: -0.4
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 3.8 DV: -0.2
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 3.8 DV: -0.2
 08:08 SP: 4.0 PV: 3.6 DV: -0.4
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 3.6 DV: -0.4
 05:08 SP: 4.0 PV: 3.6 DV: -0.4
 04:08 SP: 4.0 PV: 3.6 DV: -0.4
 03:08 SP: 4.0 PV: 3.6 DV: -0.4
 02:08 SP: 4.0 PV: 3.6 DV: -0.4
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 3.7 DV: -0.3
 2024-08-29

23:08 SP: 4.0 PV: 3.7 DV: -0.3
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 4.1 DV: 0.1
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 3.8 DV: -0.2
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.1 DV: 0.1
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.0 DV: 0.0
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2024-08-31

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.1 DV: 0.1
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 3.7 DV: -0.3
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 3.9 DV: -0.1
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.4 DV: -0.6
 10:08 SP: 4.0 PV: 3.4 DV: -0.6
 09:08 SP: 4.0 PV: 3.4 DV: -0.6
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 4.0 DV: 0.0
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 3.6 DV: -0.4
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2024-08-30

SP Setting Point
 PV Present Value
 DV Difference Value