

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.8
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.08.01	9:30	18.7	18.9	18.8		
2024.08.02	10:00	18.7	18.9	18.8		
2024.08.03						
2024.08.04						
2024.08.05	9:30	18.6	18.8	18.8		
2024.08.06	9:30	18.6	18.9	18.7		
2024.08.07	9:30	18.6	18.9	18.8		
2024.08.08	9:30	18.6	18.9	18.8		
2024.08.09	9:30	18.7	18.9	18.7		
2024.08.10						
2024.08.11						
2024.08.12	9:30	18.6	18.9	18.8		
2024.08.13	9:30	18.7	18.9	18.8		
2024.08.14	9:30	18.7	18.9	18.8		
2024.08.15						광복절
2024.08.16	9:30	18.6	18.9	18.7		
2024.08.17						
2024.08.18						
2024.08.19	9:30	18.6	18.9	18.8		
2024.08.20	9:30	18.7	18.9	18.7		
2024.08.21	9:30	18.7	18.9	18.8		
2024.08.22	9:30	18.7	18.9	18.8		
2024.08.23	9:30	18.7	18.9	18.8		
2024.08.24						
2024.08.25						
2024.08.26	9:30	18.7	18.9	18.8		
2024.08.27	9:30	18.7	18.9	18.9		
2024.08.28	9:30	18.6	18.9	18.8		
2024.08.29	9:30	18.7	18.9	18.8		
2024.08.30	9:30	18.7	18.9	18.9		
2024.08.31						

date of copying

sign