

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.9
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	DeltaTrak		
	Model Name	12238-02		
	Serial Number	10039708		
	Date of Calibration	2022.10.3(valid for 2y)	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2024.10.3	Alarm Method	Call to Pharmacist
	Certificate No.	100397082022103	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.09.01						
2024.09.02	11:30	18.7	18.8	18.8	ml	
2024.09.03	P:30	18.7	18.9	18.8	ml	
2024.09.04	P:30	18.7	18.9	18.8	ml	
2024.09.05	P:30	18.7	18.9	18.8	ml	
2024.09.06	P:30	18.7	18.9	18.8	ml	
2024.09.07						
2024.09.08						
2024.09.09	P:30	18.7	18.9	18.9	ml	
2024.09.10	P:30	18.7	18.9	18.9	ml	
2024.09.11	P:30	18.7	19.0	18.8	ml	
2024.09.12	P:30	18.7	19.0	18.8	ml	
2024.09.13	P:30	18.7	18.9	18.8	ml	
2024.09.14						
2024.09.15						
2024.09.16						추석연휴
2024.09.17						추석
2024.09.18						추석연휴
2024.09.19	P:30	18.7	19.0	18.9	ml	
2024.09.20	P:30	18.7	18.9	18.8	ml	
2024.09.21						
2024.09.22						
2024.09.23	P:30	18.7	19.0	18.8	ml	
2024.09.24	P:30	18.7	18.9	18.9	ml	
2024.09.25	P:30	18.7	18.9	18.8	ml	
2024.09.26	P:30	18.7	18.9	18.8	ml	
2024.09.27	P:30	18.7	18.9	18.8	ml	DET 7P 062.
2024.09.28						
2024.09.29						
2024.09.30	P:30	18.7	18.9	18.8	ml	

date of copying

sign