

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2024.10
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2025.3.7	Alarm Method	Call to Pharmacist
	Certificate No.	24005617-1	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2024.10.01						국군의날, 임시공휴일
2024.10.02	P:30	18	20	20	김홍민	온도계 교체
2024.10.03						개천절
2024.10.04	P:30	18	20	19	ant	
2024.10.05						
2024.10.06						
2024.10.07	P:30	18	20	20	ant	
2024.10.08	P:30	17	20	20	ant	
2024.10.09						한글날
2024.10.10	P:30	18	20	19	ant	
2024.10.11	P:30	18	20	20	ant	
2024.10.12						
2024.10.13						
2024.10.14	P:30	18	20	18	ant	
2024.10.15	P:30	18	20	20	ant	
2024.10.16	P:30	18	20	19	ant	
2024.10.17	P:30	18	20	19	ant	
2024.10.18	P:30	18	20	19	ant	
2024.10.19						
2024.10.20						
2024.10.21	P:30	17	20	19	ant	
2024.10.22	P:30	18	20	20	ant	
2024.10.23	P:30	18	20	19	ant	
2024.10.24	P:30	17	20	19	ant	
2024.10.25	P:30	18	20	19	ant	
2024.10.26						
2024.10.27						
2024.10.28	P:30	18	20	19	ant	
2024.10.29	P:30	18	20	20	ant	
2024.10.30	P:30	18	20	20	ant	
2024.10.31	P:30	18	20	19	ant	

date of copying

sign