

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고
	Manufacturer	Giant Star	Storage Period	2025.5
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	833780		
	Date of Calibration	2025.2.25	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-7	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.05.01						근로자의 날
2025.05.02	P:30	3	5	5	anl	
2025.05.03						
2025.05.04						
2025.05.05						어린이날, 부처님오신날
2025.05.06						대체공휴일
2025.05.07	P:30	3	5	4	anl	
2025.05.08	P:30	3	5	4	anl	
2025.05.09	P:30	3	5	4	anl	
2025.05.10						
2025.05.11						
2025.05.12	P:30	3	5	3	anl	
2025.05.13	P:30	3	5	4	anl	
2025.05.14	P:30	3	5	3	anl	
2025.05.15	P:30	3	5	4	anl	
2025.05.16	P:30	3	6	3	anl	
2025.05.17						
2025.05.18						
2025.05.19	P:30	3	5	5	anl	
2025.05.20	P:30	3	5	4	anl	
2025.05.21	P:30	3	5	4	anl	
2025.05.22	P:30	3	5	4	anl	
2025.05.23	P:30	3	5	4	anl	CTC 이전(3F→9F)
2025.05.24	P:20	3	8	4	anl	
2025.05.25						
2025.05.26	P:00	3	5	5	anl	
2025.05.27	P:30	3	5	4	anl	
2025.05.28	P:25	3	5	4	anl	
2025.05.29	P:30	3	5	4	anl	
2025.05.30	P:30	3	5	3	anl	
2025.05.31						

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:08 SP: 4.0 PV: 4.1 DV: 0.1
 22:08 SP: 4.0 PV: 4.3 DV: 0.3
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 4.4 DV: 0.4
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.4 DV: 0.4
 05:08 SP: 4.0 PV: 3.9 DV: -0.1
 04:08 SP: 4.0 PV: 3.6 DV: -0.4
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 3.9 DV: -0.1
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 3.9 DV: -0.1
 2025-05-02
 23:08 SP: 4.0 PV: 3.7 DV: -0.3
 3
 22:08 SP: 4.0 PV: 4.3 DV: 0.
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.9 DV: -0.1
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 3.9 DV: -0.1
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 3.9 DV: -0.1
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.5 DV: 0.5
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 3.8 DV: -0.2
 2025-05-01

23:08 SP: 4.0 PV: 4.3 DV: 0.3
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 3.9 DV: -0.1
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.2 DV: 0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 5
 10:08 SP: 4.0 PV: 4.5 DV: 0.
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 4.5 DV: 0.5
 2025-05-04
 23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.2 DV: 0.2
 20:08 SP: 4.0 PV: 3.9 DV: -0.1
 19:08 SP: 4.0 PV: 4.1 DV: 0.1
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.5 DV: 0.5
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.3 DV: 0.3
 12:08 SP: 4.0 PV: 3.8 DV: -0.2
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.7 DV: 0.7
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.8 DV: -0.2
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.0 DV: 0.0
 00:08 SP: 4.0 PV: 4.1 DV: 0.1
 2025-05-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:08 SP: 4.0 PV: 4.7 DV: 0.7
 22:08 SP: 4.0 PV: 4.7 DV: 0.7
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.5 DV: 0.5
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.5 DV: 0.5
 16:08 SP: 4.0 PV: 4.3 DV: 0.3
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 4.5 DV: 0.5
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.5 DV: 0.5
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2025-05-06

23:08 SP: 4.0 PV: 4.2 DV: 0.2
 22:08 SP: 4.0 PV: 4.3 DV: 0.3

1

21:08 SP: 4.0 PV: 4.1 DV: 0.
 20:08 SP: 4.0 PV: 4.3 DV: 0.3
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.7 DV: 0.7
 17:08 SP: 4.0 PV: 4.6 DV: 0.6
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.5 DV: 0.5
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 4.3 DV: 0.3
 10:08 SP: 4.0 PV: 4.0 DV: 0.0
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.3 DV: 0.3
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 4.1 DV: 0.1
 05:08 SP: 4.0 PV: 4.1 DV: 0.1
 04:08 SP: 4.0 PV: 4.3 DV: 0.3
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.1 DV: 0.1

2025-05-05

SP Setting Point
 PV Present Value
 DV Difference Value

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 4.5 DV: 0.5
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.7 DV: 0.7
 16:08 SP: 4.0 PV: 4.6 DV: 0.6
 15:08 SP: 4.0 PV: 4.4 DV: 0.4
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 3.7 DV: -0.3
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 3.7 DV: -0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.4 DV: 0.4
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2025-05-08

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.0 DV: 0.0
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.0 DV: 0.0
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 3.9 DV: -0.1
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 4.4 DV: 0.4
 11:08 SP: 4.0 PV: 3.7 DV: -0.3

2

10:08 SP: 4.0 PV: 4.2 DV: 0.
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 4.4 DV: 0.4
 07:08 SP: 4.0 PV: 4.3 DV: 0.3

3

06:08 SP: 4.0 PV: 4.3 DV: 0.
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.7 DV: 0.7

2025-05-07

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:08 SP: 4.0 PV: 4.4 DV: 0.4
 22:08 SP: 4.0 PV: 3.8 DV: -0.2
 21:08 SP: 4.0 PV: 4.3 DV: 0.3
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.1 DV: 0.1
 17:08 SP: 4.0 PV: 4.4 DV: 0.4
 16:08 SP: 4.0 PV: 4.1 DV: 0.1
 15:08 SP: 4.0 PV: 3.6 DV: -0.4
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.0 DV: 0.0
 12:08 SP: 4.0 PV: 4.3 DV: 0.3
 11:08 SP: 4.0 PV: 4.1 DV: 0.1
 10:08 SP: 4.0 PV: 4.5 DV: 0.5
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.2 DV: 0.2
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 3.7 DV: -0.3
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2025-05-10

23:08 SP: 4.0 PV: 3.9 DV: -0.1
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 3.6 DV: -0.4
 20:08 SP: 4.0 PV: 4.2 DV: 0.2
 19:08 SP: 4.0 PV: 3.8 DV: -0.2
 18:08 SP: 4.0 PV: 3.6 DV: -0.4
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 4.3 DV: 0.3
 14:08 SP: 4.0 PV: 4.4 DV: 0.4
 13:08 SP: 4.0 PV: 4.6 DV: 0.6
 12:08 SP: 4.0 PV: 4.7 DV: 0.7
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.7 DV: 0.7
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 4.7 DV: 0.7
 07:08 SP: 4.0 PV: 3.8 DV: -0.2
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.4 DV: 0.4
 04:08 SP: 4.0 PV: 4.5 DV: 0.5
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 3.9 DV: -0.1
 00:08 SP: 4.0 PV: 3.9 DV: -0.1

2025-05-09

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.5 DV: 0.5

2

20:08 SP: 4.0 PV: 3.8 DV: -0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2

4

18:08 SP: 4.0 PV: 3.6 DV: -0.1
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 4.4 DV: 0.4
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.2 DV: 0.2
 13:08 SP: 4.0 PV: 4.5 DV: 0.5
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.6 DV: 0.6
 09:08 SP: 4.0 PV: 4.6 DV: 0.6
 08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 4.2 DV: 0.2
 06:08 SP: 4.0 PV: 4.5 DV: 0.5
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 4.4 DV: 0.4
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 3.7 DV: -0.3
 00:08 SP: 4.0 PV: 4.2 DV: 0.2

2025-05-12

23:08 SP: 4.0 PV: 4.0 DV: 0.0
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 4.6 DV: 0.6
 17:08 SP: 4.0 PV: 4.5 DV: 0.5

6

16:08 SP: 4.0 PV: 4.6 DV: 0.1
 15:08 SP: 4.0 PV: 3.8 DV: -0.2
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 4.2 DV: 0.2
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.2 DV: 0.2
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.4 DV: 0.4
 06:08 SP: 4.0 PV: 3.9 DV: -0.1
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.6 DV: -0.4
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 3.8 DV: -0.2
 01:08 SP: 4.0 PV: 4.3 DV: 0.3
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2025-05-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

3

23:08 SP: 4.0 PV: 4.3 DV: 0.
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 3.9 DV: -0.1
 20:08 SP: 4.0 PV: 4.1 DV: 0.1
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 3.6 DV: -0.4
 17:08 SP: 4.0 PV: 4.2 DV: 0.2
 16:08 SP: 4.0 PV: 3.8 DV: -0.2
 15:08 SP: 4.0 PV: 3.7 DV: -0.3
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 3.8 DV: -0.2
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.2 DV: 0.2
 10:08 SP: 4.0 PV: 4.3 DV: 0.3
 09:08 SP: 4.0 PV: 4.5 DV: 0.5
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.7 DV: 0.7
 04:08 SP: 4.0 PV: 4.6 DV: 0.6
 03:08 SP: 4.0 PV: 4.6 DV: 0.6
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.6 DV: 0.6
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2025-05-14

23:08 SP: 4.0 PV: 4.5 DV: 0.5
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.0 DV: 0.0
 20:08 SP: 4.0 PV: 4.4 DV: 0.4
 19:08 SP: 4.0 PV: 4.6 DV: 0.6
 18:08 SP: 4.0 PV: 3.7 DV: -0.3
 17:08 SP: 4.0 PV: 4.3 DV: 0.3
 16:08 SP: 4.0 PV: 3.9 DV: -0.1
 15:08 SP: 4.0 PV: 4.0 DV: 0.0
 14:08 SP: 4.0 PV: 4.3 DV: 0.3
 13:08 SP: 4.0 PV: 4.4 DV: 0.4
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 4.5 DV: 0.5
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 4.7 DV: 0.7
 03:08 SP: 4.0 PV: 4.6 DV: 0.6

7

02:08 SP: 4.0 PV: 4.7 DV: 0.
 01:08 SP: 4.0 PV: 4.7 DV: 0.7
 00:08 SP: 4.0 PV: 4.6 DV: 0.6

2025-05-13

SP Setting Point
 PV Present Value
 DV Difference Value

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.6 DV: 0.6
 20:08 SP: 4.0 PV: 4.7 DV: 0.7
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.6 DV: 0.6

6

17:08 SP: 4.0 PV: 4.6 DV: 0.
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.7 DV: 0.7
 14:08 SP: 4.0 PV: 4.7 DV: 0.7
 13:08 SP: 4.0 PV: 3.9 DV: -0.2
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 4.4 DV: 0.4
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 4.4 DV: 0.4
 08:08 SP: 4.0 PV: 4.6 DV: 0.6
 07:08 SP: 4.0 PV: 4.6 DV: 0.6
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 4.2 DV: 0.2
 02:08 SP: 4.0 PV: 4.2 DV: 0.2
 01:08 SP: 4.0 PV: 3.6 DV: -0.4
 00:08 SP: 4.0 PV: 4.3 DV: 0.3

2025-05-16

23:08 SP: 4.0 PV: 4.6 DV: 0.6
 22:08 SP: 4.0 PV: 4.6 DV: 0.6
 21:08 SP: 4.0 PV: 4.7 DV: 0.7
 20:08 SP: 4.0 PV: 4.6 DV: 0.6
 19:08 SP: 4.0 PV: 4.7 DV: 0.7
 18:08 SP: 4.0 PV: 4.7 DV: 0.7

6

17:08 SP: 4.0 PV: 4.6 DV: 0.
 16:08 SP: 4.0 PV: 4.7 DV: 0.7
 15:08 SP: 4.0 PV: 4.6 DV: 0.6
 14:08 SP: 4.0 PV: 4.0 DV: 0.0
 13:08 SP: 4.0 PV: 3.7 DV: -0.3
 12:08 SP: 4.0 PV: 3.9 DV: -0.1
 11:08 SP: 4.0 PV: 3.9 DV: -0.1
 10:08 SP: 4.0 PV: 4.1 DV: 0.1
 09:08 SP: 4.0 PV: 4.1 DV: 0.1
 08:08 SP: 4.0 PV: 3.7 DV: -0.3
 07:08 SP: 4.0 PV: 4.3 DV: 0.3
 06:08 SP: 4.0 PV: 3.7 DV: -0.3
 05:08 SP: 4.0 PV: 4.3 DV: 0.3
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.4 DV: 0.4
 02:08 SP: 4.0 PV: 4.6 DV: 0.6
 01:08 SP: 4.0 PV: 4.4 DV: 0.4
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2025-05-15

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:08 SP: 4.0 PV: 4.3 DV: 0.3	23:08 SP: 4.0 PV: 4.6 DV: 0.6
22:08 SP: 4.0 PV: 4.4 DV: 0.4	22:08 SP: 4.0 PV: 4.7 DV: 0.7
21:08 SP: 4.0 PV: 4.1 DV: 0.1	21:08 SP: 4.0 PV: 3.8 DV: -0.2
20:08 SP: 4.0 PV: 4.3 DV: 0.3	20:08 SP: 4.0 PV: 4.0 DV: 0.0
19:08 SP: 4.0 PV: 4.3 DV: 0.3	19:08 SP: 4.0 PV: 4.2 DV: 0.2
18:08 SP: 4.0 PV: 4.4 DV: 0.4	18:08 SP: 4.0 PV: 3.6 DV: -0.4
17:08 SP: 4.0 PV: 4.6 DV: 0.6	17:08 SP: 4.0 PV: 4.3 DV: 0.3
16:08 SP: 4.0 PV: 3.9 DV: -0.1	16:08 SP: 4.0 PV: 3.8 DV: -0.2
15:08 SP: 4.0 PV: 4.2 DV: 0.2	15:08 SP: 4.0 PV: 3.9 DV: -0.1
14:08 SP: 4.0 PV: 4.4 DV: 0.4	14:08 SP: 4.0 PV: 4.6 DV: 0.6
13:08 SP: 4.0 PV: 4.5 DV: 0.5	13:08 SP: 4.0 PV: 3.9 DV: -0.1
12:08 SP: 4.0 PV: 3.9 DV: -0.1	12:08 SP: 4.0 PV: 4.4 DV: 0.4
11:08 SP: 4.0 PV: 4.4 DV: 0.4	11:08 SP: 4.0 PV: 4.6 DV: 0.6
10:08 SP: 4.0 PV: 3.8 DV: -0.2	10:08 SP: 4.0 PV: 4.2 DV: 0.2
09:08 SP: 4.0 PV: 4.4 DV: 0.4	09:08 SP: 4.0 PV: 3.6 DV: -0.4
08:08 SP: 4.0 PV: 4.2 DV: 0.2	08:08 SP: 4.0 PV: 4.6 DV: 0.6
07:08 SP: 4.0 PV: 4.4 DV: 0.4	07:08 SP: 4.0 PV: 4.6 DV: 0.6
06:08 SP: 4.0 PV: 4.5 DV: 0.5	06:08 SP: 4.0 PV: 4.7 DV: 0.7
6	05:08 SP: 4.0 PV: 4.6 DV: 0.6
05:08 SP: 4.0 PV: 4.6 DV: 0.6	04:08 SP: 4.0 PV: 4.4 DV: 0.4
04:08 SP: 4.0 PV: 4.6 DV: 0.6	03:08 SP: 4.0 PV: 4.2 DV: 0.2
03:08 SP: 4.0 PV: 4.6 DV: 0.6	02:08 SP: 4.0 PV: 4.6 DV: 0.6
02:08 SP: 4.0 PV: 4.6 DV: 0.6	01:08 SP: 4.0 PV: 4.6 DV: 0.6
01:08 SP: 4.0 PV: 4.7 DV: 0.7	00:08 SP: 4.0 PV: 4.0 DV: 0.0
00:08 SP: 4.0 PV: 4.6 DV: 0.6	2025-05-20
2025-05-18	23:08 SP: 4.0 PV: 3.7 DV: -0.3
23:08 SP: 4.0 PV: 4.6 DV: 0.6	22:08 SP: 4.0 PV: 4.3 DV: 0.3
22:08 SP: 4.0 PV: 4.6 DV: 0.6	21:08 SP: 4.0 PV: 4.6 DV: 0.6
21:08 SP: 4.0 PV: 4.1 DV: 0.1	20:08 SP: 4.0 PV: 4.0 DV: 0.0
20:08 SP: 4.0 PV: 4.2 DV: 0.2	19:08 SP: 4.0 PV: 4.5 DV: 0.5
19:08 SP: 4.0 PV: 4.3 DV: 0.3	18:08 SP: 4.0 PV: 4.6 DV: 0.6
18:08 SP: 4.0 PV: 4.5 DV: 0.5	17:08 SP: 4.0 PV: 3.9 DV: -0.1
17:08 SP: 4.0 PV: 4.3 DV: 0.3	16:08 SP: 4.0 PV: 4.2 DV: 0.2
16:08 SP: 4.0 PV: 4.6 DV: 0.6	15:08 SP: 4.0 PV: 3.7 DV: -0.3
15:08 SP: 4.0 PV: 4.6 DV: 0.6	14:08 SP: 4.0 PV: 4.4 DV: 0.4
14:08 SP: 4.0 PV: 4.1 DV: 0.1	13:08 SP: 4.0 PV: 4.6 DV: 0.6
13:08 SP: 4.0 PV: 4.1 DV: 0.1	12:08 SP: 4.0 PV: 4.3 DV: 0.3
12:08 SP: 4.0 PV: 4.4 DV: 0.4	11:08 SP: 4.0 PV: 4.3 DV: 0.3
11:08 SP: 4.0 PV: 4.4 DV: 0.4	10:08 SP: 4.0 PV: 4.4 DV: 0.4
10:08 SP: 4.0 PV: 4.5 DV: 0.5	09:08 SP: 4.0 PV: 4.1 DV: 0.1
09:08 SP: 4.0 PV: 4.6 DV: 0.6	08:08 SP: 4.0 PV: 4.5 DV: 0.5
08:08 SP: 4.0 PV: 4.6 DV: 0.6	07:08 SP: 4.0 PV: 4.4 DV: 0.4
07:08 SP: 4.0 PV: 4.6 DV: 0.6	06:08 SP: 4.0 PV: 4.5 DV: 0.5
06:08 SP: 4.0 PV: 4.7 DV: 0.7	05:08 SP: 4.0 PV: 4.3 DV: 0.3
05:08 SP: 4.0 PV: 4.6 DV: 0.6	04:08 SP: 4.0 PV: 4.3 DV: 0.3
04:08 SP: 4.0 PV: 4.6 DV: 0.6	03:08 SP: 4.0 PV: 4.4 DV: 0.4
03:08 SP: 4.0 PV: 4.6 DV: 0.6	02:08 SP: 4.0 PV: 4.4 DV: 0.4
02:08 SP: 4.0 PV: 4.1 DV: 0.1	01:08 SP: 4.0 PV: 4.5 DV: 0.5
01:08 SP: 4.0 PV: 4.5 DV: 0.5	00:08 SP: 4.0 PV: 4.1 DV: 0.1
00:08 SP: 4.0 PV: 4.6 DV: 0.6	2025-05-19
2025-05-17	

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:08 SP: 4.0 PV: 3.6 DV: -0.4
 22:08 SP: 4.0 PV: 4.4 DV: 0.4
 21:08 SP: 4.0 PV: 3.8 DV: -0.2
 20:08 SP: 4.0 PV: 3.6 DV: -0.4
 19:08 SP: 4.0 PV: 4.2 DV: 0.2
 18:08 SP: 4.0 PV: 3.8 DV: -0.2
 17:08 SP: 4.0 PV: 3.6 DV: -0.4
 16:08 SP: 4.0 PV: 4.2 DV: 0.2
 15:08 SP: 4.0 PV: 3.9 DV: -0.1
 14:08 SP: 4.0 PV: 3.8 DV: -0.2

4

13:08 SP: 4.0 PV: 3.6 DV: -0.4
 12:08 SP: 4.0 PV: 3.6 DV: -0.4
 11:08 SP: 4.0 PV: 3.6 DV: -0.4
 10:08 SP: 4.0 PV: 3.7 DV: -0.3
 09:08 SP: 4.0 PV: 4.3 DV: 0.3
 08:08 SP: 4.0 PV: 3.9 DV: -0.1
 07:08 SP: 4.0 PV: 3.7 DV: -0.3
 06:08 SP: 4.0 PV: 4.3 DV: 0.3
 05:08 SP: 4.0 PV: 4.6 DV: 0.6
 04:08 SP: 4.0 PV: 3.9 DV: -0.1
 03:08 SP: 4.0 PV: 4.0 DV: 0.0
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 4.2 DV: 0.2
 00:08 SP: 4.0 PV: 3.8 DV: -0.2

2025-05-22

23:08 SP: 4.0 PV: 3.8 DV: -0.2
 22:08 SP: 4.0 PV: 4.0 DV: 0.0
 21:08 SP: 4.0 PV: 4.4 DV: 0.4
 20:08 SP: 4.0 PV: 3.8 DV: -0.2
 19:08 SP: 4.0 PV: 4.0 DV: 0.0
 18:08 SP: 4.0 PV: 4.0 DV: 0.0
 17:08 SP: 4.0 PV: 3.7 DV: -0.3
 16:08 SP: 4.0 PV: 4.0 DV: 0.0
 15:08 SP: 4.0 PV: 4.1 DV: 0.1
 14:08 SP: 4.0 PV: 4.6 DV: 0.6
 13:08 SP: 4.0 PV: 4.2 DV: 0.2
 12:08 SP: 4.0 PV: 3.7 DV: -0.3
 11:08 SP: 4.0 PV: 4.6 DV: 0.6
 10:08 SP: 4.0 PV: 4.2 DV: 0.2
 09:08 SP: 4.0 PV: 3.9 DV: -0.1
 08:08 SP: 4.0 PV: 4.0 DV: 0.0
 07:08 SP: 4.0 PV: 3.9 DV: -0.1
 06:08 SP: 4.0 PV: 4.6 DV: 0.6
 05:08 SP: 4.0 PV: 3.7 DV: -0.3
 04:08 SP: 4.0 PV: 4.1 DV: 0.1
 03:08 SP: 4.0 PV: 3.9 DV: -0.1
 02:08 SP: 4.0 PV: 4.1 DV: 0.1
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 4.0 DV: 0.0

2025-05-21

08:08 SP: 4.0 PV: 3.8 DV: -0.2
 07:08 SP: 4.0 PV: 3.6 DV: -0.4
 06:08 SP: 4.0 PV: 4.2 DV: 0.2
 05:08 SP: 4.0 PV: 3.8 DV: -0.2
 04:08 SP: 4.0 PV: 3.8 DV: -0.2
 03:08 SP: 4.0 PV: 3.7 DV: -0.3
 02:08 SP: 4.0 PV: 4.3 DV: 0.3
 01:08 SP: 4.0 PV: 3.8 DV: -0.2
 00:08 SP: 4.0 PV: 3.7 DV: -0.3

2025-05-23

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 3.9 DV: -0.1
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.0 DV: 0.0
 15:54 SP: 4.0 PV: 4.0 DV: 0.0
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 3.8 DV: -0.2
 12:54 SP: 4.0 PV: 3.6 DV: -0.4
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 3.6 DV: -0.4
 09:54 SP: 4.0 PV: 3.6 DV: -0.4
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 4.1 DV: 0.1

2025-05-24

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.2 DV: 0.2
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 4.0 DV: 0.0
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 3.6 DV: -0.4
 12:54 SP: 4.0 PV: 3.6 DV: -0.4
 11:54 SP: 4.0 PV: 3.6 DV: -0.4
 10:54 SP: 4.0 PV: 4.2 DV: 0.2

Condensor1 Temp#1. OK...
 Compressor#1 Output OK...
 System Checking Process...

09:54:05

2025-05-23

주소창에 한글로 혈액냉장고

www.gsmedical.co.kr

CBR-1000SS

GIANT STAR Ltd.

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

23:54 SP: 4.0 PV: 4.1 DV: 0.1
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.1 DV: 0.1
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 4.1 DV: 0.1
 10:54 SP: 4.0 PV: 3.8 DV: -0.2
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.6 DV: 0.6
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 4.1 DV: 0.1
 2025-05-26
 23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 3.6 DV: -0.4
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 3.7 DV: -0.3
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 3.6 DV: -0.4
 2025-05-25

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.5 DV: 0.5
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 3.8 DV: -0.2
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 4.1 DV: 0.1
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.6 DV: 0.6
 2025-05-28
 23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.8 DV: 0.8
 18:54 SP: 4.0 PV: 4.5 DV: 0.5
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.0 DV: 0.0
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.7 DV: -0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 4.0 DV: 0.0
 2025-05-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.5

13:54 SP: 4.0 PV: 4.4 DV: 0.4	
12:54 SP: 4.0 PV: 4.1 DV: 0.1	
11:54 SP: 4.0 PV: 4.4 DV: 0.4	
10:54 SP: 4.0 PV: 4.4 DV: 0.4	
09:54 SP: 4.0 PV: 4.7 DV: 0.7	
08:54 SP: 4.0 PV: 4.7 DV: 0.7	
07:54 SP: 4.0 PV: 4.1 DV: 0.1	
06:54 SP: 4.0 PV: 3.7 DV: -0.3	
05:54 SP: 4.0 PV: 4.4 DV: 0.4	
04:54 SP: 4.0 PV: 4.7 DV: 0.7	
03:54 SP: 4.0 PV: 4.7 DV: 0.7	
02:54 SP: 4.0 PV: 4.6 DV: 0.6	
01:54 SP: 4.0 PV: 4.4 DV: 0.4	
00:54 SP: 4.0 PV: 4.2 DV: 0.2	
2025-05-30	
23:54 SP: 4.0 PV: 4.5 DV: 0.5	
22:54 SP: 4.0 PV: 4.7 DV: 0.7	
21:54 SP: 4.0 PV: 4.6 DV: 0.6	
20:54 SP: 4.0 PV: 4.5 DV: 0.5	
19:54 SP: 4.0 PV: 4.2 DV: 0.2	
18:54 SP: 4.0 PV: 4.1 DV: 0.1	
17:54 SP: 4.0 PV: 4.5 DV: 0.5	
16:54 SP: 4.0 PV: 4.6 DV: 0.6	
15:54 SP: 4.0 PV: 4.7 DV: 0.7	
14:54 SP: 4.0 PV: 4.8 DV: 0.8	
13:54 SP: 4.0 PV: 4.8 DV: 0.8	
7	
12:54 SP: 4.0 PV: 4.7 DV: 0.7	
11:54 SP: 4.0 PV: 4.7 DV: 0.7	
10:54 SP: 4.0 PV: 3.9 DV: -0.1	
09:54 SP: 4.0 PV: 3.6 DV: -0.4	
08:54 SP: 4.0 PV: 4.5 DV: 0.5	
07:54 SP: 4.0 PV: 3.7 DV: -0.3	
06:54 SP: 4.0 PV: 4.4 DV: 0.4	
05:54 SP: 4.0 PV: 4.1 DV: 0.1	
04:54 SP: 4.0 PV: 3.7 DV: -0.3	
03:54 SP: 4.0 PV: 4.3 DV: 0.3	
02:54 SP: 4.0 PV: 3.9 DV: -0.1	
01:54 SP: 4.0 PV: 4.3 DV: 0.3	
00:54 SP: 4.0 PV: 3.6 DV: -0.4	
2025-05-29	
23:54 SP: 4.0 PV: 4.6 DV: 0.6	
22:54 SP: 4.0 PV: 3.9 DV: -0.1	
21:54 SP: 4.0 PV: 4.5 DV: 0.5	
20:54 SP: 4.0 PV: 4.4 DV: 0.4	
19:54 SP: 4.0 PV: 4.6 DV: 0.6	
18:54 SP: 4.0 PV: 4.6 DV: 0.6	
17:54 SP: 4.0 PV: 3.9 DV: -0.1	
16:54 SP: 4.0 PV: 4.5 DV: 0.5	
15:54 SP: 4.0 PV: 4.6 DV: 0.6	
14:54 SP: 4.0 PV: 3.7 DV: -0.3	
13:54 SP: 4.0 PV: 4.5 DV: 0.5	
12:54 SP: 4.0 PV: 4.4 DV: 0.4	
11:54 SP: 4.0 PV: 3.9 DV: -0.1	
10:54 SP: 4.0 PV: 4.5 DV: 0.5	
09:54 SP: 4.0 PV: 4.6 DV: 0.6	
08:54 SP: 4.0 PV: 4.4 DV: 0.4	
07:54 SP: 4.0 PV: 3.7 DV: -0.3	
06:54 SP: 4.0 PV: 4.4 DV: 0.4	
05:54 SP: 4.0 PV: 4.6 DV: 0.6	
04:54 SP: 4.0 PV: 4.5 DV: 0.5	
03:54 SP: 4.0 PV: 3.8 DV: -0.2	
02:54 SP: 4.0 PV: 4.5 DV: 0.5	
01:54 SP: 4.0 PV: 4.6 DV: 0.6	
00:54 SP: 4.0 PV: 4.2 DV: 0.2	
2025-05-31	
23:54 SP: 4.0 PV: 4.3 DV: 0.3	
22:54 SP: 4.0 PV: 4.6 DV: 0.6	
21:54 SP: 4.0 PV: 4.4 DV: 0.4	
20:54 SP: 4.0 PV: 3.7 DV: -0.3	
19:54 SP: 4.0 PV: 4.4 DV: 0.4	
18:54 SP: 4.0 PV: 4.6 DV: 0.6	
17:54 SP: 4.0 PV: 4.3 DV: 0.3	
16:54 SP: 4.0 PV: 4.6 DV: 0.6	
15:54 SP: 4.0 PV: 5.6 DV: 1.6	
14:54 SP: 4.0 PV: 4.6 DV: 0.6	

SP Setting Point
PV Present Value
DV Difference Value

Storage Subject	Description	약품냉장고	Storage Space	약제팀 냉장고(B1)
	Manufacturer		Storage Period	2025.5
	Model Name		Temperature Range	2~8°C
	Serial Number		Setting Point	4.0°C
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	855626		
	Date of Calibration	2025.2.25	Alarm Range	NA
	Date of Calibration Due	2026.2.25	Alarm Method	NA
	Certificate No.	SC2502-05400-3	Alarm Clear	NA

[illegible]

sign