

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.5
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

[illegible]

sign

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.5
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-4	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.05.01						근로자의 날
2025.05.02	P:30	17	20	20		
2025.05.03						
2025.05.04						
2025.05.05						어린이날, 부처님오신날
2025.05.06						대체공휴일
2025.05.07	P:30	17	20	19		
2025.05.08	P:30	17	20	20		
2025.05.09	P:30	17	20	19		11W316243017P.102
2025.05.10						
2025.05.11						
2025.05.12	P:30	17	20	19		
2025.05.13	P:30	17	20	20		
2025.05.14	P:30	17	20	20		
2025.05.15	P:30	17	20	19		
2025.05.16	P:30	17	20	19		
2025.05.17						
2025.05.18						
2025.05.19	P:30	17	20	19		
2025.05.20	P:30	17	20	19		휴게시간
2025.05.21						
2025.05.22						
2025.05.23						CTC 이전(3F→9F)
2025.05.24						
2025.05.25						
2025.05.26						
2025.05.27						
2025.05.28						
2025.05.29						
2025.05.30						
2025.05.31						

date of copying

sign