

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.1
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2025.3.7	Alarm Method	Call to Pharmacist
	Certificate No.	24005617-1	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.01.01						신청
2025.01.02	P:50	17	20	19	anf	
2025.01.03	P:30	17	20	18	anf	
2025.01.04						
2025.01.05						
2025.01.06	P:30	17	20	20	anf	
2025.01.07	P:30	18	20	19	anf	
2025.01.08	P:30	17	20	20	anf	
2025.01.09	P:30	17	20	19	anf	
2025.01.10	P:30	17	20	20	anf	
2025.01.11						
2025.01.12						
2025.01.13	10:30	17	20	19	anf	
2025.01.14	P:30	17	20	20	anf	
2025.01.15	P:30	17	20	20	anf	
2025.01.16	P:30	17	20	19	anf	
2025.01.17	P:30	17	20	19	anf	
2025.01.18						
2025.01.19						
2025.01.20	P:30	17	20	20	anf	
2025.01.21	P:30	17	20	20	anf	
2025.01.22	P:30	17	20	19	anf	PW3164301 IP042
2025.01.23	P:30	17	20	19	anf	
2025.01.24	P:30	17	20	19	anf	
2025.01.25						
2025.01.26						
2025.01.27						이비강염
2025.01.28						설날연휴
2025.01.29						설날
2025.01.30						설날연휴
2025.01.31	P:30	17	20	18	anf	

date of copying

sign