

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고 9134Room
	Manufacturer	Giant Star	Storage Period	2025.6
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	833780		
	Date of Calibration	2025.2.25	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-7	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.06.01						
2025.06.02	P:30	3	5	3	anl	
2025.06.03						대통령선거
2025.06.04	P:30	3	5	4	anl	
2025.06.05	P:30	3	6	4	anl	
2025.06.06						현충일
2025.06.07						
2025.06.08						
2025.06.09	Q:40	2	5	3	anl	
2025.06.10	Q:30	2	5	3	anl	
2025.06.11	Q:30	2	5	3	anl	
2025.06.12	Q:30	2	5	3	anl	
2025.06.13	P:00	2	5	4	anl	
2025.06.14						
2025.06.15						
2025.06.16	P:00	3	5	3	anl	
2025.06.17	Q:30	2	5	3	anl	
2025.06.18	P:15	3	5	3	anl	
2025.06.19	P:00	2	5	3	anl	MS IP 0102
2025.06.20	P:30	3	5	3	anl	
2025.06.21						
2025.06.22						
2025.06.23	P:30	2	5	3	anl	
2025.06.24	P:30	2	5	5	anl	
2025.06.25	P:30	3	5	3	anl	
2025.06.26	P:30	2	5	3	anl	2025.06.26 IP 0102
2025.06.27	P:30	2	5	3	anl	
2025.06.28						
2025.06.29						
2025.06.30	P:30	2	5	3	anl	

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 3.8 DV: -0.2
 22:54 SP: 4.0 PV: 3.8 DV: -0.2
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.8 DV: -0.2

2

13:54 SP: 4.0 PV: 3.8 DV: -0.1
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.6 DV: 0.6
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.6 DV: 0.6
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-06-02

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.5 DV: 0.5
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 3.8 DV: -0.2

4

15:54 SP: 4.0 PV: 4.4 DV: 0.1
 14:54 SP: 4.0 PV: 4.7 DV: 0.7
 13:54 SP: 4.0 PV: 4.6 DV: 0.6
 12:54 SP: 4.0 PV: 4.0 DV: 0.0
 11:54 SP: 4.0 PV: 4.5 DV: 0.5
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 3.6 DV: -0.4
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.6 DV: 0.6
 05:54 SP: 4.0 PV: 4.1 DV: 0.1
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.5 DV: 0.5
 02:54 SP: 4.0 PV: 4.0 DV: 0.0
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.7 DV: 0.7

2025-06-01

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.0 DV: 0.0
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 4.5 DV: 0.5
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 5.7 DV: 1.7

08:54 SP: 4.0 PV: 3.8 DV: -0.2
 07:54 SP: 4.0 PV: 4.0 DV: 0.0
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 4.0 DV: 0.0

3

03:54 SP: 4.0 PV: 4.3 DV: 0.1
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-04

23:54 SP: 4.0 PV: 3.8 DV: -0.2
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 4.1 DV: 0.1
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 3.9 DV: -0.1
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-06-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.0 DV: 0.0
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.6 DV: 0.6
 12:54 SP: 4.0 PV: 4.5 DV: 0.5
 11:54 SP: 4.0 PV: 4.5 DV: 0.5

4

10:54 SP: 4.0 PV: 4.4 DV: 0.
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.0 DV: 0.0
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.5 DV: 0.5
 03:54 SP: 4.0 PV: 4.5 DV: 0.5
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-06-06

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.5 DV: 0.5
 13:54 SP: 4.0 PV: 4.6 DV: 0.6
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 3.7 DV: -0.3
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 4.2 DV: 0.2
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-06-05

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.5 DV: 0.5
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.5 DV: 0.5
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.6 DV: 0.6

1

00:54 SP: 4.0 PV: 3.9 DV: -0.

2025-06-08

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.5 DV: 0.5
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.5 DV: 0.5
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.5 DV: 0.5
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.0 DV: 0.0
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.0 DV: 0.0
 03:54 SP: 4.0 PV: 4.5 DV: 0.5
 02:54 SP: 4.0 PV: 4.5 DV: 0.5

6

01:54 SP: 4.0 PV: 4.6 DV: 0.

00:54 SP: 4.0 PV: 4.6 DV: 0.6

2025-06-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 4.1 DV: 0.1
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.1 DV: 0.1
 05:54 SP: 4.0 PV: 3.8 DV: -0.2
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.6 DV: 0.6
 00:54 SP: 4.0 PV: 4.6 DV: 0.6
 2025-06-10
 23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 4.5 DV: 0.5
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.5 DV: 0.5
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 4.0 DV: 0.0
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.5 DV: 0.5
 00:54 SP: 4.0 PV: 4.0 DV: 0.0
 2025-06-09

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 6
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 1
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 1
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 4
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.5 DV: 0.5
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-06-12
 23:54 SP: 4.0 PV: 4.1 DV: 0.1
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.5 DV: 0.5
 17:54 SP: 4.0 PV: 4.6 DV: 0.6
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.5 DV: 0.5
 14:54 SP: 4.0 PV: 4.5 DV: 0.5
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 4.0 DV: 0.0
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.5 DV: 0.5
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.1 DV: 0.1
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.6 DV: 0.6
 2025-06-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.5 DV: 0.5
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 4.5 DV: 0.5
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.6 DV: 0.6
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 4.0 DV: 0.0

2025-06-14

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.6 DV: 0.6
 17:54 SP: 4.0 PV: 4.6 DV: 0.6
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 3.9 DV: -0.1
 06:54 SP: 4.0 PV: 3.7 DV: -0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.5 DV: 0.5
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.6 DV: 0.6
 00:54 SP: 4.0 PV: 4.6 DV: 0.6

2025-06-13

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.5 DV: 0.5
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.5 DV: 0.5
 10:54 SP: 4.0 PV: 4.5 DV: 0.5
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-06-16

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.6 DV: 0.6
 17:54 SP: 4.0 PV: 4.4 DV: 0.4

6

16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.5 DV: 0.5

2025-06-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

3

23:54 SP: 4.0 PV: 4.3 DV: 0.
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.6 DV: 0.6
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.6 DV: 0.6
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.5 DV: 0.5
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-06-18

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.5 DV: 0.5
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.5 DV: 0.5
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.5 DV: 0.5
 10:54 SP: 4.0 PV: 4.7 DV: 0.7
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.5 DV: 0.5
 01:54 SP: 4.0 PV: 4.3 DV: 0.3

6

00:54 SP: 4.0 PV: 4.6 DV: 0.
 2025-06-17

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 3.9 DV: -0.2
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.1 DV: 0.1
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.2 DV: 0.2

0

10:54 SP: 4.0 PV: 4.0 DV: 0.
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.1 DV: 0.1
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 4.2 DV: 0.2
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-20

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.5 DV: 0.5
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 4.2 DV: 0.2
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-06-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 4.2 DV: 0.2
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 4.0 DV: 0.0
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.0 DV: 0.0
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 4.0 DV: 0.0
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.1 DV: 0.1
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 4.0 DV: 0.0
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-22

23:54 SP: 4.0 PV: 3.8 DV: -0.2
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 4.0 DV: 0.0
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 4.2 DV: 0.2
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 3.9 DV: -0.1
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.0 DV: 0.0
 06:54 SP: 4.0 PV: 3.7 DV: -0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3

2

04:54 SP: 4.0 PV: 3.8 DV: -0.1
 03:54 SP: 4.0 PV: 4.1 DV: 0.1
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-21

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.0 DV: 0.0
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 3.9 DV: -0.1
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 3.9 DV: -0.1
 06:54 SP: 4.0 PV: 4.1 DV: 0.1
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 3.7 DV: -0.3

2025-06-24

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.2 DV: 0.2
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 3.6 DV: -0.4
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 3.8 DV: -0.2
 12:54 SP: 4.0 PV: 4.1 DV: 0.1
 11:54 SP: 4.0 PV: 3.8 DV: -0.2
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 4.1 DV: 0.1
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 3.6 DV: -0.4
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-06-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.1 DV: 0.1
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.6 DV: 0.6
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 3.8 DV: -0.2
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 4.0 DV: 0.0
 03:54 SP: 4.0 PV: 4.1 DV: 0.1
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-06-26

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 3.8 DV: -0.2
 21:54 SP: 4.0 PV: 4.2 DV: 0.2
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.1 DV: 0.1
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 4.2 DV: 0.2
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-06-25

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 4.0 DV: 0.0
 21:54 SP: 4.0 PV: 4.2 DV: 0.2
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 3.7 DV: -0.3
 02:54 SP: 4.0 PV: 4.1 DV: 0.1
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-28

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 4.0 DV: 0.0
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.1 DV: 0.1
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.1 DV: 0.1
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-06-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.6

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 3.9 DV: -0.1
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.0 DV: 0.0
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 4.1 DV: 0.1
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 4.0 DV: 0.0
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-06-30

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 3.9 DV: -0.1
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 3.8 DV: -0.2
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 3.8 DV: -0.2
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 3.8 DV: -0.2
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 4.0 DV: 0.0

2025-06-29

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

date of copying

sign