

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.3
	Model Name	MPR-720R	Temperature Range	15~25°C
	Serial Number	07120004	Setting Point	19.0°C
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2025.2.25	Alarm Range	<15°C, >25°C, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-4	Alarm Clear	000


Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2025.03.01						
2025.03.02						
2025.03.03						삼일절 대체공휴일
2025.03.04	14:50	20	20	20	anf	온도계 교체
2025.03.05	P:30	17	20	20	anf	
2025.03.06	P:30	17	20	19	anf	
2025.03.07	P:30	17	20	18	anf	
2025.03.08						
2025.03.09						
2025.03.10	P:30	17	20	18	anf	
2025.03.11	P:30	17	20	18	anf	
2025.03.12	P:30	17	20	19	anf	
2025.03.13	P:30	17	20	18	anf	
2025.03.14	P:30	17	20	17	anf	
2025.03.15						
2025.03.16						
2025.03.17	P:30	17	20	20	anf	
2025.03.18	P:30	17	20	19	anf	
2025.03.19	P:30	17	20	19	anf	
2025.03.20	P:30	17	20	17	anf	
2025.03.21	P:30	17	20	17	anf	
2025.03.22						
2025.03.23						
2025.03.24	P:30	17	20	20	anf	HB2202 IP 온도
2025.03.25	P:30	17	20	18	anf	
2025.03.26	P:30	17	20	19	anf	
2025.03.27	P:30	17	20	20	anf	
2025.03.28	P:30	17	19	19	anf	
2025.03.29						
2025.03.30						
2025.03.31	P:30	17	20	18	anf	

date of copying

sign

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.3
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	2102992		
	Date of Calibration	2024.3.7	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2025.3.7	Alarm Method	Call to Pharmacist
	Certificate No.	24005617-1	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.03.01						
2025.03.02						
2025.03.03						삼일절 대체공휴일
2025.03.04	9:30	17	20	19		
2025.03.05	14:50	17	20	20		
2025.03.06						
2025.03.07						
2025.03.08						
2025.03.09						
2025.03.10						
2025.03.11						
2025.03.12						
2025.03.13						
2025.03.14						
2025.03.15						
2025.03.16						
2025.03.17						
2025.03.18						
2025.03.19						
2025.03.20						
2025.03.21						
2025.03.22						
2025.03.23						
2025.03.24						
2025.03.25						
2025.03.26						
2025.03.27						
2025.03.28						
2025.03.29						
2025.03.30						
2025.03.31						

date of copying

sign