

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1
	Manufacturer	SANYO	Storage Period	2025.4
	Model Name	MPR-720R	Temperature Range	15~25°C
	Serial Number	07120004	Setting Point	19.0°C
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	848811		
	Date of Calibration	2025.2.25	Alarm Range	<15°C, >25°C, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-4	Alarm Clear	000

Date	Time	Temperature(°C)			Verified by	Comment
		Min	Max	Current		
2025.04.01	P: 30	17	20	19	anf	
2025.04.02	P: 30	17	20	20	anf	
2025.04.03	P: 30	17	20	18	anf	
2025.04.04	P: 30	17	20	19	anf	
2025.04.05						
2025.04.06						
2025.04.07	P: 30	17	20	20	anf	
2025.04.08	P: 30	17	20	19	anf	
2025.04.09	P: 30	17	20	18	anf	
2025.04.10	P: 30	17	20	19	anf	
2025.04.11	P: 30	17	20	18	anf	
2025.04.12						
2025.04.13						
2025.04.14	P: 30	17	20	19	anf	
2025.04.15	P: 30	17	20	19	anf	
2025.04.16	P: 30	17	20	18	anf	
2025.04.17	P: 40	17	20	18	anf	
2025.04.18	P: 30	17	20	17	anf	
2025.04.19						
2025.04.20						
2025.04.21	P: 40	17	20	20	anf	
2025.04.22	P: 30	17	20	17	anf	
2025.04.23	P: 30	17	20	20	anf	
2025.04.24	P: 30	17	20	19	anf	
2025.04.25	P: 30	17	20	19	anf	
2025.04.26						
2025.04.27						
2025.04.28	P: 30	17	20	20	anf	
2025.04.29	P: 30	17	20	20	anf	
2025.04.30	P: 30	17	20	17	anf	

date of copying

sign