

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고 9134Room
	Manufacturer	Giant Star	Storage Period	2025.7
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	833780		
	Date of Calibration	2025.2.25	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-7	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.07.01	P: 50	2	5	3	ml	
2025.07.02	P: 40	2	5	4	ml	
2025.07.03	P: 30	2	5	4	ml	
2025.07.04	P: 30	2	5	4	ml	
2025.07.05						
2025.07.06						
2025.07.07	P: 30	2	5	4	ml	
2025.07.08	P: 30	3	5	4	ml	
2025.07.09	P: 30	3	5	4	ml	
2025.07.10	P: 20	3	5	3	ml	
2025.07.11	P: 30	3	5	3	ml	
2025.07.12						
2025.07.13						
2025.07.14	P: 30	3	5	3	ml	
2025.07.15	P: 30	3	5	4	ml	
2025.07.16	P: 30	3	5	3	ml	
2025.07.17	P: 30	3	5	3	ml	
2025.07.18	P: 30	3	5	3	ml	
2025.07.19						
2025.07.20						
2025.07.21	P: 30	3	5	4	ml	
2025.07.22	P: 00	3	5	4	ml	
2025.07.23	P: 30	3	5	4	ml	
2025.07.24	P: 40	3	5	3	ml	
2025.07.25	P: 30	3	5	4	ml	
2025.07.26						
2025.07.27						
2025.07.28	P: 30	2	5	3	ml	
2025.07.29	P: 30	3	5	3	ml	
2025.07.30	P: 30	2	5	3	ml	
2025.07.31	P: 20	2	5	3	ml	

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.4 DV: 0.4
2
22:54 SP: 4.0 PV: 3.8 DV: -0.1
21:54 SP: 4.0 PV: 4.3 DV: 0.3
20:54 SP: 4.0 PV: 3.8 DV: -0.2
19:54 SP: 4.0 PV: 3.6 DV: -0.4
18:54 SP: 4.0 PV: 4.2 DV: 0.2
17:54 SP: 4.0 PV: 3.8 DV: -0.2
16:54 SP: 4.0 PV: 4.1 DV: 0.1
15:54 SP: 4.0 PV: 4.1 DV: 0.1
14:54 SP: 4.0 PV: 4.0 DV: 0.0
13:54 SP: 4.0 PV: 3.7 DV: -0.3
12:54 SP: 4.0 PV: 4.2 DV: 0.2
11:54 SP: 4.0 PV: 3.7 DV: -0.3
10:54 SP: 4.0 PV: 4.2 DV: 0.2
09:54 SP: 4.0 PV: 3.7 DV: -0.3
08:54 SP: 4.0 PV: 4.2 DV: 0.2
07:54 SP: 4.0 PV: 3.7 DV: -0.3
06:54 SP: 4.0 PV: 4.1 DV: 0.1
05:54 SP: 4.0 PV: 3.8 DV: -0.2
04:54 SP: 4.0 PV: 4.3 DV: 0.3
03:54 SP: 4.0 PV: 3.9 DV: -0.1
02:54 SP: 4.0 PV: 4.1 DV: 0.1
01:54 SP: 4.0 PV: 4.0 DV: 0.0
00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-07-02

23:54 SP: 4.0 PV: 3.7 DV: -0.3
22:54 SP: 4.0 PV: 4.2 DV: 0.2
21:54 SP: 4.0 PV: 3.9 DV: -0.1
20:54 SP: 4.0 PV: 4.1 DV: 0.1
19:54 SP: 4.0 PV: 4.0 DV: 0.0
18:54 SP: 4.0 PV: 4.3 DV: 0.3
17:54 SP: 4.0 PV: 3.8 DV: -0.2
16:54 SP: 4.0 PV: 4.3 DV: 0.3
15:54 SP: 4.0 PV: 3.9 DV: -0.1
14:54 SP: 4.0 PV: 4.1 DV: 0.1
13:54 SP: 4.0 PV: 3.7 DV: -0.3
12:54 SP: 4.0 PV: 4.3 DV: 0.3
11:54 SP: 4.0 PV: 3.8 DV: -0.2
10:54 SP: 4.0 PV: 4.4 DV: 0.4
09:54 SP: 4.0 PV: 3.8 DV: -0.2
08:54 SP: 4.0 PV: 4.3 DV: 0.3
07:54 SP: 4.0 PV: 3.8 DV: -0.2
06:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 3.9 DV: -0.1
04:54 SP: 4.0 PV: 4.0 DV: 0.0
03:54 SP: 4.0 PV: 4.1 DV: 0.1
02:54 SP: 4.0 PV: 4.1 DV: 0.1
01:54 SP: 4.0 PV: 4.0 DV: 0.0
00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-07-01

23:54 SP: 4.0 PV: 3.9 DV: -0.1
3
22:54 SP: 4.0 PV: 3.7 DV: -0.1
21:54 SP: 4.0 PV: 4.0 DV: 0.0
20:54 SP: 4.0 PV: 3.8 DV: -0.2
19:54 SP: 4.0 PV: 3.7 DV: -0.3
18:54 SP: 4.0 PV: 3.7 DV: -0.3
17:54 SP: 4.0 PV: 4.0 DV: 0.0
16:54 SP: 4.0 PV: 3.8 DV: -0.2
15:54 SP: 4.0 PV: 3.6 DV: -0.4
14:54 SP: 4.0 PV: 3.6 DV: -0.4
13:54 SP: 4.0 PV: 4.2 DV: 0.2
12:54 SP: 4.0 PV: 4.6 DV: 0.6
11:54 SP: 4.0 PV: 4.6 DV: 0.6
10:54 SP: 4.0 PV: 4.4 DV: 0.4
09:54 SP: 4.0 PV: 4.1 DV: 0.1
08:54 SP: 4.0 PV: 3.9 DV: -0.1
07:54 SP: 4.0 PV: 3.7 DV: -0.3
06:54 SP: 4.0 PV: 4.1 DV: 0.1
05:54 SP: 4.0 PV: 3.8 DV: -0.2
04:54 SP: 4.0 PV: 3.7 DV: -0.3
03:54 SP: 4.0 PV: 3.6 DV: -0.4
02:54 SP: 4.0 PV: 3.7 DV: -0.3
01:54 SP: 4.0 PV: 4.1 DV: 0.1
00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-07-04

23:54 SP: 4.0 PV: 3.9 DV: -0.1
22:54 SP: 4.0 PV: 3.9 DV: -0.1
21:54 SP: 4.0 PV: 3.9 DV: -0.1
20:54 SP: 4.0 PV: 3.8 DV: -0.2
19:54 SP: 4.0 PV: 3.8 DV: -0.2
18:54 SP: 4.0 PV: 3.7 DV: -0.3
17:54 SP: 4.0 PV: 3.7 DV: -0.3
16:54 SP: 4.0 PV: 4.1 DV: 0.1
15:54 SP: 4.0 PV: 3.9 DV: -0.1
14:54 SP: 4.0 PV: 3.7 DV: -0.3
13:54 SP: 4.0 PV: 4.1 DV: 0.1
12:54 SP: 4.0 PV: 3.8 DV: -0.2
11:54 SP: 4.0 PV: 3.6 DV: -0.4
10:54 SP: 4.0 PV: 3.6 DV: -0.4
09:54 SP: 4.0 PV: 4.6 DV: 0.6
08:54 SP: 4.0 PV: 3.7 DV: -0.3
07:54 SP: 4.0 PV: 3.6 DV: -0.4
06:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 3.9 DV: -0.1
04:54 SP: 4.0 PV: 4.1 DV: 0.1
03:54 SP: 4.0 PV: 3.7 DV: -0.3
02:54 SP: 4.0 PV: 4.2 DV: 0.2
01:54 SP: 4.0 PV: 3.9 DV: -0.1
00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-07-03

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.1 DV: 0.1
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.0 DV: 0.0
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.0 DV: 0.0
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.0 DV: 0.0
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 3.8 DV: -0.2
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-07-06

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 3.8 DV: -0.2
 21:54 SP: 4.0 PV: 4.0 DV: 0.0
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.1 DV: 0.1
 11:54 SP: 4.0 PV: 3.8 DV: -0.2
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 3.9 DV: -0.1
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.0 DV: 0.0
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 4.0 DV: 0.0
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-07-05

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 3.8 DV: -0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.0 DV: 0.0
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 3.8 DV: -0.2
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.1 DV: 0.1
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 3.9 DV: -0.1
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.9 DV: -0.1
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 3.8 DV: -0.2
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 3.7 DV: -0.3

2025-07-08

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.0 DV: 0.0
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.0 DV: 0.0
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.0 DV: 0.0
 05:54 SP: 4.0 PV: 3.7 DV: -0.3
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.1 DV: 0.1
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 3.7 DV: -0.3

2025-07-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.3 DV: 0.3
3
22:54 SP: 4.0 PV: 4.3 DV: 0.
21:54 SP: 4.0 PV: 4.4 DV: 0.4
20:54 SP: 4.0 PV: 3.9 DV: -0.1
19:54 SP: 4.0 PV: 4.2 DV: 0.2
18:54 SP: 4.0 PV: 4.3 DV: 0.3
17:54 SP: 4.0 PV: 4.4 DV: 0.4
16:54 SP: 4.0 PV: 4.4 DV: 0.4
15:54 SP: 4.0 PV: 4.6 DV: 0.6
14:54 SP: 4.0 PV: 4.6 DV: 0.6
13:54 SP: 4.0 PV: 4.4 DV: 0.4
12:54 SP: 4.0 PV: 4.5 DV: 0.5
11:54 SP: 4.0 PV: 4.4 DV: 0.4
10:54 SP: 4.0 PV: 4.5 DV: 0.5
09:54 SP: 4.0 PV: 4.4 DV: 0.4
08:54 SP: 4.0 PV: 4.4 DV: 0.4
07:54 SP: 4.0 PV: 4.3 DV: 0.3
06:54 SP: 4.0 PV: 4.3 DV: 0.3
05:54 SP: 4.0 PV: 4.4 DV: 0.4
04:54 SP: 4.0 PV: 4.4 DV: 0.4
03:54 SP: 4.0 PV: 4.3 DV: 0.3
02:54 SP: 4.0 PV: 4.3 DV: 0.3
3
01:54 SP: 4.0 PV: 4.3 DV: 0.
00:54 SP: 4.0 PV: 4.4 DV: 0.4
2025-07-10
23:54 SP: 4.0 PV: 4.4 DV: 0.4
22:54 SP: 4.0 PV: 4.3 DV: 0.3
21:54 SP: 4.0 PV: 4.4 DV: 0.4
20:54 SP: 4.0 PV: 4.3 DV: 0.3
19:54 SP: 4.0 PV: 4.3 DV: 0.3
18:54 SP: 4.0 PV: 4.3 DV: 0.3
17:54 SP: 4.0 PV: 4.4 DV: 0.4
16:54 SP: 4.0 PV: 4.4 DV: 0.4
15:54 SP: 4.0 PV: 3.9 DV: -0.1
14:54 SP: 4.0 PV: 4.3 DV: 0.3
13:54 SP: 4.0 PV: 3.9 DV: -0.1
12:54 SP: 4.0 PV: 3.7 DV: -0.3
11:54 SP: 4.0 PV: 4.2 DV: 0.2
10:54 SP: 4.0 PV: 3.8 DV: -0.2
09:54 SP: 4.0 PV: 3.7 DV: -0.3
08:54 SP: 4.0 PV: 4.1 DV: 0.1
07:54 SP: 4.0 PV: 3.8 DV: -0.2
06:54 SP: 4.0 PV: 4.0 DV: 0.0
05:54 SP: 4.0 PV: 3.8 DV: -0.2
04:54 SP: 4.0 PV: 3.7 DV: -0.3
03:54 SP: 4.0 PV: 4.0 DV: 0.0
02:54 SP: 4.0 PV: 3.7 DV: -0.3
01:54 SP: 4.0 PV: 4.0 DV: 0.0
00:54 SP: 4.0 PV: 3.8 DV: -0.2
2025-07-09

23:54 SP: 4.0 PV: 4.4 DV: 0.4
22:54 SP: 4.0 PV: 4.4 DV: 0.4
21:54 SP: 4.0 PV: 4.4 DV: 0.4
20:54 SP: 4.0 PV: 4.5 DV: 0.5
19:54 SP: 4.0 PV: 4.4 DV: 0.4
18:54 SP: 4.0 PV: 4.3 DV: 0.3
17:54 SP: 4.0 PV: 4.1 DV: 0.1
16:54 SP: 4.0 PV: 4.2 DV: 0.2
15:54 SP: 4.0 PV: 4.4 DV: 0.4
14:54 SP: 4.0 PV: 3.9 DV: -0.1
13:54 SP: 4.0 PV: 4.2 DV: 0.2
12:54 SP: 4.0 PV: 4.4 DV: 0.4
11:54 SP: 4.0 PV: 4.4 DV: 0.4
3
10:54 SP: 4.0 PV: 4.3 DV: 0.
09:54 SP: 4.0 PV: 4.3 DV: 0.3
08:54 SP: 4.0 PV: 4.3 DV: 0.3
07:54 SP: 4.0 PV: 4.3 DV: 0.3
06:54 SP: 4.0 PV: 4.1 DV: 0.1
05:54 SP: 4.0 PV: 4.3 DV: 0.3
04:54 SP: 4.0 PV: 4.1 DV: 0.1
03:54 SP: 4.0 PV: 4.2 DV: 0.2
02:54 SP: 4.0 PV: 4.3 DV: 0.3
01:54 SP: 4.0 PV: 4.4 DV: 0.4
00:54 SP: 4.0 PV: 4.5 DV: 0.5
2025-07-12
23:54 SP: 4.0 PV: 4.4 DV: 0.4
22:54 SP: 4.0 PV: 4.4 DV: 0.4
21:54 SP: 4.0 PV: 4.4 DV: 0.4
20:54 SP: 4.0 PV: 4.4 DV: 0.4
19:54 SP: 4.0 PV: 4.4 DV: 0.4
18:54 SP: 4.0 PV: 4.3 DV: 0.3
17:54 SP: 4.0 PV: 4.3 DV: 0.3
16:54 SP: 4.0 PV: 4.1 DV: 0.1
15:54 SP: 4.0 PV: 4.2 DV: 0.2
14:54 SP: 4.0 PV: 4.3 DV: 0.3
13:54 SP: 4.0 PV: 4.2 DV: 0.2
12:54 SP: 4.0 PV: 4.3 DV: 0.3
11:54 SP: 4.0 PV: 4.3 DV: 0.3
10:54 SP: 4.0 PV: 4.3 DV: 0.3
09:54 SP: 4.0 PV: 4.3 DV: 0.3
08:54 SP: 4.0 PV: 4.3 DV: 0.3
2
07:54 SP: 4.0 PV: 4.2 DV: 0.
06:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 4.2 DV: 0.2
04:54 SP: 4.0 PV: 4.0 DV: 0.0
03:54 SP: 4.0 PV: 4.2 DV: 0.2
02:54 SP: 4.0 PV: 4.3 DV: 0.3
01:54 SP: 4.0 PV: 4.2 DV: 0.2
00:54 SP: 4.0 PV: 4.3 DV: 0.3
2025-07-11

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.4 DV: 0.4	-
22:54 SP: 4.0 PV: 4.5 DV: 0.5	23:54 SP: 4.0 PV: 4.4 DV: 0.4
21:54 SP: 4.0 PV: 4.4 DV: 0.4	22:54 SP: 4.0 PV: 4.6 DV: 0.6
20:54 SP: 4.0 PV: 4.4 DV: 0.4	21:54 SP: 4.0 PV: 4.5 DV: 0.5
19:54 SP: 4.0 PV: 4.3 DV: 0.3	20:54 SP: 4.0 PV: 4.4 DV: 0.4
18:54 SP: 4.0 PV: 4.4 DV: 0.4	19:54 SP: 4.0 PV: 4.4 DV: 0.4
17:54 SP: 4.0 PV: 4.4 DV: 0.4	18:54 SP: 4.0 PV: 4.3 DV: 0.3
16:54 SP: 4.0 PV: 4.4 DV: 0.4	17:54 SP: 4.0 PV: 4.3 DV: 0.3
15:54 SP: 4.0 PV: 4.4 DV: 0.4	16:54 SP: 4.0 PV: 4.3 DV: 0.3
14:54 SP: 4.0 PV: 4.3 DV: 0.3	15:54 SP: 4.0 PV: 4.4 DV: 0.4
13:54 SP: 4.0 PV: 4.4 DV: 0.4	14:54 SP: 4.0 PV: 3.9 DV: -0.1
12:54 SP: 4.0 PV: 4.4 DV: 0.4	13:54 SP: 4.0 PV: 3.8 DV: -0.2
11:54 SP: 4.0 PV: 4.6 DV: 0.6	12:54 SP: 4.0 PV: 4.3 DV: 0.3
10:54 SP: 4.0 PV: 4.4 DV: 0.4	11:54 SP: 4.0 PV: 4.5 DV: 0.5
09:54 SP: 4.0 PV: 4.4 DV: 0.4	10:54 SP: 4.0 PV: 4.3 DV: 0.3
08:54 SP: 4.0 PV: 4.4 DV: 0.4	09:54 SP: 4.0 PV: 4.4 DV: 0.4
07:54 SP: 4.0 PV: 4.4 DV: 0.4	08:54 SP: 4.0 PV: 4.5 DV: 0.5
06:54 SP: 4.0 PV: 4.4 DV: 0.4	07:54 SP: 4.0 PV: 4.5 DV: 0.5
05:54 SP: 4.0 PV: 4.5 DV: 0.5	06:54 SP: 4.0 PV: 4.5 DV: 0.5
04:54 SP: 4.0 PV: 4.3 DV: 0.3	05:54 SP: 4.0 PV: 4.4 DV: 0.4
03:54 SP: 4.0 PV: 4.3 DV: 0.3	04:54 SP: 4.0 PV: 4.4 DV: 0.4
02:54 SP: 4.0 PV: 4.3 DV: 0.3	03:54 SP: 4.0 PV: 4.4 DV: 0.4
01:54 SP: 4.0 PV: 4.3 DV: 0.3	02:54 SP: 4.0 PV: 4.4 DV: 0.4
00:54 SP: 4.0 PV: 4.3 DV: 0.3	01:54 SP: 4.0 PV: 4.6 DV: 0.6
2025-07-14	00:54 SP: 4.0 PV: 4.5 DV: 0.5
23:54 SP: 4.0 PV: 4.3 DV: 0.3	2025-07-16
22:54 SP: 4.0 PV: 4.3 DV: 0.3	23:54 SP: 4.0 PV: 4.4 DV: 0.4
21:54 SP: 4.0 PV: 4.3 DV: 0.3	22:54 SP: 4.0 PV: 4.4 DV: 0.4
20:54 SP: 4.0 PV: 4.4 DV: 0.4	21:54 SP: 4.0 PV: 4.1 DV: 0.1
19:54 SP: 4.0 PV: 4.5 DV: 0.5	20:54 SP: 4.0 PV: 4.2 DV: 0.2
18:54 SP: 4.0 PV: 4.6 DV: 0.6	19:54 SP: 4.0 PV: 4.4 DV: 0.4
17:54 SP: 4.0 PV: 4.4 DV: 0.4	18:54 SP: 4.0 PV: 4.5 DV: 0.5
16:54 SP: 4.0 PV: 4.4 DV: 0.4	17:54 SP: 4.0 PV: 4.6 DV: 0.6
15:54 SP: 4.0 PV: 4.6 DV: 0.6	16:54 SP: 4.0 PV: 4.5 DV: 0.5
14:54 SP: 4.0 PV: 4.4 DV: 0.4	15:54 SP: 4.0 PV: 4.5 DV: 0.5
13:54 SP: 4.0 PV: 4.4 DV: 0.4	14:54 SP: 4.0 PV: 4.5 DV: 0.5
12:54 SP: 4.0 PV: 4.3 DV: 0.3	13:54 SP: 4.0 PV: 4.5 DV: 0.5
11:54 SP: 4.0 PV: 4.4 DV: 0.4	12:54 SP: 4.0 PV: 4.5 DV: 0.5
10:54 SP: 4.0 PV: 4.4 DV: 0.4	11:54 SP: 4.0 PV: 4.5 DV: 0.5
09:54 SP: 4.0 PV: 4.4 DV: 0.4	10:54 SP: 4.0 PV: 4.4 DV: 0.4
08:54 SP: 4.0 PV: 4.4 DV: 0.4	09:54 SP: 4.0 PV: 4.2 DV: 0.2
07:54 SP: 4.0 PV: 4.4 DV: 0.4	08:54 SP: 4.0 PV: 4.3 DV: 0.3
06:54 SP: 4.0 PV: 4.4 DV: 0.4	07:54 SP: 4.0 PV: 4.6 DV: 0.6
05:54 SP: 4.0 PV: 4.4 DV: 0.4	06:54 SP: 4.0 PV: 4.4 DV: 0.4
04:54 SP: 4.0 PV: 4.4 DV: 0.4	05:54 SP: 4.0 PV: 4.5 DV: 0.5
03:54 SP: 4.0 PV: 4.4 DV: 0.4	04:54 SP: 4.0 PV: 4.4 DV: 0.4
02:54 SP: 4.0 PV: 4.3 DV: 0.3	03:54 SP: 4.0 PV: 4.5 DV: 0.5
01:54 SP: 4.0 PV: 4.4 DV: 0.4	02:54 SP: 4.0 PV: 4.4 DV: 0.4
00:54 SP: 4.0 PV: 4.5 DV: 0.5	01:54 SP: 4.0 PV: 4.4 DV: 0.4
2025-07-13	00:54 SP: 4.0 PV: 4.4 DV: 0.4
	2025-07-15

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.5 DV: 0.5
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-07-18
 23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 2
 15:54 SP: 4.0 PV: 4.2 DV: 0.
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 4
 09:54 SP: 4.0 PV: 4.4 DV: 0.
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.6 DV: 0.6
 04:54 SP: 4.0 PV: 4.5 DV: 0.5
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 4.0 DV: 0.0
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 5
 00:54 SP: 4.0 PV: 4.5 DV: 0.
 2025-07-17

2025-07-19
 23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 4.5 DV: 0.5
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 4
 15:54 SP: 4.0 PV: 4.4 DV: 0.
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.6 DV: 0.6
 12:54 SP: 4.0 PV: 4.6 DV: 0.6
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.5 DV: 0.5
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.2 DV: 0.2
 3
 03:54 SP: 4.0 PV: 4.3 DV: 0.
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-07-20
 23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.5 DV: 0.5
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 4.5 DV: 0.5
 11:54 SP: 4.0 PV: 4.5 DV: 0.5
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.4 DV: 0.4
 2025-07-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.5 DV: 0.5
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 4.6 DV: 0.6
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.4 DV: 0.4
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.5 DV: 0.5
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-07-22

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.6 DV: 0.6
 17:54 SP: 4.0 PV: 4.6 DV: 0.6
 16:54 SP: 4.0 PV: 4.5 DV: 0.5
 15:54 SP: 4.0 PV: 4.5 DV: 0.5
 14:54 SP: 4.0 PV: 4.5 DV: 0.5
 13:54 SP: 4.0 PV: 4.5 DV: 0.5
 12:54 SP: 4.0 PV: 4.5 DV: 0.5
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.5 DV: 0.5
 04:54 SP: 4.0 PV: 4.6 DV: 0.6
 03:54 SP: 4.0 PV: 4.5 DV: 0.5
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.5 DV: 0.5
 00:54 SP: 4.0 PV: 4.4 DV: 0.4

2025-07-21

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.6 DV: 0.6

2025-07-24

23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.6 DV: 0.6
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.0 DV: 0.0
 19:54 SP: 4.0 PV: 4.2 DV: 0.2
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 3.8 DV: -0.2
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.4 DV: 0.4

5

11:54 SP: 4.0 PV: 4.5 DV: 0.
 10:54 SP: 4.0 PV: 4.5 DV: 0.5
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.2 DV: 0.2
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.5 DV: 0.5
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.5 DV: 0.5

2025-07-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 4.5 DV: 0.5
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.1 DV: 0.1
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 3
 15:54 SP: 4.0 PV: 4.3 DV: 0.
 14:54 SP: 4.0 PV: 4.4 DV: 0.4
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.4 DV: 0.4
 08:54 SP: 4.0 PV: 4.5 DV: 0.5
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.4 DV: 0.4
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.4 DV: 0.4
 2025-07-26
 23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.4 DV: 0.4
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.1 DV: 0.1
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.5 DV: 0.5
 05:54 SP: 4.0 PV: 4.4 DV: 0.4
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-07-25

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.1 DV: 0.1
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-07-28
 23:54 SP: 4.0 PV: 4.4 DV: 0.4
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.3 DV: 0.3
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.0 DV: 0.0
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.5 DV: 0.5
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.4 DV: 0.4
 2025-07-27

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.7

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.3 DV: 0.3
 20:54 SP: 4.0 PV: 4.2 DV: 0.2
 19:54 SP: 4.0 PV: 4.2 DV: 0.2
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 4.3 DV: 0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 4.4 DV: 0.4
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.4 DV: 0.4
 10:54 SP: 4.0 PV: 4.5 DV: 0.5
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-07-29

SP Setting Point
 PV Present Value
 DV Difference Value

8 of 8

23:54 SP: 4.0 PV: 4.1 DV: 0.1
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.0 DV: 0.0
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.1 DV: 0.1
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 3.6 DV: -0.4
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.3 DV: 0.3
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 4.3 DV: 0.3
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.4 DV: 0.4
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.0 DV: 0.0

2025-07-31

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 4.3 DV: 0.3

4

21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 4.5 DV: 0.5
 16:54 SP: 4.0 PV: 4.4 DV: 0.4
 15:54 SP: 4.0 PV: 4.4 DV: 0.4
 14:54 SP: 4.0 PV: 3.9 DV: -0.1

13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 4.4 DV: 0.4
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 4.4 DV: 0.4
 09:54 SP: 4.0 PV: 4.5 DV: 0.5
 08:54 SP: 4.0 PV: 4.4 DV: 0.4
 07:54 SP: 4.0 PV: 4.5 DV: 0.5
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 4.4 DV: 0.4
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 4.2 DV: 0.2
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-07-30

date of copying

sign