








IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1 9134Room
	Manufacturer	SANYO	Storage Period	2025.7
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.07.01	P:50	17	20	20		
2025.07.02	P:40	17	20	18		
2025.07.03	P:30	17	20	18		
2025.07.04	P:30	17	20	19		
2025.07.05						
2025.07.06						
2025.07.07	P:30	17	20	19		
2025.07.08	P:30	17	20	20		
2025.07.09	P:30	17	20	18		
2025.07.10	P:20	17	20	20		
2025.07.11	P:30	17	20	19		
2025.07.12						
2025.07.13						
2025.07.14	P:30	17	20	18		
2025.07.15	P:30	17	20	20		
2025.07.16	P:30	17	20	18		
2025.07.17	P:30	17	20	18		
2025.07.18	P:30	17	20	18		
2025.07.19						
2025.07.20						
2025.07.21	P:30	17	20	20		
2025.07.22	P:30	17	20	20		
2025.07.23	P:30	17	20	19		
2025.07.24	P:40	17	20	19		
2025.07.25	P:30	17	20	20		
2025.07.26						
2025.07.27						
2025.07.28	P:30	17	20	17		
2025.07.29	P:30	17	20	18		
2025.07.30	P:30	17	20	19		
2025.07.31	P:20	17	20	18		

date of copying

sign