

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고 9134Room
	Manufacturer	Giant Star	Storage Period	2025.10
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	833780		
	Date of Calibration	2025.2.25	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-7	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.10.01	P:30	2	5	5	[Signature]	
2025.10.02	P:30	2	5	5		
2025.10.03						개천절
2025.10.04						
2025.10.05						
2025.10.06						추석
2025.10.07						추석연휴
2025.10.08						대체공휴일
2025.10.09						한글날
2025.10.10	P:45	2	5	4	[Signature]	
2025.10.11						
2025.10.12						
2025.10.13	P:35	3	5	5	[Signature]	
2025.10.14	P:30	3	5	4	[Signature]	
2025.10.15	P:30	3	5	3	[Signature]	
2025.10.16	P:30	3	5	3	[Signature]	
2025.10.17	P:30	3	5	3	[Signature]	
2025.10.18						
2025.10.19						
2025.10.20	P:30	2	6	3	[Signature]	
2025.10.21	P:30	2	5	3	[Signature]	
2025.10.22	P:30	2	5	3	[Signature]	
2025.10.23	P:00	2	5	5	[Signature]	
2025.10.24	P:10	2	5	4	[Signature]	
2025.10.25						
2025.10.26						
2025.10.27	10:30	2	5	3	[Signature]	냉장고 교체
2025.10.28						
2025.10.29						
2025.10.30						
2025.10.31						

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 3.9 DV: -0.1
 22:24 SP: 4.0 PV: 3.9 DV: -0.1
 21:24 SP: 4.0 PV: 3.9 DV: -0.1
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.9 DV: -0.1
 18:24 SP: 4.0 PV: 3.7 DV: -0.3
 17:24 SP: 4.0 PV: 3.7 DV: -0.3
 16:24 SP: 4.0 PV: 4.2 DV: 0.2
 15:24 SP: 4.0 PV: 3.8 DV: -0.2
 14:24 SP: 4.0 PV: 4.0 DV: 0.0
 13:24 SP: 4.0 PV: 3.8 DV: -0.2
 12:24 SP: 4.0 PV: 4.1 DV: 0.1
 11:24 SP: 4.0 PV: 3.6 DV: -0.4
 10:24 SP: 4.0 PV: 4.2 DV: 0.2
 09:24 SP: 4.0 PV: 3.8 DV: -0.2
 08:24 SP: 4.0 PV: 3.7 DV: -0.3
 07:24 SP: 4.0 PV: 4.2 DV: 0.2
 06:24 SP: 4.0 PV: 3.8 DV: -0.2
 05:24 SP: 4.0 PV: 3.7 DV: -0.3
 04:24 SP: 4.0 PV: 4.1 DV: 0.1
 03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.9 DV: -0.1
 01:24 SP: 4.0 PV: 3.9 DV: -0.1
 00:24 SP: 4.0 PV: 4.0 DV: 0.0

2025-10-02

23:24 SP: 4.0 PV: 3.8 DV: -0.2
 22:24 SP: 4.0 PV: 3.9 DV: -0.1
 21:24 SP: 4.0 PV: 3.8 DV: -0.2
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.7 DV: -0.3
 18:24 SP: 4.0 PV: 4.3 DV: 0.3
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 3.7 DV: -0.3
 15:24 SP: 4.0 PV: 4.2 DV: 0.2
 14:24 SP: 4.0 PV: 3.9 DV: -0.1
 13:24 SP: 4.0 PV: 3.7 DV: -0.3
 12:24 SP: 4.0 PV: 4.1 DV: 0.1
 11:24 SP: 4.0 PV: 3.8 DV: -0.2
 10:24 SP: 4.0 PV: 4.2 DV: 0.2
 09:24 SP: 4.0 PV: 3.9 DV: -0.1
 08:24 SP: 4.0 PV: 4.1 DV: 0.1
 07:24 SP: 4.0 PV: 3.9 DV: -0.1
 06:24 SP: 4.0 PV: 4.0 DV: 0.0
 05:24 SP: 4.0 PV: 3.9 DV: -0.1
 04:24 SP: 4.0 PV: 4.1 DV: 0.1
 03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.6 DV: -0.4
 01:24 SP: 4.0 PV: 4.2 DV: 0.2
 00:24 SP: 4.0 PV: 3.9 DV: -0.1

2025-10-01

23:24 SP: 4.0 PV: 3.7 DV: -0.3
 22:24 SP: 4.0 PV: 3.6 DV: -0.4
 21:24 SP: 4.0 PV: 3.6 DV: -0.4
 20:24 SP: 4.0 PV: 3.4 DV: -0.6
 19:24 SP: 4.0 PV: 4.2 DV: 0.2
 18:24 SP: 4.0 PV: 3.8 DV: -0.2
 17:24 SP: 4.0 PV: 4.0 DV: 0.0
 16:24 SP: 4.0 PV: 3.8 DV: -0.1
 15:24 SP: 4.0 PV: 4.1 DV: 0.1
 14:24 SP: 4.0 PV: 3.8 DV: -0.2
 13:24 SP: 4.0 PV: 4.0 DV: 0.0
 12:24 SP: 4.0 PV: 3.9 DV: -0.1
 11:24 SP: 4.0 PV: 3.7 DV: -0.3
 10:24 SP: 4.0 PV: 4.4 DV: 0.4
 09:24 SP: 4.0 PV: 3.8 DV: -0.2
 08:24 SP: 4.0 PV: 3.7 DV: -0.3
 07:24 SP: 4.0 PV: 4.1 DV: 0.1
 06:24 SP: 4.0 PV: 3.9 DV: -0.1
 05:24 SP: 4.0 PV: 3.9 DV: -0.1
 04:24 SP: 4.0 PV: 3.8 DV: -0.2
 03:24 SP: 4.0 PV: 3.6 DV: -0.4
 02:24 SP: 4.0 PV: 4.3 DV: 0.3
 01:24 SP: 4.0 PV: 3.9 DV: -0.1
 00:24 SP: 4.0 PV: 3.8 DV: -0.2

2025-10-04

23:24 SP: 4.0 PV: 4.0 DV: 0.0
 22:24 SP: 4.0 PV: 3.9 DV: -0.1
 21:24 SP: 4.0 PV: 3.9 DV: -0.1
 20:24 SP: 4.0 PV: 3.8 DV: -0.2
 19:24 SP: 4.0 PV: 3.7 DV: -0.3
 18:24 SP: 4.0 PV: 4.3 DV: 0.3
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 3.7 DV: -0.3
 15:24 SP: 4.0 PV: 4.0 DV: 0.0
 14:24 SP: 4.0 PV: 3.8 DV: -0.2
 13:24 SP: 4.0 PV: 4.1 DV: 0.1
 12:24 SP: 4.0 PV: 3.7 DV: -0.3
 11:24 SP: 4.0 PV: 4.2 DV: 0.2
 10:24 SP: 4.0 PV: 3.9 DV: -0.1
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.8 DV: -0.2
 07:24 SP: 4.0 PV: 3.6 DV: -0.4
 06:24 SP: 4.0 PV: 4.2 DV: 0.2
 05:24 SP: 4.0 PV: 3.9 DV: -0.1
 04:24 SP: 4.0 PV: 3.7 DV: -0.3
 03:24 SP: 4.0 PV: 3.7 DV: -0.3
 02:24 SP: 4.0 PV: 4.2 DV: 0.2
 01:24 SP: 4.0 PV: 3.8 DV: -0.2
 00:24 SP: 4.0 PV: 4.1 DV: 0.1

2025-10-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 3.6 DV: -0.4
 22:24 SP: 4.0 PV: 4.2 DV: 0.2
 21:24 SP: 4.0 PV: 3.9 DV: -0.1
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.9 DV: -0.1
 18:24 SP: 4.0 PV: 3.8 DV: -0.2
 17:24 SP: 4.0 PV: 3.8 DV: -0.2
 16:24 SP: 4.0 PV: 3.7 DV: -0.3
 15:24 SP: 4.0 PV: 3.6 DV: -0.4
 14:24 SP: 4.0 PV: 4.2 DV: 0.2
 13:24 SP: 4.0 PV: 3.8 DV: -0.2
 12:24 SP: 4.0 PV: 3.9 DV: -0.1
 11:24 SP: 4.0 PV: 3.9 DV: -0.1
 10:24 SP: 4.0 PV: 3.9 DV: -0.1
 09:24 SP: 4.0 PV: 3.9 DV: -0.1
 08:24 SP: 4.0 PV: 3.9 DV: -0.1
 07:24 SP: 4.0 PV: 3.9 DV: -0.1
 06:24 SP: 4.0 PV: 3.7 DV: -0.3
 05:24 SP: 4.0 PV: 4.0 DV: 0.0
 04:24 SP: 4.0 PV: 3.8 DV: -0.2
 03:24 SP: 4.0 PV: 4.1 DV: 0.1
 02:24 SP: 4.0 PV: 3.9 DV: -0.1
 01:24 SP: 4.0 PV: 4.0 DV: 0.0
 00:24 SP: 4.0 PV: 3.8 DV: -0.2
 2025-10-06
 23:24 SP: 4.0 PV: 3.9 DV: -0.1
 22:24 SP: 4.0 PV: 3.8 DV: -0.2
 21:24 SP: 4.0 PV: 4.0 DV: 0.0
 20:24 SP: 4.0 PV: 3.8 DV: -0.2
 19:24 SP: 4.0 PV: 4.0 DV: 0.0
 18:24 SP: 4.0 PV: 3.9 DV: -0.1
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 3.9 DV: -0.1
 15:24 SP: 4.0 PV: 4.0 DV: 0.0
 14:24 SP: 4.0 PV: 3.8 DV: -0.2
 13:24 SP: 4.0 PV: 4.1 DV: 0.1
 12:24 SP: 4.0 PV: 3.9 DV: -0.1
 11:24 SP: 4.0 PV: 3.9 DV: -0.1
 10:24 SP: 4.0 PV: 3.9 DV: -0.1
 09:24 SP: 4.0 PV: 4.0 DV: -0.1
 08:24 SP: 4.0 PV: 4.0 DV: 0.0
 07:24 SP: 4.0 PV: 3.7 DV: -0.3
 06:24 SP: 4.0 PV: 3.6 DV: -0.4
 05:24 SP: 4.0 PV: 4.2 DV: 0.2
 04:24 SP: 4.0 PV: 3.9 DV: -0.1
 03:24 SP: 4.0 PV: 4.0 DV: 0.0
 02:24 SP: 4.0 PV: 3.7 DV: -0.3
 01:24 SP: 4.0 PV: 4.3 DV: 0.3
 00:24 SP: 4.0 PV: 3.9 DV: -0.1
 2025-10-05

23:24 SP: 4.0 PV: 3.9 DV: -0.1
 22:24 SP: 4.0 PV: 3.8 DV: -0.2
 21:24 SP: 4.0 PV: 4.2 DV: 0.2
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 4.1 DV: 0.1
 18:24 SP: 4.0 PV: 3.8 DV: -0.2
 17:24 SP: 4.0 PV: 3.8 DV: -0.2
 16:24 SP: 4.0 PV: 4.1 DV: 0.1
 15:24 SP: 4.0 PV: 3.9 DV: -0.1
 14:24 SP: 4.0 PV: 3.8 DV: -0.2
 13:24 SP: 4.0 PV: 4.1 DV: 0.1
 12:24 SP: 4.0 PV: 3.9 DV: -0.1
 11:24 SP: 4.0 PV: 3.8 DV: -0.2
 10:24 SP: 4.0 PV: 4.0 DV: 0.0
 09:24 SP: 4.0 PV: 3.9 DV: -0.1
 08:24 SP: 4.0 PV: 3.9 DV: -0.1
 07:24 SP: 4.0 PV: 3.8 DV: -0.2
 06:24 SP: 4.0 PV: 3.8 DV: -0.2
 05:24 SP: 4.0 PV: 4.4 DV: 0.4
 04:24 SP: 4.0 PV: 3.9 DV: -0.1
 03:24 SP: 4.0 PV: 3.7 DV: -0.3
 02:24 SP: 4.0 PV: 4.0 DV: 0.0
 01:24 SP: 4.0 PV: 3.8 DV: -0.2
 00:24 SP: 4.0 PV: 3.6 DV: -0.4
 2025-10-06
 23:24 SP: 4.0 PV: 4.2 DV: 0.2
 22:24 SP: 4.0 PV: 3.9 DV: -0.1
 21:24 SP: 4.0 PV: 4.1 DV: 0.1
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.8 DV: -0.2
 18:24 SP: 4.0 PV: 3.9 DV: -0.1
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 3.7 DV: -0.3
 15:24 SP: 4.0 PV: 3.6 DV: -0.4
 14:24 SP: 4.0 PV: 4.2 DV: 0.2
 13:24 SP: 4.0 PV: 3.9 DV: -0.1
 0
 12:24 SP: 4.0 PV: 4.0 DV: 0.0
 11:24 SP: 4.0 PV: 3.9 DV: -0.1
 10:24 SP: 4.0 PV: 3.8 DV: -0.2
 09:24 SP: 4.0 PV: 3.8 DV: -0.2
 08:24 SP: 4.0 PV: 3.7 DV: -0.3
 07:24 SP: 4.0 PV: 4.0 DV: 0.0
 06:24 SP: 4.0 PV: 3.9 DV: -0.1
 05:24 SP: 4.0 PV: 3.8 DV: -0.2
 04:24 SP: 4.0 PV: 4.0 DV: 0.0
 03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.7 DV: -0.3
 01:24 SP: 4.0 PV: 3.7 DV: -0.3
 00:24 SP: 4.0 PV: 3.6 DV: -0.4
 2025-10-07

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 3.8 DV: -0.2
 22:24 SP: 4.0 PV: 3.8 DV: -0.2
 21:24 SP: 4.0 PV: 4.1 DV: 0.1
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 4.0 DV: 0.0
 18:24 SP: 4.0 PV: 3.9 DV: -0.1
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 4.1 DV: 0.1
 15:24 SP: 4.0 PV: 3.8 DV: -0.2
 14:24 SP: 4.0 PV: 4.1 DV: 0.1
 13:24 SP: 4.0 PV: 3.9 DV: -0.1
 12:24 SP: 4.0 PV: 4.0 DV: 0.0
 11:24 SP: 4.0 PV: 3.7 DV: -0.3
 10:24 SP: 4.0 PV: 4.1 DV: 0.1
 09:24 SP: 4.0 PV: 3.9 DV: -0.1
 08:24 SP: 4.0 PV: 3.7 DV: -0.3
 07:24 SP: 4.0 PV: 4.1 DV: 0.1
 06:24 SP: 4.0 PV: 3.9 DV: -0.1
 05:24 SP: 4.0 PV: 3.8 DV: -0.2
 04:24 SP: 4.0 PV: 4.0 DV: 0.0
 03:24 SP: 4.0 PV: 3.7 DV: -0.3
 02:24 SP: 4.0 PV: 4.1 DV: 0.1
 01:24 SP: 4.0 PV: 3.9 DV: -0.1
 00:24 SP: 4.0 PV: 3.9 DV: -0.1

2025-10-10

23:24 SP: 4.0 PV: 3.8 DV: -0.2
 22:24 SP: 4.0 PV: 3.7 DV: -0.3
 21:24 SP: 4.0 PV: 4.3 DV: 0.3
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.7 DV: -0.3
 18:24 SP: 4.0 PV: 4.0 DV: 0.0
 17:24 SP: 4.0 PV: 3.9 DV: -0.1
 16:24 SP: 4.0 PV: 3.9 DV: -0.1
 15:24 SP: 4.0 PV: 3.9 DV: -0.1
 14:24 SP: 4.0 PV: 3.7 DV: -0.3
 13:24 SP: 4.0 PV: 4.1 DV: 0.1
 12:24 SP: 4.0 PV: 3.9 DV: -0.1
 11:24 SP: 4.0 PV: 3.7 DV: -0.3
 10:24 SP: 4.0 PV: 3.7 DV: -0.3
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.8 DV: -0.2
 07:24 SP: 4.0 PV: 4.1 DV: 0.1
 06:24 SP: 4.0 PV: 3.8 DV: -0.2
 05:24 SP: 4.0 PV: 4.0 DV: 0.0
 04:24 SP: 4.0 PV: 3.9 DV: -0.1
 03:24 SP: 4.0 PV: 3.6 DV: -0.4
 02:24 SP: 4.0 PV: 4.2 DV: 0.2
 01:24 SP: 4.0 PV: 3.9 DV: -0.1
 00:24 SP: 4.0 PV: 4.0 DV: 0.0

2025-10-09

23:24 SP: 4.0 PV: 3.7 DV: -0.3
 22:24 SP: 4.0 PV: 4.2 DV: 0.2
 21:24 SP: 4.0 PV: 3.8 DV: -0.2
 20:24 SP: 4.0 PV: 3.8 DV: -0.2
 19:24 SP: 4.0 PV: 3.9 DV: -0.1
 18:24 SP: 4.0 PV: 3.9 DV: -0.1
 17:24 SP: 4.0 PV: 3.8 DV: -0.2
 16:24 SP: 4.0 PV: 3.7 DV: -0.3
 15:24 SP: 4.0 PV: 4.1 DV: 0.1
 14:24 SP: 4.0 PV: 3.9 DV: -0.1
 13:24 SP: 4.0 PV: 4.0 DV: 0.0
 12:24 SP: 4.0 PV: 3.8 DV: -0.2
 11:24 SP: 4.0 PV: 4.0 DV: 0.0
 10:24 SP: 4.0 PV: 3.9 DV: -0.1
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.9 DV: -0.1
 07:24 SP: 4.0 PV: 3.8 DV: -0.2
 06:24 SP: 4.0 PV: 4.0 DV: 0.0
 05:24 SP: 4.0 PV: 3.8 DV: -0.2
 04:24 SP: 4.0 PV: 4.1 DV: 0.1
 03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.7 DV: -0.3
 01:24 SP: 4.0 PV: 3.6 DV: -0.4
 00:24 SP: 4.0 PV: 4.2 DV: 0.2

2025-10-12

23:24 SP: 4.0 PV: 3.9 DV: -0.1
 22:24 SP: 4.0 PV: 3.7 DV: -0.3
 21:24 SP: 4.0 PV: 3.7 DV: -0.3
 20:24 SP: 4.0 PV: 4.0 DV: 0.0
 19:24 SP: 4.0 PV: 3.8 DV: -0.2
 18:24 SP: 4.0 PV: 4.0 DV: 0.0
 17:24 SP: 4.0 PV: 3.8 DV: -0.2
 16:24 SP: 4.0 PV: 3.6 DV: -0.4
 15:24 SP: 4.0 PV: 4.2 DV: 0.2
 14:24 SP: 4.0 PV: 3.9 DV: -0.1
 13:24 SP: 4.0 PV: 4.1 DV: 0.1
 12:24 SP: 4.0 PV: 3.8 DV: -0.2
 11:24 SP: 4.0 PV: 4.2 DV: 0.2
 10:24 SP: 4.0 PV: 3.9 DV: -0.1
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.9 DV: -0.1
 07:24 SP: 4.0 PV: 3.9 DV: -0.1
 06:24 SP: 4.0 PV: 4.0 DV: 0.0
 05:24 SP: 4.0 PV: 3.9 DV: -0.1

2

04:24 SP: 4.0 PV: 3.8 DV: -0.1

1

03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.9 DV: -0.1
 01:24 SP: 4.0 PV: 3.9 DV: -0.1
 00:24 SP: 4.0 PV: 4.0 DV: 0.0

2025-10-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 4.2 DV: 0.2
 22:24 SP: 4.0 PV: 3.8 DV: -0.2
 21:24 SP: 4.0 PV: 4.2 DV: 0.2
 20:24 SP: 4.0 PV: 3.9 DV: -0.1
 19:24 SP: 4.0 PV: 3.6 DV: -0.4
 18:24 SP: 4.0 PV: 4.2 DV: 0.2
 17:24 SP: 4.0 PV: 3.7 DV: -0.3
 16:24 SP: 4.0 PV: 3.6 DV: -0.4
 15:24 SP: 4.0 PV: 3.7 DV: -0.3
 14:24 SP: 4.0 PV: 3.7 DV: -0.3
 13:24 SP: 4.0 PV: 3.8 DV: -0.2
 12:24 SP: 4.0 PV: 4.1 DV: 0.1
 11:24 SP: 4.0 PV: 3.9 DV: -0.1
 10:24 SP: 4.0 PV: 3.8 DV: -0.2
 09:24 SP: 4.0 PV: 3.8 DV: -0.2
 08:24 SP: 4.0 PV: 3.6 DV: -0.4

2

07:24 SP: 4.0 PV: 4.2 DV: 0.
 06:24 SP: 4.0 PV: 3.9 DV: -0.1
 05:24 SP: 4.0 PV: 3.7 DV: -0.3
 04:24 SP: 4.0 PV: 4.2 DV: 0.2
 03:24 SP: 4.0 PV: 3.9 DV: -0.1
 02:24 SP: 4.0 PV: 3.8 DV: -0.2
 01:24 SP: 4.0 PV: 3.8 DV: -0.2
 00:24 SP: 4.0 PV: 3.7 DV: -0.3

2025-10-14

23:24 SP: 4.0 PV: 4.1 DV: 0.1
 22:24 SP: 4.0 PV: 4.0 DV: 0.0

2

21:24 SP: 4.0 PV: 3.8 DV: -0.
 20:24 SP: 4.0 PV: 4.1 DV: 0.1
 19:24 SP: 4.0 PV: 3.8 DV: -0.2
 18:24 SP: 4.0 PV: 3.7 DV: -0.3
 17:24 SP: 4.0 PV: 3.7 DV: -0.3
 16:24 SP: 4.0 PV: 4.1 DV: 0.1
 15:24 SP: 4.0 PV: 3.9 DV: -0.1
 14:24 SP: 4.0 PV: 3.9 DV: -0.1
 13:24 SP: 4.0 PV: 3.8 DV: -0.2
 12:24 SP: 4.0 PV: 3.8 DV: -0.2
 11:24 SP: 4.0 PV: 3.7 DV: -0.3
 10:24 SP: 4.0 PV: 4.0 DV: 0.0
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.8 DV: -0.2
 07:24 SP: 4.0 PV: 3.6 DV: -0.4
 06:24 SP: 4.0 PV: 4.1 DV: 0.1
 05:24 SP: 4.0 PV: 3.9 DV: -0.1
 04:24 SP: 4.0 PV: 4.0 DV: 0.0
 03:24 SP: 4.0 PV: 3.8 DV: -0.2
 02:24 SP: 4.0 PV: 3.8 DV: -0.2
 01:24 SP: 4.0 PV: 4.0 DV: 0.0
 00:24 SP: 4.0 PV: 3.8 DV: -0.2

2025-10-13

23:24 SP: 4.0 PV: 3.3 DV: -0.7
 22:24 SP: 4.0 PV: 4.9 DV: 0.9
 21:24 SP: 4.0 PV: 4.6 DV: 0.6
 20:24 SP: 4.0 PV: 3.3 DV: -0.7
 19:24 SP: 4.0 PV: 4.9 DV: 0.9
 18:24 SP: 4.0 PV: 4.7 DV: 0.7
 17:24 SP: 4.0 PV: 3.3 DV: -0.7
 16:24 SP: 4.0 PV: 4.8 DV: 0.8
 15:24 SP: 4.0 PV: 4.6 DV: 0.6
 14:24 SP: 4.0 PV: 4.2 DV: 0.2
 13:24 SP: 4.0 PV: 3.3 DV: -0.7
 12:24 SP: 4.0 PV: 5.0 DV: 1.0
 11:24 SP: 4.0 PV: 4.4 DV: 0.4
 10:24 SP: 4.0 PV: 3.8 DV: -0.2
 09:24 SP: 4.0 PV: 4.0 DV: 0.0
 08:24 SP: 4.0 PV: 3.8 DV: -0.2
 07:24 SP: 4.0 PV: 3.6 DV: -0.4
 06:24 SP: 4.0 PV: 3.6 DV: -0.4
 05:24 SP: 4.0 PV: 4.3 DV: 0.3
 04:24 SP: 4.0 PV: 3.7 DV: -0.3
 03:24 SP: 4.0 PV: 4.2 DV: 0.2
 02:24 SP: 4.0 PV: 3.8 DV: -0.2
 01:24 SP: 4.0 PV: 4.0 DV: 0.0
 00:24 SP: 4.0 PV: 4.0 DV: 0.0

2025-10-16

23:24 SP: 4.0 PV: 3.7 DV: -0.3
 22:24 SP: 4.0 PV: 4.2 DV: 0.2
 21:24 SP: 4.0 PV: 3.8 DV: -0.2
 20:24 SP: 4.0 PV: 3.7 DV: -0.3
 19:24 SP: 4.0 PV: 4.2 DV: 0.2
 18:24 SP: 4.0 PV: 3.7 DV: -0.3
 17:24 SP: 4.0 PV: 3.6 DV: -0.4
 16:24 SP: 4.0 PV: 4.2 DV: 0.2
 15:24 SP: 4.0 PV: 3.8 DV: -0.2
 14:24 SP: 4.0 PV: 4.1 DV: 0.1
 13:24 SP: 4.0 PV: 3.7 DV: -0.3
 12:24 SP: 4.0 PV: 4.2 DV: 0.2
 11:24 SP: 4.0 PV: 3.7 DV: -0.3
 10:24 SP: 4.0 PV: 4.2 DV: 0.2
 09:24 SP: 4.0 PV: 3.7 DV: -0.3
 08:24 SP: 4.0 PV: 4.2 DV: 0.2
 07:24 SP: 4.0 PV: 3.8 DV: -0.2
 06:24 SP: 4.0 PV: 4.2 DV: 0.2
 05:24 SP: 4.0 PV: 3.9 DV: -0.1
 04:24 SP: 4.0 PV: 4.1 DV: 0.1
 03:24 SP: 4.0 PV: 3.8 DV: -0.2
 02:24 SP: 4.0 PV: 4.1 DV: 0.1
 01:24 SP: 4.0 PV: 3.7 DV: -0.3
 00:24 SP: 4.0 PV: 3.6 DV: -0.4

2025-10-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 3.8 DV: -0.2
 22:24 SP: 4.0 PV: 4.3 DV: 0.3
 21:24 SP: 4.0 PV: 4.9 DV: 0.9
 20:24 SP: 4.0 PV: 3.3 DV: -0.7
 19:24 SP: 4.0 PV: 4.8 DV: 0.8
 18:24 SP: 4.0 PV: 4.8 DV: 0.8
 17:24 SP: 4.0 PV: 3.6 DV: -0.4
 16:24 SP: 4.0 PV: 4.4 DV: 0.4
 15:24 SP: 4.0 PV: 4.8 DV: 0.8
 14:24 SP: 4.0 PV: 3.6 DV: -0.4
 13:24 SP: 4.0 PV: 4.0 DV: 0.0
 12:24 SP: 4.0 PV: 5.1 DV: 1.1
 11:24 SP: 4.0 PV: 3.6 DV: -0.4
 10:24 SP: 4.0 PV: 4.0 DV: 0.0
 09:24 SP: 4.0 PV: 5.0 DV: 1.0
 08:24 SP: 4.0 PV: 4.0 DV: 0.0
 07:24 SP: 4.0 PV: 4.2 DV: 0.2
 06:24 SP: 4.0 PV: 4.9 DV: 0.9
 05:24 SP: 4.0 PV: 3.6 DV: -0.4
 04:24 SP: 4.0 PV: 3.8 DV: -0.2
 03:24 SP: 4.0 PV: 5.1 DV: 1.1
 02:24 SP: 4.0 PV: 3.7 DV: -0.3
 01:24 SP: 4.0 PV: 4.2 DV: 0.2
 00:24 SP: 4.0 PV: 4.9 DV: 0.9

2025-10-18

23:24 SP: 4.0 PV: 3.7 DV: -0.3
 22:24 SP: 4.0 PV: 4.3 DV: 0.3
 21:24 SP: 4.0 PV: 4.9 DV: 0.9
 20:24 SP: 4.0 PV: 3.8 DV: -0.2
 19:24 SP: 4.0 PV: 3.4 DV: -0.6
 18:24 SP: 4.0 PV: 4.9 DV: 0.9
 17:24 SP: 4.0 PV: 4.6 DV: 0.6
 16:24 SP: 4.0 PV: 3.3 DV: -0.7
 15:24 SP: 4.0 PV: 4.9 DV: 0.9
 14:24 SP: 4.0 PV: 4.6 DV: 0.6
 13:24 SP: 4.0 PV: 3.4 DV: -0.6
 12:24 SP: 4.0 PV: 3.7 DV: -0.3
 11:24 SP: 4.0 PV: 4.9 DV: 0.9
 10:24 SP: 4.0 PV: 4.3 DV: 0.3
 09:24 SP: 4.0 PV: 3.2 DV: -0.8
 08:24 SP: 4.0 PV: 4.9 DV: 0.9
 07:24 SP: 4.0 PV: 4.7 DV: 0.7
 06:24 SP: 4.0 PV: 3.4 DV: -0.6
 05:24 SP: 4.0 PV: 4.6 DV: 0.6
 04:24 SP: 4.0 PV: 4.8 DV: 0.8
 03:24 SP: 4.0 PV: 3.8 DV: -0.2
 02:24 SP: 4.0 PV: 3.4 DV: -0.6
 01:24 SP: 4.0 PV: 5.0 DV: 1.0
 00:24 SP: 4.0 PV: 4.6 DV: 0.6

2025-10-17

23:24 SP: 4.0 PV: 3.3 DV: -0.7

9

22:24 SP: 4.0 PV: 4.9 DV: 0.
 21:24 SP: 4.0 PV: 4.6 DV: 0.6
 20:24 SP: 4.0 PV: 3.3 DV: -0.7
 19:24 SP: 4.0 PV: 4.9 DV: 0.9
 18:24 SP: 4.0 PV: 4.6 DV: 0.6
 17:24 SP: 4.0 PV: 4.2 DV: 0.2
 16:24 SP: 4.0 PV: 4.3 DV: 0.3
 15:24 SP: 4.0 PV: 4.3 DV: 0.3
 14:24 SP: 4.0 PV: 3.3 DV: -0.7
 13:24 SP: 4.0 PV: 4.9 DV: 0.9
 12:24 SP: 4.0 PV: 4.4 DV: 0.4
 11:24 SP: 4.0 PV: 3.3 DV: -0.7
 10:24 SP: 4.0 PV: 4.8 DV: 0.8

8

09:24 SP: 4.0 PV: 4.8 DV: 0.
 08:24 SP: 4.0 PV: 4.1 DV: 0.1
 07:24 SP: 4.0 PV: 3.3 DV: -0.7
 06:24 SP: 4.0 PV: 4.9 DV: 0.9
 05:24 SP: 4.0 PV: 4.5 DV: 0.5
 04:24 SP: 4.0 PV: 3.3 DV: -0.7
 03:24 SP: 4.0 PV: 4.8 DV: 0.8
 02:24 SP: 4.0 PV: 4.7 DV: 0.7
 01:24 SP: 4.0 PV: 3.4 DV: -0.6
 00:24 SP: 4.0 PV: 3.8 DV: -0.2

2025-10-20

23:24 SP: 4.0 PV: 5.0 DV: 1.0
 22:24 SP: 4.0 PV: 3.8 DV: -0.2
 21:24 SP: 4.0 PV: 3.6 DV: -0.4
 20:24 SP: 4.0 PV: 4.9 DV: 0.9
 19:24 SP: 4.0 PV: 4.4 DV: 0.4
 18:24 SP: 4.0 PV: 3.3 DV: -0.7
 17:24 SP: 4.0 PV: 4.9 DV: 0.9
 16:24 SP: 4.0 PV: 4.6 DV: 0.6
 15:24 SP: 4.0 PV: 3.3 DV: -0.7
 14:24 SP: 4.0 PV: 4.9 DV: 0.9
 13:24 SP: 4.0 PV: 4.6 DV: 0.6
 12:24 SP: 4.0 PV: 4.2 DV: 0.2
 11:24 SP: 4.0 PV: 3.3 DV: -0.7
 10:24 SP: 4.0 PV: 4.9 DV: 0.9
 09:24 SP: 4.0 PV: 4.7 DV: 0.7
 08:24 SP: 4.0 PV: 3.3 DV: -0.7
 07:24 SP: 4.0 PV: 4.9 DV: 0.9
 06:24 SP: 4.0 PV: 4.8 DV: 0.8
 05:24 SP: 4.0 PV: 3.4 DV: -0.6
 04:24 SP: 4.0 PV: 4.5 DV: 0.5
 03:24 SP: 4.0 PV: 4.9 DV: 0.9
 02:24 SP: 4.0 PV: 3.4 DV: -0.6
 01:24 SP: 4.0 PV: 3.8 DV: -0.2
 00:24 SP: 4.0 PV: 5.1 DV: 1.1

2025-10-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:24 SP: 4.0 PV: 3.8 DV: -0.2
 22:24 SP: 4.0 PV: 3.4 DV: -0.6
 21:24 SP: 4.0 PV: 4.9 DV: 0.9
 20:24 SP: 4.0 PV: 4.6 DV: 0.6
 19:24 SP: 4.0 PV: 3.4 DV: -0.6
 18:24 SP: 4.0 PV: 3.9 DV: -0.1
 17:24 SP: 4.0 PV: 4.9 DV: 0.9
 16:24 SP: 4.0 PV: 4.3 DV: 0.3
 15:24 SP: 4.0 PV: 4.3 DV: 0.3
 14:24 SP: 4.0 PV: 3.3 DV: -0.7
 13:24 SP: 4.0 PV: 5.0 DV: 1.0
 12:24 SP: 4.0 PV: 4.7 DV: 0.7
 11:24 SP: 4.0 PV: 3.3 DV: -0.7
 10:24 SP: 4.0 PV: 4.8 DV: 0.8
 09:24 SP: 4.0 PV: 4.7 DV: 0.7
 08:24 SP: 4.0 PV: 3.9 DV: -0.1
 07:24 SP: 4.0 PV: 3.4 DV: -0.6
 06:24 SP: 4.0 PV: 4.8 DV: 0.8
 05:24 SP: 4.0 PV: 4.6 DV: 0.6
 04:24 SP: 4.0 PV: 3.4 DV: -0.6
 03:24 SP: 4.0 PV: 4.8 DV: 0.8
 02:24 SP: 4.0 PV: 4.7 DV: 0.7
 01:24 SP: 4.0 PV: 3.4 DV: -0.6
 00:24 SP: 4.0 PV: 4.0 DV: 0.0
 2025-10-22
 23:24 SP: 4.0 PV: 5.0 DV: 1.0
 22:24 SP: 4.0 PV: 3.6 DV: -0.4
 21:24 SP: 4.0 PV: 4.4 DV: 0.4
 20:24 SP: 4.0 PV: 4.8 DV: 0.8
 19:24 SP: 4.0 PV: 3.6 DV: -0.4
 18:24 SP: 4.0 PV: 4.4 DV: 0.4
 17:24 SP: 4.0 PV: 4.9 DV: 0.9
 16:24 SP: 4.0 PV: 4.1 DV: 0.1
 15:24 SP: 4.0 PV: 3.6 DV: -0.4
 14:24 SP: 4.0 PV: 4.9 DV: 0.9
 13:24 SP: 4.0 PV: 4.4 DV: 0.4
 12:24 SP: 4.0 PV: 3.4 DV: -0.6
 11:24 SP: 4.0 PV: 4.8 DV: 0.8
 10:24 SP: 4.0 PV: 4.7 DV: 0.7
 09:24 SP: 4.0 PV: 4.2 DV: 0.2
 08:24 SP: 4.0 PV: 3.3 DV: -0.7
 07:24 SP: 4.0 PV: 5.0 DV: 1.0
 06:24 SP: 4.0 PV: 4.4 DV: 0.4
 05:24 SP: 4.0 PV: 3.3 DV: -0.7
 04:24 SP: 4.0 PV: 4.8 DV: 0.8
 03:24 SP: 4.0 PV: 4.7 DV: 0.7
 02:24 SP: 4.0 PV: 3.3 DV: -0.7
 01:24 SP: 4.0 PV: 4.8 DV: 0.8
 00:24 SP: 4.0 PV: 4.8 DV: 0.8
 2025-10-21

SP Setting Point
 PV Present Value
 DV Difference Value

주소창에 한글로 혈액냉장고
 www.gsmedical.co.kr

CBR-1000SS
 GIANT STAR Ltd.

Condensor1 Temp#1. OK...
 Compressor#1 Output OK...
 System Checking Process...
 11:21:27

2025-10-24

주소창에 한글로 혈액냉장고
 www.gsmedical.co.kr

CBR-1000SS
 GIANT STAR Ltd.

10:24 SP: 4.0 PV: 4.9 DV: 0.9
 09:24 SP: 4.0 PV: 4.5 DV: 0.5
 08:24 SP: 4.0 PV: 3.3 DV: -0.7
 07:24 SP: 4.0 PV: 4.9 DV: 0.9
 06:24 SP: 4.0 PV: 4.6 DV: 0.6
 05:24 SP: 4.0 PV: 4.3 DV: 0.3
 04:24 SP: 4.0 PV: 3.3 DV: -0.7
 03:24 SP: 4.0 PV: 5.1 DV: 1.1
 02:24 SP: 4.0 PV: 4.4 DV: 0.4
 01:24 SP: 4.0 PV: 3.3 DV: -0.7
 00:24 SP: 4.0 PV: 4.9 DV: 0.9

2025-10-24

23:24 SP: 4.0 PV: 4.5 DV: 0.5
 22:24 SP: 4.0 PV: 4.3 DV: 0.3
 21:24 SP: 4.0 PV: 3.3 DV: -0.7
 20:24 SP: 4.0 PV: 4.8 DV: 0.8
 19:24 SP: 4.0 PV: 4.7 DV: 0.7
 18:24 SP: 4.0 PV: 3.4 DV: -0.6
 17:24 SP: 4.0 PV: 4.6 DV: 0.6
 16:24 SP: 4.0 PV: 4.9 DV: 0.9
 15:24 SP: 4.0 PV: 3.7 DV: -0.3
 14:24 SP: 4.0 PV: 4.2 DV: 0.2
 13:24 SP: 4.0 PV: 4.9 DV: 0.9
 12:24 SP: 4.0 PV: 4.4 DV: 0.4
 11:24 SP: 4.0 PV: 3.4 DV: -0.6
 10:24 SP: 4.0 PV: 4.8 DV: 0.8
 09:24 SP: 4.0 PV: 4.8 DV: 0.8
 08:24 SP: 4.0 PV: 3.6 DV: -0.4
 07:24 SP: 4.0 PV: 4.3 DV: 0.3
 06:24 SP: 4.0 PV: 4.9 DV: 0.9
 05:24 SP: 4.0 PV: 3.7 DV: -0.3
 04:24 SP: 4.0 PV: 4.1 DV: 0.1
 03:24 SP: 4.0 PV: 4.9 DV: 0.9
 02:24 SP: 4.0 PV: 3.6 DV: -0.4
 01:24 SP: 4.0 PV: 4.3 DV: 0.3
 00:24 SP: 4.0 PV: 5.0 DV: 1.0

2025-10-23

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.10

23:36 SP: 4.0 PV: 4.2 DV: 0.2
 22:36 SP: 4.0 PV: 4.2 DV: 0.2
 21:36 SP: 4.0 PV: 5.0 DV: 1.0
 20:36 SP: 4.0 PV: 3.9 DV: -0.1
 19:36 SP: 4.0 PV: 4.6 DV: 0.6
 18:36 SP: 4.0 PV: 4.9 DV: 0.9
 17:36 SP: 4.0 PV: 3.6 DV: -0.4
 16:36 SP: 4.0 PV: 5.1 DV: 1.1
 15:36 SP: 4.0 PV: 4.6 DV: 0.6
 14:36 SP: 4.0 PV: 4.3 DV: 0.3
 13:36 SP: 4.0 PV: 4.5 DV: 0.5
 12:36 SP: 4.0 PV: 4.9 DV: 0.9
 11:36 SP: 4.0 PV: 3.9 DV: -0.1
 10:36 SP: 4.0 PV: 3.6 DV: -0.4
 09:36 SP: 4.0 PV: 5.0 DV: 1.0
 08:36 SP: 4.0 PV: 4.4 DV: 0.4
 07:36 SP: 4.0 PV: 3.3 DV: -0.7
 06:36 SP: 4.0 PV: 5.0 DV: 1.0
 05:36 SP: 4.0 PV: 4.6 DV: 0.6
 04:36 SP: 4.0 PV: 4.2 DV: 0.2
 03:36 SP: 4.0 PV: 3.3 DV: -0.7
 02:36 SP: 4.0 PV: 4.9 DV: 0.9
 01:36 SP: 4.0 PV: 4.4 DV: 0.4
 00:36 SP: 4.0 PV: 4.3 DV: 0.3

2025-10-25

23:36 SP: 4.0 PV: 3.3 DV: -0.7
 22:36 SP: 4.0 PV: 4.8 DV: 0.8
 21:36 SP: 4.0 PV: 4.8 DV: 0.8
 20:36 SP: 4.0 PV: 3.6 DV: -0.4
 19:36 SP: 4.0 PV: 3.6 DV: -0.4
 18:36 SP: 4.0 PV: 4.9 DV: 0.9
 17:36 SP: 4.0 PV: 4.5 DV: 0.5
 16:36 SP: 4.0 PV: 4.2 DV: 0.2
 15:36 SP: 4.0 PV: 3.2 DV: -0.8
 14:36 SP: 4.0 PV: 4.9 DV: 0.9
 13:36 SP: 4.0 PV: 4.6 DV: 0.6
 12:36 SP: 4.0 PV: 3.4 DV: -0.6

Condensor1 Temp#1. OK...

Compressor#1 Output OK...

System Checking Process...

11:36:04

2025-10-24

09:36 SP: 4.0 PV: 4.5 DV: 0.5
 08:36 SP: 4.0 PV: 3.4 DV: -0.6
 07:36 SP: 4.0 PV: 3.8 DV: -0.2
 06:36 SP: 4.0 PV: 5.1 DV: 1.1
 05:36 SP: 4.0 PV: 4.0 DV: 0.0
 04:36 SP: 4.0 PV: 4.2 DV: 0.2
 03:36 SP: 4.0 PV: 4.9 DV: 0.9
 02:36 SP: 4.0 PV: 3.7 DV: -0.3
 01:36 SP: 4.0 PV: 4.7 DV: 0.7
 00:36 SP: 4.0 PV: 4.9 DV: 0.9

2025-10-27

23:36 SP: 4.0 PV: 4.2 DV: 0.2
 22:36 SP: 4.0 PV: 4.2 DV: 0.2
 21:36 SP: 4.0 PV: 5.0 DV: 1.0
 20:36 SP: 4.0 PV: 3.7 DV: -0.3
 19:36 SP: 4.0 PV: 4.7 DV: 0.7
 18:36 SP: 4.0 PV: 4.8 DV: 0.8
 17:36 SP: 4.0 PV: 4.2 DV: 0.2
 16:36 SP: 4.0 PV: 4.2 DV: 0.2
 15:36 SP: 4.0 PV: 4.9 DV: 0.9
 14:36 SP: 4.0 PV: 3.9 DV: -0.1
 13:36 SP: 4.0 PV: 4.5 DV: 0.5
 12:36 SP: 4.0 PV: 4.9 DV: 0.9
 11:36 SP: 4.0 PV: 4.2 DV: 0.2
 10:36 SP: 4.0 PV: 4.3 DV: 0.3
 09:36 SP: 4.0 PV: 5.1 DV: 1.1
 08:36 SP: 4.0 PV: 3.8 DV: -0.2
 07:36 SP: 4.0 PV: 4.4 DV: 0.4
 06:36 SP: 4.0 PV: 5.0 DV: 1.0
 05:36 SP: 4.0 PV: 3.8 DV: -0.2
 04:36 SP: 4.0 PV: 4.2 DV: 0.2
 03:36 SP: 4.0 PV: 5.1 DV: 1.1
 02:36 SP: 4.0 PV: 3.9 DV: -0.1
 01:36 SP: 4.0 PV: 4.8 DV: 0.8
 00:36 SP: 4.0 PV: 4.8 DV: 0.8

2025-10-26

SP Setting Point
 PV Present Value
 DV Difference Value

7 of 7

date of copying

sign