

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1 9134Room
	Manufacturer	SANYO	Storage Period	2025.10
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.10.01	P:30	17	20	19	[Signature]	
2025.10.02	P:30	17	20	17		
2025.10.03						개천절
2025.10.04						
2025.10.05						
2025.10.06						추석
2025.10.07						추석연휴
2025.10.08						대체공휴일
2025.10.09						한글날
2025.10.10	P:45	17	20	19	[Signature]	
2025.10.11						
2025.10.12						
2025.10.13	P:35	17	20	18	[Signature]	
2025.10.14	P:30	17	20	17	[Signature]	
2025.10.15	P:30	17	20	20	[Signature]	ATB 2P 한도
2025.10.16	P:30	17	20	19	[Signature]	
2025.10.17	P:30	17	20	18	[Signature]	
2025.10.18						
2025.10.19						
2025.10.20	P:30	17	20	19	[Signature]	
2025.10.21	P:30	17	20	19	[Signature]	
2025.10.22	P:30	17	20	19	[Signature]	
2025.10.23	P:00	17	20	18	[Signature]	
2025.10.24	P:10	17	20	19	[Signature]	DW31621301 2P 한도
2025.10.25						
2025.10.26						
2025.10.27	P:45	17	20	19	[Signature]	
2025.10.28	P:30	17	20	19	[Signature]	
2025.10.29	P:30	17	20	19	[Signature]	
2025.10.30	P:00	17	20	19	[Signature]	
2025.10.31	P:10	17	20	19	[Signature]	

date of copying

sign