

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1 9134Room
	Manufacturer	SANYO	Storage Period	2025.12
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.12.01	8:30	17	20	19	ant	
2025.12.02	8:30	17	20	19	ant	
2025.12.03	8:40	17	20	19	ant	
2025.12.04	8:30	17	20	19	ant	
2025.12.05	8:30	17	20	19	ant	
2025.12.06						
2025.12.07						
2025.12.08	8:30	17	20	20	ant	
2025.12.09	8:30	17	20	19	ant	
2025.12.10	8:30	17	20	19	ant	
2025.12.11	8:30	17	20	20	ant	
2025.12.12	8:30	17	20	19	ant	
2025.12.13						
2025.12.14						
2025.12.15	8:30	19	20	19	ant	
2025.12.16	8:30	17	20	19	ant	
2025.12.17	8:30	17	20	19	ant	
2025.12.18	8:30	17	20	18	ant	
2025.12.19	8:20	17	20	19	ant	
2025.12.20						
2025.12.21						
2025.12.22	8:30	17	20	19	ant	
2025.12.23	8:30	17	20	18	ant	
2025.12.24	8:40	17	20	19	ant	
2025.12.25						성탄절
2025.12.26	8:00	17	20	18	ant	
2025.12.27						
2025.12.28						
2025.12.29	8:50	17	20	19	ant	
2025.12.30	8:30	17	20	19	ant	
2025.12.31	10:00	17	20	18	ant	

date of copying

sign