








IP STORAGE TEMPERATURE LOG

Storage Subject	Description	약품냉장고	Storage Space	냉장고 9134Room
	Manufacturer	Giant Star	Storage Period	2025.8
	Model Name	CBR1000SS	Temperature Range	2~8℃
	Serial Number	GS17090102-10	Setting Point	4.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	833780		
	Date of Calibration	2025.2.25	Alarm Range	<2℃, >8℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-7	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.08.01	P:30	2	5	5		
2025.08.02						
2025.08.03						
2025.08.04	P:40	2	5	4		
2025.08.05	P:30	3	5	4		
2025.08.06	P:30	3	5	4		
2025.08.07	P:30	3	5	5		
2025.08.08	P:30	3	5	5		
2025.08.09						
2025.08.10						
2025.08.11	P:30	2	5	4		
2025.08.12	P:30	3	5	4		
2025.08.13	P:15	3	5	4		
2025.08.14	P:30	3	5	4		
2025.08.15						광복절
2025.08.16						
2025.08.17						
2025.08.18	P:20	3	5	4		
2025.08.19	P:40	3	5	3		
2025.08.20	P:30	2	5	4		
2025.08.21	P:30	3	5	4		
2025.08.22	P:20	3	5	3		
2025.08.23						
2025.08.24						
2025.08.25	P:30	3	5	5		
2025.08.26	P:30	3	5	4		
2025.08.27	P:20	3	5	5		
2025.08.28	P:40	3	5	4		
2025.08.29	P:30	3	5	4		
2025.08.30						
2025.08.31						

date of copying

sign

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 4.0 DV: 0.0
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 3.6 DV: -0.4
 16:54 SP: 4.0 PV: 4.0 DV: 0.2
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.3 DV: 0.3
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.0 DV: 0.0
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 3.6 DV: -0.4
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 4.8 DV: 0.8
 01:54 SP: 4.0 PV: 4.6 DV: 0.6
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-08-02
 23:54 SP: 4.0 PV: 3.8 DV: -0.2
 6
 22:54 SP: 4.0 PV: 3.4 DV: -0.6
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 4.7 DV: 0.7
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.6 DV: -0.4
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 3.6 DV: -0.4
 12:54 SP: 4.0 PV: 3.6 DV: -0.4
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.6 DV: 0.6
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.6 DV: -0.4
 06:54 SP: 4.0 PV: 4.3 DV: 0.3
 05:54 SP: 4.0 PV: 3.8 DV: -0.2
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 3.6 DV: -0.4
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 3.9 DV: -0.1
 2025-08-01

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.1 DV: 0.1
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.2 DV: 0.2
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 3.6 DV: -0.4
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 3.6 DV: -0.4
 14:54 SP: 4.0 PV: 3.6 DV: -0.4
 13:54 SP: 4.0 PV: 4.7 DV: 0.7
 7
 12:54 SP: 4.0 PV: 4.7 DV: 0.7
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 3.7 DV: -0.3
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.1 DV: 0.1
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 3.9 DV: -0.1
 2025-08-04
 23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 4.0 DV: 0.0
 21:54 SP: 4.0 PV: 3.6 DV: -0.4
 20:54 SP: 4.0 PV: 4.2 DV: 0.2
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 3.6 DV: -0.4
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 3.4 DV: -0.6
 14:54 SP: 4.0 PV: 4.6 DV: 0.6
 13:54 SP: 4.0 PV: 4.7 DV: 0.7
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 4.0 DV: 0.0
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.0 DV: 0.0
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 3.9 DV: -0.1
 2025-08-03

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 4.0 DV: 0.0	23:54 SP: 4.0 PV: 3.6 DV: -0.4
22:54 SP: 4.0 PV: 3.9 DV: -0.1	22:54 SP: 4.0 PV: 4.3 DV: 0.3
21:54 SP: 4.0 PV: 4.1 DV: 0.1	21:54 SP: 4.0 PV: 3.9 DV: -0.1
20:54 SP: 4.0 PV: 3.7 DV: -0.3	20:54 SP: 4.0 PV: 3.6 DV: -0.4
19:54 SP: 4.0 PV: 4.1 DV: 0.1	19:54 SP: 4.0 PV: 3.6 DV: -0.4
18:54 SP: 4.0 PV: 4.0 DV: 0.0	18:54 SP: 4.0 PV: 3.6 DV: -0.4
17:54 SP: 4.0 PV: 4.1 DV: 0.1	17:54 SP: 4.0 PV: 4.6 DV: 0.6
16:54 SP: 4.0 PV: 4.0 DV: 0.0	16:54 SP: 4.0 PV: 4.7 DV: 0.7
15:54 SP: 4.0 PV: 3.7 DV: -0.3	15:54 SP: 4.0 PV: 4.3 DV: 0.3
14:54 SP: 4.0 PV: 4.2 DV: 0.2	14:54 SP: 4.0 PV: 3.8 DV: -0.2
13:54 SP: 4.0 PV: 4.0 DV: 0.0	13:54 SP: 4.0 PV: 3.6 DV: -0.4
12:54 SP: 4.0 PV: 3.6 DV: -0.4	12:54 SP: 4.0 PV: 3.6 DV: -0.4
11:54 SP: 4.0 PV: 3.6 DV: -0.4	11:54 SP: 4.0 PV: 4.6 DV: 0.6
10:54 SP: 4.0 PV: 4.2 DV: 0.2	10:54 SP: 4.0 PV: 4.7 DV: 0.7
09:54 SP: 4.0 PV: 3.9 DV: -0.1	3
08:54 SP: 4.0 PV: 3.6 DV: -0.4	09:54 SP: 4.0 PV: 4.3 DV: 0.
07:54 SP: 4.0 PV: 4.2 DV: 0.2	08:54 SP: 4.0 PV: 3.8 DV: -0.2
06:54 SP: 4.0 PV: 3.8 DV: -0.2	07:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 3.6 DV: -0.4	06:54 SP: 4.0 PV: 3.9 DV: -0.1
04:54 SP: 4.0 PV: 4.2 DV: 0.2	05:54 SP: 4.0 PV: 3.6 DV: -0.4
03:54 SP: 4.0 PV: 3.8 DV: -0.2	04:54 SP: 4.0 PV: 4.3 DV: 0.3
02:54 SP: 4.0 PV: 3.6 DV: -0.4	03:54 SP: 4.0 PV: 3.8 DV: -0.2
01:54 SP: 4.0 PV: 4.2 DV: 0.2	02:54 SP: 4.0 PV: 3.6 DV: -0.4
00:54 SP: 4.0 PV: 3.9 DV: -0.1	01:54 SP: 4.0 PV: 3.6 DV: -0.4
2025-08-06	00:54 SP: 4.0 PV: 4.3 DV: 0.3
23:54 SP: 4.0 PV: 3.6 DV: -0.4	2025-08-08
2	23:54 SP: 4.0 PV: 3.3 DV: -0.2
22:54 SP: 4.0 PV: 4.2 DV: 0.	22:54 SP: 4.0 PV: 4.2 DV: 0.2
21:54 SP: 4.0 PV: 3.9 DV: -0.1	21:54 SP: 4.0 PV: 3.9 DV: -0.1
20:54 SP: 4.0 PV: 3.7 DV: -0.3	20:54 SP: 4.0 PV: 3.9 DV: -0.1
19:54 SP: 4.0 PV: 4.1 DV: 0.1	19:54 SP: 4.0 PV: 4.0 DV: 0.0
18:54 SP: 4.0 PV: 3.7 DV: -0.3	18:54 SP: 4.0 PV: 3.9 DV: -0.1
17:54 SP: 4.0 PV: 3.6 DV: -0.4	17:54 SP: 4.0 PV: 3.7 DV: -0.3
16:54 SP: 4.0 PV: 4.3 DV: 0.3	16:54 SP: 4.0 PV: 4.1 DV: 0.1
15:54 SP: 4.0 PV: 3.9 DV: -0.1	15:54 SP: 4.0 PV: 3.8 DV: -0.2
14:54 SP: 4.0 PV: 3.6 DV: -0.4	14:54 SP: 4.0 PV: 3.6 DV: -0.4
13:54 SP: 4.0 PV: 3.4 DV: -0.6	13:54 SP: 4.0 PV: 3.7 DV: -0.3
12:54 SP: 4.0 PV: 4.6 DV: 0.6	12:54 SP: 4.0 PV: 3.6 DV: -0.4
11:54 SP: 4.0 PV: 4.8 DV: 0.8	11:54 SP: 4.0 PV: 4.3 DV: 0.3
10:54 SP: 4.0 PV: 4.4 DV: 0.4	.2
09:54 SP: 4.0 PV: 3.8 DV: -0.2	10:54 SP: 4.0 PV: 3.8 DV: -0.
08:54 SP: 4.0 PV: 3.7 DV: -0.3	09:54 SP: 4.0 PV: 4.2 DV: 0.2
07:54 SP: 4.0 PV: 4.2 DV: 0.2	08:54 SP: 4.0 PV: 3.8 DV: -0.2
06:54 SP: 4.0 PV: 3.8 DV: -0.2	07:54 SP: 4.0 PV: 4.1 DV: 0.1
05:54 SP: 4.0 PV: 4.2 DV: 0.2	06:54 SP: 4.0 PV: 3.9 DV: -0.1
04:54 SP: 4.0 PV: 3.7 DV: -0.3	05:54 SP: 4.0 PV: 3.7 DV: -0.3
03:54 SP: 4.0 PV: 3.7 DV: -0.3	04:54 SP: 4.0 PV: 4.1 DV: 0.1
02:54 SP: 4.0 PV: 4.3 DV: 0.3	03:54 SP: 4.0 PV: 3.8 DV: -0.2
01:54 SP: 4.0 PV: 3.9 DV: -0.1	02:54 SP: 4.0 PV: 4.0 DV: 0.0
00:54 SP: 4.0 PV: 4.1 DV: 0.1	01:54 SP: 4.0 PV: 4.1 DV: 0.1
2025-08-05	00:54 SP: 4.0 PV: 3.7 DV: -0.3
	2025-08-07

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 3.6 DV: -0.4
 22:54 SP: 4.0 PV: 3.6 DV: -0.4
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.7 DV: 0.7
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 4.1 DV: 0.1
 15:54 SP: 4.0 PV: 4.0 DV: 0.0
 14:54 SP: 4.0 PV: 3.6 DV: -0.4
 13:54 SP: 4.0 PV: 3.6 DV: -0.4
 12:54 SP: 4.0 PV: 4.8 DV: 0.8
 11:54 SP: 4.0 PV: 4.6 DV: 0.6
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.0 DV: 0.0
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 3.9 DV: -0.1
 2025-08-10
 23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 3.9 DV: -0.1
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 4.1 DV: 0.1
 14:54 SP: 4.0 PV: 4.0 DV: 0.0
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.1 DV: 0.1
 11:54 SP: 4.0 PV: 3.9 DV: -0.1
 10:54 SP: 4.0 PV: 4.0 DV: 0.0
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 3.6 DV: -0.4
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 3.7 DV: -0.3
 04:54 SP: 4.0 PV: 4.1 DV: 0.1
 03:54 SP: 4.0 PV: 3.9 DV: -0.1
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 3.7 DV: -0.3
 2025-08-09

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.1 DV: 0.1
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 3
 18:54 SP: 4.0 PV: 3.7 DV: -0.1
 17:54 SP: 4.0 PV: 3.6 DV: -0.4
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 4.1 DV: 0.1
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 3.7 DV: -0.3
 09:54 SP: 4.0 PV: 4.2 DV: 0.2
 08:54 SP: 4.0 PV: 3.8 DV: -0.2
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 3.6 DV: -0.4
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 3.6 DV: -0.4
 00:54 SP: 4.0 PV: 4.2 DV: 0.2
 2025-08-12
 23:54 SP: 4.0 PV: 3.8 DV: -0.2
 22:54 SP: 4.0 PV: 4.0 DV: 0.0
 0
 21:54 SP: 4.0 PV: 4.0 DV: 0.0
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.0 DV: 0.0
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 3.6 DV: -0.4
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 4.1 DV: 0.1
 08:54 SP: 4.0 PV: 3.8 DV: -0.2
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 3.7 DV: -0.3
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 3.6 DV: -0.2
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 3.9 DV: -0.1
 2025-08-11

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.8 DV: -0.2
 20:54 SP: 4.0 PV: 3.6 DV: -0.4
 19:54 SP: 4.0 PV: 4.3 DV: 0.3
 18:54 SP: 4.0 PV: 3.8 DV: -0.2
 17:54 SP: 4.0 PV: 3.6 DV: -0.4
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.9 DV: -0.1
 14:54 SP: 4.0 PV: 3.7 DV: -0.3
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.8 DV: -0.2
 11:54 SP: 4.0 PV: 3.6 DV: -0.4
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 3.9 DV: -0.1
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.7 DV: -0.3
 06:54 SP: 4.0 PV: 4.2 DV: 0.2
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 3.6 DV: -0.4
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 3.8 DV: -0.2
 00:54 SP: 4.0 PV: 3.6 DV: -0.4

2025-08-14

23:54 SP: 4.0 PV: 4.2 DV: 0.2
 22:54 SP: 4.0 PV: 3.8 DV: -0.2
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.9 DV: -0.1
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 3.6 DV: -0.4
 15:54 SP: 4.0 PV: 3.6 DV: -0.4
 14:54 SP: 4.0 PV: 4.1 DV: 0.1
 13:54 SP: 4.0 PV: 4.7 DV: 0.7
 12:54 SP: 4.0 PV: 4.5 DV: 0.5
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.1 DV: 0.1
 08:54 SP: 4.0 PV: 3.9 DV: -0.1
 07:54 SP: 4.0 PV: 3.9 DV: -0.1
 06:54 SP: 4.0 PV: 3.6 DV: -0.4
 05:54 SP: 4.0 PV: 3.6 DV: -0.4
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.7 DV: 0.7
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-08-13

~~

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 3.6 DV: -0.4
 21:54 SP: 4.0 PV: 3.4 DV: -0.6
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.3 DV: 0.3
 17:54 SP: 4.0 PV: 3.8 DV: -0.2
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.8 DV: -0.2
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 4.0 DV: 0.0
 12:54 SP: 4.0 PV: 4.0 DV: 0.0
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.1 DV: 0.1
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 4.2 DV: 0.2
 07:54 SP: 4.0 PV: 3.8 DV: -0.2
 06:54 SP: 4.0 PV: 4.1 DV: 0.1
 05:54 SP: 4.0 PV: 3.7 DV: -0.3
 04:54 SP: 4.0 PV: 4.2 DV: 0.2
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 4.2 DV: 0.2
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-08-16

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 4.0 DV: 0.0
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.2 DV: 0.2
 18:54 SP: 4.0 PV: 3.7 DV: -0.3
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 3.6 DV: -0.4
 10:54 SP: 4.0 PV: 4.3 DV: 0.3
 09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 4.1 DV: 0.1
 07:54 SP: 4.0 PV: 4.1 DV: 0.1
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.0 DV: 0.0
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-08-15

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 3.4 DV: -0.6
 20:54 SP: 4.0 PV: 4.5 DV: 0.5
 19:54 SP: 4.0 PV: 4.7 DV: 0.7
 18:54 SP: 4.0 PV: 4.4 DV: 0.4
 17:54 SP: 4.0 PV: 3.9 DV: -0.1
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.8 DV: -0.2
 13:54 SP: 4.0 PV: 3.6 DV: -0.4
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.7 DV: -0.3
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 4.0 DV: 0.0
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.1 DV: 0.1
 00:54 SP: 4.0 PV: 3.9 DV: -0.1

2025-08-18

23:54 SP: 4.0 PV: 4.0 DV: 0.0
 22:54 SP: 4.0 PV: 3.7 DV: -0.3
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.8 DV: -0.2
 19:54 SP: 4.0 PV: 3.9 DV: -0.1
 18:54 SP: 4.0 PV: 4.0 DV: 0.0
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.2 DV: 0.2
 15:54 SP: 4.0 PV: 3.7 DV: -0.3
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 3.9 DV: -0.1
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 3.6 DV: -0.4
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 3.7 DV: -0.3
 08:54 SP: 4.0 PV: 3.6 DV: -0.4
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.1 DV: 0.1
 04:54 SP: 4.0 PV: 3.9 DV: -0.1
 03:54 SP: 4.0 PV: 3.8 DV: -0.2
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 3.6 DV: -0.4
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-08-17

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.7 DV: -0.3
 20:54 SP: 4.0 PV: 4.2 DV: 0.2
 19:54 SP: 4.0 PV: 3.7 DV: -0.3
 18:54 SP: 4.0 PV: 4.2 DV: 0.2
 17:54 SP: 4.0 PV: 3.7 DV: -0.3
 16:54 SP: 4.0 PV: 4.1 DV: 0.1
 15:54 SP: 4.0 PV: 4.1 DV: 0.1
 14:54 SP: 4.0 PV: 4.2 DV: 0.2
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.3 DV: 0.3
 10:54 SP: 4.0 PV: 3.7 DV: -0.3
 09:54 SP: 4.0 PV: 4.1 DV: 0.1
 08:54 SP: 4.0 PV: 4.0 DV: 0.0
 07:54 SP: 4.0 PV: 4.1 DV: 0.1
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 3.9 DV: -0.1
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 3.7 DV: -0.3
 01:54 SP: 4.0 PV: 4.3 DV: 0.3
 00:54 SP: 4.0 PV: 3.8 DV: -0.2

2025-08-20

23:54 SP: 4.0 PV: 3.7 DV: -0.3
 22:54 SP: 4.0 PV: 3.4 DV: -0.6
 21:54 SP: 4.0 PV: 4.4 DV: 0.4
 20:54 SP: 4.0 PV: 4.6 DV: 0.6
 19:54 SP: 4.0 PV: 4.4 DV: 0.4
 18:54 SP: 4.0 PV: 3.9 DV: -0.1
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.8 DV: -0.2
 15:54 SP: 4.0 PV: 4.2 DV: 0.2
 14:54 SP: 4.0 PV: 3.8 DV: -0.2
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.2 DV: 0.2
 11:54 SP: 4.0 PV: 3.8 DV: -0.2
 10:54 SP: 4.0 PV: 3.4 DV: -0.6
 09:54 SP: 4.0 PV: 4.3 DV: 0.3
 08:54 SP: 4.0 PV: 4.8 DV: 0.8
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 3.4 DV: -0.6
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.6 DV: 0.6
 02:54 SP: 4.0 PV: 4.6 DV: 0.6
 01:54 SP: 4.0 PV: 4.4 DV: 0.4
 00:54 SP: 4.0 PV: 4.3 DV: 0.3

2025-08-19

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 4.6 DV: 0.6
 22:54 SP: 4.0 PV: 4.3 DV: 0.3
 21:54 SP: 4.0 PV: 3.9 DV: -0.1
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 4.2 DV: 0.2
 18:54 SP: 4.0 PV: 3.6 DV: -0.4
 17:54 SP: 4.0 PV: 4.2 DV: 0.2
 16:54 SP: 4.0 PV: 3.9 DV: -0.1
 15:54 SP: 4.0 PV: 3.6 DV: -0.4
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 3.7 DV: -0.3
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 3.4 DV: -0.6
 09:54 SP: 4.0 PV: 4.6 DV: 0.6
 08:54 SP: 4.0 PV: 4.7 DV: 0.7
 07:54 SP: 4.0 PV: 4.4 DV: 0.4
 06:54 SP: 4.0 PV: 3.7 DV: -0.3
 05:54 SP: 4.0 PV: 3.6 DV: -0.4
 04:54 SP: 4.0 PV: 4.2 DV: 0.2
 03:54 SP: 4.0 PV: 3.7 DV: -0.3
 02:54 SP: 4.0 PV: 4.3 DV: 0.3
 01:54 SP: 4.0 PV: 3.7 DV: -0.3
 00:54 SP: 4.0 PV: 4.3 DV: 0.3
 2025-08-22
 23:54 SP: 4.0 PV: 3.9 DV: -0.1
 22:54 SP: 4.0 PV: 4.2 DV: 0.2
 21:54 SP: 4.0 PV: 3.6 DV: -0.4
 20:54 SP: 4.0 PV: 3.4 DV: -0.6
 19:54 SP: 4.0 PV: 4.6 DV: 0.6
 18:54 SP: 4.0 PV: 4.8 DV: 0.8
 17:54 SP: 4.0 PV: 4.3 DV: 0.3
 16:54 SP: 4.0 PV: 3.7 DV: -0.3
 15:54 SP: 4.0 PV: 4.3 DV: 0.3

14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.7 DV: -0.3
 11:54 SP: 4.0 PV: 4.2 DV: 0.2
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 3.6 DV: -0.4
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 4.6 DV: 0.6
 06:54 SP: 4.0 PV: 4.7 DV: 0.7
 05:54 SP: 4.0 PV: 4.3 DV: 0.3
 04:54 SP: 4.0 PV: 3.8 DV: -0.2
 03:54 SP: 4.0 PV: 4.2 DV: 0.2
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 4.2 DV: 0.2
 00:54 SP: 4.0 PV: 4.1 DV: 0.1
 2025-08-21

23:54 SP: 4.0 PV: 3.6 DV: -0.4
 22:54 SP: 4.0 PV: 4.4 DV: 0.4
 21:54 SP: 4.0 PV: 4.8 DV: 0.8
 20:54 SP: 4.0 PV: 4.4 DV: 0.4
 19:54 SP: 4.0 PV: 3.8 DV: -0.2
 18:54 SP: 4.0 PV: 3.6 DV: -0.4
 17:54 SP: 4.0 PV: 3.6 DV: -0.4
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.6 DV: 0.6
 14:54 SP: 4.0 PV: 4.3 DV: 0.3
 13:54 SP: 4.0 PV: 3.8 DV: -0.2
 12:54 SP: 4.0 PV: 4.3 DV: 0.3
 11:54 SP: 4.0 PV: 3.7 DV: -0.3
 10:54 SP: 4.0 PV: 4.2 DV: 0.2
 09:54 SP: 4.0 PV: 4.0 DV: 0.0
 08:54 SP: 4.0 PV: 3.7 DV: -0.3
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.9 DV: -0.1
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 3.6 DV: -0.4
 02:54 SP: 4.0 PV: 3.8 DV: -0.2
 01:54 SP: 4.0 PV: 4.8 DV: 0.8
 00:54 SP: 4.0 PV: 4.6 DV: 0.6

2025-08-24

23:54 SP: 4.0 PV: 4.3 DV: 0.3
 22:54 SP: 4.0 PV: 3.9 DV: -0.1
 21:54 SP: 4.0 PV: 4.1 DV: 0.1
 20:54 SP: 4.0 PV: 3.7 DV: -0.3
 19:54 SP: 4.0 PV: 3.6 DV: -0.4
 18:54 SP: 4.0 PV: 3.6 DV: -0.4
 17:54 SP: 4.0 PV: 4.8 DV: 0.8
 16:54 SP: 4.0 PV: 4.6 DV: 0.6
 15:54 SP: 4.0 PV: 4.3 DV: 0.3
 14:54 SP: 4.0 PV: 3.9 DV: -0.1
 13:54 SP: 4.0 PV: 4.2 DV: 0.2
 12:54 SP: 4.0 PV: 3.9 DV: -0.1
 11:54 SP: 4.0 PV: 4.1 DV: 0.1
 10:54 SP: 4.0 PV: 3.9 DV: -0.1
 09:54 SP: 4.0 PV: 4.1 DV: 0.1
 08:54 SP: 4.0 PV: 3.6 DV: -0.4
 07:54 SP: 4.0 PV: 4.2 DV: 0.2
 06:54 SP: 4.0 PV: 3.8 DV: -0.2
 05:54 SP: 4.0 PV: 4.2 DV: 0.2
 04:54 SP: 4.0 PV: 3.7 DV: -0.3
 03:54 SP: 4.0 PV: 3.6 DV: -0.4
 02:54 SP: 4.0 PV: 3.6 DV: -0.4
 01:54 SP: 4.0 PV: 3.9 DV: -0.1
 00:54 SP: 4.0 PV: 4.8 DV: 0.8

2025-08-23

SP Setting Point
 PV Present Value
 DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:54 SP: 4.0 PV: 3.7 DV: -0.3	23:54 SP: 4.0 PV: 4.8 DV: 0.8
22:54 SP: 4.0 PV: 3.6 DV: -0.4	22:54 SP: 4.0 PV: 4.4 DV: 0.4
21:54 SP: 4.0 PV: 4.4 DV: 0.4	21:54 SP: 4.0 PV: 4.3 DV: 0.3
20:54 SP: 4.0 PV: 4.7 DV: 0.7	20:54 SP: 4.0 PV: 3.8 DV: -0.2
19:54 SP: 4.0 PV: 4.6 DV: 0.6	19:54 SP: 4.0 PV: 4.1 DV: 0.1
18:54 SP: 4.0 PV: 4.4 DV: 0.4	18:54 SP: 4.0 PV: 4.0 DV: 0.0
17:54 SP: 4.0 PV: 3.9 DV: -0.1	17:54 SP: 4.0 PV: 3.7 DV: -0.3
16:54 SP: 4.0 PV: 3.4 DV: -0.6	16:54 SP: 4.0 PV: 4.2 DV: 0.2
15:54 SP: 4.0 PV: 4.5 DV: 0.5	15:54 SP: 4.0 PV: 3.6 DV: -0.4
14:54 SP: 4.0 PV: 4.7 DV: 0.7	14:54 SP: 4.0 PV: 4.2 DV: 0.2
13:54 SP: 4.0 PV: 4.5 DV: 0.5	13:54 SP: 4.0 PV: 4.8 DV: 0.8
12:54 SP: 4.0 PV: 3.7 DV: -0.3	12:54 SP: 4.0 PV: 4.4 DV: 0.4
11:54 SP: 4.0 PV: 3.6 DV: -0.4	3
10:54 SP: 4.0 PV: 4.3 DV: 0.3	11:54 SP: 4.0 PV: 4.3 DV: 0.
09:54 SP: 4.0 PV: 3.8 DV: -0.2	10:54 SP: 4.0 PV: 4.0 DV: 0.0
08:54 SP: 4.0 PV: 4.2 DV: 0.2	09:54 SP: 4.0 PV: 4.2 DV: 0.2
07:54 SP: 4.0 PV: 3.7 DV: -0.3	08:54 SP: 4.0 PV: 3.7 DV: -0.3
06:54 SP: 4.0 PV: 3.6 DV: -0.4	07:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 4.3 DV: 0.3	06:54 SP: 4.0 PV: 3.9 DV: -0.1
04:54 SP: 4.0 PV: 3.9 DV: -0.1	05:54 SP: 4.0 PV: 3.6 DV: -0.4
03:54 SP: 4.0 PV: 4.2 DV: 0.2	04:54 SP: 4.0 PV: 4.3 DV: 0.3
02:54 SP: 4.0 PV: 3.7 DV: -0.3	03:54 SP: 4.0 PV: 3.7 DV: -0.3
01:54 SP: 4.0 PV: 3.6 DV: -0.4	02:54 SP: 4.0 PV: 4.2 DV: 0.2
00:54 SP: 4.0 PV: 3.8 DV: -0.2	01:54 SP: 4.0 PV: 3.9 DV: -0.1
2025-08-26	00:54 SP: 4.0 PV: 4.2 DV: 0.2
23:54 SP: 4.0 PV: 4.8 DV: 0.8	2025-08-28
22:54 SP: 4.0 PV: 4.5 DV: 0.5	23:54 SP: 4.0 PV: 3.6 DV: -0.4
21:54 SP: 4.0 PV: 4.3 DV: 0.3	22:54 SP: 4.0 PV: 4.3 DV: 0.3
20:54 SP: 4.0 PV: 3.8 DV: -0.2	21:54 SP: 4.0 PV: 4.0 DV: 0.0
19:54 SP: 4.0 PV: 4.2 DV: 0.2	20:54 SP: 4.0 PV: 3.6 DV: -0.4
18:54 SP: 4.0 PV: 3.6 DV: -0.4	19:54 SP: 4.0 PV: 3.8 DV: -0.2
17:54 SP: 4.0 PV: 3.4 DV: -0.6	18:54 SP: 4.0 PV: 4.8 DV: 0.8
16:54 SP: 4.0 PV: 4.5 DV: 0.5	17:54 SP: 4.0 PV: 4.7 DV: 0.7
15:54 SP: 4.0 PV: 4.8 DV: 0.8	16:54 SP: 4.0 PV: 4.3 DV: 0.3
14:54 SP: 4.0 PV: 4.3 DV: 0.3	15:54 SP: 4.0 PV: 3.7 DV: -0.3
13:54 SP: 4.0 PV: 3.9 DV: -0.1	14:54 SP: 4.0 PV: 3.6 DV: -0.4
12:54 SP: 4.0 PV: 4.2 DV: 0.2	13:54 SP: 4.0 PV: 4.8 DV: 0.8
11:54 SP: 4.0 PV: 3.8 DV: -0.2	12:54 SP: 4.0 PV: 4.6 DV: 0.6
10:54 SP: 4.0 PV: 4.2 DV: 0.2	11:54 SP: 4.0 PV: 4.3 DV: 0.3
09:54 SP: 4.0 PV: 4.1 DV: 0.1	10:54 SP: 4.0 PV: 3.7 DV: -0.3
08:54 SP: 4.0 PV: 3.7 DV: -0.3	09:54 SP: 4.0 PV: 4.2 DV: 0.2
07:54 SP: 4.0 PV: 4.1 DV: 0.1	08:54 SP: 4.0 PV: 3.7 DV: -0.3
06:54 SP: 4.0 PV: 4.1 DV: 0.1	07:54 SP: 4.0 PV: 4.2 DV: 0.2
05:54 SP: 4.0 PV: 3.6 DV: -0.4	06:54 SP: 4.0 PV: 4.0 DV: 0.0
04:54 SP: 4.0 PV: 4.3 DV: 0.3	05:54 SP: 4.0 PV: 4.1 DV: 0.1
03:54 SP: 4.0 PV: 3.8 DV: -0.2	04:54 SP: 4.0 PV: 4.0 DV: 0.0
02:54 SP: 4.0 PV: 4.2 DV: 0.2	03:54 SP: 4.0 PV: 3.7 DV: -0.3
01:54 SP: 4.0 PV: 3.7 DV: -0.3	02:54 SP: 4.0 PV: 4.2 DV: 0.2
00:54 SP: 4.0 PV: 3.6 DV: -0.4	01:54 SP: 4.0 PV: 3.8 DV: -0.2
2025-08-25	00:54 SP: 4.0 PV: 4.1 DV: 0.1
	2025-08-27

SP Setting Point
PV Present Value
DV Difference Value

부록8. 서식II-01-08, 온도기록지(IP 보관)-자동온도 기록 레코더형

2025.8

23:49 SP: 4.0 PV: 3.8 DV: -0.2
 22:49 SP: 4.0 PV: 3.4 DV: -0.6
 21:49 SP: 4.0 PV: 3.6 DV: -0.4
 20:49 SP: 4.0 PV: 4.8 DV: 0.8
 19:49 SP: 4.0 PV: 4.6 DV: 0.6
 18:49 SP: 4.0 PV: 3.8 DV: -0.2
 17:49 SP: 4.0 PV: 3.4 DV: -0.6
 16:49 SP: 4.0 PV: 4.5 DV: 0.5
 15:49 SP: 4.0 PV: 4.7 DV: 0.7
 14:49 SP: 4.0 PV: 4.4 DV: 0.4
 13:49 SP: 4.0 PV: 3.7 DV: -0.3
 12:49 SP: 4.0 PV: 3.4 DV: -0.6
 11:49 SP: 4.0 PV: 4.6 DV: 0.6

Condensor1 Temp#1. OK...

Compressor#1 Output OK...

System Checking Process...

10:49:55

2025-08-29

주소창에 한글로 혈액냉장고

www.gsmedical.co.kr

CBR-1000SS

GIANT STAR Ltd.

09:54 SP: 4.0 PV: 3.8 DV: -0.2
 08:54 SP: 4.0 PV: 3.6 DV: -0.4
 07:54 SP: 4.0 PV: 4.3 DV: 0.3
 06:54 SP: 4.0 PV: 4.0 DV: 0.0
 05:54 SP: 4.0 PV: 3.6 DV: -0.4
 04:54 SP: 4.0 PV: 3.6 DV: -0.4
 03:54 SP: 4.0 PV: 4.3 DV: 0.3
 02:54 SP: 4.0 PV: 3.9 DV: -0.1
 01:54 SP: 4.0 PV: 3.6 DV: -0.4
 00:54 SP: 4.0 PV: 4.2 DV: 0.2

2025-08-29

SP Setting Point
 PV Present Value
 DV Difference Value

23:49 SP: 4.0 PV: 4.2 DV: 0.2
 22:49 SP: 4.0 PV: 3.7 DV: -0.3
 21:49 SP: 4.0 PV: 4.3 DV: 0.3
 20:49 SP: 4.0 PV: 3.7 DV: -0.3
 19:49 SP: 4.0 PV: 4.3 DV: 0.3
 18:49 SP: 4.0 PV: 3.9 DV: -0.1
 17:49 SP: 4.0 PV: 4.1 DV: 0.1
 16:49 SP: 4.0 PV: 3.9 DV: -0.1
 15:49 SP: 4.0 PV: 3.6 DV: -0.4
 14:49 SP: 4.0 PV: 4.3 DV: 0.3
 13:49 SP: 4.0 PV: 3.9 DV: -0.1
 12:49 SP: 4.0 PV: 3.6 DV: -0.4
 11:49 SP: 4.0 PV: 3.6 DV: -0.4
 10:49 SP: 4.0 PV: 3.8 DV: -0.2
 09:49 SP: 4.0 PV: 4.7 DV: 0.7
 08:49 SP: 4.0 PV: 4.6 DV: 0.6
 07:49 SP: 4.0 PV: 4.3 DV: 0.3
 06:49 SP: 4.0 PV: 3.9 DV: -0.1
 05:49 SP: 4.0 PV: 3.7 DV: -0.3
 04:49 SP: 4.0 PV: 4.3 DV: 0.3
 03:49 SP: 4.0 PV: 3.8 DV: -0.2
 02:49 SP: 4.0 PV: 4.2 DV: 0.2
 01:49 SP: 4.0 PV: 4.0 DV: 0.0
 00:49 SP: 4.0 PV: 4.1 DV: 0.1
 2025-08-31

23:49 SP: 4.0 PV: 3.7 DV: -0.3
 22:49 SP: 4.0 PV: 4.2 DV: 0.2
 21:49 SP: 4.0 PV: 3.9 DV: -0.1
 20:49 SP: 4.0 PV: 4.0 DV: 0.0
 19:49 SP: 4.0 PV: 4.0 DV: 0.0
 18:49 SP: 4.0 PV: 4.1 DV: 0.1
 17:49 SP: 4.0 PV: 4.0 DV: 0.0
 16:49 SP: 4.0 PV: 3.7 DV: -0.3
 15:49 SP: 4.0 PV: 4.2 DV: 0.2
 14:49 SP: 4.0 PV: 3.7 DV: -0.3
 13:49 SP: 4.0 PV: 4.2 DV: 0.2
 12:49 SP: 4.0 PV: 4.0 DV: 0.0
 11:49 SP: 4.0 PV: 4.1 DV: 0.1
 10:49 SP: 4.0 PV: 3.7 DV: -0.3
 09:49 SP: 4.0 PV: 4.2 DV: 0.2
 08:49 SP: 4.0 PV: 3.8 DV: -0.2
 07:49 SP: 4.0 PV: 3.7 DV: -0.3
 06:49 SP: 4.0 PV: 4.2 DV: 0.2
 05:49 SP: 4.0 PV: 3.8 DV: -0.2
 04:49 SP: 4.0 PV: 3.6 DV: -0.4
 03:49 SP: 4.0 PV: 3.6 DV: -0.4
 02:49 SP: 4.0 PV: 4.5 DV: 0.5
 01:49 SP: 4.0 PV: 4.8 DV: 0.8
 00:49 SP: 4.0 PV: 4.3 DV: 0.3
 2025-08-30