

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1 9134Room
	Manufacturer	SANYO	Storage Period	2025.8
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.08.01	P:30	17	20	19	ml	
2025.08.02						
2025.08.03						
2025.08.04	P:40	17	20	18	ml	
2025.08.05	P:30	17	20	19	ml	
2025.08.06	P:30	17	20	19	ml	
2025.08.07	P:30	17	20	19	ml	
2025.08.08	P:30	17	20	19	ml	
2025.08.09						
2025.08.10						
2025.08.11	P:30	17	20	18	ml	
2025.08.12	P:30	17	20	19	ml	
2025.08.13	P:15	17	20	18	ml	
2025.08.14	P:30	17	20	20	ml	
2025.08.15						광복절
2025.08.16						
2025.08.17						
2025.08.18	9:20	17	20	19	ml	
2025.08.19	P:40	17	20	19	ml	
2025.08.20	P:30	17	20	20	ml	
2025.08.21	P:30	17	20	19	ml	
2025.08.22	P:20	17	20	18	ml	
2025.08.23						
2025.08.24						
2025.08.25	P:30	17	20	19	ml	
2025.08.26	P:30	17	20	19	ml	
2025.08.27	P:20	17	20	18	ml	
2025.08.28	P:40	17	20	20	ml	
2025.08.29	P:30	17	20	19	ml	
2025.08.30						
2025.08.31						

date of copying

sign