

IP STORAGE TEMPERATURE LOG

Storage Subject	Description	전기냉장냉동기	Storage Space	항온고1 9134Room
	Manufacturer	SANYO	Storage Period	2025.9
	Model Name	MPR-720R	Temperature Range	15~25℃
	Serial Number	07120004	Setting Point	19.0℃
Calibration Subject	Description	Digital Thermometer	Min/Max Record	온도기록 후 reset하여 그 시점 이후 Min/Max 새로 기록됨
	Manufacturer	SATO		
	Model Name	PC-3500		
	Serial Number	825287		
	Date of Calibration	2025.2.25	Alarm Range	<15℃, >25℃, 5분 후
	Date of Calibration Due	2026.2.25	Alarm Method	Call to Pharmacist
	Certificate No.	SC2502-05400-6	Alarm Clear	000

Date	Time	Temperature(℃)			Verified by	Comment
		Min	Max	Current		
2025.09.01	P:30	17	20	19	ant	
2025.09.02	P:30	17	20	19	ant	
2025.09.03	P:30	17	20	19	ant	
2025.09.04	P:30	17	20	19	ant	
2025.09.05	P:40	17	20	19	ant	
2025.09.06						
2025.09.07						
2025.09.08	P:00	17	20	19	ant	
2025.09.09	P:30	17	20	19	ant	
2025.09.10	P:30	17	20	19	ant	
2025.09.11	P:30	17	20	19	ant	
2025.09.12	P:30	17	20	20	ant	
2025.09.13						
2025.09.14						
2025.09.15	P:30	17	20	19	ant	
2025.09.16	P:40	17	20	19	ant	
2025.09.17	P:30	17	20	19	ant	
2025.09.18	P:20	17	20	20	ant	
2025.09.19	P:20	17	20	19	ant	
2025.09.20						
2025.09.21						
2025.09.22	P:30	17	20	19	ant	
2025.09.23	P:30	17	20	19	ant	
2025.09.24	P:30	17	20	19	ant	
2025.09.25	P:20	17	20	19	ant	
2025.09.26	P:10	17	20	19	ant	
2025.09.27						
2025.09.28						
2025.09.29	P:30	17	20	19	ant	
2025.09.30	P:40	17	20	19	ant	

date of copying

sign