

IP STORAGE TEMPERATURE LOG

Period of Record	Year, Month	2026.6
Storage	Location	동국대학교일산병원 임상시험센터 약국
	Description	Pharmaceutical Refrigerator
	Manufacturer	SGM
	Model Name	CS-1300P
	Serial Number	K1300P0292510M
	Temperature Range	2~8℃
	Setting Point	4.0℃
Thermometer (including alarm system)	Description	Wi-Fi data logger
	Manufacturer	Testo Korea
	Model Name	Saveris2-T3
	Serial Number	0054857013
	Sensor Channel	Channel 1
	Min/Max Record	Monthly report 2page / This month Min (4.9)℃, Max(5.8)℃
Thermometer Calibration	Date of Calibration	2025.10.1
	Date of Calibration Due	2026.10.1
	Certificate No.	A25-1985-001
Alarm System	Alarm Threshold	Lower : 2.5℃, Upper : 7.5℃, immediate alarm without delay
	Alarm Method	Mail, SMS to Pharmacists

Date	Time	Temperature(℃)	Deviation	Verified by	Comment
2026.06.01	P:30	5.5	☐ Y / ☒ N		
2026.06.02	P:30	5.5	☐ Y / ☒ N		
2026.06.03			☐ Y / ☐ N		2026 지방선거
2026.06.04	10:00	5.5	☐ Y / ☒ N		
2026.06.05	P:30	5.4	☐ Y / ☒ N		
2026.06.06			☐ Y / ☐ N		
2026.06.07			☐ Y / ☐ N		
2026.06.08	10:00	5.2	☐ Y / ☒ N		
2026.06.09	P:10	5.2	☐ Y / ☒ N		
2026.06.10	P:20	5.4	☐ Y / ☒ N		
2026.06.11	P:40	5.2	☐ Y / ☒ N		
2026.06.12	P:40	5.2	☐ Y / ☒ N		
2026.06.13			☐ Y / ☐ N		
2026.06.14			☐ Y / ☐ N		
2026.06.15	P:30	5.1	☐ Y / ☒ N		
2026.06.16	P:35	5.1	☐ Y / ☒ N		
2026.06.17	P:40	5.1	☐ Y / ☒ N		
2026.06.18	P:20	5.3	☐ Y / ☒ N		
2026.06.19	P:30	5.2	☐ Y / ☒ N		
2026.06.20			☐ Y / ☐ N		
2026.06.21			☐ Y / ☐ N		
2026.06.22	P:30	5.2	☐ Y / ☒ N		
2026.06.23	P:20	5.7	☐ Y / ☒ N		
2026.06.24	P:15	5.2	☐ Y / ☒ N		
2026.06.25	P:20	5.3	☐ Y / ☒ N		
2026.06.26	P:10	5.2	☐ Y / ☒ N		
2026.06.27			☐ Y / ☐ N		
2026.06.28			☐ Y / ☐ N		
2026.06.29	P:30	5.2	☐ Y / ☒ N		
2026.06.30	P:20	5.1	☐ Y / ☒ N		
			☐ Y / ☐ N		

date of copying

sign



측정 테이블

측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
Min	4.9
Max	5.8
Average	5.3
26. 6. 1. 오전 12:00:00	5.6
26. 6. 1. 오전 12:30:00	5.5
26. 6. 1. 오전 1:00:00	5.4
26. 6. 1. 오전 1:30:00	5.5
26. 6. 1. 오전 2:00:00	5.5
26. 6. 1. 오전 2:30:00	5.5
26. 6. 1. 오전 3:00:00	5.5
26. 6. 1. 오전 3:30:00	5.5
26. 6. 1. 오전 4:00:00	5.5
26. 6. 1. 오전 4:30:00	5.5
26. 6. 1. 오전 5:00:00	5.5
26. 6. 1. 오전 5:30:00	5.6
26. 6. 1. 오전 6:00:00	5.5
26. 6. 1. 오전 6:30:00	5.5
26. 6. 1. 오전 7:00:00	5.5
26. 6. 1. 오전 7:30:00	5.5
26. 6. 1. 오전 8:00:00	5.5
26. 6. 1. 오전 8:30:00	5.5
26. 6. 1. 오전 9:00:00	5.5
26. 6. 1. 오전 9:30:00	5.4
26. 6. 1. 오전 10:00:00	5.5



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 1. 오전 10:30:00	5.5
26. 6. 1. 오전 11:00:00	5.5
26. 6. 1. 오전 11:30:00	5.6
26. 6. 1. 오후 12:00:00	5.6
26. 6. 1. 오후 12:30:00	5.6
26. 6. 1. 오후 1:00:00	5.6
26. 6. 1. 오후 1:30:00	5.6
26. 6. 1. 오후 2:00:00	5.6
26. 6. 1. 오후 2:30:00	5.6
26. 6. 1. 오후 3:00:00	5.6
26. 6. 1. 오후 3:30:00	5.6
26. 6. 1. 오후 4:00:00	5.5
26. 6. 1. 오후 4:30:00	5.5
26. 6. 1. 오후 5:00:00	5.5
26. 6. 1. 오후 5:30:00	5.6
26. 6. 1. 오후 6:00:00	5.5
26. 6. 1. 오후 6:30:00	5.5
26. 6. 1. 오후 7:00:00	5.5
26. 6. 1. 오후 7:30:00	5.5
26. 6. 1. 오후 8:00:00	5.5
26. 6. 1. 오후 8:30:00	5.4
26. 6. 1. 오후 9:00:00	5.4
26. 6. 1. 오후 9:30:00	5.4
26. 6. 1. 오후 10:00:00	5.4
26. 6. 1. 오후 10:30:00	5.4
26. 6. 1. 오후 11:00:00	5.4
26. 6. 1. 오후 11:30:00	5.5
26. 6. 2. 오전 12:00:00	5.5
26. 6. 2. 오전 12:30:00	5.5
26. 6. 2. 오전 1:00:00	5.5
26. 6. 2. 오전 1:30:00	5.5
26. 6. 2. 오전 2:00:00	5.5



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 2. 오전 2:30:00	5.5
26. 6. 2. 오전 3:00:00	5.5
26. 6. 2. 오전 3:30:00	5.5
26. 6. 2. 오전 4:00:00	5.5
26. 6. 2. 오전 4:30:00	5.5
26. 6. 2. 오전 5:00:00	5.5
26. 6. 2. 오전 5:30:00	5.6
26. 6. 2. 오전 6:00:00	5.6
26. 6. 2. 오전 6:30:00	5.5
26. 6. 2. 오전 7:00:00	5.5
26. 6. 2. 오전 7:30:00	5.4
26. 6. 2. 오전 8:00:00	5.5
26. 6. 2. 오전 8:30:00	5.5
26. 6. 2. 오전 9:00:00	5.5
26. 6. 2. 오전 9:30:00	5.5
26. 6. 2. 오전 10:00:00	5.5
26. 6. 2. 오전 10:30:00	5.5
26. 6. 2. 오전 11:00:00	5.5
26. 6. 2. 오전 11:30:00	5.6
26. 6. 2. 오후 12:00:00	5.6
26. 6. 2. 오후 12:30:00	5.6
26. 6. 2. 오후 1:00:00	5.6
26. 6. 2. 오후 1:30:00	5.6
26. 6. 2. 오후 2:00:00	5.6
26. 6. 2. 오후 2:30:00	5.6
26. 6. 2. 오후 3:00:00	5.6
26. 6. 2. 오후 3:30:00	5.6
26. 6. 2. 오후 4:00:00	5.5
26. 6. 2. 오후 4:30:00	5.5
26. 6. 2. 오후 5:00:00	5.5
26. 6. 2. 오후 5:30:00	5.7
26. 6. 2. 오후 6:00:00	5.6



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 2. 오후 6:30:00	5.6
26. 6. 2. 오후 7:00:00	5.5
26. 6. 2. 오후 7:30:00	5.5
26. 6. 2. 오후 8:00:00	5.5
26. 6. 2. 오후 8:30:00	5.5
26. 6. 2. 오후 9:00:00	5.4
26. 6. 2. 오후 9:30:00	5.4
26. 6. 2. 오후 10:00:00	5.4
26. 6. 2. 오후 10:30:00	5.5
26. 6. 2. 오후 11:00:00	5.5
26. 6. 2. 오후 11:30:00	5.7
26. 6. 3. 오전 12:00:00	5.6
26. 6. 3. 오전 12:30:00	5.5
26. 6. 3. 오전 1:00:00	5.5
26. 6. 3. 오전 1:30:00	5.5
26. 6. 3. 오전 2:00:00	5.5
26. 6. 3. 오전 2:30:00	5.5
26. 6. 3. 오전 3:00:00	5.5
26. 6. 3. 오전 3:30:00	5.5
26. 6. 3. 오전 4:00:00	5.5
26. 6. 3. 오전 4:30:00	5.5
26. 6. 3. 오전 5:00:00	5.5
26. 6. 3. 오전 5:30:00	5.5
26. 6. 3. 오전 6:00:00	5.6
26. 6. 3. 오전 6:30:00	5.6
26. 6. 3. 오전 7:00:00	5.5
26. 6. 3. 오전 7:30:00	5.5
26. 6. 3. 오전 8:00:00	5.6
26. 6. 3. 오전 8:30:00	5.6
26. 6. 3. 오전 9:00:00	5.5
26. 6. 3. 오전 9:30:00	5.5
26. 6. 3. 오전 10:00:00	5.5



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 3. 오전 10:30:00	5.5
26. 6. 3. 오전 11:00:00	5.5
26. 6. 3. 오전 11:30:00	5.5
26. 6. 3. 오후 12:00:00	5.6
26. 6. 3. 오후 12:30:00	5.5
26. 6. 3. 오후 1:00:00	5.5
26. 6. 3. 오후 1:30:00	5.5
26. 6. 3. 오후 2:00:00	5.5
26. 6. 3. 오후 2:30:00	5.5
26. 6. 3. 오후 3:00:00	5.5
26. 6. 3. 오후 3:30:00	5.5
26. 6. 3. 오후 4:00:00	5.5
26. 6. 3. 오후 4:30:00	5.5
26. 6. 3. 오후 5:00:00	5.5
26. 6. 3. 오후 5:30:00	5.4
26. 6. 3. 오후 6:00:00	5.5
26. 6. 3. 오후 6:30:00	5.5
26. 6. 3. 오후 7:00:00	5.4
26. 6. 3. 오후 7:30:00	5.4
26. 6. 3. 오후 8:00:00	5.3
26. 6. 3. 오후 8:30:00	5.3
26. 6. 3. 오후 9:00:00	5.4
26. 6. 3. 오후 9:30:00	5.5
26. 6. 3. 오후 10:00:00	5.4
26. 6. 3. 오후 10:30:00	5.3
26. 6. 3. 오후 11:00:00	5.3
26. 6. 3. 오후 11:30:00	5.3
26. 6. 4. 오전 12:00:00	5.3
26. 6. 4. 오전 12:30:00	5.3
26. 6. 4. 오전 1:00:00	5.3
26. 6. 4. 오전 1:30:00	5.4
26. 6. 4. 오전 2:00:00	5.5



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 4. 오전 2:30:00	5.5
26. 6. 4. 오전 3:00:00	5.5
26. 6. 4. 오전 3:30:00	5.6
26. 6. 4. 오전 4:00:00	5.6
26. 6. 4. 오전 4:30:00	5.6
26. 6. 4. 오전 5:00:00	5.6
26. 6. 4. 오전 5:30:00	5.6
26. 6. 4. 오전 6:00:00	5.8
26. 6. 4. 오전 6:30:00	5.8
26. 6. 4. 오전 7:00:00	5.7
26. 6. 4. 오전 7:30:00	5.6
26. 6. 4. 오전 8:00:00	5.5
26. 6. 4. 오전 8:30:00	5.5
26. 6. 4. 오전 9:00:00	5.5
26. 6. 4. 오전 9:30:00	5.5
26. 6. 4. 오전 10:00:00	5.5
26. 6. 4. 오전 10:30:00	5.5
26. 6. 4. 오전 11:00:00	5.5
26. 6. 4. 오전 11:30:00	5.5
26. 6. 4. 오후 12:00:00	5.6
26. 6. 4. 오후 12:30:00	5.6
26. 6. 4. 오후 1:00:00	5.6
26. 6. 4. 오후 1:30:00	5.5
26. 6. 4. 오후 2:00:00	5.5
26. 6. 4. 오후 2:30:00	5.5
26. 6. 4. 오후 3:00:00	5.5
26. 6. 4. 오후 3:30:00	5.5
26. 6. 4. 오후 4:00:00	5.5
26. 6. 4. 오후 4:30:00	5.5
26. 6. 4. 오후 5:00:00	5.5
26. 6. 4. 오후 5:30:00	5.5
26. 6. 4. 오후 6:00:00	5.5



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 4. 오후 6:30:00	5.5
26. 6. 4. 오후 7:00:00	5.4
26. 6. 4. 오후 7:30:00	5.4
26. 6. 4. 오후 8:00:00	5.4
26. 6. 4. 오후 8:30:00	5.4
26. 6. 4. 오후 9:00:00	5.4
26. 6. 4. 오후 9:30:00	5.5
26. 6. 4. 오후 10:00:00	5.5
26. 6. 4. 오후 10:30:00	5.4
26. 6. 4. 오후 11:00:00	5.4
26. 6. 4. 오후 11:30:00	5.4
26. 6. 5. 오전 12:00:00	5.4
26. 6. 5. 오전 12:30:00	5.5
26. 6. 5. 오전 1:00:00	5.5
26. 6. 5. 오전 1:30:00	5.4
26. 6. 5. 오전 2:00:00	5.5
26. 6. 5. 오전 2:30:00	5.5
26. 6. 5. 오전 3:00:00	5.5
26. 6. 5. 오전 3:30:00	5.6
26. 6. 5. 오전 4:00:00	5.6
26. 6. 5. 오전 4:30:00	5.6
26. 6. 5. 오전 5:00:00	5.5
26. 6. 5. 오전 5:30:00	5.5
26. 6. 5. 오전 6:00:00	5.6
26. 6. 5. 오전 6:30:00	5.6
26. 6. 5. 오전 7:00:00	5.5
26. 6. 5. 오전 7:30:00	5.4
26. 6. 5. 오전 8:00:00	5.4
26. 6. 5. 오전 8:30:00	5.4
26. 6. 5. 오전 9:00:00	5.4
26. 6. 5. 오전 9:30:00	5.4
26. 6. 5. 오전 10:00:00	5.4



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 5. 오전 10:30:00	5.5
26. 6. 5. 오전 11:00:00	5.5
26. 6. 5. 오전 11:30:00	5.5
26. 6. 5. 오후 12:00:00	5.5
26. 6. 5. 오후 12:30:00	5.5
26. 6. 5. 오후 1:00:00	5.4
26. 6. 5. 오후 1:30:00	5.5
26. 6. 5. 오후 2:00:00	5.5
26. 6. 5. 오후 2:30:00	5.4
26. 6. 5. 오후 3:00:00	5.4
26. 6. 5. 오후 3:30:00	5.3
26. 6. 5. 오후 4:00:00	5.2
26. 6. 5. 오후 4:30:00	5.3
26. 6. 5. 오후 5:00:00	5.4
26. 6. 5. 오후 5:30:00	5.3
26. 6. 5. 오후 6:00:00	5.4
26. 6. 5. 오후 6:30:00	5.4
26. 6. 5. 오후 7:00:00	5.4
26. 6. 5. 오후 7:30:00	5.5
26. 6. 5. 오후 8:00:00	5.4
26. 6. 5. 오후 8:30:00	5.4
26. 6. 5. 오후 9:00:00	5.4
26. 6. 5. 오후 9:30:00	5.3
26. 6. 5. 오후 10:00:00	5.3
26. 6. 5. 오후 10:30:00	5.3
26. 6. 5. 오후 11:00:00	5.3
26. 6. 5. 오후 11:30:00	5.3
26. 6. 6. 오전 12:00:00	5.5
26. 6. 6. 오전 12:30:00	5.4
26. 6. 6. 오전 1:00:00	5.3
26. 6. 6. 오전 1:30:00	5.3
26. 6. 6. 오전 2:00:00	5.4



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 6. 오전 2:30:00	5.3
26. 6. 6. 오전 3:00:00	5.3
26. 6. 6. 오전 3:30:00	5.3
26. 6. 6. 오전 4:00:00	5.3
26. 6. 6. 오전 4:30:00	5.3
26. 6. 6. 오전 5:00:00	5.3
26. 6. 6. 오전 5:30:00	5.3
26. 6. 6. 오전 6:00:00	5.4
26. 6. 6. 오전 6:30:00	5.4
26. 6. 6. 오전 7:00:00	5.3
26. 6. 6. 오전 7:30:00	5.1
26. 6. 6. 오전 8:00:00	5.1
26. 6. 6. 오전 8:30:00	5.2
26. 6. 6. 오전 9:00:00	5.1
26. 6. 6. 오전 9:30:00	5.1
26. 6. 6. 오전 10:00:00	5.1
26. 6. 6. 오전 10:30:00	5.1
26. 6. 6. 오전 11:00:00	5.0
26. 6. 6. 오전 11:30:00	5.1
26. 6. 6. 오후 12:00:00	5.0
26. 6. 6. 오후 12:30:00	5.1
26. 6. 6. 오후 1:00:00	5.0
26. 6. 6. 오후 1:30:00	5.0
26. 6. 6. 오후 2:00:00	5.1
26. 6. 6. 오후 2:30:00	5.1
26. 6. 6. 오후 3:00:00	5.1
26. 6. 6. 오후 3:30:00	5.2
26. 6. 6. 오후 4:00:00	5.1
26. 6. 6. 오후 4:30:00	5.2
26. 6. 6. 오후 5:00:00	5.1
26. 6. 6. 오후 5:30:00	5.1
26. 6. 6. 오후 6:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 6. 오후 6:30:00	5.2
26. 6. 6. 오후 7:00:00	5.2
26. 6. 6. 오후 7:30:00	5.2
26. 6. 6. 오후 8:00:00	5.2
26. 6. 6. 오후 8:30:00	5.2
26. 6. 6. 오후 9:00:00	5.2
26. 6. 6. 오후 9:30:00	5.2
26. 6. 6. 오후 10:00:00	5.2
26. 6. 6. 오후 10:30:00	5.3
26. 6. 6. 오후 11:00:00	5.2
26. 6. 6. 오후 11:30:00	5.1
26. 6. 7. 오전 12:00:00	4.9
26. 6. 7. 오전 12:30:00	5.0
26. 6. 7. 오전 1:00:00	5.0
26. 6. 7. 오전 1:30:00	5.0
26. 6. 7. 오전 2:00:00	5.0
26. 6. 7. 오전 2:30:00	5.1
26. 6. 7. 오전 3:00:00	5.1
26. 6. 7. 오전 3:30:00	5.1
26. 6. 7. 오전 4:00:00	5.1
26. 6. 7. 오전 4:30:00	5.1
26. 6. 7. 오전 5:00:00	5.1
26. 6. 7. 오전 5:30:00	5.1
26. 6. 7. 오전 6:00:00	5.1
26. 6. 7. 오전 6:30:00	5.3
26. 6. 7. 오전 7:00:00	5.3
26. 6. 7. 오전 7:30:00	5.2
26. 6. 7. 오전 8:00:00	5.2
26. 6. 7. 오전 8:30:00	5.1
26. 6. 7. 오전 9:00:00	5.1
26. 6. 7. 오전 9:30:00	5.1
26. 6. 7. 오전 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 7. 오전 10:30:00	5.1
26. 6. 7. 오전 11:00:00	5.1
26. 6. 7. 오전 11:30:00	5.1
26. 6. 7. 오후 12:00:00	5.1
26. 6. 7. 오후 12:30:00	5.2
26. 6. 7. 오후 1:00:00	5.2
26. 6. 7. 오후 1:30:00	5.1
26. 6. 7. 오후 2:00:00	5.1
26. 6. 7. 오후 2:30:00	5.2
26. 6. 7. 오후 3:00:00	5.2
26. 6. 7. 오후 3:30:00	5.1
26. 6. 7. 오후 4:00:00	5.1
26. 6. 7. 오후 4:30:00	5.1
26. 6. 7. 오후 5:00:00	5.1
26. 6. 7. 오후 5:30:00	5.3
26. 6. 7. 오후 6:00:00	5.3
26. 6. 7. 오후 6:30:00	5.3
26. 6. 7. 오후 7:00:00	5.3
26. 6. 7. 오후 7:30:00	5.3
26. 6. 7. 오후 8:00:00	5.2
26. 6. 7. 오후 8:30:00	5.3
26. 6. 7. 오후 9:00:00	5.4
26. 6. 7. 오후 9:30:00	5.3
26. 6. 7. 오후 10:00:00	5.3
26. 6. 7. 오후 10:30:00	5.3
26. 6. 7. 오후 11:00:00	5.3
26. 6. 7. 오후 11:30:00	5.3
26. 6. 8. 오전 12:00:00	5.2
26. 6. 8. 오전 12:30:00	5.2
26. 6. 8. 오전 1:00:00	5.2
26. 6. 8. 오전 1:30:00	5.2
26. 6. 8. 오전 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 8. 오전 2:30:00	5.2
26. 6. 8. 오전 3:00:00	5.1
26. 6. 8. 오전 3:30:00	5.1
26. 6. 8. 오전 4:00:00	5.1
26. 6. 8. 오전 4:30:00	5.1
26. 6. 8. 오전 5:00:00	5.1
26. 6. 8. 오전 5:30:00	5.1
26. 6. 8. 오전 6:00:00	5.1
26. 6. 8. 오전 6:30:00	5.3
26. 6. 8. 오전 7:00:00	5.4
26. 6. 8. 오전 7:30:00	5.4
26. 6. 8. 오전 8:00:00	5.2
26. 6. 8. 오전 8:30:00	5.2
26. 6. 8. 오전 9:00:00	5.2
26. 6. 8. 오전 9:30:00	5.2
26. 6. 8. 오전 10:00:00	5.2
26. 6. 8. 오전 10:30:00	5.1
26. 6. 8. 오전 11:00:00	5.2
26. 6. 8. 오전 11:30:00	5.2
26. 6. 8. 오후 12:00:00	5.2
26. 6. 8. 오후 12:30:00	5.2
26. 6. 8. 오후 1:00:00	5.3
26. 6. 8. 오후 1:30:00	5.3
26. 6. 8. 오후 2:00:00	5.3
26. 6. 8. 오후 2:30:00	5.3
26. 6. 8. 오후 3:00:00	5.3
26. 6. 8. 오후 3:30:00	5.3
26. 6. 8. 오후 4:00:00	5.3
26. 6. 8. 오후 4:30:00	5.4
26. 6. 8. 오후 5:00:00	5.3
26. 6. 8. 오후 5:30:00	5.3
26. 6. 8. 오후 6:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 8. 오후 6:30:00	5.3
26. 6. 8. 오후 7:00:00	5.3
26. 6. 8. 오후 7:30:00	5.3
26. 6. 8. 오후 8:00:00	5.3
26. 6. 8. 오후 8:30:00	5.3
26. 6. 8. 오후 9:00:00	5.3
26. 6. 8. 오후 9:30:00	5.3
26. 6. 8. 오후 10:00:00	5.3
26. 6. 8. 오후 10:30:00	5.3
26. 6. 8. 오후 11:00:00	5.4
26. 6. 8. 오후 11:30:00	5.4
26. 6. 9. 오전 12:00:00	5.4
26. 6. 9. 오전 12:30:00	5.4
26. 6. 9. 오전 1:00:00	5.3
26. 6. 9. 오전 1:30:00	5.1
26. 6. 9. 오전 2:00:00	5.1
26. 6. 9. 오전 2:30:00	5.1
26. 6. 9. 오전 3:00:00	5.0
26. 6. 9. 오전 3:30:00	5.0
26. 6. 9. 오전 4:00:00	5.1
26. 6. 9. 오전 4:30:00	5.1
26. 6. 9. 오전 5:00:00	5.1
26. 6. 9. 오전 5:30:00	5.0
26. 6. 9. 오전 6:00:00	5.1
26. 6. 9. 오전 6:30:00	5.2
26. 6. 9. 오전 7:00:00	5.2
26. 6. 9. 오전 7:30:00	5.2
26. 6. 9. 오전 8:00:00	5.2
26. 6. 9. 오전 8:30:00	5.2
26. 6. 9. 오전 9:00:00	5.2
26. 6. 9. 오전 9:30:00	5.2
26. 6. 9. 오전 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 9. 오전 10:30:00	5.1
26. 6. 9. 오전 11:00:00	5.1
26. 6. 9. 오전 11:30:00	5.1
26. 6. 9. 오후 12:00:00	5.1
26. 6. 9. 오후 12:30:00	5.1
26. 6. 9. 오후 1:00:00	5.2
26. 6. 9. 오후 1:30:00	5.2
26. 6. 9. 오후 2:00:00	5.2
26. 6. 9. 오후 2:30:00	5.3
26. 6. 9. 오후 3:00:00	5.4
26. 6. 9. 오후 3:30:00	5.4
26. 6. 9. 오후 4:00:00	5.4
26. 6. 9. 오후 4:30:00	5.4
26. 6. 9. 오후 5:00:00	5.3
26. 6. 9. 오후 5:30:00	5.2
26. 6. 9. 오후 6:00:00	5.2
26. 6. 9. 오후 6:30:00	5.2
26. 6. 9. 오후 7:00:00	5.3
26. 6. 9. 오후 7:30:00	5.2
26. 6. 9. 오후 8:00:00	5.1
26. 6. 9. 오후 8:30:00	5.1
26. 6. 9. 오후 9:00:00	5.1
26. 6. 9. 오후 9:30:00	5.2
26. 6. 9. 오후 10:00:00	5.1
26. 6. 9. 오후 10:30:00	5.1
26. 6. 9. 오후 11:00:00	5.1
26. 6. 9. 오후 11:30:00	5.1
26. 6. 10. 오전 12:00:00	5.2
26. 6. 10. 오전 12:30:00	5.2
26. 6. 10. 오전 1:00:00	5.2
26. 6. 10. 오전 1:30:00	5.1
26. 6. 10. 오전 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 10. 오전 2:30:00	5.1
26. 6. 10. 오전 3:00:00	5.1
26. 6. 10. 오전 3:30:00	5.2
26. 6. 10. 오전 4:00:00	5.2
26. 6. 10. 오전 4:30:00	5.2
26. 6. 10. 오전 5:00:00	5.2
26. 6. 10. 오전 5:30:00	5.2
26. 6. 10. 오전 6:00:00	5.2
26. 6. 10. 오전 6:30:00	5.4
26. 6. 10. 오전 7:00:00	5.5
26. 6. 10. 오전 7:30:00	5.5
26. 6. 10. 오전 8:00:00	5.5
26. 6. 10. 오전 8:30:00	5.5
26. 6. 10. 오전 9:00:00	5.4
26. 6. 10. 오전 9:30:00	5.3
26. 6. 10. 오전 10:00:00	5.3
26. 6. 10. 오전 10:30:00	5.3
26. 6. 10. 오전 11:00:00	5.3
26. 6. 10. 오전 11:30:00	5.3
26. 6. 10. 오후 12:00:00	5.3
26. 6. 10. 오후 12:30:00	5.2
26. 6. 10. 오후 1:00:00	5.4
26. 6. 10. 오후 1:30:00	5.3
26. 6. 10. 오후 2:00:00	5.4
26. 6. 10. 오후 2:30:00	5.3
26. 6. 10. 오후 3:00:00	5.3
26. 6. 10. 오후 3:30:00	5.3
26. 6. 10. 오후 4:00:00	5.2
26. 6. 10. 오후 4:30:00	5.2
26. 6. 10. 오후 5:00:00	5.1
26. 6. 10. 오후 5:30:00	5.1
26. 6. 10. 오후 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 10. 오후 6:30:00	5.2
26. 6. 10. 오후 7:00:00	5.2
26. 6. 10. 오후 7:30:00	5.2
26. 6. 10. 오후 8:00:00	5.3
26. 6. 10. 오후 8:30:00	5.2
26. 6. 10. 오후 9:00:00	5.2
26. 6. 10. 오후 9:30:00	5.2
26. 6. 10. 오후 10:00:00	5.2
26. 6. 10. 오후 10:30:00	5.2
26. 6. 10. 오후 11:00:00	5.2
26. 6. 10. 오후 11:30:00	5.2
26. 6. 11. 오전 12:00:00	5.3
26. 6. 11. 오전 12:30:00	5.2
26. 6. 11. 오전 1:00:00	5.0
26. 6. 11. 오전 1:30:00	5.0
26. 6. 11. 오전 2:00:00	5.1
26. 6. 11. 오전 2:30:00	5.1
26. 6. 11. 오전 3:00:00	5.2
26. 6. 11. 오전 3:30:00	5.2
26. 6. 11. 오전 4:00:00	5.1
26. 6. 11. 오전 4:30:00	5.2
26. 6. 11. 오전 5:00:00	5.2
26. 6. 11. 오전 5:30:00	5.2
26. 6. 11. 오전 6:00:00	5.3
26. 6. 11. 오전 6:30:00	5.3
26. 6. 11. 오전 7:00:00	5.5
26. 6. 11. 오전 7:30:00	5.4
26. 6. 11. 오전 8:00:00	5.3
26. 6. 11. 오전 8:30:00	5.2
26. 6. 11. 오전 9:00:00	5.2
26. 6. 11. 오전 9:30:00	5.2
26. 6. 11. 오전 10:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 11. 오전 10:30:00	5.2
26. 6. 11. 오전 11:00:00	5.2
26. 6. 11. 오전 11:30:00	5.1
26. 6. 11. 오후 12:00:00	5.2
26. 6. 11. 오후 12:30:00	5.3
26. 6. 11. 오후 1:00:00	5.4
26. 6. 11. 오후 1:30:00	5.4
26. 6. 11. 오후 2:00:00	5.2
26. 6. 11. 오후 2:30:00	5.2
26. 6. 11. 오후 3:00:00	5.3
26. 6. 11. 오후 3:30:00	5.4
26. 6. 11. 오후 4:00:00	5.3
26. 6. 11. 오후 4:30:00	5.3
26. 6. 11. 오후 5:00:00	5.3
26. 6. 11. 오후 5:30:00	5.3
26. 6. 11. 오후 6:00:00	5.2
26. 6. 11. 오후 6:30:00	5.2
26. 6. 11. 오후 7:00:00	5.3
26. 6. 11. 오후 7:30:00	5.2
26. 6. 11. 오후 8:00:00	5.3
26. 6. 11. 오후 8:30:00	5.3
26. 6. 11. 오후 9:00:00	5.3
26. 6. 11. 오후 9:30:00	5.2
26. 6. 11. 오후 10:00:00	5.3
26. 6. 11. 오후 10:30:00	5.3
26. 6. 11. 오후 11:00:00	5.2
26. 6. 11. 오후 11:30:00	5.3
26. 6. 12. 오전 12:00:00	5.4
26. 6. 12. 오전 12:30:00	5.3
26. 6. 12. 오전 1:00:00	5.3
26. 6. 12. 오전 1:30:00	5.3
26. 6. 12. 오전 2:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 12. 오전 2:30:00	5.3
26. 6. 12. 오전 3:00:00	5.4
26. 6. 12. 오전 3:30:00	5.4
26. 6. 12. 오전 4:00:00	5.4
26. 6. 12. 오전 4:30:00	5.4
26. 6. 12. 오전 5:00:00	5.3
26. 6. 12. 오전 5:30:00	5.4
26. 6. 12. 오전 6:00:00	5.4
26. 6. 12. 오전 6:30:00	5.5
26. 6. 12. 오전 7:00:00	5.5
26. 6. 12. 오전 7:30:00	5.6
26. 6. 12. 오전 8:00:00	5.5
26. 6. 12. 오전 8:30:00	5.4
26. 6. 12. 오전 9:00:00	5.3
26. 6. 12. 오전 9:30:00	5.2
26. 6. 12. 오전 10:00:00	5.3
26. 6. 12. 오전 10:30:00	5.2
26. 6. 12. 오전 11:00:00	5.1
26. 6. 12. 오전 11:30:00	5.0
26. 6. 12. 오후 12:00:00	5.0
26. 6. 12. 오후 12:30:00	5.1
26. 6. 12. 오후 1:00:00	5.3
26. 6. 12. 오후 1:30:00	5.4
26. 6. 12. 오후 2:00:00	5.4
26. 6. 12. 오후 2:30:00	5.3
26. 6. 12. 오후 2:54:23	5.7
26. 6. 12. 오후 3:00:00	5.4
26. 6. 12. 오후 3:30:00	5.3
26. 6. 12. 오후 4:00:00	5.3
26. 6. 12. 오후 4:30:00	5.3
26. 6. 12. 오후 5:00:00	5.3
26. 6. 12. 오후 5:30:00	5.4



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 12. 오후 6:00:00	5.3
26. 6. 12. 오후 6:30:00	5.3
26. 6. 12. 오후 7:00:00	5.2
26. 6. 12. 오후 7:30:00	5.3
26. 6. 12. 오후 8:00:00	5.4
26. 6. 12. 오후 8:30:00	5.4
26. 6. 12. 오후 9:00:00	5.4
26. 6. 12. 오후 9:30:00	5.3
26. 6. 12. 오후 10:00:00	5.2
26. 6. 12. 오후 10:30:00	5.2
26. 6. 12. 오후 11:00:00	5.3
26. 6. 12. 오후 11:30:00	5.3
26. 6. 13. 오전 12:00:00	5.3
26. 6. 13. 오전 12:30:00	5.2
26. 6. 13. 오전 1:00:00	5.1
26. 6. 13. 오전 1:30:00	5.3
26. 6. 13. 오전 2:00:00	5.2
26. 6. 13. 오전 2:30:00	5.3
26. 6. 13. 오전 3:00:00	5.3
26. 6. 13. 오전 3:30:00	5.3
26. 6. 13. 오전 4:00:00	5.4
26. 6. 13. 오전 4:30:00	5.3
26. 6. 13. 오전 5:00:00	5.2
26. 6. 13. 오전 5:30:00	5.3
26. 6. 13. 오전 6:00:00	5.2
26. 6. 13. 오전 6:30:00	5.2
26. 6. 13. 오전 7:00:00	5.2
26. 6. 13. 오전 7:30:00	5.3
26. 6. 13. 오전 8:00:00	5.2
26. 6. 13. 오전 8:30:00	5.1
26. 6. 13. 오전 9:00:00	5.1
26. 6. 13. 오전 9:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 13. 오전 10:00:00	5.2
26. 6. 13. 오전 10:30:00	5.1
26. 6. 13. 오전 11:00:00	5.1
26. 6. 13. 오전 11:30:00	5.2
26. 6. 13. 오후 12:00:00	5.2
26. 6. 13. 오후 12:30:00	5.1
26. 6. 13. 오후 1:00:00	5.2
26. 6. 13. 오후 1:30:00	5.2
26. 6. 13. 오후 2:00:00	5.2
26. 6. 13. 오후 2:30:00	5.2
26. 6. 13. 오후 3:00:00	5.2
26. 6. 13. 오후 3:30:00	5.2
26. 6. 13. 오후 4:00:00	5.3
26. 6. 13. 오후 4:30:00	5.3
26. 6. 13. 오후 5:00:00	5.3
26. 6. 13. 오후 5:30:00	5.3
26. 6. 13. 오후 6:00:00	5.3
26. 6. 13. 오후 6:30:00	5.2
26. 6. 13. 오후 7:00:00	5.2
26. 6. 13. 오후 7:30:00	5.2
26. 6. 13. 오후 8:00:00	5.2
26. 6. 13. 오후 8:30:00	5.2
26. 6. 13. 오후 9:00:00	5.1
26. 6. 13. 오후 9:30:00	5.1
26. 6. 13. 오후 10:00:00	5.1
26. 6. 13. 오후 10:30:00	5.1
26. 6. 13. 오후 11:00:00	5.2
26. 6. 13. 오후 11:30:00	5.2
26. 6. 14. 오전 12:00:00	5.2
26. 6. 14. 오전 12:30:00	5.2
26. 6. 14. 오전 1:00:00	5.2
26. 6. 14. 오전 1:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 14. 오전 2:00:00	5.1
26. 6. 14. 오전 2:30:00	5.1
26. 6. 14. 오전 3:00:00	5.1
26. 6. 14. 오전 3:30:00	5.1
26. 6. 14. 오전 4:00:00	5.1
26. 6. 14. 오전 4:30:00	5.1
26. 6. 14. 오전 5:00:00	5.1
26. 6. 14. 오전 5:30:00	5.2
26. 6. 14. 오전 6:00:00	5.2
26. 6. 14. 오전 6:30:00	5.2
26. 6. 14. 오전 7:00:00	5.3
26. 6. 14. 오전 7:30:00	5.3
26. 6. 14. 오전 8:00:00	5.3
26. 6. 14. 오전 8:30:00	5.2
26. 6. 14. 오전 9:00:00	5.1
26. 6. 14. 오전 9:30:00	5.2
26. 6. 14. 오전 10:00:00	5.2
26. 6. 14. 오전 10:30:00	5.1
26. 6. 14. 오전 11:00:00	5.1
26. 6. 14. 오전 11:30:00	5.0
26. 6. 14. 오후 12:00:00	5.0
26. 6. 14. 오후 12:30:00	5.0
26. 6. 14. 오후 1:00:00	5.0
26. 6. 14. 오후 1:30:00	5.2
26. 6. 14. 오후 2:00:00	5.2
26. 6. 14. 오후 2:30:00	5.1
26. 6. 14. 오후 3:00:00	5.1
26. 6. 14. 오후 3:30:00	5.0
26. 6. 14. 오후 4:00:00	4.9
26. 6. 14. 오후 4:30:00	5.0
26. 6. 14. 오후 5:00:00	5.0
26. 6. 14. 오후 5:30:00	5.0



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 14. 오후 6:00:00	5.0
26. 6. 14. 오후 6:30:00	5.0
26. 6. 14. 오후 7:00:00	5.0
26. 6. 14. 오후 7:30:00	4.9
26. 6. 14. 오후 8:00:00	5.0
26. 6. 14. 오후 8:30:00	5.0
26. 6. 14. 오후 9:00:00	5.1
26. 6. 14. 오후 9:30:00	5.1
26. 6. 14. 오후 10:00:00	5.1
26. 6. 14. 오후 10:30:00	5.0
26. 6. 14. 오후 11:00:00	5.0
26. 6. 14. 오후 11:30:00	5.0
26. 6. 15. 오전 12:00:00	5.1
26. 6. 15. 오전 12:30:00	5.1
26. 6. 15. 오전 1:00:00	4.9
26. 6. 15. 오전 1:30:00	5.2
26. 6. 15. 오전 2:00:00	5.0
26. 6. 15. 오전 2:30:00	5.1
26. 6. 15. 오전 3:00:00	5.1
26. 6. 15. 오전 3:30:00	5.0
26. 6. 15. 오전 4:00:00	5.0
26. 6. 15. 오전 4:30:00	5.1
26. 6. 15. 오전 5:00:00	5.0
26. 6. 15. 오전 5:30:00	5.0
26. 6. 15. 오전 6:00:00	5.0
26. 6. 15. 오전 6:30:00	5.0
26. 6. 15. 오전 7:00:00	5.0
26. 6. 15. 오전 7:30:00	5.1
26. 6. 15. 오전 8:00:00	5.2
26. 6. 15. 오전 8:30:00	5.2
26. 6. 15. 오전 9:00:00	5.1
26. 6. 15. 오전 9:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 15. 오전 10:00:00	5.1
26. 6. 15. 오전 10:30:00	5.1
26. 6. 15. 오전 11:00:00	5.1
26. 6. 15. 오전 11:30:00	5.0
26. 6. 15. 오후 12:00:00	5.1
26. 6. 15. 오후 12:30:00	5.1
26. 6. 15. 오후 1:00:00	5.1
26. 6. 15. 오후 1:30:00	5.1
26. 6. 15. 오후 2:00:00	5.2
26. 6. 15. 오후 2:30:00	5.3
26. 6. 15. 오후 3:00:00	5.2
26. 6. 15. 오후 3:30:00	5.1
26. 6. 15. 오후 4:00:00	5.3
26. 6. 15. 오후 4:30:00	5.4
26. 6. 15. 오후 5:00:00	5.3
26. 6. 15. 오후 5:30:00	5.1
26. 6. 15. 오후 6:00:00	5.2
26. 6. 15. 오후 6:30:00	5.2
26. 6. 15. 오후 7:00:00	5.2
26. 6. 15. 오후 7:30:00	5.2
26. 6. 15. 오후 8:00:00	5.3
26. 6. 15. 오후 8:30:00	5.3
26. 6. 15. 오후 9:00:00	5.3
26. 6. 15. 오후 9:30:00	5.2
26. 6. 15. 오후 10:00:00	5.2
26. 6. 15. 오후 10:30:00	5.2
26. 6. 15. 오후 11:00:00	5.1
26. 6. 15. 오후 11:30:00	5.2
26. 6. 16. 오전 12:00:00	5.2
26. 6. 16. 오전 12:30:00	5.2
26. 6. 16. 오전 1:00:00	5.1
26. 6. 16. 오전 1:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 16. 오전 2:00:00	5.3
26. 6. 16. 오전 2:30:00	5.4
26. 6. 16. 오전 3:00:00	5.3
26. 6. 16. 오전 3:30:00	5.3
26. 6. 16. 오전 4:00:00	5.3
26. 6. 16. 오전 4:30:00	5.3
26. 6. 16. 오전 5:00:00	5.2
26. 6. 16. 오전 5:30:00	5.1
26. 6. 16. 오전 6:00:00	5.1
26. 6. 16. 오전 6:30:00	5.2
26. 6. 16. 오전 7:00:00	5.3
26. 6. 16. 오전 7:30:00	5.2
26. 6. 16. 오전 8:00:00	5.1
26. 6. 16. 오전 8:30:00	5.1
26. 6. 16. 오전 9:00:00	5.1
26. 6. 16. 오전 9:30:00	5.1
26. 6. 16. 오전 10:00:00	5.1
26. 6. 16. 오전 10:30:00	5.1
26. 6. 16. 오전 11:00:00	5.1
26. 6. 16. 오전 11:30:00	5.1
26. 6. 16. 오후 12:00:00	5.3
26. 6. 16. 오후 12:30:00	5.3
26. 6. 16. 오후 1:00:00	5.2
26. 6. 16. 오후 1:30:00	5.2
26. 6. 16. 오후 2:00:00	5.3
26. 6. 16. 오후 2:30:00	5.3
26. 6. 16. 오후 3:00:00	5.3
26. 6. 16. 오후 3:30:00	5.3
26. 6. 16. 오후 4:00:00	5.2
26. 6. 16. 오후 4:30:00	5.2
26. 6. 16. 오후 5:00:00	5.1
26. 6. 16. 오후 5:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 16. 오후 6:00:00	5.2
26. 6. 16. 오후 6:30:00	5.2
26. 6. 16. 오후 7:00:00	5.2
26. 6. 16. 오후 7:30:00	5.1
26. 6. 16. 오후 8:00:00	5.2
26. 6. 16. 오후 8:30:00	5.2
26. 6. 16. 오후 9:00:00	5.2
26. 6. 16. 오후 9:30:00	5.2
26. 6. 16. 오후 10:00:00	5.2
26. 6. 16. 오후 10:30:00	5.2
26. 6. 16. 오후 11:00:00	5.1
26. 6. 16. 오후 11:30:00	5.1
26. 6. 17. 오전 12:00:00	5.2
26. 6. 17. 오전 12:30:00	5.1
26. 6. 17. 오전 1:00:00	5.0
26. 6. 17. 오전 1:30:00	5.2
26. 6. 17. 오전 2:00:00	5.2
26. 6. 17. 오전 2:30:00	5.2
26. 6. 17. 오전 3:00:00	5.2
26. 6. 17. 오전 3:30:00	5.2
26. 6. 17. 오전 4:00:00	5.3
26. 6. 17. 오전 4:30:00	5.2
26. 6. 17. 오전 5:00:00	5.3
26. 6. 17. 오전 5:30:00	5.2
26. 6. 17. 오전 6:00:00	5.2
26. 6. 17. 오전 6:30:00	5.2
26. 6. 17. 오전 7:00:00	5.3
26. 6. 17. 오전 7:30:00	5.3
26. 6. 17. 오전 8:00:00	5.3
26. 6. 17. 오전 8:30:00	5.2
26. 6. 17. 오전 9:00:00	5.2
26. 6. 17. 오전 9:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 17. 오전 10:00:00	5.2
26. 6. 17. 오전 10:30:00	5.3
26. 6. 17. 오전 11:00:00	5.3
26. 6. 17. 오전 11:30:00	5.3
26. 6. 17. 오후 12:00:00	5.3
26. 6. 17. 오후 12:30:00	5.4
26. 6. 17. 오후 1:00:00	5.4
26. 6. 17. 오후 1:30:00	5.3
26. 6. 17. 오후 2:00:00	5.4
26. 6. 17. 오후 2:30:00	5.4
26. 6. 17. 오후 3:00:00	5.3
26. 6. 17. 오후 3:30:00	5.4
26. 6. 17. 오후 4:00:00	5.4
26. 6. 17. 오후 4:30:00	5.4
26. 6. 17. 오후 5:00:00	5.3
26. 6. 17. 오후 5:30:00	5.3
26. 6. 17. 오후 6:00:00	5.3
26. 6. 17. 오후 6:30:00	5.2
26. 6. 17. 오후 7:00:00	5.2
26. 6. 17. 오후 7:30:00	5.2
26. 6. 17. 오후 8:00:00	5.3
26. 6. 17. 오후 8:30:00	5.4
26. 6. 17. 오후 9:00:00	5.4
26. 6. 17. 오후 9:30:00	5.4
26. 6. 17. 오후 10:00:00	5.3
26. 6. 17. 오후 10:30:00	5.4
26. 6. 17. 오후 11:00:00	5.3
26. 6. 17. 오후 11:30:00	5.2
26. 6. 18. 오전 12:00:00	5.2
26. 6. 18. 오전 12:30:00	5.2
26. 6. 18. 오전 1:00:00	5.2
26. 6. 18. 오전 1:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 18. 오전 2:00:00	5.4
26. 6. 18. 오전 2:30:00	5.3
26. 6. 18. 오전 3:00:00	5.2
26. 6. 18. 오전 3:30:00	5.3
26. 6. 18. 오전 4:00:00	5.2
26. 6. 18. 오전 4:30:00	5.2
26. 6. 18. 오전 5:00:00	5.3
26. 6. 18. 오전 5:30:00	5.3
26. 6. 18. 오전 6:00:00	5.4
26. 6. 18. 오전 6:30:00	5.4
26. 6. 18. 오전 7:00:00	5.3
26. 6. 18. 오전 7:30:00	5.3
26. 6. 18. 오전 8:00:00	5.3
26. 6. 18. 오전 8:30:00	5.3
26. 6. 18. 오전 9:00:00	5.3
26. 6. 18. 오전 9:30:00	5.4
26. 6. 18. 오전 10:00:00	5.4
26. 6. 18. 오전 10:30:00	5.4
26. 6. 18. 오전 11:00:00	5.4
26. 6. 18. 오전 11:30:00	5.3
26. 6. 18. 오후 12:00:00	5.2
26. 6. 18. 오후 12:30:00	5.3
26. 6. 18. 오후 1:00:00	5.2
26. 6. 18. 오후 1:30:00	5.1
26. 6. 18. 오후 2:00:00	5.3
26. 6. 18. 오후 2:30:00	5.3
26. 6. 18. 오후 3:00:00	5.2
26. 6. 18. 오후 3:30:00	5.2
26. 6. 18. 오후 4:00:00	5.1
26. 6. 18. 오후 4:30:00	5.2
26. 6. 18. 오후 5:00:00	5.2
26. 6. 18. 오후 5:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 18. 오후 6:00:00	5.1
26. 6. 18. 오후 6:30:00	5.1
26. 6. 18. 오후 7:00:00	5.2
26. 6. 18. 오후 7:30:00	5.2
26. 6. 18. 오후 8:00:00	5.2
26. 6. 18. 오후 8:30:00	5.3
26. 6. 18. 오후 9:00:00	5.4
26. 6. 18. 오후 9:30:00	5.3
26. 6. 18. 오후 10:00:00	5.3
26. 6. 18. 오후 10:30:00	5.3
26. 6. 18. 오후 11:00:00	5.3
26. 6. 18. 오후 11:30:00	5.2
26. 6. 19. 오전 12:00:00	5.2
26. 6. 19. 오전 12:30:00	5.2
26. 6. 19. 오전 1:00:00	5.3
26. 6. 19. 오전 1:30:00	5.3
26. 6. 19. 오전 2:00:00	5.2
26. 6. 19. 오전 2:30:00	5.3
26. 6. 19. 오전 3:00:00	5.3
26. 6. 19. 오전 3:30:00	5.3
26. 6. 19. 오전 4:00:00	5.4
26. 6. 19. 오전 4:30:00	5.3
26. 6. 19. 오전 5:00:00	5.3
26. 6. 19. 오전 5:30:00	5.3
26. 6. 19. 오전 6:00:00	5.4
26. 6. 19. 오전 6:30:00	5.4
26. 6. 19. 오전 7:00:00	5.2
26. 6. 19. 오전 7:30:00	5.2
26. 6. 19. 오전 8:00:00	5.2
26. 6. 19. 오전 8:30:00	5.2
26. 6. 19. 오전 9:00:00	5.2
26. 6. 19. 오전 9:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 19. 오전 10:00:00	5.3
26. 6. 19. 오전 10:30:00	5.2
26. 6. 19. 오전 11:00:00	5.2
26. 6. 19. 오전 11:30:00	5.2
26. 6. 19. 오후 12:00:00	5.2
26. 6. 19. 오후 12:30:00	5.3
26. 6. 19. 오후 1:00:00	5.3
26. 6. 19. 오후 1:30:00	5.4
26. 6. 19. 오후 2:00:00	5.5
26. 6. 19. 오후 2:30:00	5.5
26. 6. 19. 오후 3:00:00	5.4
26. 6. 19. 오후 3:30:00	5.4
26. 6. 19. 오후 4:00:00	5.4
26. 6. 19. 오후 4:30:00	5.4
26. 6. 19. 오후 5:00:00	5.3
26. 6. 19. 오후 5:30:00	5.3
26. 6. 19. 오후 6:00:00	5.3
26. 6. 19. 오후 6:30:00	5.2
26. 6. 19. 오후 7:00:00	5.2
26. 6. 19. 오후 7:30:00	5.3
26. 6. 19. 오후 8:00:00	5.2
26. 6. 19. 오후 8:30:00	5.3
26. 6. 19. 오후 9:00:00	5.2
26. 6. 19. 오후 9:30:00	5.2
26. 6. 19. 오후 10:00:00	5.2
26. 6. 19. 오후 10:30:00	5.2
26. 6. 19. 오후 11:00:00	5.2
26. 6. 19. 오후 11:30:00	5.2
26. 6. 20. 오전 12:00:00	5.2
26. 6. 20. 오전 12:30:00	5.1
26. 6. 20. 오전 1:00:00	5.2
26. 6. 20. 오전 1:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 20. 오전 2:00:00	5.2
26. 6. 20. 오전 2:30:00	5.4
26. 6. 20. 오전 3:00:00	5.4
26. 6. 20. 오전 3:30:00	5.3
26. 6. 20. 오전 4:00:00	5.3
26. 6. 20. 오전 4:30:00	5.3
26. 6. 20. 오전 5:00:00	5.3
26. 6. 20. 오전 5:30:00	5.3
26. 6. 20. 오전 6:00:00	5.3
26. 6. 20. 오전 6:30:00	5.2
26. 6. 20. 오전 7:00:00	5.2
26. 6. 20. 오전 7:30:00	5.3
26. 6. 20. 오전 8:00:00	5.3
26. 6. 20. 오전 8:30:00	5.2
26. 6. 20. 오전 9:00:00	5.2
26. 6. 20. 오전 9:30:00	5.2
26. 6. 20. 오전 10:00:00	5.2
26. 6. 20. 오전 10:30:00	5.1
26. 6. 20. 오전 11:00:00	5.1
26. 6. 20. 오전 11:30:00	5.2
26. 6. 20. 오후 12:00:00	5.2
26. 6. 20. 오후 12:30:00	5.1
26. 6. 20. 오후 1:00:00	5.2
26. 6. 20. 오후 1:30:00	5.2
26. 6. 20. 오후 2:00:00	5.2
26. 6. 20. 오후 2:30:00	5.3
26. 6. 20. 오후 3:00:00	5.2
26. 6. 20. 오후 3:30:00	5.2
26. 6. 20. 오후 4:00:00	5.2
26. 6. 20. 오후 4:30:00	5.2
26. 6. 20. 오후 5:00:00	5.4
26. 6. 20. 오후 5:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 20. 오후 6:00:00	5.3
26. 6. 20. 오후 6:30:00	5.4
26. 6. 20. 오후 7:00:00	5.3
26. 6. 20. 오후 7:30:00	5.3
26. 6. 20. 오후 8:00:00	5.3
26. 6. 20. 오후 8:30:00	5.4
26. 6. 20. 오후 9:00:00	5.2
26. 6. 20. 오후 9:30:00	5.2
26. 6. 20. 오후 10:00:00	5.2
26. 6. 20. 오후 10:30:00	5.1
26. 6. 20. 오후 11:00:00	5.1
26. 6. 20. 오후 11:30:00	5.1
26. 6. 21. 오전 12:00:00	5.2
26. 6. 21. 오전 12:30:00	5.1
26. 6. 21. 오전 1:00:00	5.1
26. 6. 21. 오전 1:30:00	5.2
26. 6. 21. 오전 2:00:00	5.1
26. 6. 21. 오전 2:30:00	5.1
26. 6. 21. 오전 3:00:00	5.2
26. 6. 21. 오전 3:30:00	5.2
26. 6. 21. 오전 4:00:00	5.2
26. 6. 21. 오전 4:30:00	5.2
26. 6. 21. 오전 5:00:00	5.2
26. 6. 21. 오전 5:30:00	5.1
26. 6. 21. 오전 6:00:00	5.2
26. 6. 21. 오전 6:30:00	5.3
26. 6. 21. 오전 7:00:00	5.2
26. 6. 21. 오전 7:30:00	5.2
26. 6. 21. 오전 8:00:00	5.2
26. 6. 21. 오전 8:30:00	5.2
26. 6. 21. 오전 9:00:00	5.2
26. 6. 21. 오전 9:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 21. 오전 10:00:00	5.2
26. 6. 21. 오전 10:30:00	5.2
26. 6. 21. 오전 11:00:00	5.1
26. 6. 21. 오전 11:30:00	5.1
26. 6. 21. 오후 12:00:00	5.2
26. 6. 21. 오후 12:30:00	5.1
26. 6. 21. 오후 1:00:00	5.1
26. 6. 21. 오후 1:30:00	5.1
26. 6. 21. 오후 2:00:00	5.2
26. 6. 21. 오후 2:30:00	5.2
26. 6. 21. 오후 3:00:00	5.2
26. 6. 21. 오후 3:30:00	5.3
26. 6. 21. 오후 4:00:00	5.3
26. 6. 21. 오후 4:30:00	5.3
26. 6. 21. 오후 5:00:00	5.2
26. 6. 21. 오후 5:30:00	5.2
26. 6. 21. 오후 6:00:00	5.2
26. 6. 21. 오후 6:30:00	5.2
26. 6. 21. 오후 7:00:00	5.2
26. 6. 21. 오후 7:30:00	5.3
26. 6. 21. 오후 8:00:00	5.3
26. 6. 21. 오후 8:30:00	5.2
26. 6. 21. 오후 9:00:00	5.2
26. 6. 21. 오후 9:30:00	5.2
26. 6. 21. 오후 10:00:00	5.2
26. 6. 21. 오후 10:30:00	5.1
26. 6. 21. 오후 11:00:00	5.1
26. 6. 21. 오후 11:30:00	5.2
26. 6. 22. 오전 12:00:00	5.2
26. 6. 22. 오전 12:30:00	5.1
26. 6. 22. 오전 1:00:00	5.2
26. 6. 22. 오전 1:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 22. 오전 2:00:00	5.1
26. 6. 22. 오전 2:30:00	5.1
26. 6. 22. 오전 3:00:00	5.3
26. 6. 22. 오전 3:30:00	5.3
26. 6. 22. 오전 4:00:00	5.2
26. 6. 22. 오전 4:30:00	5.4
26. 6. 22. 오전 5:00:00	5.4
26. 6. 22. 오전 5:30:00	5.4
26. 6. 22. 오전 6:00:00	5.3
26. 6. 22. 오전 6:30:00	5.4
26. 6. 22. 오전 7:00:00	5.4
26. 6. 22. 오전 7:30:00	5.3
26. 6. 22. 오전 8:00:00	5.2
26. 6. 22. 오전 8:30:00	5.2
26. 6. 22. 오전 9:00:00	5.2
26. 6. 22. 오전 9:30:00	5.2
26. 6. 22. 오전 10:00:00	5.3
26. 6. 22. 오전 10:30:00	5.3
26. 6. 22. 오전 11:00:00	5.3
26. 6. 22. 오전 11:30:00	5.3
26. 6. 22. 오후 12:00:00	5.3
26. 6. 22. 오후 12:30:00	5.3
26. 6. 22. 오후 1:00:00	5.4
26. 6. 22. 오후 1:30:00	5.3
26. 6. 22. 오후 2:00:00	5.3
26. 6. 22. 오후 2:30:00	5.3
26. 6. 22. 오후 3:00:00	5.3
26. 6. 22. 오후 3:30:00	5.3
26. 6. 22. 오후 4:00:00	5.3
26. 6. 22. 오후 4:30:00	5.3
26. 6. 22. 오후 5:00:00	5.2
26. 6. 22. 오후 5:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 22. 오후 6:00:00	5.2
26. 6. 22. 오후 6:30:00	5.2
26. 6. 22. 오후 7:00:00	5.2
26. 6. 22. 오후 7:30:00	5.2
26. 6. 22. 오후 8:00:00	5.3
26. 6. 22. 오후 8:30:00	5.2
26. 6. 22. 오후 9:00:00	5.3
26. 6. 22. 오후 9:30:00	5.2
26. 6. 22. 오후 10:00:00	5.2
26. 6. 22. 오후 10:30:00	5.3
26. 6. 22. 오후 11:00:00	5.3
26. 6. 22. 오후 11:30:00	5.3
26. 6. 23. 오전 12:00:00	5.4
26. 6. 23. 오전 12:30:00	5.4
26. 6. 23. 오전 1:00:00	5.3
26. 6. 23. 오전 1:30:00	5.3
26. 6. 23. 오전 2:00:00	5.4
26. 6. 23. 오전 2:30:00	5.4
26. 6. 23. 오전 3:00:00	5.4
26. 6. 23. 오전 3:30:00	5.4
26. 6. 23. 오전 4:00:00	5.4
26. 6. 23. 오전 4:30:00	5.4
26. 6. 23. 오전 5:00:00	5.2
26. 6. 23. 오전 5:30:00	5.3
26. 6. 23. 오전 6:00:00	5.4
26. 6. 23. 오전 6:30:00	5.5
26. 6. 23. 오전 7:00:00	5.5
26. 6. 23. 오전 7:30:00	5.4
26. 6. 23. 오전 8:00:00	5.4
26. 6. 23. 오전 8:30:00	5.4
26. 6. 23. 오전 9:00:00	5.7
26. 6. 23. 오전 9:30:00	5.4



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 23. 오전 10:00:00	5.3
26. 6. 23. 오전 10:30:00	5.4
26. 6. 23. 오전 11:00:00	5.4
26. 6. 23. 오전 11:30:00	5.4
26. 6. 23. 오후 12:00:00	5.5
26. 6. 23. 오후 12:30:00	5.4
26. 6. 23. 오후 1:00:00	5.3
26. 6. 23. 오후 1:30:00	5.3
26. 6. 23. 오후 2:00:00	5.3
26. 6. 23. 오후 2:30:00	5.3
26. 6. 23. 오후 3:00:00	5.3
26. 6. 23. 오후 3:30:00	5.4
26. 6. 23. 오후 4:00:00	5.4
26. 6. 23. 오후 4:30:00	5.3
26. 6. 23. 오후 5:00:00	5.3
26. 6. 23. 오후 5:30:00	5.3
26. 6. 23. 오후 6:00:00	5.2
26. 6. 23. 오후 6:30:00	5.3
26. 6. 23. 오후 7:00:00	5.3
26. 6. 23. 오후 7:30:00	5.2
26. 6. 23. 오후 8:00:00	5.2
26. 6. 23. 오후 8:30:00	5.2
26. 6. 23. 오후 9:00:00	5.3
26. 6. 23. 오후 9:30:00	5.3
26. 6. 23. 오후 10:00:00	5.3
26. 6. 23. 오후 10:30:00	5.3
26. 6. 23. 오후 11:00:00	5.4
26. 6. 23. 오후 11:30:00	5.2
26. 6. 24. 오전 12:00:00	5.3
26. 6. 24. 오전 12:30:00	5.2
26. 6. 24. 오전 1:00:00	5.2
26. 6. 24. 오전 1:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 24. 오전 2:00:00	5.3
26. 6. 24. 오전 2:30:00	5.4
26. 6. 24. 오전 3:00:00	5.3
26. 6. 24. 오전 3:30:00	5.4
26. 6. 24. 오전 4:00:00	5.4
26. 6. 24. 오전 4:30:00	5.3
26. 6. 24. 오전 5:00:00	5.2
26. 6. 24. 오전 5:30:00	5.3
26. 6. 24. 오전 6:00:00	5.3
26. 6. 24. 오전 6:30:00	5.3
26. 6. 24. 오전 7:00:00	5.3
26. 6. 24. 오전 7:30:00	5.3
26. 6. 24. 오전 8:00:00	5.2
26. 6. 24. 오전 8:30:00	5.2
26. 6. 24. 오전 9:00:00	5.2
26. 6. 24. 오전 9:30:00	5.3
26. 6. 24. 오전 10:00:00	5.3
26. 6. 24. 오전 10:30:00	5.3
26. 6. 24. 오전 11:00:00	5.3
26. 6. 24. 오전 11:30:00	5.3
26. 6. 24. 오후 12:00:00	5.3
26. 6. 24. 오후 12:30:00	5.3
26. 6. 24. 오후 1:00:00	5.2
26. 6. 24. 오후 1:30:00	5.2
26. 6. 24. 오후 2:00:00	5.2
26. 6. 24. 오후 2:30:00	5.2
26. 6. 24. 오후 3:00:00	5.3
26. 6. 24. 오후 3:30:00	5.3
26. 6. 24. 오후 4:00:00	5.2
26. 6. 24. 오후 4:30:00	5.2
26. 6. 24. 오후 5:00:00	5.2
26. 6. 24. 오후 5:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 24. 오후 6:00:00	5.3
26. 6. 24. 오후 6:30:00	5.2
26. 6. 24. 오후 7:00:00	5.1
26. 6. 24. 오후 7:30:00	5.2
26. 6. 24. 오후 8:00:00	5.2
26. 6. 24. 오후 8:30:00	5.1
26. 6. 24. 오후 9:00:00	5.2
26. 6. 24. 오후 9:30:00	5.3
26. 6. 24. 오후 10:00:00	5.2
26. 6. 24. 오후 10:30:00	5.2
26. 6. 24. 오후 11:00:00	5.2
26. 6. 24. 오후 11:30:00	5.3
26. 6. 25. 오전 12:00:00	5.4
26. 6. 25. 오전 12:30:00	5.2
26. 6. 25. 오전 1:00:00	5.3
26. 6. 25. 오전 1:30:00	5.3
26. 6. 25. 오전 2:00:00	5.3
26. 6. 25. 오전 2:30:00	5.3
26. 6. 25. 오전 3:00:00	5.4
26. 6. 25. 오전 3:30:00	5.5
26. 6. 25. 오전 4:00:00	5.3
26. 6. 25. 오전 4:30:00	5.2
26. 6. 25. 오전 5:00:00	5.3
26. 6. 25. 오전 5:30:00	5.3
26. 6. 25. 오전 6:00:00	5.2
26. 6. 25. 오전 6:30:00	5.3
26. 6. 25. 오전 7:00:00	5.3
26. 6. 25. 오전 7:30:00	5.3
26. 6. 25. 오전 8:00:00	5.3
26. 6. 25. 오전 8:30:00	5.3
26. 6. 25. 오전 9:00:00	5.3
26. 6. 25. 오전 9:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 25. 오전 10:00:00	5.3
26. 6. 25. 오전 10:30:00	5.2
26. 6. 25. 오전 11:00:00	5.2
26. 6. 25. 오전 11:30:00	5.2
26. 6. 25. 오후 12:00:00	5.2
26. 6. 25. 오후 12:30:00	5.1
26. 6. 25. 오후 1:00:00	5.2
26. 6. 25. 오후 1:30:00	5.2
26. 6. 25. 오후 2:00:00	5.2
26. 6. 25. 오후 2:30:00	5.1
26. 6. 25. 오후 3:00:00	5.2
26. 6. 25. 오후 3:30:00	5.2
26. 6. 25. 오후 4:00:00	5.1
26. 6. 25. 오후 4:30:00	5.1
26. 6. 25. 오후 5:00:00	5.1
26. 6. 25. 오후 5:30:00	5.1
26. 6. 25. 오후 6:00:00	5.2
26. 6. 25. 오후 6:30:00	5.2
26. 6. 25. 오후 7:00:00	5.2
26. 6. 25. 오후 7:30:00	5.1
26. 6. 25. 오후 8:00:00	5.2
26. 6. 25. 오후 8:30:00	5.2
26. 6. 25. 오후 9:00:00	5.2
26. 6. 25. 오후 9:30:00	5.3
26. 6. 25. 오후 10:00:00	5.3
26. 6. 25. 오후 10:30:00	5.2
26. 6. 25. 오후 11:00:00	5.3
26. 6. 25. 오후 11:30:00	5.2
26. 6. 26. 오전 12:00:00	5.2
26. 6. 26. 오전 12:30:00	5.2
26. 6. 26. 오전 1:00:00	5.2
26. 6. 26. 오전 1:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 26. 오전 2:00:00	5.2
26. 6. 26. 오전 2:30:00	5.2
26. 6. 26. 오전 3:00:00	5.2
26. 6. 26. 오전 3:30:00	5.4
26. 6. 26. 오전 4:00:00	5.3
26. 6. 26. 오전 4:30:00	5.3
26. 6. 26. 오전 5:00:00	5.3
26. 6. 26. 오전 5:30:00	5.3
26. 6. 26. 오전 6:00:00	5.3
26. 6. 26. 오전 6:30:00	5.3
26. 6. 26. 오전 7:00:00	5.4
26. 6. 26. 오전 7:30:00	5.3
26. 6. 26. 오전 8:00:00	5.3
26. 6. 26. 오전 8:30:00	5.3
26. 6. 26. 오전 9:00:00	5.2
26. 6. 26. 오전 9:30:00	5.4
26. 6. 26. 오전 10:00:00	5.3
26. 6. 26. 오전 10:30:00	5.2
26. 6. 26. 오전 11:00:00	5.2
26. 6. 26. 오전 11:30:00	5.2
26. 6. 26. 오후 12:00:00	5.2
26. 6. 26. 오후 12:30:00	5.1
26. 6. 26. 오후 1:00:00	5.1
26. 6. 26. 오후 1:30:00	5.1
26. 6. 26. 오후 2:00:00	5.1
26. 6. 26. 오후 2:30:00	5.1
26. 6. 26. 오후 3:00:00	5.1
26. 6. 26. 오후 3:30:00	5.2
26. 6. 26. 오후 4:00:00	5.2
26. 6. 26. 오후 4:30:00	5.3
26. 6. 26. 오후 5:00:00	5.3
26. 6. 26. 오후 5:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 26. 오후 6:00:00	5.3
26. 6. 26. 오후 6:30:00	5.4
26. 6. 26. 오후 7:00:00	5.3
26. 6. 26. 오후 7:30:00	5.2
26. 6. 26. 오후 8:00:00	5.2
26. 6. 26. 오후 8:30:00	5.3
26. 6. 26. 오후 9:00:00	5.2
26. 6. 26. 오후 9:30:00	5.2
26. 6. 26. 오후 10:00:00	5.3
26. 6. 26. 오후 10:30:00	5.3
26. 6. 26. 오후 11:00:00	5.2
26. 6. 26. 오후 11:30:00	5.3
26. 6. 27. 오전 12:00:00	5.3
26. 6. 27. 오전 12:30:00	5.3
26. 6. 27. 오전 1:00:00	5.3
26. 6. 27. 오전 1:30:00	5.3
26. 6. 27. 오전 2:00:00	5.3
26. 6. 27. 오전 2:30:00	5.3
26. 6. 27. 오전 3:00:00	5.3
26. 6. 27. 오전 3:30:00	5.3
26. 6. 27. 오전 4:00:00	5.5
26. 6. 27. 오전 4:30:00	5.4
26. 6. 27. 오전 5:00:00	5.3
26. 6. 27. 오전 5:30:00	5.3
26. 6. 27. 오전 6:00:00	5.3
26. 6. 27. 오전 6:30:00	5.3
26. 6. 27. 오전 7:00:00	5.1
26. 6. 27. 오전 7:30:00	5.2
26. 6. 27. 오전 8:00:00	5.2
26. 6. 27. 오전 8:30:00	5.3
26. 6. 27. 오전 9:00:00	5.3
26. 6. 27. 오전 9:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 27. 오전 10:00:00	5.4
26. 6. 27. 오전 10:30:00	5.4
26. 6. 27. 오전 11:00:00	5.3
26. 6. 27. 오전 11:30:00	5.2
26. 6. 27. 오후 12:00:00	5.3
26. 6. 27. 오후 12:30:00	5.3
26. 6. 27. 오후 1:00:00	5.2
26. 6. 27. 오후 1:30:00	5.1
26. 6. 27. 오후 2:00:00	5.2
26. 6. 27. 오후 2:30:00	5.2
26. 6. 27. 오후 3:00:00	5.1
26. 6. 27. 오후 3:30:00	5.1
26. 6. 27. 오후 4:00:00	5.2
26. 6. 27. 오후 4:30:00	5.2
26. 6. 27. 오후 5:00:00	5.3
26. 6. 27. 오후 5:30:00	5.2
26. 6. 27. 오후 6:00:00	5.2
26. 6. 27. 오후 6:30:00	5.2
26. 6. 27. 오후 7:00:00	5.3
26. 6. 27. 오후 7:30:00	5.3
26. 6. 27. 오후 8:00:00	5.3
26. 6. 27. 오후 8:30:00	5.2
26. 6. 27. 오후 9:00:00	5.3
26. 6. 27. 오후 9:30:00	5.3
26. 6. 27. 오후 10:00:00	5.3
26. 6. 27. 오후 10:30:00	5.3
26. 6. 27. 오후 11:00:00	5.3
26. 6. 27. 오후 11:30:00	5.2
26. 6. 28. 오전 12:00:00	5.2
26. 6. 28. 오전 12:30:00	5.2
26. 6. 28. 오전 1:00:00	5.1
26. 6. 28. 오전 1:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 28. 오전 2:00:00	5.2
26. 6. 28. 오전 2:30:00	5.2
26. 6. 28. 오전 3:00:00	5.2
26. 6. 28. 오전 3:30:00	5.1
26. 6. 28. 오전 4:00:00	5.2
26. 6. 28. 오전 4:30:00	5.2
26. 6. 28. 오전 5:00:00	5.2
26. 6. 28. 오전 5:30:00	5.2
26. 6. 28. 오전 6:00:00	5.2
26. 6. 28. 오전 6:30:00	5.3
26. 6. 28. 오전 7:00:00	5.2
26. 6. 28. 오전 7:30:00	5.2
26. 6. 28. 오전 8:00:00	5.2
26. 6. 28. 오전 8:30:00	5.3
26. 6. 28. 오전 9:00:00	5.1
26. 6. 28. 오전 9:30:00	5.2
26. 6. 28. 오전 10:00:00	5.2
26. 6. 28. 오전 10:30:00	5.2
26. 6. 28. 오전 11:00:00	5.3
26. 6. 28. 오전 11:30:00	5.3
26. 6. 28. 오후 12:00:00	5.3
26. 6. 28. 오후 12:30:00	5.3
26. 6. 28. 오후 1:00:00	5.4
26. 6. 28. 오후 1:30:00	5.3
26. 6. 28. 오후 2:00:00	5.3
26. 6. 28. 오후 2:30:00	5.3
26. 6. 28. 오후 3:00:00	5.2
26. 6. 28. 오후 3:30:00	5.2
26. 6. 28. 오후 4:00:00	5.4
26. 6. 28. 오후 4:30:00	5.4
26. 6. 28. 오후 5:00:00	5.4
26. 6. 28. 오후 5:30:00	5.3



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 28. 오후 6:00:00	5.3
26. 6. 28. 오후 6:30:00	5.3
26. 6. 28. 오후 7:00:00	5.3
26. 6. 28. 오후 7:30:00	5.2
26. 6. 28. 오후 8:00:00	5.2
26. 6. 28. 오후 8:30:00	5.3
26. 6. 28. 오후 9:00:00	5.2
26. 6. 28. 오후 9:30:00	5.1
26. 6. 28. 오후 10:00:00	5.2
26. 6. 28. 오후 10:30:00	5.2
26. 6. 28. 오후 11:00:00	5.2
26. 6. 28. 오후 11:30:00	5.1
26. 6. 29. 오전 12:00:00	5.2
26. 6. 29. 오전 12:30:00	5.2
26. 6. 29. 오전 1:00:00	5.2
26. 6. 29. 오전 1:30:00	5.2
26. 6. 29. 오전 2:00:00	5.2
26. 6. 29. 오전 2:30:00	5.3
26. 6. 29. 오전 3:00:00	5.3
26. 6. 29. 오전 3:30:00	5.3
26. 6. 29. 오전 4:00:00	5.5
26. 6. 29. 오전 4:30:00	5.4
26. 6. 29. 오전 5:00:00	5.3
26. 6. 29. 오전 5:30:00	5.3
26. 6. 29. 오전 6:00:00	5.3
26. 6. 29. 오전 6:30:00	5.3
26. 6. 29. 오전 7:00:00	5.3
26. 6. 29. 오전 7:30:00	5.3
26. 6. 29. 오전 8:00:00	5.3
26. 6. 29. 오전 8:30:00	5.2
26. 6. 29. 오전 9:00:00	5.2
26. 6. 29. 오전 9:30:00	5.1



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 29. 오전 10:00:00	5.2
26. 6. 29. 오전 10:30:00	5.2
26. 6. 29. 오전 11:00:00	5.1
26. 6. 29. 오전 11:30:00	5.2
26. 6. 29. 오후 12:00:00	5.2
26. 6. 29. 오후 12:30:00	5.3
26. 6. 29. 오후 1:00:00	5.3
26. 6. 29. 오후 1:30:00	5.3
26. 6. 29. 오후 2:00:00	5.3
26. 6. 29. 오후 2:30:00	5.3
26. 6. 29. 오후 3:00:00	5.3
26. 6. 29. 오후 3:30:00	5.2
26. 6. 29. 오후 4:00:00	5.2
26. 6. 29. 오후 4:30:00	5.3
26. 6. 29. 오후 5:00:00	5.3
26. 6. 29. 오후 5:30:00	5.2
26. 6. 29. 오후 6:00:00	5.3
26. 6. 29. 오후 6:30:00	5.2
26. 6. 29. 오후 7:00:00	5.3
26. 6. 29. 오후 7:30:00	5.3
26. 6. 29. 오후 8:00:00	5.3
26. 6. 29. 오후 8:30:00	5.3
26. 6. 29. 오후 9:00:00	5.3
26. 6. 29. 오후 9:30:00	5.4
26. 6. 29. 오후 10:00:00	5.4
26. 6. 29. 오후 10:30:00	5.4
26. 6. 29. 오후 11:00:00	5.4
26. 6. 29. 오후 11:30:00	5.3
26. 6. 30. 오전 12:00:00	5.3
26. 6. 30. 오전 12:30:00	5.2
26. 6. 30. 오전 1:00:00	5.3
26. 6. 30. 오전 1:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 30. 오전 2:00:00	5.3
26. 6. 30. 오전 2:30:00	5.4
26. 6. 30. 오전 3:00:00	5.4
26. 6. 30. 오전 3:30:00	5.3
26. 6. 30. 오전 4:00:00	5.3
26. 6. 30. 오전 4:30:00	5.4
26. 6. 30. 오전 5:00:00	5.4
26. 6. 30. 오전 5:30:00	5.4
26. 6. 30. 오전 6:00:00	5.3
26. 6. 30. 오전 6:30:00	5.2
26. 6. 30. 오전 7:00:00	5.2
26. 6. 30. 오전 7:30:00	5.2
26. 6. 30. 오전 8:00:00	5.1
26. 6. 30. 오전 8:30:00	5.1
26. 6. 30. 오전 9:00:00	5.1
26. 6. 30. 오전 9:30:00	5.1
26. 6. 30. 오전 10:00:00	5.0
26. 6. 30. 오전 10:30:00	5.0
26. 6. 30. 오전 11:00:00	5.1
26. 6. 30. 오전 11:30:00	5.1
26. 6. 30. 오후 12:00:00	5.1
26. 6. 30. 오후 12:30:00	5.1
26. 6. 30. 오후 1:00:00	5.1
26. 6. 30. 오후 1:30:00	5.1
26. 6. 30. 오후 2:00:00	5.2
26. 6. 30. 오후 2:30:00	5.4
26. 6. 30. 오후 3:00:00	5.4
26. 6. 30. 오후 3:30:00	5.3
26. 6. 30. 오후 4:00:00	5.2
26. 6. 30. 오후 4:30:00	5.3
26. 6. 30. 오후 5:00:00	5.4
26. 6. 30. 오후 5:30:00	5.4



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 6. 30. 오후 6:00:00	5.4
26. 6. 30. 오후 6:30:00	5.3
26. 6. 30. 오후 7:00:00	5.4
26. 6. 30. 오후 7:30:00	5.4
26. 6. 30. 오후 8:00:00	5.4
26. 6. 30. 오후 8:30:00	5.3
26. 6. 30. 오후 9:00:00	5.3
26. 6. 30. 오후 9:30:00	5.3
26. 6. 30. 오후 10:00:00	5.2
26. 6. 30. 오후 10:30:00	5.2
26. 6. 30. 오후 11:00:00	5.2
26. 6. 30. 오후 11:30:00	5.2



IP Storage Temperature Log_Monthly report

2026년 6월 1일 오전 12시 0분 0초 KST - 2026년 7월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 7월 1일 오전 6시 2분 59초 KST

알람 테이블

다음에서 발생 [Asia/Seoul]	마감일 [Asia/Seoul]	측정 포인트	알림 유형	상세 정보
26. 6. 9. 오전 1:36:00	26. 6. 9. 오전 1:36:05	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용의약품(냉장고A 2~8℃)_CH1 동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용의약품(냉장고B 2~8℃)_CH2	시스템 알림	연결 중단 일련번호: 54857013