

IP STORAGE TEMPERATURE LOG

Period of Record	Year. Month	2026.7	
Storage	Location	동국대학교일산병원 임상시험센터 약국	9134Room
	Description	Pharmaceutical Refrigerator	냉장고A
	Manufacturer	SGM	
	Model Name	CS-1300P	
	Serial Number	K1300P0292510M	
	Temperature Range	2~8℃	
	Setting Point	4.0℃	
Thermometer (including alarm system)	Description	Wi-Fi data logger	
	Manufacturer	Testo Korea	
	Model Name	Saveris2-T3	
	Serial Number	0054857013	
	Sensor Channel	Channel 1	
Thermometer Calibration Certification	Min/Max Record	Monthly report	This month Min (4.8)℃, Max(5.6)℃
	Date of Calibration	2025.10.1	
	Date of Calibration Due	2026.10.1	
Alarm System	Certificate No.	A25-1985-001	
	Alarm Threshold	Lower : 2.5℃, Upper : 7.5℃, immediate alarm without delay	
	Alarm Method	Mail, SMS to Pharmacists	

Date	Time	Temperature(℃)	Deviation	Verified by	Comment
2026.07.01	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.02	p:25	5.2	☐ Y / ☒ N	ml	
2026.07.03	p:10	5.2	☐ Y / ☒ N	ml	
2026.07.04			☐ Y / ☐ N		
2026.07.05			☐ Y / ☐ N		
2026.07.06	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.07	p:20	5.2	☐ Y / ☒ N	ml	
2026.07.08	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.09	10:20	5.3	☐ Y / ☒ N	ml	
2026.07.10	p:20	5.2	☐ Y / ☒ N	ml	
2026.07.11			☐ Y / ☐ N		
2026.07.12			☐ Y / ☐ N		
2026.07.13	p:30	5.1	☐ Y / ☒ N	김영환	
2026.07.14	p:40	5.2	☐ Y / ☒ N	ml	
2026.07.15	9:50	5.2	☐ Y / ☒ N	ml	
2026.07.16	p:20	5.1	☐Y / ☒N	ml	218022
2026.07.17			☐ Y / ☐ N		
2026.07.18			☐ Y / ☐ N		
2026.07.19			☐ Y / ☐ N		
2026.07.20	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.21	p:20	5.0	☐ Y / ☒ N	ml	
2026.07.22	p:20	5.0	☐ Y / ☒ N	ml	
2026.07.23	10:10	5.0	☐ Y / ☒ N	ml	
2026.07.24	p:20	5.0	☐ Y / ☒ N	ml	
2026.07.25			☐ Y / ☐ N		
2026.07.26			☐ Y / ☐ N		
2026.07.27	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.28	p:20	5.0	☐ Y / ☒ N	ml	
2026.07.29	p:20	5.1	☐ Y / ☒ N	ml	
2026.07.30	p:20	5.0	☐ Y / ☒ N	ml	
2026.07.31	p:20	5.2	☐ Y / ☒ N	ml	

date of copying


sign



내용

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1

(A 2 ~ 8) _ C H 1





IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타입 스탬프 [Asia/Seoul]	온도 [°C]
Min	4.8
Max	5.6
Average	5.2
26. 7. 1. 오전 12:00:00	5.2
26. 7. 1. 오전 12:30:00	5.2
26. 7. 1. 오전 1:00:00	5.2
26. 7. 1. 오전 1:30:00	5.2
26. 7. 1. 오전 2:00:00	5.2
26. 7. 1. 오전 2:30:00	5.2
26. 7. 1. 오전 3:00:00	5.2
26. 7. 1. 오전 3:30:00	5.1
26. 7. 1. 오전 4:00:00	5.2
26. 7. 1. 오전 4:30:00	5.2
26. 7. 1. 오전 5:00:00	5.2
26. 7. 1. 오전 5:30:00	5.2
26. 7. 1. 오전 6:00:00	5.1
26. 7. 1. 오전 6:30:00	5.2
26. 7. 1. 오전 7:00:00	5.2
26. 7. 1. 오전 7:30:00	5.1
26. 7. 1. 오전 8:00:00	5.1
26. 7. 1. 오전 8:30:00	5.1
26. 7. 1. 오전 9:00:00	5.1
26. 7. 1. 오전 9:30:00	5.1
26. 7. 1. 오전 10:00:00	5.1
26. 7. 1. 오전 10:30:00	5.4
26. 7. 1. 오전 11:00:00	5.4
26. 7. 1. 오전 11:30:00	5.4
26. 7. 1. 오후 12:00:00	5.4
26. 7. 1. 오후 12:30:00	5.4
26. 7. 1. 오후 1:00:00	5.5
26. 7. 1. 오후 1:30:00	5.5
26. 7. 1. 오후 2:00:00	5.4



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 1. 오후 2:30:00	5.4
26. 7. 1. 오후 3:00:00	5.4
26. 7. 1. 오후 3:30:00	5.4
26. 7. 1. 오후 4:00:00	5.4
26. 7. 1. 오후 4:30:00	5.6
26. 7. 1. 오후 5:00:00	5.4
26. 7. 1. 오후 5:30:00	5.4
26. 7. 1. 오후 6:00:00	5.3
26. 7. 1. 오후 6:30:00	5.4
26. 7. 1. 오후 7:00:00	5.3
26. 7. 1. 오후 7:30:00	5.3
26. 7. 1. 오후 8:00:00	5.4
26. 7. 1. 오후 8:30:00	5.3
26. 7. 1. 오후 9:00:00	5.3
26. 7. 1. 오후 9:30:00	5.4
26. 7. 1. 오후 10:00:00	5.3
26. 7. 1. 오후 10:30:00	5.3
26. 7. 1. 오후 11:00:00	5.2
26. 7. 1. 오후 11:30:00	5.2
26. 7. 2. 오전 12:00:00	5.2
26. 7. 2. 오전 12:30:00	5.2
26. 7. 2. 오전 1:00:00	5.2
26. 7. 2. 오전 1:30:00	5.2
26. 7. 2. 오전 2:00:00	5.2
26. 7. 2. 오전 2:30:00	5.2
26. 7. 2. 오전 3:00:00	5.2
26. 7. 2. 오전 3:30:00	5.2
26. 7. 2. 오전 4:00:00	5.2
26. 7. 2. 오전 4:30:00	5.3
26. 7. 2. 오전 5:00:00	5.3
26. 7. 2. 오전 5:30:00	5.3
26. 7. 2. 오전 6:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 2. 오전 6:30:00	5.3
26. 7. 2. 오전 7:00:00	5.2
26. 7. 2. 오전 7:30:00	5.3
26. 7. 2. 오전 8:00:00	5.2
26. 7. 2. 오전 8:30:00	5.1
26. 7. 2. 오전 9:00:00	5.2
26. 7. 2. 오전 9:30:00	5.2
26. 7. 2. 오전 10:00:00	5.1
26. 7. 2. 오전 10:30:00	5.2
26. 7. 2. 오전 11:00:00	5.3
26. 7. 2. 오전 11:30:00	5.3
26. 7. 2. 오후 12:00:00	5.2
26. 7. 2. 오후 12:30:00	5.3
26. 7. 2. 오후 1:00:00	5.2
26. 7. 2. 오후 1:30:00	5.2
26. 7. 2. 오후 2:00:00	5.3
26. 7. 2. 오후 2:30:00	5.3
26. 7. 2. 오후 3:00:00	5.2
26. 7. 2. 오후 3:30:00	5.2
26. 7. 2. 오후 4:00:00	5.3
26. 7. 2. 오후 4:30:00	5.3
26. 7. 2. 오후 5:00:00	5.3
26. 7. 2. 오후 5:30:00	5.3
26. 7. 2. 오후 6:00:00	5.4
26. 7. 2. 오후 6:30:00	5.4
26. 7. 2. 오후 7:00:00	5.4
26. 7. 2. 오후 7:30:00	5.5
26. 7. 2. 오후 8:00:00	5.5
26. 7. 2. 오후 8:30:00	5.4
26. 7. 2. 오후 9:00:00	5.4
26. 7. 2. 오후 9:30:00	5.3
26. 7. 2. 오후 10:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 2. 오후 10:30:00	5.3
26. 7. 2. 오후 11:00:00	5.3
26. 7. 2. 오후 11:30:00	5.2
26. 7. 3. 오전 12:00:00	5.2
26. 7. 3. 오전 12:30:00	5.2
26. 7. 3. 오전 1:00:00	5.2
26. 7. 3. 오전 1:30:00	5.3
26. 7. 3. 오전 2:00:00	5.2
26. 7. 3. 오전 2:30:00	5.2
26. 7. 3. 오전 3:00:00	5.2
26. 7. 3. 오전 3:30:00	5.2
26. 7. 3. 오전 4:00:00	5.2
26. 7. 3. 오전 4:30:00	5.2
26. 7. 3. 오전 5:00:00	5.2
26. 7. 3. 오전 5:30:00	5.3
26. 7. 3. 오전 6:00:00	5.2
26. 7. 3. 오전 6:30:00	5.3
26. 7. 3. 오전 7:00:00	5.4
26. 7. 3. 오전 7:30:00	5.3
26. 7. 3. 오전 8:00:00	5.3
26. 7. 3. 오전 8:30:00	5.3
26. 7. 3. 오전 9:00:00	5.2
26. 7. 3. 오전 9:30:00	5.2
26. 7. 3. 오전 10:00:00	5.3
26. 7. 3. 오전 10:30:00	5.3
26. 7. 3. 오전 11:00:00	5.3
26. 7. 3. 오전 11:30:00	5.3
26. 7. 3. 오후 12:00:00	5.3
26. 7. 3. 오후 12:30:00	5.3
26. 7. 3. 오후 1:00:00	5.3
26. 7. 3. 오후 1:30:00	5.4
26. 7. 3. 오후 2:00:00	5.4



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 3. 오후 2:30:00	5.4
26. 7. 3. 오후 3:00:00	5.3
26. 7. 3. 오후 3:30:00	5.4
26. 7. 3. 오후 4:00:00	5.4
26. 7. 3. 오후 4:30:00	5.4
26. 7. 3. 오후 5:00:00	5.4
26. 7. 3. 오후 5:30:00	5.4
26. 7. 3. 오후 6:00:00	5.4
26. 7. 3. 오후 6:30:00	5.4
26. 7. 3. 오후 7:00:00	5.4
26. 7. 3. 오후 7:30:00	5.3
26. 7. 3. 오후 8:00:00	5.3
26. 7. 3. 오후 8:30:00	5.3
26. 7. 3. 오후 9:00:00	5.3
26. 7. 3. 오후 9:30:00	5.3
26. 7. 3. 오후 10:00:00	5.3
26. 7. 3. 오후 10:30:00	5.2
26. 7. 3. 오후 11:00:00	5.1
26. 7. 3. 오후 11:30:00	5.2
26. 7. 4. 오전 12:00:00	5.2
26. 7. 4. 오전 12:30:00	5.2
26. 7. 4. 오전 1:00:00	5.3
26. 7. 4. 오전 1:30:00	5.3
26. 7. 4. 오전 2:00:00	5.3
26. 7. 4. 오전 2:30:00	5.4
26. 7. 4. 오전 3:00:00	5.4
26. 7. 4. 오전 3:30:00	5.4
26. 7. 4. 오전 4:00:00	5.5
26. 7. 4. 오전 4:30:00	5.4
26. 7. 4. 오전 5:00:00	5.4
26. 7. 4. 오전 5:30:00	5.3
26. 7. 4. 오전 6:00:00	5.4



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 4. 오전 6:30:00	5.3
26. 7. 4. 오전 7:00:00	5.3
26. 7. 4. 오전 7:30:00	5.3
26. 7. 4. 오전 8:00:00	5.3
26. 7. 4. 오전 8:30:00	5.2
26. 7. 4. 오전 9:00:00	5.3
26. 7. 4. 오전 9:30:00	5.3
26. 7. 4. 오전 10:00:00	5.2
26. 7. 4. 오전 10:30:00	5.2
26. 7. 4. 오전 11:00:00	5.4
26. 7. 4. 오전 11:30:00	5.3
26. 7. 4. 오후 12:00:00	5.2
26. 7. 4. 오후 12:30:00	5.2
26. 7. 4. 오후 1:00:00	5.3
26. 7. 4. 오후 1:30:00	5.2
26. 7. 4. 오후 2:00:00	5.1
26. 7. 4. 오후 2:30:00	5.2
26. 7. 4. 오후 3:00:00	5.2
26. 7. 4. 오후 3:30:00	5.1
26. 7. 4. 오후 4:00:00	5.1
26. 7. 4. 오후 4:30:00	5.2
26. 7. 4. 오후 5:00:00	5.3
26. 7. 4. 오후 5:30:00	5.2
26. 7. 4. 오후 6:00:00	5.3
26. 7. 4. 오후 6:30:00	5.2
26. 7. 4. 오후 7:00:00	5.2
26. 7. 4. 오후 7:30:00	5.3
26. 7. 4. 오후 8:00:00	5.2
26. 7. 4. 오후 8:30:00	5.1
26. 7. 4. 오후 9:00:00	5.1
26. 7. 4. 오후 9:30:00	5.1
26. 7. 4. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 4. 오후 10:30:00	5.2
26. 7. 4. 오후 11:00:00	5.3
26. 7. 4. 오후 11:30:00	5.2
26. 7. 5. 오전 12:00:00	5.2
26. 7. 5. 오전 12:30:00	5.2
26. 7. 5. 오전 1:00:00	5.2
26. 7. 5. 오전 1:30:00	5.3
26. 7. 5. 오전 2:00:00	5.2
26. 7. 5. 오전 2:30:00	5.2
26. 7. 5. 오전 3:00:00	5.2
26. 7. 5. 오전 3:30:00	5.2
26. 7. 5. 오전 4:00:00	5.1
26. 7. 5. 오전 4:30:00	5.2
26. 7. 5. 오전 5:00:00	5.3
26. 7. 5. 오전 5:30:00	5.2
26. 7. 5. 오전 6:00:00	5.2
26. 7. 5. 오전 6:30:00	5.3
26. 7. 5. 오전 7:00:00	5.3
26. 7. 5. 오전 7:30:00	5.3
26. 7. 5. 오전 8:00:00	5.4
26. 7. 5. 오전 8:30:00	5.4
26. 7. 5. 오전 9:00:00	5.3
26. 7. 5. 오전 9:30:00	5.3
26. 7. 5. 오전 10:00:00	5.4
26. 7. 5. 오전 10:30:00	5.4
26. 7. 5. 오전 11:00:00	5.4
26. 7. 5. 오전 11:30:00	5.4
26. 7. 5. 오후 12:00:00	5.4
26. 7. 5. 오후 12:30:00	5.3
26. 7. 5. 오후 1:00:00	5.2
26. 7. 5. 오후 1:30:00	5.3
26. 7. 5. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 5. 오후 2:30:00	5.2
26. 7. 5. 오후 3:00:00	5.2
26. 7. 5. 오후 3:30:00	5.3
26. 7. 5. 오후 4:00:00	5.3
26. 7. 5. 오후 4:30:00	5.3
26. 7. 5. 오후 5:00:00	5.3
26. 7. 5. 오후 5:30:00	5.3
26. 7. 5. 오후 6:00:00	5.3
26. 7. 5. 오후 6:30:00	5.4
26. 7. 5. 오후 7:00:00	5.3
26. 7. 5. 오후 7:30:00	5.3
26. 7. 5. 오후 8:00:00	5.3
26. 7. 5. 오후 8:30:00	5.3
26. 7. 5. 오후 9:00:00	5.3
26. 7. 5. 오후 9:30:00	5.4
26. 7. 5. 오후 10:00:00	5.4
26. 7. 5. 오후 10:30:00	5.4
26. 7. 5. 오후 11:00:00	5.4
26. 7. 5. 오후 11:30:00	5.4
26. 7. 6. 오전 12:00:00	5.3
26. 7. 6. 오전 12:30:00	5.2
26. 7. 6. 오전 1:00:00	5.2
26. 7. 6. 오전 1:30:00	5.3
26. 7. 6. 오전 2:00:00	5.3
26. 7. 6. 오전 2:30:00	5.3
26. 7. 6. 오전 3:00:00	5.4
26. 7. 6. 오전 3:30:00	5.2
26. 7. 6. 오전 4:00:00	5.2
26. 7. 6. 오전 4:30:00	5.3
26. 7. 6. 오전 5:00:00	5.3
26. 7. 6. 오전 5:30:00	5.2
26. 7. 6. 오전 6:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 6. 오전 6:30:00	5.3
26. 7. 6. 오전 7:00:00	5.3
26. 7. 6. 오전 7:30:00	5.2
26. 7. 6. 오전 8:00:00	5.2
26. 7. 6. 오전 8:30:00	5.2
26. 7. 6. 오전 9:00:00	5.1
26. 7. 6. 오전 9:30:00	5.1
26. 7. 6. 오전 10:00:00	5.1
26. 7. 6. 오전 10:30:00	5.2
26. 7. 6. 오전 11:00:00	5.2
26. 7. 6. 오전 11:30:00	5.2
26. 7. 6. 오후 12:00:00	5.2
26. 7. 6. 오후 12:30:00	5.2
26. 7. 6. 오후 1:00:00	5.2
26. 7. 6. 오후 1:30:00	5.2
26. 7. 6. 오후 2:00:00	5.2
26. 7. 6. 오후 2:30:00	5.3
26. 7. 6. 오후 3:00:00	5.3
26. 7. 6. 오후 3:30:00	5.3
26. 7. 6. 오후 4:00:00	5.3
26. 7. 6. 오후 4:30:00	5.3
26. 7. 6. 오후 5:00:00	5.2
26. 7. 6. 오후 5:30:00	5.3
26. 7. 6. 오후 6:00:00	5.2
26. 7. 6. 오후 6:30:00	5.3
26. 7. 6. 오후 7:00:00	5.3
26. 7. 6. 오후 7:30:00	5.2
26. 7. 6. 오후 8:00:00	5.2
26. 7. 6. 오후 8:30:00	5.2
26. 7. 6. 오후 9:00:00	5.3
26. 7. 6. 오후 9:30:00	5.2
26. 7. 6. 오후 10:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 6. 오후 10:30:00	5.3
26. 7. 6. 오후 11:00:00	5.2
26. 7. 6. 오후 11:30:00	5.3
26. 7. 7. 오전 12:00:00	5.2
26. 7. 7. 오전 12:30:00	5.2
26. 7. 7. 오전 1:00:00	5.2
26. 7. 7. 오전 1:30:00	5.2
26. 7. 7. 오전 2:00:00	5.2
26. 7. 7. 오전 2:30:00	5.2
26. 7. 7. 오전 3:00:00	5.2
26. 7. 7. 오전 3:30:00	5.2
26. 7. 7. 오전 4:00:00	5.2
26. 7. 7. 오전 4:30:00	5.3
26. 7. 7. 오전 5:00:00	5.2
26. 7. 7. 오전 5:30:00	5.3
26. 7. 7. 오전 6:00:00	5.2
26. 7. 7. 오전 6:30:00	5.2
26. 7. 7. 오전 7:00:00	5.2
26. 7. 7. 오전 7:30:00	5.2
26. 7. 7. 오전 8:00:00	5.2
26. 7. 7. 오전 8:30:00	5.2
26. 7. 7. 오전 9:00:00	5.2
26. 7. 7. 오전 9:30:00	5.2
26. 7. 7. 오전 10:00:00	5.1
26. 7. 7. 오전 10:30:00	5.2
26. 7. 7. 오전 11:00:00	5.2
26. 7. 7. 오전 11:30:00	5.3
26. 7. 7. 오후 12:00:00	5.2
26. 7. 7. 오후 12:30:00	5.2
26. 7. 7. 오후 1:00:00	5.2
26. 7. 7. 오후 1:30:00	5.2
26. 7. 7. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 7. 오후 2:30:00	5.2
26. 7. 7. 오후 3:00:00	5.1
26. 7. 7. 오후 3:30:00	5.2
26. 7. 7. 오후 4:00:00	5.2
26. 7. 7. 오후 4:30:00	5.3
26. 7. 7. 오후 5:00:00	5.2
26. 7. 7. 오후 5:30:00	5.3
26. 7. 7. 오후 6:00:00	5.3
26. 7. 7. 오후 6:30:00	5.3
26. 7. 7. 오후 7:00:00	5.2
26. 7. 7. 오후 7:30:00	5.2
26. 7. 7. 오후 8:00:00	5.2
26. 7. 7. 오후 8:30:00	5.3
26. 7. 7. 오후 9:00:00	5.2
26. 7. 7. 오후 9:30:00	5.2
26. 7. 7. 오후 10:00:00	5.2
26. 7. 7. 오후 10:30:00	5.2
26. 7. 7. 오후 11:00:00	5.2
26. 7. 7. 오후 11:30:00	5.2
26. 7. 8. 오전 12:00:00	5.2
26. 7. 8. 오전 12:30:00	5.3
26. 7. 8. 오전 1:00:00	5.3
26. 7. 8. 오전 1:30:00	5.3
26. 7. 8. 오전 2:00:00	5.2
26. 7. 8. 오전 2:30:00	5.2
26. 7. 8. 오전 3:00:00	5.3
26. 7. 8. 오전 3:30:00	5.2
26. 7. 8. 오전 4:00:00	5.2
26. 7. 8. 오전 4:30:00	5.2
26. 7. 8. 오전 5:00:00	5.3
26. 7. 8. 오전 5:30:00	5.2
26. 7. 8. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 8. 오전 6:30:00	5.2
26. 7. 8. 오전 7:00:00	5.2
26. 7. 8. 오전 7:30:00	5.2
26. 7. 8. 오전 8:00:00	5.2
26. 7. 8. 오전 8:30:00	5.1
26. 7. 8. 오전 9:00:00	5.1
26. 7. 8. 오전 9:30:00	5.2
26. 7. 8. 오전 10:00:00	5.1
26. 7. 8. 오전 10:30:00	5.2
26. 7. 8. 오전 11:00:00	5.2
26. 7. 8. 오전 11:30:00	5.2
26. 7. 8. 오후 12:00:00	5.2
26. 7. 8. 오후 12:30:00	5.2
26. 7. 8. 오후 1:00:00	5.2
26. 7. 8. 오후 1:30:00	5.1
26. 7. 8. 오후 2:00:00	5.2
26. 7. 8. 오후 2:30:00	5.3
26. 7. 8. 오후 3:00:00	5.3
26. 7. 8. 오후 3:30:00	5.3
26. 7. 8. 오후 4:00:00	5.5
26. 7. 8. 오후 4:30:00	5.5
26. 7. 8. 오후 5:00:00	5.4
26. 7. 8. 오후 5:30:00	5.4
26. 7. 8. 오후 6:00:00	5.4
26. 7. 8. 오후 6:30:00	5.3
26. 7. 8. 오후 7:00:00	5.3
26. 7. 8. 오후 7:30:00	5.2
26. 7. 8. 오후 8:00:00	5.3
26. 7. 8. 오후 8:30:00	5.3
26. 7. 8. 오후 9:00:00	5.3
26. 7. 8. 오후 9:30:00	5.2
26. 7. 8. 오후 10:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 8. 오후 10:30:00	5.2
26. 7. 8. 오후 11:00:00	5.2
26. 7. 8. 오후 11:30:00	5.2
26. 7. 9. 오전 12:00:00	5.3
26. 7. 9. 오전 12:30:00	5.2
26. 7. 9. 오전 1:00:00	5.3
26. 7. 9. 오전 1:30:00	5.3
26. 7. 9. 오전 2:00:00	5.2
26. 7. 9. 오전 2:30:00	5.3
26. 7. 9. 오전 3:00:00	5.3
26. 7. 9. 오전 3:30:00	5.2
26. 7. 9. 오전 4:00:00	5.2
26. 7. 9. 오전 4:30:00	5.2
26. 7. 9. 오전 5:00:00	5.2
26. 7. 9. 오전 5:30:00	5.2
26. 7. 9. 오전 6:00:00	5.3
26. 7. 9. 오전 6:30:00	5.2
26. 7. 9. 오전 7:00:00	5.2
26. 7. 9. 오전 7:30:00	5.2
26. 7. 9. 오전 8:00:00	5.3
26. 7. 9. 오전 8:30:00	5.2
26. 7. 9. 오전 9:00:00	5.2
26. 7. 9. 오전 9:30:00	5.2
26. 7. 9. 오전 10:00:00	5.3
26. 7. 9. 오전 10:30:00	5.2
26. 7. 9. 오전 11:00:00	5.2
26. 7. 9. 오전 11:30:00	5.2
26. 7. 9. 오후 12:00:00	5.3
26. 7. 9. 오후 12:30:00	5.3
26. 7. 9. 오후 1:00:00	5.3
26. 7. 9. 오후 1:30:00	5.2
26. 7. 9. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 9. 오후 2:30:00	5.2
26. 7. 9. 오후 3:00:00	5.2
26. 7. 9. 오후 3:30:00	5.3
26. 7. 9. 오후 4:00:00	5.3
26. 7. 9. 오후 4:30:00	5.2
26. 7. 9. 오후 5:00:00	5.2
26. 7. 9. 오후 5:30:00	5.2
26. 7. 9. 오후 6:00:00	5.2
26. 7. 9. 오후 6:30:00	5.3
26. 7. 9. 오후 7:00:00	5.3
26. 7. 9. 오후 7:30:00	5.2
26. 7. 9. 오후 8:00:00	5.2
26. 7. 9. 오후 8:30:00	5.2
26. 7. 9. 오후 9:00:00	5.2
26. 7. 9. 오후 9:30:00	5.2
26. 7. 9. 오후 10:00:00	5.3
26. 7. 9. 오후 10:30:00	5.2
26. 7. 9. 오후 11:00:00	5.2
26. 7. 9. 오후 11:30:00	5.3
26. 7. 10. 오전 12:00:00	5.3
26. 7. 10. 오전 12:30:00	5.2
26. 7. 10. 오전 1:00:00	5.2
26. 7. 10. 오전 1:30:00	5.2
26. 7. 10. 오전 2:00:00	5.2
26. 7. 10. 오전 2:30:00	5.2
26. 7. 10. 오전 3:00:00	5.2
26. 7. 10. 오전 3:30:00	5.1
26. 7. 10. 오전 4:00:00	5.2
26. 7. 10. 오전 4:30:00	5.2
26. 7. 10. 오전 5:00:00	5.2
26. 7. 10. 오전 5:30:00	5.1
26. 7. 10. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 10. 오전 6:30:00	5.3
26. 7. 10. 오전 7:00:00	5.3
26. 7. 10. 오전 7:30:00	5.3
26. 7. 10. 오전 8:00:00	5.3
26. 7. 10. 오전 8:30:00	5.2
26. 7. 10. 오전 9:00:00	5.2
26. 7. 10. 오전 9:30:00	5.3
26. 7. 10. 오전 10:00:00	5.3
26. 7. 10. 오전 10:30:00	5.2
26. 7. 10. 오전 11:00:00	5.3
26. 7. 10. 오전 11:30:00	5.3
26. 7. 10. 오후 12:00:00	5.3
26. 7. 10. 오후 12:30:00	5.3
26. 7. 10. 오후 1:00:00	5.3
26. 7. 10. 오후 1:30:00	5.3
26. 7. 10. 오후 2:00:00	5.3
26. 7. 10. 오후 2:30:00	5.3
26. 7. 10. 오후 3:00:00	5.3
26. 7. 10. 오후 3:30:00	5.4
26. 7. 10. 오후 4:00:00	5.4
26. 7. 10. 오후 4:30:00	5.4
26. 7. 10. 오후 5:00:00	5.5
26. 7. 10. 오후 5:30:00	5.5
26. 7. 10. 오후 6:00:00	5.5
26. 7. 10. 오후 6:30:00	5.5
26. 7. 10. 오후 7:00:00	5.5
26. 7. 10. 오후 7:30:00	5.4
26. 7. 10. 오후 8:00:00	5.4
26. 7. 10. 오후 8:30:00	5.3
26. 7. 10. 오후 9:00:00	5.2
26. 7. 10. 오후 9:30:00	5.3
26. 7. 10. 오후 10:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 10. 오후 10:30:00	5.2
26. 7. 10. 오후 11:00:00	5.3
26. 7. 10. 오후 11:30:00	5.3
26. 7. 11. 오전 12:00:00	5.3
26. 7. 11. 오전 12:30:00	5.4
26. 7. 11. 오전 1:00:00	5.4
26. 7. 11. 오전 1:30:00	5.3
26. 7. 11. 오전 2:00:00	5.2
26. 7. 11. 오전 2:30:00	5.3
26. 7. 11. 오전 3:00:00	5.2
26. 7. 11. 오전 3:30:00	5.1
26. 7. 11. 오전 4:00:00	5.2
26. 7. 11. 오전 4:30:00	5.2
26. 7. 11. 오전 5:00:00	5.2
26. 7. 11. 오전 5:30:00	5.3
26. 7. 11. 오전 6:00:00	5.3
26. 7. 11. 오전 6:30:00	5.3
26. 7. 11. 오전 7:00:00	5.2
26. 7. 11. 오전 7:30:00	5.2
26. 7. 11. 오전 8:00:00	5.2
26. 7. 11. 오전 8:30:00	5.2
26. 7. 11. 오전 9:00:00	5.1
26. 7. 11. 오전 9:30:00	5.1
26. 7. 11. 오전 10:00:00	5.2
26. 7. 11. 오전 10:30:00	5.2
26. 7. 11. 오전 11:00:00	5.2
26. 7. 11. 오전 11:30:00	5.1
26. 7. 11. 오후 12:00:00	5.1
26. 7. 11. 오후 12:30:00	5.2
26. 7. 11. 오후 1:00:00	5.2
26. 7. 11. 오후 1:30:00	5.2
26. 7. 11. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 11. 오후 2:30:00	5.2
26. 7. 11. 오후 3:00:00	5.2
26. 7. 11. 오후 3:30:00	5.2
26. 7. 11. 오후 4:00:00	5.1
26. 7. 11. 오후 4:30:00	5.1
26. 7. 11. 오후 5:00:00	5.2
26. 7. 11. 오후 5:30:00	5.2
26. 7. 11. 오후 6:00:00	5.2
26. 7. 11. 오후 6:30:00	5.3
26. 7. 11. 오후 7:00:00	5.2
26. 7. 11. 오후 7:30:00	5.2
26. 7. 11. 오후 8:00:00	5.2
26. 7. 11. 오후 8:30:00	5.2
26. 7. 11. 오후 9:00:00	5.2
26. 7. 11. 오후 9:30:00	5.2
26. 7. 11. 오후 10:00:00	5.3
26. 7. 11. 오후 10:30:00	5.4
26. 7. 11. 오후 11:00:00	5.3
26. 7. 11. 오후 11:30:00	5.3
26. 7. 12. 오전 12:00:00	5.3
26. 7. 12. 오전 12:30:00	5.3
26. 7. 12. 오전 1:00:00	5.3
26. 7. 12. 오전 1:30:00	5.3
26. 7. 12. 오전 2:00:00	5.3
26. 7. 12. 오전 2:30:00	5.3
26. 7. 12. 오전 3:00:00	5.3
26. 7. 12. 오전 3:30:00	5.3
26. 7. 12. 오전 4:00:00	5.3
26. 7. 12. 오전 4:30:00	5.3
26. 7. 12. 오전 5:00:00	5.2
26. 7. 12. 오전 5:30:00	5.2
26. 7. 12. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 12. 오전 6:30:00	5.3
26. 7. 12. 오전 7:00:00	5.3
26. 7. 12. 오전 7:30:00	5.3
26. 7. 12. 오전 8:00:00	5.2
26. 7. 12. 오전 8:30:00	5.2
26. 7. 12. 오전 9:00:00	5.2
26. 7. 12. 오전 9:30:00	5.3
26. 7. 12. 오전 10:00:00	5.2
26. 7. 12. 오전 10:30:00	5.2
26. 7. 12. 오전 11:00:00	5.3
26. 7. 12. 오전 11:30:00	5.2
26. 7. 12. 오후 12:00:00	5.2
26. 7. 12. 오후 12:30:00	5.3
26. 7. 12. 오후 1:00:00	5.3
26. 7. 12. 오후 1:30:00	5.3
26. 7. 12. 오후 2:00:00	5.3
26. 7. 12. 오후 2:30:00	5.3
26. 7. 12. 오후 3:00:00	5.3
26. 7. 12. 오후 3:30:00	5.3
26. 7. 12. 오후 4:00:00	5.3
26. 7. 12. 오후 4:30:00	5.4
26. 7. 12. 오후 5:00:00	5.4
26. 7. 12. 오후 5:30:00	5.3
26. 7. 12. 오후 6:00:00	5.2
26. 7. 12. 오후 6:30:00	5.4
26. 7. 12. 오후 7:00:00	5.3
26. 7. 12. 오후 7:30:00	5.3
26. 7. 12. 오후 8:00:00	5.3
26. 7. 12. 오후 8:30:00	5.2
26. 7. 12. 오후 9:00:00	5.2
26. 7. 12. 오후 9:30:00	5.2
26. 7. 12. 오후 10:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 12. 오후 10:30:00	5.2
26. 7. 12. 오후 11:00:00	5.2
26. 7. 12. 오후 11:30:00	5.2
26. 7. 13. 오전 12:00:00	5.2
26. 7. 13. 오전 12:30:00	5.2
26. 7. 13. 오전 1:00:00	5.2
26. 7. 13. 오전 1:30:00	5.3
26. 7. 13. 오전 2:00:00	5.2
26. 7. 13. 오전 2:30:00	5.1
26. 7. 13. 오전 3:00:00	5.2
26. 7. 13. 오전 3:30:00	5.2
26. 7. 13. 오전 4:00:00	5.2
26. 7. 13. 오전 4:30:00	5.2
26. 7. 13. 오전 5:00:00	5.2
26. 7. 13. 오전 5:30:00	5.2
26. 7. 13. 오전 6:00:00	5.1
26. 7. 13. 오전 6:30:00	5.2
26. 7. 13. 오전 7:00:00	5.2
26. 7. 13. 오전 7:30:00	5.2
26. 7. 13. 오전 8:00:00	5.2
26. 7. 13. 오전 8:30:00	5.2
26. 7. 13. 오전 9:00:00	5.2
26. 7. 13. 오전 9:30:00	5.1
26. 7. 13. 오전 10:00:00	5.2
26. 7. 13. 오전 10:30:00	5.2
26. 7. 13. 오전 11:00:00	5.3
26. 7. 13. 오전 11:30:00	5.2
26. 7. 13. 오후 12:00:00	5.2
26. 7. 13. 오후 12:30:00	5.2
26. 7. 13. 오후 1:00:00	5.3
26. 7. 13. 오후 1:30:00	5.2
26. 7. 13. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 13. 오후 2:30:00	5.2
26. 7. 13. 오후 3:00:00	5.2
26. 7. 13. 오후 3:30:00	5.3
26. 7. 13. 오후 4:00:00	5.2
26. 7. 13. 오후 4:30:00	5.2
26. 7. 13. 오후 5:00:00	5.2
26. 7. 13. 오후 5:30:00	5.3
26. 7. 13. 오후 6:00:00	5.3
26. 7. 13. 오후 6:30:00	5.2
26. 7. 13. 오후 7:00:00	5.2
26. 7. 13. 오후 7:30:00	5.3
26. 7. 13. 오후 8:00:00	5.2
26. 7. 13. 오후 8:30:00	5.2
26. 7. 13. 오후 9:00:00	5.3
26. 7. 13. 오후 9:30:00	5.3
26. 7. 13. 오후 10:00:00	5.2
26. 7. 13. 오후 10:30:00	5.1
26. 7. 13. 오후 11:00:00	5.1
26. 7. 13. 오후 11:30:00	5.2
26. 7. 14. 오전 12:00:00	5.2
26. 7. 14. 오전 12:30:00	5.1
26. 7. 14. 오전 1:00:00	5.2
26. 7. 14. 오전 1:30:00	5.2
26. 7. 14. 오전 2:00:00	5.2
26. 7. 14. 오전 2:30:00	5.1
26. 7. 14. 오전 3:00:00	5.2
26. 7. 14. 오전 3:30:00	5.2
26. 7. 14. 오전 4:00:00	5.2
26. 7. 14. 오전 4:30:00	5.2
26. 7. 14. 오전 5:00:00	5.2
26. 7. 14. 오전 5:30:00	5.2
26. 7. 14. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 14. 오전 6:30:00	5.2
26. 7. 14. 오전 7:00:00	5.4
26. 7. 14. 오전 7:30:00	5.3
26. 7. 14. 오전 8:00:00	5.2
26. 7. 14. 오전 8:30:00	5.2
26. 7. 14. 오전 9:00:00	5.2
26. 7. 14. 오전 9:30:00	5.2
26. 7. 14. 오전 10:00:00	5.2
26. 7. 14. 오전 10:30:00	5.1
26. 7. 14. 오전 11:00:00	5.2
26. 7. 14. 오전 11:30:00	5.2
26. 7. 14. 오후 12:00:00	5.2
26. 7. 14. 오후 12:30:00	5.2
26. 7. 14. 오후 1:00:00	5.3
26. 7. 14. 오후 1:30:00	5.3
26. 7. 14. 오후 2:00:00	5.3
26. 7. 14. 오후 2:30:00	5.3
26. 7. 14. 오후 3:00:00	5.2
26. 7. 14. 오후 3:30:00	5.2
26. 7. 14. 오후 4:00:00	5.2
26. 7. 14. 오후 4:30:00	5.2
26. 7. 14. 오후 5:00:00	5.2
26. 7. 14. 오후 5:30:00	5.1
26. 7. 14. 오후 6:00:00	5.1
26. 7. 14. 오후 6:30:00	5.2
26. 7. 14. 오후 7:00:00	5.2
26. 7. 14. 오후 7:30:00	5.2
26. 7. 14. 오후 8:00:00	5.2
26. 7. 14. 오후 8:30:00	5.2
26. 7. 14. 오후 9:00:00	5.2
26. 7. 14. 오후 9:30:00	5.2
26. 7. 14. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 14. 오후 10:30:00	5.2
26. 7. 14. 오후 11:00:00	5.2
26. 7. 14. 오후 11:30:00	5.2
26. 7. 15. 오전 12:00:00	5.2
26. 7. 15. 오전 12:30:00	5.3
26. 7. 15. 오전 1:00:00	5.4
26. 7. 15. 오전 1:30:00	5.3
26. 7. 15. 오전 2:00:00	5.2
26. 7. 15. 오전 2:30:00	5.3
26. 7. 15. 오전 3:00:00	5.3
26. 7. 15. 오전 3:30:00	5.3
26. 7. 15. 오전 4:00:00	5.2
26. 7. 15. 오전 4:30:00	5.2
26. 7. 15. 오전 5:00:00	5.2
26. 7. 15. 오전 5:30:00	5.2
26. 7. 15. 오전 6:00:00	5.2
26. 7. 15. 오전 6:30:00	5.2
26. 7. 15. 오전 7:00:00	5.2
26. 7. 15. 오전 7:30:00	5.3
26. 7. 15. 오전 8:00:00	5.3
26. 7. 15. 오전 8:30:00	5.2
26. 7. 15. 오전 9:00:00	5.3
26. 7. 15. 오전 9:30:00	5.3
26. 7. 15. 오전 10:00:00	5.3
26. 7. 15. 오전 10:30:00	5.3
26. 7. 15. 오전 11:00:00	5.3
26. 7. 15. 오전 11:30:00	5.2
26. 7. 15. 오후 12:00:00	5.1
26. 7. 15. 오후 12:30:00	5.1
26. 7. 15. 오후 1:00:00	5.3
26. 7. 15. 오후 1:30:00	5.3
26. 7. 15. 오후 2:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 15. 오후 2:30:00	5.1
26. 7. 15. 오후 3:00:00	5.1
26. 7. 15. 오후 3:30:00	5.1
26. 7. 15. 오후 4:00:00	5.1
26. 7. 15. 오후 4:30:00	5.1
26. 7. 15. 오후 5:00:00	5.1
26. 7. 15. 오후 5:30:00	5.1
26. 7. 15. 오후 6:00:00	5.1
26. 7. 15. 오후 6:30:00	5.0
26. 7. 15. 오후 7:00:00	5.1
26. 7. 15. 오후 7:30:00	5.1
26. 7. 15. 오후 8:00:00	5.1
26. 7. 15. 오후 8:30:00	5.0
26. 7. 15. 오후 9:00:00	5.0
26. 7. 15. 오후 9:30:00	5.0
26. 7. 15. 오후 10:00:00	5.1
26. 7. 15. 오후 10:30:00	5.0
26. 7. 15. 오후 11:00:00	5.0
26. 7. 15. 오후 11:30:00	5.1
26. 7. 16. 오전 12:00:00	5.1
26. 7. 16. 오전 12:30:00	5.1
26. 7. 16. 오전 1:00:00	5.2
26. 7. 16. 오전 1:30:00	5.3
26. 7. 16. 오전 2:00:00	5.1
26. 7. 16. 오전 2:30:00	5.1
26. 7. 16. 오전 3:00:00	5.2
26. 7. 16. 오전 3:30:00	5.2
26. 7. 16. 오전 4:00:00	5.1
26. 7. 16. 오전 4:30:00	5.1
26. 7. 16. 오전 5:00:00	5.1
26. 7. 16. 오전 5:30:00	5.0
26. 7. 16. 오전 6:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 16. 오전 6:30:00	5.2
26. 7. 16. 오전 7:00:00	5.2
26. 7. 16. 오전 7:30:00	5.2
26. 7. 16. 오전 8:00:00	5.1
26. 7. 16. 오전 8:30:00	5.1
26. 7. 16. 오전 9:00:00	5.1
26. 7. 16. 오전 9:30:00	5.2
26. 7. 16. 오전 10:00:00	5.2
26. 7. 16. 오전 10:30:00	5.2
26. 7. 16. 오전 11:00:00	5.2
26. 7. 16. 오전 11:30:00	5.1
26. 7. 16. 오후 12:00:00	5.1
26. 7. 16. 오후 12:30:00	5.1
26. 7. 16. 오후 1:00:00	5.2
26. 7. 16. 오후 1:30:00	5.0
26. 7. 16. 오후 2:00:00	5.1
26. 7. 16. 오후 2:30:00	5.1
26. 7. 16. 오후 3:00:00	5.1
26. 7. 16. 오후 3:30:00	5.1
26. 7. 16. 오후 4:00:00	5.2
26. 7. 16. 오후 4:30:00	5.2
26. 7. 16. 오후 5:00:00	5.2
26. 7. 16. 오후 5:30:00	5.2
26. 7. 16. 오후 6:00:00	5.2
26. 7. 16. 오후 6:30:00	5.2
26. 7. 16. 오후 7:00:00	5.1
26. 7. 16. 오후 7:30:00	5.1
26. 7. 16. 오후 8:00:00	5.0
26. 7. 16. 오후 8:30:00	5.1
26. 7. 16. 오후 9:00:00	5.1
26. 7. 16. 오후 9:30:00	5.0
26. 7. 16. 오후 10:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 16. 오후 10:30:00	5.1
26. 7. 16. 오후 11:00:00	5.1
26. 7. 16. 오후 11:30:00	5.0
26. 7. 17. 오전 12:00:00	5.1
26. 7. 17. 오전 12:30:00	5.1
26. 7. 17. 오전 1:00:00	5.0
26. 7. 17. 오전 1:30:00	5.1
26. 7. 17. 오전 2:00:00	5.0
26. 7. 17. 오전 2:30:00	5.1
26. 7. 17. 오전 3:00:00	5.2
26. 7. 17. 오전 3:30:00	5.1
26. 7. 17. 오전 4:00:00	5.0
26. 7. 17. 오전 4:30:00	5.1
26. 7. 17. 오전 5:00:00	5.1
26. 7. 17. 오전 5:30:00	5.0
26. 7. 17. 오전 6:00:00	5.1
26. 7. 17. 오전 6:30:00	5.1
26. 7. 17. 오전 7:00:00	5.0
26. 7. 17. 오전 7:30:00	5.1
26. 7. 17. 오전 8:00:00	5.2
26. 7. 17. 오전 8:30:00	5.1
26. 7. 17. 오전 9:00:00	5.1
26. 7. 17. 오전 9:30:00	5.1
26. 7. 17. 오전 10:00:00	5.0
26. 7. 17. 오전 10:30:00	5.0
26. 7. 17. 오전 11:00:00	5.1
26. 7. 17. 오전 11:30:00	5.2
26. 7. 17. 오후 12:00:00	5.2
26. 7. 17. 오후 12:30:00	5.1
26. 7. 17. 오후 1:00:00	5.1
26. 7. 17. 오후 1:30:00	5.2
26. 7. 17. 오후 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 17. 오후 2:30:00	5.0
26. 7. 17. 오후 3:00:00	5.0
26. 7. 17. 오후 3:30:00	5.1
26. 7. 17. 오후 4:00:00	5.1
26. 7. 17. 오후 4:30:00	5.0
26. 7. 17. 오후 5:00:00	5.1
26. 7. 17. 오후 5:30:00	5.1
26. 7. 17. 오후 6:00:00	5.0
26. 7. 17. 오후 6:30:00	5.1
26. 7. 17. 오후 7:00:00	5.2
26. 7. 17. 오후 7:30:00	5.2
26. 7. 17. 오후 8:00:00	5.2
26. 7. 17. 오후 8:30:00	5.3
26. 7. 17. 오후 9:00:00	5.3
26. 7. 17. 오후 9:30:00	5.3
26. 7. 17. 오후 10:00:00	5.3
26. 7. 17. 오후 10:30:00	5.4
26. 7. 17. 오후 11:00:00	5.3
26. 7. 17. 오후 11:30:00	5.3
26. 7. 18. 오전 12:00:00	5.4
26. 7. 18. 오전 12:30:00	5.3
26. 7. 18. 오전 1:00:00	5.3
26. 7. 18. 오전 1:30:00	5.3
26. 7. 18. 오전 2:00:00	5.2
26. 7. 18. 오전 2:30:00	5.2
26. 7. 18. 오전 3:00:00	5.2
26. 7. 18. 오전 3:30:00	5.1
26. 7. 18. 오전 4:00:00	5.1
26. 7. 18. 오전 4:30:00	5.1
26. 7. 18. 오전 5:00:00	5.1
26. 7. 18. 오전 5:30:00	5.1
26. 7. 18. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 18. 오전 6:30:00	5.1
26. 7. 18. 오전 7:00:00	5.0
26. 7. 18. 오전 7:30:00	5.2
26. 7. 18. 오전 8:00:00	5.0
26. 7. 18. 오전 8:30:00	4.9
26. 7. 18. 오전 9:00:00	5.0
26. 7. 18. 오전 9:30:00	5.0
26. 7. 18. 오전 10:00:00	5.0
26. 7. 18. 오전 10:30:00	4.9
26. 7. 18. 오전 11:00:00	5.0
26. 7. 18. 오전 11:30:00	5.0
26. 7. 18. 오후 12:00:00	5.0
26. 7. 18. 오후 12:30:00	5.0
26. 7. 18. 오후 1:00:00	5.1
26. 7. 18. 오후 1:30:00	5.1
26. 7. 18. 오후 2:00:00	5.1
26. 7. 18. 오후 2:30:00	5.1
26. 7. 18. 오후 3:00:00	5.0
26. 7. 18. 오후 3:30:00	5.1
26. 7. 18. 오후 4:00:00	5.1
26. 7. 18. 오후 4:30:00	5.0
26. 7. 18. 오후 5:00:00	5.0
26. 7. 18. 오후 5:30:00	5.1
26. 7. 18. 오후 6:00:00	5.0
26. 7. 18. 오후 6:30:00	5.0
26. 7. 18. 오후 7:00:00	5.1
26. 7. 18. 오후 7:30:00	5.2
26. 7. 18. 오후 8:00:00	5.1
26. 7. 18. 오후 8:30:00	5.1
26. 7. 18. 오후 9:00:00	5.2
26. 7. 18. 오후 9:30:00	5.3
26. 7. 18. 오후 10:00:00	5.3



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 18. 오후 10:30:00	5.2
26. 7. 18. 오후 11:00:00	5.1
26. 7. 18. 오후 11:30:00	5.1
26. 7. 19. 오전 12:00:00	5.1
26. 7. 19. 오전 12:30:00	5.0
26. 7. 19. 오전 1:00:00	5.0
26. 7. 19. 오전 1:30:00	5.1
26. 7. 19. 오전 2:00:00	5.0
26. 7. 19. 오전 2:30:00	5.0
26. 7. 19. 오전 3:00:00	5.0
26. 7. 19. 오전 3:30:00	5.0
26. 7. 19. 오전 4:00:00	5.0
26. 7. 19. 오전 4:30:00	5.0
26. 7. 19. 오전 5:00:00	5.0
26. 7. 19. 오전 5:30:00	5.0
26. 7. 19. 오전 6:00:00	5.0
26. 7. 19. 오전 6:30:00	5.1
26. 7. 19. 오전 7:00:00	5.1
26. 7. 19. 오전 7:30:00	5.0
26. 7. 19. 오전 8:00:00	5.1
26. 7. 19. 오전 8:30:00	5.1
26. 7. 19. 오전 9:00:00	5.1
26. 7. 19. 오전 9:30:00	5.0
26. 7. 19. 오전 10:00:00	5.0
26. 7. 19. 오전 10:30:00	5.1
26. 7. 19. 오전 11:00:00	5.0
26. 7. 19. 오전 11:30:00	5.0
26. 7. 19. 오후 12:00:00	5.1
26. 7. 19. 오후 12:30:00	5.1
26. 7. 19. 오후 1:00:00	5.0
26. 7. 19. 오후 1:30:00	5.0
26. 7. 19. 오후 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 19. 오후 2:30:00	5.0
26. 7. 19. 오후 3:00:00	4.9
26. 7. 19. 오후 3:30:00	5.0
26. 7. 19. 오후 4:00:00	5.1
26. 7. 19. 오후 4:30:00	4.9
26. 7. 19. 오후 5:00:00	5.0
26. 7. 19. 오후 5:30:00	5.1
26. 7. 19. 오후 6:00:00	4.9
26. 7. 19. 오후 6:30:00	5.0
26. 7. 19. 오후 7:00:00	5.1
26. 7. 19. 오후 7:30:00	5.0
26. 7. 19. 오후 8:00:00	5.0
26. 7. 19. 오후 8:30:00	5.1
26. 7. 19. 오후 9:00:00	5.1
26. 7. 19. 오후 9:30:00	5.0
26. 7. 19. 오후 10:00:00	5.1
26. 7. 19. 오후 10:30:00	5.1
26. 7. 19. 오후 11:00:00	5.0
26. 7. 19. 오후 11:30:00	5.1
26. 7. 20. 오전 12:00:00	5.1
26. 7. 20. 오전 12:30:00	5.1
26. 7. 20. 오전 1:00:00	5.1
26. 7. 20. 오전 1:30:00	5.1
26. 7. 20. 오전 2:00:00	5.2
26. 7. 20. 오전 2:30:00	5.1
26. 7. 20. 오전 3:00:00	5.1
26. 7. 20. 오전 3:30:00	5.2
26. 7. 20. 오전 4:00:00	5.1
26. 7. 20. 오전 4:30:00	5.1
26. 7. 20. 오전 5:00:00	5.2
26. 7. 20. 오전 5:30:00	5.1
26. 7. 20. 오전 6:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 20. 오전 6:30:00	5.1
26. 7. 20. 오전 7:00:00	5.2
26. 7. 20. 오전 7:30:00	5.1
26. 7. 20. 오전 8:00:00	5.1
26. 7. 20. 오전 8:30:00	5.0
26. 7. 20. 오전 9:00:00	5.1
26. 7. 20. 오전 9:30:00	5.1
26. 7. 20. 오전 10:00:00	5.1
26. 7. 20. 오전 10:30:00	5.0
26. 7. 20. 오전 11:00:00	5.0
26. 7. 20. 오전 11:30:00	5.1
26. 7. 20. 오후 12:00:00	5.1
26. 7. 20. 오후 12:30:00	5.0
26. 7. 20. 오후 1:00:00	5.0
26. 7. 20. 오후 1:30:00	5.1
26. 7. 20. 오후 2:00:00	5.1
26. 7. 20. 오후 2:30:00	5.1
26. 7. 20. 오후 3:00:00	5.1
26. 7. 20. 오후 3:30:00	5.0
26. 7. 20. 오후 4:00:00	5.0
26. 7. 20. 오후 4:30:00	5.1
26. 7. 20. 오후 5:00:00	5.1
26. 7. 20. 오후 5:30:00	5.1
26. 7. 20. 오후 6:00:00	5.1
26. 7. 20. 오후 6:30:00	5.0
26. 7. 20. 오후 7:00:00	5.0
26. 7. 20. 오후 7:30:00	5.1
26. 7. 20. 오후 8:00:00	5.1
26. 7. 20. 오후 8:30:00	5.1
26. 7. 20. 오후 9:00:00	4.9
26. 7. 20. 오후 9:30:00	5.0
26. 7. 20. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 20. 오후 10:30:00	4.9
26. 7. 20. 오후 11:00:00	4.9
26. 7. 20. 오후 11:30:00	5.0
26. 7. 21. 오전 12:00:00	5.1
26. 7. 21. 오전 12:30:00	4.9
26. 7. 21. 오전 1:00:00	4.9
26. 7. 21. 오전 1:30:00	5.0
26. 7. 21. 오전 2:00:00	5.0
26. 7. 21. 오전 2:30:00	5.0
26. 7. 21. 오전 3:00:00	5.0
26. 7. 21. 오전 3:30:00	5.1
26. 7. 21. 오전 4:00:00	5.1
26. 7. 21. 오전 4:30:00	5.1
26. 7. 21. 오전 5:00:00	5.1
26. 7. 21. 오전 5:30:00	5.1
26. 7. 21. 오전 6:00:00	5.0
26. 7. 21. 오전 6:30:00	5.0
26. 7. 21. 오전 7:00:00	5.0
26. 7. 21. 오전 7:30:00	5.1
26. 7. 21. 오전 8:00:00	5.0
26. 7. 21. 오전 8:30:00	5.0
26. 7. 21. 오전 9:00:00	5.0
26. 7. 21. 오전 9:30:00	5.1
26. 7. 21. 오전 10:00:00	5.1
26. 7. 21. 오전 10:30:00	5.0
26. 7. 21. 오전 11:00:00	5.1
26. 7. 21. 오전 11:30:00	5.2
26. 7. 21. 오후 12:00:00	5.1
26. 7. 21. 오후 12:30:00	5.1
26. 7. 21. 오후 1:00:00	5.1
26. 7. 21. 오후 1:30:00	5.1
26. 7. 21. 오후 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 21. 오후 2:30:00	5.0
26. 7. 21. 오후 3:00:00	5.0
26. 7. 21. 오후 3:30:00	4.9
26. 7. 21. 오후 4:00:00	4.9
26. 7. 21. 오후 4:30:00	4.8
26. 7. 21. 오후 5:00:00	4.9
26. 7. 21. 오후 5:30:00	5.0
26. 7. 21. 오후 6:00:00	4.9
26. 7. 21. 오후 6:30:00	5.0
26. 7. 21. 오후 7:00:00	5.0
26. 7. 21. 오후 7:30:00	4.9
26. 7. 21. 오후 8:00:00	4.9
26. 7. 21. 오후 8:30:00	5.0
26. 7. 21. 오후 9:00:00	4.9
26. 7. 21. 오후 9:30:00	4.9
26. 7. 21. 오후 10:00:00	4.9
26. 7. 21. 오후 10:30:00	4.9
26. 7. 21. 오후 11:00:00	4.8
26. 7. 21. 오후 11:30:00	4.9
26. 7. 22. 오전 12:00:00	4.9
26. 7. 22. 오전 12:30:00	4.9
26. 7. 22. 오전 1:00:00	4.9
26. 7. 22. 오전 1:30:00	4.9
26. 7. 22. 오전 2:00:00	4.9
26. 7. 22. 오전 2:30:00	4.9
26. 7. 22. 오전 3:00:00	4.9
26. 7. 22. 오전 3:30:00	5.0
26. 7. 22. 오전 4:00:00	4.9
26. 7. 22. 오전 4:30:00	4.9
26. 7. 22. 오전 5:00:00	5.0
26. 7. 22. 오전 5:30:00	5.0
26. 7. 22. 오전 6:00:00	4.9



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 22. 오전 6:30:00	4.9
26. 7. 22. 오전 7:00:00	5.0
26. 7. 22. 오전 7:30:00	4.9
26. 7. 22. 오전 8:00:00	4.9
26. 7. 22. 오전 8:30:00	5.0
26. 7. 22. 오전 9:00:00	5.0
26. 7. 22. 오전 9:30:00	4.9
26. 7. 22. 오전 10:00:00	4.8
26. 7. 22. 오전 10:30:00	5.0
26. 7. 22. 오전 11:00:00	5.0
26. 7. 22. 오전 11:30:00	5.0
26. 7. 22. 오후 12:00:00	4.9
26. 7. 22. 오후 12:30:00	4.9
26. 7. 22. 오후 1:00:00	5.0
26. 7. 22. 오후 1:30:00	4.9
26. 7. 22. 오후 2:00:00	4.9
26. 7. 22. 오후 2:30:00	5.0
26. 7. 22. 오후 3:00:00	5.0
26. 7. 22. 오후 3:30:00	4.9
26. 7. 22. 오후 4:00:00	4.9
26. 7. 22. 오후 4:30:00	5.0
26. 7. 22. 오후 5:00:00	5.0
26. 7. 22. 오후 5:30:00	4.9
26. 7. 22. 오후 6:00:00	5.0
26. 7. 22. 오후 6:30:00	5.1
26. 7. 22. 오후 7:00:00	4.9
26. 7. 22. 오후 7:30:00	4.9
26. 7. 22. 오후 8:00:00	5.0
26. 7. 22. 오후 8:30:00	5.1
26. 7. 22. 오후 9:00:00	4.9
26. 7. 22. 오후 9:30:00	5.0
26. 7. 22. 오후 10:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 22. 오후 10:30:00	4.9
26. 7. 22. 오후 11:00:00	5.0
26. 7. 22. 오후 11:30:00	5.0
26. 7. 23. 오전 12:00:00	4.9
26. 7. 23. 오전 12:30:00	5.0
26. 7. 23. 오전 1:00:00	5.0
26. 7. 23. 오전 1:30:00	4.9
26. 7. 23. 오전 2:00:00	4.9
26. 7. 23. 오전 2:30:00	5.0
26. 7. 23. 오전 3:00:00	5.0
26. 7. 23. 오전 3:30:00	5.0
26. 7. 23. 오전 4:00:00	4.9
26. 7. 23. 오전 4:30:00	5.0
26. 7. 23. 오전 5:00:00	4.9
26. 7. 23. 오전 5:30:00	4.9
26. 7. 23. 오전 6:00:00	4.9
26. 7. 23. 오전 6:30:00	5.0
26. 7. 23. 오전 7:00:00	4.9
26. 7. 23. 오전 7:30:00	5.0
26. 7. 23. 오전 8:00:00	5.0
26. 7. 23. 오전 8:30:00	5.0
26. 7. 23. 오전 9:00:00	5.0
26. 7. 23. 오전 9:30:00	5.1
26. 7. 23. 오전 10:00:00	5.0
26. 7. 23. 오전 10:30:00	4.8
26. 7. 23. 오전 11:00:00	4.9
26. 7. 23. 오전 11:30:00	5.0
26. 7. 23. 오후 12:00:00	5.1
26. 7. 23. 오후 12:30:00	5.0
26. 7. 23. 오후 1:00:00	5.0
26. 7. 23. 오후 1:30:00	5.1
26. 7. 23. 오후 2:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 23. 오후 2:30:00	5.1
26. 7. 23. 오후 3:00:00	5.0
26. 7. 23. 오후 3:30:00	5.0
26. 7. 23. 오후 4:00:00	5.0
26. 7. 23. 오후 4:30:00	5.0
26. 7. 23. 오후 5:00:00	5.0
26. 7. 23. 오후 5:30:00	5.0
26. 7. 23. 오후 6:00:00	5.1
26. 7. 23. 오후 6:30:00	5.1
26. 7. 23. 오후 7:00:00	5.0
26. 7. 23. 오후 7:30:00	5.0
26. 7. 23. 오후 8:00:00	5.0
26. 7. 23. 오후 8:30:00	5.1
26. 7. 23. 오후 9:00:00	5.0
26. 7. 23. 오후 9:30:00	4.9
26. 7. 23. 오후 10:00:00	5.0
26. 7. 23. 오후 10:30:00	5.1
26. 7. 23. 오후 11:00:00	5.0
26. 7. 23. 오후 11:30:00	5.0
26. 7. 24. 오전 12:00:00	5.0
26. 7. 24. 오전 12:30:00	5.0
26. 7. 24. 오전 1:00:00	4.9
26. 7. 24. 오전 1:30:00	5.0
26. 7. 24. 오전 2:00:00	5.1
26. 7. 24. 오전 2:30:00	5.2
26. 7. 24. 오전 3:00:00	5.0
26. 7. 24. 오전 3:30:00	5.1
26. 7. 24. 오전 4:00:00	5.0
26. 7. 24. 오전 4:30:00	5.1
26. 7. 24. 오전 5:00:00	5.1
26. 7. 24. 오전 5:30:00	5.0
26. 7. 24. 오전 6:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 24. 오전 6:30:00	5.0
26. 7. 24. 오전 7:00:00	5.0
26. 7. 24. 오전 7:30:00	5.0
26. 7. 24. 오전 8:00:00	5.1
26. 7. 24. 오전 8:30:00	5.1
26. 7. 24. 오전 9:00:00	5.0
26. 7. 24. 오전 9:30:00	5.1
26. 7. 24. 오전 10:00:00	5.1
26. 7. 24. 오전 10:30:00	5.0
26. 7. 24. 오전 11:00:00	5.1
26. 7. 24. 오전 11:30:00	5.1
26. 7. 24. 오후 12:00:00	5.0
26. 7. 24. 오후 12:30:00	5.0
26. 7. 24. 오후 1:00:00	5.0
26. 7. 24. 오후 1:30:00	5.1
26. 7. 24. 오후 2:00:00	5.0
26. 7. 24. 오후 2:30:00	5.0
26. 7. 24. 오후 3:00:00	5.0
26. 7. 24. 오후 3:30:00	5.0
26. 7. 24. 오후 4:00:00	5.0
26. 7. 24. 오후 4:30:00	5.1
26. 7. 24. 오후 5:00:00	5.0
26. 7. 24. 오후 5:30:00	5.0
26. 7. 24. 오후 6:00:00	5.1
26. 7. 24. 오후 6:30:00	5.0
26. 7. 24. 오후 7:00:00	5.1
26. 7. 24. 오후 7:30:00	5.2
26. 7. 24. 오후 8:00:00	5.1
26. 7. 24. 오후 8:30:00	5.1
26. 7. 24. 오후 9:00:00	5.2
26. 7. 24. 오후 9:30:00	5.1
26. 7. 24. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 24. 오후 10:30:00	5.1
26. 7. 24. 오후 11:00:00	5.1
26. 7. 24. 오후 11:30:00	5.1
26. 7. 25. 오전 12:00:00	5.1
26. 7. 25. 오전 12:30:00	5.1
26. 7. 25. 오전 1:00:00	5.0
26. 7. 25. 오전 1:30:00	5.0
26. 7. 25. 오전 2:00:00	5.1
26. 7. 25. 오전 2:30:00	5.0
26. 7. 25. 오전 3:00:00	5.1
26. 7. 25. 오전 3:30:00	5.1
26. 7. 25. 오전 4:00:00	5.0
26. 7. 25. 오전 4:30:00	5.0
26. 7. 25. 오전 5:00:00	5.1
26. 7. 25. 오전 5:30:00	5.1
26. 7. 25. 오전 6:00:00	5.0
26. 7. 25. 오전 6:30:00	5.0
26. 7. 25. 오전 7:00:00	5.1
26. 7. 25. 오전 7:30:00	5.0
26. 7. 25. 오전 8:00:00	5.1
26. 7. 25. 오전 8:30:00	5.1
26. 7. 25. 오전 9:00:00	5.0
26. 7. 25. 오전 9:30:00	5.0
26. 7. 25. 오전 10:00:00	5.1
26. 7. 25. 오전 10:30:00	5.1
26. 7. 25. 오전 11:00:00	5.0
26. 7. 25. 오전 11:30:00	5.0
26. 7. 25. 오후 12:00:00	5.1
26. 7. 25. 오후 12:30:00	5.1
26. 7. 25. 오후 1:00:00	5.0
26. 7. 25. 오후 1:30:00	5.1
26. 7. 25. 오후 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 25. 오후 2:30:00	5.1
26. 7. 25. 오후 3:00:00	5.1
26. 7. 25. 오후 3:30:00	5.0
26. 7. 25. 오후 4:00:00	5.1
26. 7. 25. 오후 4:30:00	5.1
26. 7. 25. 오후 5:00:00	5.1
26. 7. 25. 오후 5:30:00	5.0
26. 7. 25. 오후 6:00:00	5.1
26. 7. 25. 오후 6:30:00	5.1
26. 7. 25. 오후 7:00:00	5.1
26. 7. 25. 오후 7:30:00	5.1
26. 7. 25. 오후 8:00:00	5.1
26. 7. 25. 오후 8:30:00	5.1
26. 7. 25. 오후 9:00:00	5.0
26. 7. 25. 오후 9:30:00	5.1
26. 7. 25. 오후 10:00:00	5.1
26. 7. 25. 오후 10:30:00	5.1
26. 7. 25. 오후 11:00:00	5.0
26. 7. 25. 오후 11:30:00	5.1
26. 7. 26. 오전 12:00:00	5.1
26. 7. 26. 오전 12:30:00	5.1
26. 7. 26. 오전 1:00:00	5.1
26. 7. 26. 오전 1:30:00	5.1
26. 7. 26. 오전 2:00:00	5.1
26. 7. 26. 오전 2:30:00	5.1
26. 7. 26. 오전 3:00:00	5.0
26. 7. 26. 오전 3:30:00	5.1
26. 7. 26. 오전 4:00:00	5.1
26. 7. 26. 오전 4:30:00	5.1
26. 7. 26. 오전 5:00:00	5.1
26. 7. 26. 오전 5:30:00	5.1
26. 7. 26. 오전 6:00:00	5.2



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 26. 오전 6:30:00	5.1
26. 7. 26. 오전 7:00:00	5.2
26. 7. 26. 오전 7:30:00	5.2
26. 7. 26. 오전 8:00:00	5.1
26. 7. 26. 오전 8:30:00	5.0
26. 7. 26. 오전 9:00:00	5.2
26. 7. 26. 오전 9:30:00	5.1
26. 7. 26. 오전 10:00:00	5.0
26. 7. 26. 오전 10:30:00	5.1
26. 7. 26. 오전 11:00:00	5.1
26. 7. 26. 오전 11:30:00	5.0
26. 7. 26. 오후 12:00:00	5.0
26. 7. 26. 오후 12:30:00	5.1
26. 7. 26. 오후 1:00:00	5.1
26. 7. 26. 오후 1:30:00	5.1
26. 7. 26. 오후 2:00:00	5.1
26. 7. 26. 오후 2:30:00	5.2
26. 7. 26. 오후 3:00:00	5.1
26. 7. 26. 오후 3:30:00	5.1
26. 7. 26. 오후 4:00:00	5.0
26. 7. 26. 오후 4:30:00	5.1
26. 7. 26. 오후 5:00:00	5.1
26. 7. 26. 오후 5:30:00	5.0
26. 7. 26. 오후 6:00:00	5.0
26. 7. 26. 오후 6:30:00	5.1
26. 7. 26. 오후 7:00:00	5.1
26. 7. 26. 오후 7:30:00	5.0
26. 7. 26. 오후 8:00:00	5.1
26. 7. 26. 오후 8:30:00	5.1
26. 7. 26. 오후 9:00:00	5.2
26. 7. 26. 오후 9:30:00	5.1
26. 7. 26. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 26. 오후 10:30:00	5.1
26. 7. 26. 오후 11:00:00	5.0
26. 7. 26. 오후 11:30:00	5.1
26. 7. 27. 오전 12:00:00	5.0
26. 7. 27. 오전 12:30:00	5.0
26. 7. 27. 오전 1:00:00	5.0
26. 7. 27. 오전 1:30:00	5.0
26. 7. 27. 오전 2:00:00	5.0
26. 7. 27. 오전 2:30:00	5.1
26. 7. 27. 오전 3:00:00	5.1
26. 7. 27. 오전 3:30:00	5.0
26. 7. 27. 오전 4:00:00	5.0
26. 7. 27. 오전 4:30:00	5.1
26. 7. 27. 오전 5:00:00	5.0
26. 7. 27. 오전 5:30:00	5.0
26. 7. 27. 오전 6:00:00	5.0
26. 7. 27. 오전 6:30:00	5.1
26. 7. 27. 오전 7:00:00	5.0
26. 7. 27. 오전 7:30:00	5.0
26. 7. 27. 오전 8:00:00	5.1
26. 7. 27. 오전 8:30:00	5.0
26. 7. 27. 오전 9:00:00	5.1
26. 7. 27. 오전 9:30:00	5.1
26. 7. 27. 오전 10:00:00	5.0
26. 7. 27. 오전 10:30:00	5.0
26. 7. 27. 오전 11:00:00	5.0
26. 7. 27. 오전 11:30:00	5.0
26. 7. 27. 오후 12:00:00	4.9
26. 7. 27. 오후 12:30:00	5.0
26. 7. 27. 오후 1:00:00	5.1
26. 7. 27. 오후 1:30:00	5.0
26. 7. 27. 오후 2:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 27. 오후 2:30:00	5.0
26. 7. 27. 오후 3:00:00	5.1
26. 7. 27. 오후 3:30:00	5.1
26. 7. 27. 오후 4:00:00	5.2
26. 7. 27. 오후 4:30:00	5.1
26. 7. 27. 오후 5:00:00	5.0
26. 7. 27. 오후 5:30:00	5.1
26. 7. 27. 오후 6:00:00	5.2
26. 7. 27. 오후 6:30:00	5.1
26. 7. 27. 오후 7:00:00	5.0
26. 7. 27. 오후 7:30:00	5.1
26. 7. 27. 오후 8:00:00	5.1
26. 7. 27. 오후 8:30:00	5.0
26. 7. 27. 오후 9:00:00	5.1
26. 7. 27. 오후 9:30:00	5.1
26. 7. 27. 오후 10:00:00	5.1
26. 7. 27. 오후 10:30:00	5.2
26. 7. 27. 오후 11:00:00	5.1
26. 7. 27. 오후 11:30:00	5.0
26. 7. 28. 오전 12:00:00	5.1
26. 7. 28. 오전 12:30:00	5.1
26. 7. 28. 오전 1:00:00	5.0
26. 7. 28. 오전 1:30:00	5.1
26. 7. 28. 오전 2:00:00	5.1
26. 7. 28. 오전 2:30:00	5.1
26. 7. 28. 오전 3:00:00	5.0
26. 7. 28. 오전 3:30:00	5.0
26. 7. 28. 오전 4:00:00	5.1
26. 7. 28. 오전 4:30:00	5.2
26. 7. 28. 오전 5:00:00	5.0
26. 7. 28. 오전 5:30:00	5.0
26. 7. 28. 오전 6:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 28. 오전 6:30:00	5.0
26. 7. 28. 오전 7:00:00	5.0
26. 7. 28. 오전 7:30:00	5.1
26. 7. 28. 오전 8:00:00	5.1
26. 7. 28. 오전 8:30:00	5.0
26. 7. 28. 오전 9:00:00	5.0
26. 7. 28. 오전 9:30:00	5.1
26. 7. 28. 오전 10:00:00	5.0
26. 7. 28. 오전 10:30:00	5.0
26. 7. 28. 오전 11:00:00	5.1
26. 7. 28. 오전 11:30:00	5.1
26. 7. 28. 오후 12:00:00	5.1
26. 7. 28. 오후 12:30:00	5.0
26. 7. 28. 오후 1:00:00	5.1
26. 7. 28. 오후 1:30:00	5.1
26. 7. 28. 오후 2:00:00	5.0
26. 7. 28. 오후 2:30:00	5.0
26. 7. 28. 오후 3:00:00	5.2
26. 7. 28. 오후 3:30:00	5.2
26. 7. 28. 오후 4:00:00	5.2
26. 7. 28. 오후 4:30:00	5.1
26. 7. 28. 오후 5:00:00	5.1
26. 7. 28. 오후 5:30:00	5.2
26. 7. 28. 오후 6:00:00	5.2
26. 7. 28. 오후 6:30:00	5.1
26. 7. 28. 오후 7:00:00	5.1
26. 7. 28. 오후 7:30:00	5.2
26. 7. 28. 오후 8:00:00	5.2
26. 7. 28. 오후 8:30:00	5.1
26. 7. 28. 오후 9:00:00	5.1
26. 7. 28. 오후 9:30:00	5.1
26. 7. 28. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 28. 오후 10:30:00	5.1
26. 7. 28. 오후 11:00:00	5.1
26. 7. 28. 오후 11:30:00	5.2
26. 7. 29. 오전 12:00:00	5.0
26. 7. 29. 오전 12:30:00	5.0
26. 7. 29. 오전 1:00:00	5.1
26. 7. 29. 오전 1:30:00	5.1
26. 7. 29. 오전 2:00:00	5.0
26. 7. 29. 오전 2:30:00	5.1
26. 7. 29. 오전 3:00:00	5.1
26. 7. 29. 오전 3:30:00	5.0
26. 7. 29. 오전 4:00:00	5.2
26. 7. 29. 오전 4:30:00	5.1
26. 7. 29. 오전 5:00:00	5.0
26. 7. 29. 오전 5:30:00	5.1
26. 7. 29. 오전 6:00:00	5.1
26. 7. 29. 오전 6:30:00	5.0
26. 7. 29. 오전 7:00:00	5.0
26. 7. 29. 오전 7:30:00	5.1
26. 7. 29. 오전 8:00:00	5.0
26. 7. 29. 오전 8:30:00	5.0
26. 7. 29. 오전 9:00:00	5.1
26. 7. 29. 오전 9:30:00	5.0
26. 7. 29. 오전 10:00:00	5.1
26. 7. 29. 오전 10:30:00	5.1
26. 7. 29. 오전 11:00:00	5.1
26. 7. 29. 오전 11:30:00	5.1
26. 7. 29. 오후 12:00:00	5.1
26. 7. 29. 오후 12:30:00	5.1
26. 7. 29. 오후 1:00:00	5.1
26. 7. 29. 오후 1:30:00	5.2
26. 7. 29. 오후 2:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 29. 오후 2:30:00	5.1
26. 7. 29. 오후 3:00:00	5.2
26. 7. 29. 오후 3:30:00	5.3
26. 7. 29. 오후 4:00:00	5.1
26. 7. 29. 오후 4:30:00	5.2
26. 7. 29. 오후 5:00:00	5.1
26. 7. 29. 오후 5:30:00	5.0
26. 7. 29. 오후 6:00:00	5.0
26. 7. 29. 오후 6:30:00	5.2
26. 7. 29. 오후 7:00:00	5.1
26. 7. 29. 오후 7:30:00	5.0
26. 7. 29. 오후 8:00:00	5.0
26. 7. 29. 오후 8:30:00	5.1
26. 7. 29. 오후 9:00:00	5.0
26. 7. 29. 오후 9:30:00	5.0
26. 7. 29. 오후 10:00:00	5.1
26. 7. 29. 오후 10:30:00	5.1
26. 7. 29. 오후 11:00:00	5.1
26. 7. 29. 오후 11:30:00	5.1
26. 7. 30. 오전 12:00:00	5.1
26. 7. 30. 오전 12:30:00	5.0
26. 7. 30. 오전 1:00:00	5.0
26. 7. 30. 오전 1:30:00	5.1
26. 7. 30. 오전 2:00:00	5.1
26. 7. 30. 오전 2:30:00	5.0
26. 7. 30. 오전 3:00:00	5.1
26. 7. 30. 오전 3:30:00	5.1
26. 7. 30. 오전 4:00:00	5.0
26. 7. 30. 오전 4:30:00	5.1
26. 7. 30. 오전 5:00:00	5.1
26. 7. 30. 오전 5:30:00	5.0
26. 7. 30. 오전 6:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 30. 오전 6:30:00	5.1
26. 7. 30. 오전 7:00:00	5.0
26. 7. 30. 오전 7:30:00	5.1
26. 7. 30. 오전 8:00:00	5.1
26. 7. 30. 오전 8:30:00	5.0
26. 7. 30. 오전 9:00:00	5.0
26. 7. 30. 오전 9:30:00	5.1
26. 7. 30. 오전 10:00:00	5.1
26. 7. 30. 오전 10:30:00	5.0
26. 7. 30. 오전 11:00:00	5.1
26. 7. 30. 오전 11:30:00	5.1
26. 7. 30. 오후 12:00:00	5.0
26. 7. 30. 오후 12:30:00	5.0
26. 7. 30. 오후 1:00:00	5.0
26. 7. 30. 오후 1:30:00	5.1
26. 7. 30. 오후 2:00:00	5.0
26. 7. 30. 오후 2:30:00	5.1
26. 7. 30. 오후 3:00:00	5.1
26. 7. 30. 오후 3:30:00	5.1
26. 7. 30. 오후 4:00:00	5.0
26. 7. 30. 오후 4:30:00	5.0
26. 7. 30. 오후 5:00:00	5.1
26. 7. 30. 오후 5:30:00	5.1
26. 7. 30. 오후 6:00:00	5.1
26. 7. 30. 오후 6:30:00	5.1
26. 7. 30. 오후 7:00:00	5.1
26. 7. 30. 오후 7:30:00	5.0
26. 7. 30. 오후 8:00:00	5.1
26. 7. 30. 오후 8:30:00	5.1
26. 7. 30. 오후 9:00:00	5.0
26. 7. 30. 오후 9:30:00	5.1
26. 7. 30. 오후 10:00:00	5.1



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 30. 오후 10:30:00	5.1
26. 7. 30. 오후 11:00:00	5.2
26. 7. 30. 오후 11:30:00	5.1
26. 7. 31. 오전 12:00:00	5.0
26. 7. 31. 오전 12:30:00	5.1
26. 7. 31. 오전 1:00:00	5.1
26. 7. 31. 오전 1:30:00	5.0
26. 7. 31. 오전 2:00:00	5.1
26. 7. 31. 오전 2:30:00	5.1
26. 7. 31. 오전 3:00:00	5.1
26. 7. 31. 오전 3:30:00	5.1
26. 7. 31. 오전 4:00:00	5.2
26. 7. 31. 오전 4:30:00	5.1
26. 7. 31. 오전 5:00:00	5.1
26. 7. 31. 오전 5:30:00	5.2
26. 7. 31. 오전 6:00:00	5.1
26. 7. 31. 오전 6:30:00	5.0
26. 7. 31. 오전 7:00:00	5.1
26. 7. 31. 오전 7:30:00	5.2
26. 7. 31. 오전 8:00:00	5.1
26. 7. 31. 오전 8:30:00	5.1
26. 7. 31. 오전 9:00:00	5.2
26. 7. 31. 오전 9:30:00	5.1
26. 7. 31. 오전 10:00:00	5.0
26. 7. 31. 오전 10:30:00	5.1
26. 7. 31. 오전 11:00:00	5.1
26. 7. 31. 오전 11:30:00	5.2
26. 7. 31. 오후 12:00:00	5.1
26. 7. 31. 오후 12:30:00	5.0
26. 7. 31. 오후 1:00:00	5.1
26. 7. 31. 오후 1:30:00	5.1
26. 7. 31. 오후 2:00:00	5.0



IP Storage Temperature Log_Monthly report

2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고A 2~8°C)_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 7. 31. 오후 2:30:00	5.1
26. 7. 31. 오후 3:00:00	5.1
26. 7. 31. 오후 3:30:00	5.1
26. 7. 31. 오후 4:00:00	5.0
26. 7. 31. 오후 4:30:00	5.1
26. 7. 31. 오후 5:00:00	5.1
26. 7. 31. 오후 5:30:00	5.0
26. 7. 31. 오후 6:00:00	5.0
26. 7. 31. 오후 6:30:00	5.1
26. 7. 31. 오후 7:00:00	5.1
26. 7. 31. 오후 7:30:00	5.0
26. 7. 31. 오후 8:00:00	5.0
26. 7. 31. 오후 8:30:00	5.1
26. 7. 31. 오후 9:00:00	5.0
26. 7. 31. 오후 9:30:00	5.0
26. 7. 31. 오후 10:00:00	5.2
26. 7. 31. 오후 10:30:00	5.2
26. 7. 31. 오후 11:00:00	5.0
26. 7. 31. 오후 11:30:00	5.1



IP Storage Temperature Log_Monthly report
2026년 7월 1일 오전 12시 0분 0초 KST - 2026년 8월 1일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 8월 1일 오전 6시 2분 46초 KST

알람 테이블

다음에서 발생 [Asia/Seoul]	마감일 [Asia/Seoul]	측정 포인트	알림 유형	상세 정보
26. 7. 26. 오전 4:41:54	26. 7. 26. 오전 5:07:07	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용의약품(냉장고A 2~8℃)_CH1 동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용의약품(냉장고B 2~8℃)_CH2	시스템 알림	연결 중단 일련번호: 54857013