

IP STORAGE TEMPERATURE LOG

Period of Record	Year. Month	2026.9	
Storage	Location	동국대학교일산병원 임상시험센터 약국	9134Room
	Description	Pharmaceutical Refrigerator	냉장고B
	Manufacturer	SGM	
	Model Name	CS-600P	
	Serial Number	K600P0302510M	
	Temperature Range	2~8℃	
	Setting Point	4.0℃	
Thermometer (including alarm system)	Description	Wi-Fi data logger	
	Manufacturer	Testo Korea	
	Model Name	Saveris2-T3	
	Serial Number	0054704396	
	Sensor Channel	Channel 1	
Thermometer Calibration Certification	Min/Max Record	Monthly report	This month Min (3.6)℃, Max(4.3)℃
	Date of Calibration	2025.11.25	
	Date of Calibration Due	2026.11.25	
Alarm System	Certificate No.	A25-2356-001	
	Alarm threshold	Lower : 2.5℃, Upper : 7.5℃, immediate alarm without delay	
	Alarm Method	Mail, SMS to Pharmacists	


Date	Time	Temperature(℃)	Deviation	Verified by	Comment
2026.09.01	p:20	3.7	☐ Y / ☒ N	ml	
2026.09.02	p:10	3.8	☐ Y / ☒ N	ml	
2026.09.03	p:20	3.8	☐ Y / ☒ N	ml	
2026.09.04	p:20	3.9	☐ Y / ☒ N	ml	
2026.09.05			☐ Y / ☐ N		
2026.09.06			☐ Y / ☐ N		
2026.09.07	p:20	3.9	☐ Y / ☒ N	765ml	
2026.09.08	p:20	4.0	☐ Y / ☒ N	ml	
2026.09.09	p:20	4.0	☐ Y / ☒ N	ml	
2026.09.10	p:20	3.9	☐ Y / ☒ N	ml	
2026.09.11	p:20	3.7	☐ Y / ☒ N	ml	
2026.09.12			☐ Y / ☐ N		
2026.09.13			☐ Y / ☐ N		
2026.09.14	p:20	3.8	☐ Y / ☒ N	ml	
2026.09.15	p:20	3.7	☐ Y / ☒ N	ml	
2026.09.16	p:20	3.7	☐ Y / ☒ N	ml	
2026.09.17	p:20	3.6	☐ Y / ☒ N	ml	
2026.09.18	p:20	3.7	☐ Y / ☒ N	ml	
2026.09.19			☐ Y / ☐ N		
2026.09.20			☐ Y / ☐ N		
2026.09.21	p:20	3.8	☐ Y / ☒ N	ml	
2026.09.22	p:15	3.9	☐ Y / ☒ N	ml	
2026.09.23	p:20	4.0	☐ Y / ☒ N	ml	
2026.09.24			☐ Y / ☐ N		추석연휴
2026.09.25			☐ Y / ☐ N		추석
2026.09.26			☐ Y / ☐ N		
2026.09.27			☐ Y / ☐ N		
2026.09.28	p:20	3.8	☐ Y / ☒ N	ml	
2026.09.29	p:20	3.9	☐ Y / ☒ N	ml	
2026.09.30	p:10	4.1	☐ Y / ☒ N	ml	
			☐ Y / ☐ N		



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 1일 오전 12시 0분 0초 KST - 2026년 9월 2일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 2일 오전 6시 0분 17초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.8
26. 9. 1. 오전 12:00:00	3.8
26. 9. 1. 오전 12:30:00	3.8
26. 9. 1. 오전 1:00:00	3.9
26. 9. 1. 오전 1:30:00	3.7
26. 9. 1. 오전 2:00:00	3.7
26. 9. 1. 오전 2:30:00	3.8
26. 9. 1. 오전 3:00:00	3.9
26. 9. 1. 오전 3:30:00	3.9
26. 9. 1. 오전 4:00:00	3.8
26. 9. 1. 오전 4:30:00	3.7
26. 9. 1. 오전 5:00:00	3.8
26. 9. 1. 오전 5:30:00	3.9
26. 9. 1. 오전 6:00:00	3.8
26. 9. 1. 오전 6:30:00	3.7
26. 9. 1. 오전 7:00:00	3.7
26. 9. 1. 오전 7:30:00	3.8
26. 9. 1. 오전 8:00:00	3.8
26. 9. 1. 오전 8:30:00	3.7
26. 9. 1. 오전 9:00:00	3.7
26. 9. 1. 오전 9:30:00	3.8
26. 9. 1. 오전 10:00:00	3.8
26. 9. 1. 오전 10:30:00	3.7
26. 9. 1. 오전 11:00:00	3.7
26. 9. 1. 오전 11:30:00	3.8
26. 9. 1. 오후 12:00:00	3.8
26. 9. 1. 오후 12:30:00	4.0
26. 9. 1. 오후 1:00:00	3.9
26. 9. 1. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 1일 오전 12시 0분 0초 KST - 2026년 9월 2일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 2일 오전 6시 0분 17초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 1. 오후 2:00:00	3.8
26. 9. 1. 오후 2:30:00	3.9
26. 9. 1. 오후 3:00:00	3.7
26. 9. 1. 오후 3:30:00	3.7
26. 9. 1. 오후 4:00:00	3.8
26. 9. 1. 오후 4:30:00	3.9
26. 9. 1. 오후 5:00:00	3.8
26. 9. 1. 오후 5:30:00	3.7
26. 9. 1. 오후 6:00:00	3.8
26. 9. 1. 오후 6:30:00	3.9
26. 9. 1. 오후 7:00:00	3.8
26. 9. 1. 오후 7:30:00	3.7
26. 9. 1. 오후 8:00:00	3.8
26. 9. 1. 오후 8:30:00	3.8
26. 9. 1. 오후 9:00:00	3.8
26. 9. 1. 오후 9:30:00	3.8
26. 9. 1. 오후 10:00:00	3.7
26. 9. 1. 오후 10:30:00	3.8
26. 9. 1. 오후 11:00:00	3.8
26. 9. 1. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 2일 오전 12시 0분 0초 KST - 2026년 9월 3일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 3일 오전 6시 0분 15초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	3.9
평균	3.8
26. 9. 2. 오전 12:00:00	3.7
26. 9. 2. 오전 12:30:00	3.8
26. 9. 2. 오전 1:00:00	3.9
26. 9. 2. 오전 1:30:00	3.7
26. 9. 2. 오전 2:00:00	3.7
26. 9. 2. 오전 2:30:00	3.9
26. 9. 2. 오전 3:00:00	3.9
26. 9. 2. 오전 3:30:00	3.7
26. 9. 2. 오전 4:00:00	3.7
26. 9. 2. 오전 4:30:00	3.7
26. 9. 2. 오전 5:00:00	3.9
26. 9. 2. 오전 5:30:00	3.8
26. 9. 2. 오전 6:00:00	3.7
26. 9. 2. 오전 6:30:00	3.8
26. 9. 2. 오전 7:00:00	3.9
26. 9. 2. 오전 7:30:00	3.8
26. 9. 2. 오전 8:00:00	3.7
26. 9. 2. 오전 8:30:00	3.7
26. 9. 2. 오전 9:00:00	3.8
26. 9. 2. 오전 9:30:00	3.8
26. 9. 2. 오전 10:00:00	3.8
26. 9. 2. 오전 10:30:00	3.7
26. 9. 2. 오전 11:00:00	3.7
26. 9. 2. 오전 11:30:00	3.8
26. 9. 2. 오후 12:00:00	3.8
26. 9. 2. 오후 12:30:00	3.7
26. 9. 2. 오후 1:00:00	3.7
26. 9. 2. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 2일 오전 12시 0분 0초 KST - 2026년 9월 3일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 3일 오전 6시 0분 15초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 2. 오후 2:00:00	3.8
26. 9. 2. 오후 2:30:00	3.8
26. 9. 2. 오후 3:00:00	3.7
26. 9. 2. 오후 3:30:00	3.7
26. 9. 2. 오후 4:00:00	3.8
26. 9. 2. 오후 4:30:00	3.7
26. 9. 2. 오후 5:00:00	3.7
26. 9. 2. 오후 5:30:00	3.8
26. 9. 2. 오후 6:00:00	3.8
26. 9. 2. 오후 6:30:00	3.9
26. 9. 2. 오후 7:00:00	3.9
26. 9. 2. 오후 7:30:00	3.9
26. 9. 2. 오후 8:00:00	3.9
26. 9. 2. 오후 8:30:00	3.8
26. 9. 2. 오후 9:00:00	3.8
26. 9. 2. 오후 9:30:00	3.8
26. 9. 2. 오후 10:00:00	3.9
26. 9. 2. 오후 10:30:00	3.9
26. 9. 2. 오후 11:00:00	3.8
26. 9. 2. 오후 11:30:00	3.7



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 3일 오전 12시 0분 0초 KST - 2026년 9월 4일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 4일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 3. 오전 12:00:00	3.8
26. 9. 3. 오전 12:30:00	3.8
26. 9. 3. 오전 1:00:00	3.8
26. 9. 3. 오전 1:30:00	3.8
26. 9. 3. 오전 2:00:00	3.7
26. 9. 3. 오전 2:30:00	3.8
26. 9. 3. 오전 3:00:00	3.9
26. 9. 3. 오전 3:30:00	3.9
26. 9. 3. 오전 4:00:00	3.8
26. 9. 3. 오전 4:30:00	3.8
26. 9. 3. 오전 5:00:00	3.8
26. 9. 3. 오전 5:30:00	3.9
26. 9. 3. 오전 6:00:00	3.8
26. 9. 3. 오전 6:30:00	3.8
26. 9. 3. 오전 7:00:00	3.8
26. 9. 3. 오전 7:30:00	3.8
26. 9. 3. 오전 8:00:00	3.7
26. 9. 3. 오전 8:30:00	3.7
26. 9. 3. 오전 9:00:00	3.8
26. 9. 3. 오전 9:30:00	3.8
26. 9. 3. 오전 10:00:00	3.7
26. 9. 3. 오전 10:30:00	3.8
26. 9. 3. 오전 11:00:00	3.8
26. 9. 3. 오전 11:30:00	3.9
26. 9. 3. 오후 12:00:00	3.8
26. 9. 3. 오후 12:30:00	3.8
26. 9. 3. 오후 1:00:00	3.9
26. 9. 3. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 3일 오전 12시 0분 0초 KST - 2026년 9월 4일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 4일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 3. 오후 2:00:00	3.8
26. 9. 3. 오후 2:30:00	3.9
26. 9. 3. 오후 3:00:00	3.8
26. 9. 3. 오후 3:30:00	3.8
26. 9. 3. 오후 4:00:00	3.9
26. 9. 3. 오후 4:30:00	3.9
26. 9. 3. 오후 5:00:00	4.0
26. 9. 3. 오후 5:30:00	4.0
26. 9. 3. 오후 6:00:00	4.0
26. 9. 3. 오후 6:30:00	4.0
26. 9. 3. 오후 7:00:00	4.1
26. 9. 3. 오후 7:30:00	4.1
26. 9. 3. 오후 8:00:00	4.1
26. 9. 3. 오후 8:30:00	4.1
26. 9. 3. 오후 9:00:00	4.0
26. 9. 3. 오후 9:30:00	4.0
26. 9. 3. 오후 10:00:00	4.0
26. 9. 3. 오후 10:30:00	4.0
26. 9. 3. 오후 11:00:00	4.0
26. 9. 3. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 4일 오전 12시 0분 0초 KST - 2026년 9월 5일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 5일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	4.0
26. 9. 4. 오전 12:00:00	4.0
26. 9. 4. 오전 12:30:00	3.9
26. 9. 4. 오전 1:00:00	4.0
26. 9. 4. 오전 1:30:00	3.9
26. 9. 4. 오전 2:00:00	4.0
26. 9. 4. 오전 2:30:00	4.0
26. 9. 4. 오전 3:00:00	4.0
26. 9. 4. 오전 3:30:00	4.0
26. 9. 4. 오전 4:00:00	4.0
26. 9. 4. 오전 4:30:00	4.0
26. 9. 4. 오전 5:00:00	4.0
26. 9. 4. 오전 5:30:00	4.0
26. 9. 4. 오전 6:00:00	4.0
26. 9. 4. 오전 6:30:00	4.0
26. 9. 4. 오전 7:00:00	3.9
26. 9. 4. 오전 7:30:00	4.0
26. 9. 4. 오전 8:00:00	4.0
26. 9. 4. 오전 8:30:00	4.0
26. 9. 4. 오전 9:00:00	4.0
26. 9. 4. 오전 9:30:00	4.0
26. 9. 4. 오전 10:00:00	4.0
26. 9. 4. 오전 10:30:00	4.0
26. 9. 4. 오전 11:00:00	4.1
26. 9. 4. 오전 11:30:00	3.9
26. 9. 4. 오후 12:00:00	3.9
26. 9. 4. 오후 12:30:00	3.9
26. 9. 4. 오후 1:00:00	3.8
26. 9. 4. 오후 1:30:00	3.7



IP Storage Temperature Log_Daily report

2026년 9월 4일 오전 12시 0분 0초 KST - 2026년 9월 5일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 5일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 4. 오후 2:00:00	3.8
26. 9. 4. 오후 2:30:00	3.9
26. 9. 4. 오후 3:00:00	3.8
26. 9. 4. 오후 3:30:00	3.8
26. 9. 4. 오후 4:00:00	3.9
26. 9. 4. 오후 4:30:00	3.8
26. 9. 4. 오후 5:00:00	3.8
26. 9. 4. 오후 5:30:00	4.0
26. 9. 4. 오후 6:00:00	4.1
26. 9. 4. 오후 6:30:00	4.1
26. 9. 4. 오후 7:00:00	4.1
26. 9. 4. 오후 7:30:00	4.0
26. 9. 4. 오후 8:00:00	4.0
26. 9. 4. 오후 8:30:00	4.0
26. 9. 4. 오후 9:00:00	4.0
26. 9. 4. 오후 9:30:00	4.0
26. 9. 4. 오후 10:00:00	4.0
26. 9. 4. 오후 10:30:00	4.0
26. 9. 4. 오후 11:00:00	4.0
26. 9. 4. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 5일 오전 12시 0분 0초 KST - 2026년 9월 6일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 6일 오전 6시 0분 16초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.9
최대값	4.1
평균	4.0
26. 9. 5. 오전 12:00:00	4.0
26. 9. 5. 오전 12:30:00	4.1
26. 9. 5. 오전 1:00:00	4.1
26. 9. 5. 오전 1:30:00	4.1
26. 9. 5. 오전 2:00:00	4.0
26. 9. 5. 오전 2:30:00	4.0
26. 9. 5. 오전 3:00:00	4.0
26. 9. 5. 오전 3:30:00	3.9
26. 9. 5. 오전 4:00:00	4.0
26. 9. 5. 오전 4:30:00	4.0
26. 9. 5. 오전 5:00:00	4.0
26. 9. 5. 오전 5:30:00	4.0
26. 9. 5. 오전 6:00:00	4.0
26. 9. 5. 오전 6:30:00	4.0
26. 9. 5. 오전 7:00:00	4.1
26. 9. 5. 오전 7:30:00	4.1
26. 9. 5. 오전 8:00:00	4.1
26. 9. 5. 오전 8:30:00	4.0
26. 9. 5. 오전 9:00:00	4.0
26. 9. 5. 오전 9:30:00	3.9
26. 9. 5. 오전 10:00:00	3.9
26. 9. 5. 오전 10:30:00	4.0
26. 9. 5. 오전 11:00:00	4.1
26. 9. 5. 오전 11:30:00	4.0
26. 9. 5. 오후 12:00:00	4.0
26. 9. 5. 오후 12:30:00	3.9
26. 9. 5. 오후 1:00:00	3.9
26. 9. 5. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 5일 오전 12시 0분 0초 KST - 2026년 9월 6일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 6일 오전 6시 0분 16초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 5. 오후 2:00:00	3.9
26. 9. 5. 오후 2:30:00	4.1
26. 9. 5. 오후 3:00:00	4.1
26. 9. 5. 오후 3:30:00	4.0
26. 9. 5. 오후 4:00:00	4.1
26. 9. 5. 오후 4:30:00	4.0
26. 9. 5. 오후 5:00:00	4.1
26. 9. 5. 오후 5:30:00	4.1
26. 9. 5. 오후 6:00:00	4.1
26. 9. 5. 오후 6:30:00	3.9
26. 9. 5. 오후 7:00:00	3.9
26. 9. 5. 오후 7:30:00	4.0
26. 9. 5. 오후 8:00:00	3.9
26. 9. 5. 오후 8:30:00	3.9
26. 9. 5. 오후 9:00:00	3.9
26. 9. 5. 오후 9:30:00	3.9
26. 9. 5. 오후 10:00:00	3.9
26. 9. 5. 오후 10:30:00	3.9
26. 9. 5. 오후 11:00:00	3.9
26. 9. 5. 오후 11:30:00	3.9



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용 의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 6일 오전 12시 0분 0초 KST - 2026년 9월 7일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 7일 오전 6시 1분 49초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.8
최대값	4.1
평균	4.0
26. 9. 6. 오전 12:00:00	3.9
26. 9. 6. 오전 12:30:00	4.0
26. 9. 6. 오전 1:00:00	4.0
26. 9. 6. 오전 1:30:00	4.0
26. 9. 6. 오전 2:00:00	4.0
26. 9. 6. 오전 2:30:00	4.0
26. 9. 6. 오전 3:00:00	3.9
26. 9. 6. 오전 3:30:00	3.9
26. 9. 6. 오전 4:00:00	3.9
26. 9. 6. 오전 4:30:00	3.9
26. 9. 6. 오전 5:00:00	3.8
26. 9. 6. 오전 5:30:00	3.9
26. 9. 6. 오전 6:00:00	3.9
26. 9. 6. 오전 6:30:00	4.0
26. 9. 6. 오전 7:00:00	4.0
26. 9. 6. 오전 7:30:00	4.0
26. 9. 6. 오전 8:00:00	4.0
26. 9. 6. 오전 8:30:00	4.0
26. 9. 6. 오전 9:00:00	3.9
26. 9. 6. 오전 9:30:00	3.9
26. 9. 6. 오전 10:00:00	3.9
26. 9. 6. 오전 10:30:00	3.9
26. 9. 6. 오전 11:00:00	4.0
26. 9. 6. 오전 11:30:00	4.0
26. 9. 6. 오후 12:00:00	3.9
26. 9. 6. 오후 12:30:00	3.9
26. 9. 6. 오후 1:00:00	4.0
26. 9. 6. 오후 1:30:00	4.0



IP Storage Temperature Log_Daily report

2026년 9월 6일 오전 12시 0분 0초 KST - 2026년 9월 7일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 7일 오전 6시 1분 49초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 6. 오후 2:00:00	4.0
26. 9. 6. 오후 2:30:00	4.0
26. 9. 6. 오후 3:00:00	4.0
26. 9. 6. 오후 3:30:00	4.0
26. 9. 6. 오후 4:00:00	3.9
26. 9. 6. 오후 4:30:00	3.9
26. 9. 6. 오후 5:00:00	4.1
26. 9. 6. 오후 5:30:00	4.0
26. 9. 6. 오후 6:00:00	4.0
26. 9. 6. 오후 6:30:00	4.0
26. 9. 6. 오후 7:00:00	4.0
26. 9. 6. 오후 7:30:00	4.0
26. 9. 6. 오후 8:00:00	4.1
26. 9. 6. 오후 8:30:00	4.0
26. 9. 6. 오후 9:00:00	3.8
26. 9. 6. 오후 9:30:00	3.9
26. 9. 6. 오후 10:00:00	3.9
26. 9. 6. 오후 10:30:00	3.8
26. 9. 6. 오후 11:00:00	3.8
26. 9. 6. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 7일 오전 12시 0분 0초 KST - 2026년 9월 8일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 8일 오전 6시 0분 15초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.2
평균	3.9
26. 9. 7. 오전 12:00:00	4.0
26. 9. 7. 오전 12:30:00	3.9
26. 9. 7. 오전 1:00:00	3.9
26. 9. 7. 오전 1:30:00	3.9
26. 9. 7. 오전 2:00:00	3.9
26. 9. 7. 오전 2:30:00	4.0
26. 9. 7. 오전 3:00:00	4.0
26. 9. 7. 오전 3:30:00	4.0
26. 9. 7. 오전 4:00:00	4.1
26. 9. 7. 오전 4:30:00	4.0
26. 9. 7. 오전 5:00:00	4.2
26. 9. 7. 오전 5:30:00	4.1
26. 9. 7. 오전 6:00:00	4.0
26. 9. 7. 오전 6:30:00	3.8
26. 9. 7. 오전 7:00:00	3.9
26. 9. 7. 오전 7:30:00	3.9
26. 9. 7. 오전 8:00:00	3.8
26. 9. 7. 오전 8:30:00	3.8
26. 9. 7. 오전 9:00:00	3.8
26. 9. 7. 오전 9:30:00	3.7
26. 9. 7. 오전 10:00:00	3.8
26. 9. 7. 오전 10:30:00	3.9
26. 9. 7. 오전 11:00:00	3.8
26. 9. 7. 오전 11:30:00	3.8
26. 9. 7. 오후 12:00:00	3.8
26. 9. 7. 오후 12:30:00	3.9
26. 9. 7. 오후 1:00:00	4.0
26. 9. 7. 오후 1:30:00	4.0



IP Storage Temperature Log_Daily report

2026년 9월 7일 오전 12시 0분 0초 KST - 2026년 9월 8일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 8일 오전 6시 0분 15초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 7. 오후 2:00:00	4.0
26. 9. 7. 오후 2:30:00	4.0
26. 9. 7. 오후 3:00:00	4.0
26. 9. 7. 오후 3:30:00	3.9
26. 9. 7. 오후 4:00:00	3.9
26. 9. 7. 오후 4:30:00	3.9
26. 9. 7. 오후 5:00:00	4.0
26. 9. 7. 오후 5:30:00	4.0
26. 9. 7. 오후 6:00:00	4.0
26. 9. 7. 오후 6:30:00	3.9
26. 9. 7. 오후 7:00:00	4.0
26. 9. 7. 오후 7:30:00	3.9
26. 9. 7. 오후 8:00:00	4.0
26. 9. 7. 오후 8:30:00	4.0
26. 9. 7. 오후 9:00:00	4.0
26. 9. 7. 오후 9:30:00	3.8
26. 9. 7. 오후 10:00:00	3.8
26. 9. 7. 오후 10:30:00	3.9
26. 9. 7. 오후 11:00:00	3.7
26. 9. 7. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 8일 오전 12시 0분 0초 KST - 2026년 9월 9일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 9일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 8. 오전 12:00:00	3.9
26. 9. 8. 오전 12:30:00	3.9
26. 9. 8. 오전 1:00:00	3.9
26. 9. 8. 오전 1:30:00	3.9
26. 9. 8. 오전 2:00:00	3.9
26. 9. 8. 오전 2:30:00	3.9
26. 9. 8. 오전 3:00:00	4.0
26. 9. 8. 오전 3:30:00	4.0
26. 9. 8. 오전 4:00:00	4.1
26. 9. 8. 오전 4:30:00	4.0
26. 9. 8. 오전 5:00:00	3.8
26. 9. 8. 오전 5:30:00	3.9
26. 9. 8. 오전 6:00:00	4.0
26. 9. 8. 오전 6:30:00	4.0
26. 9. 8. 오전 7:00:00	3.9
26. 9. 8. 오전 7:30:00	3.9
26. 9. 8. 오전 8:00:00	3.9
26. 9. 8. 오전 8:30:00	3.9
26. 9. 8. 오전 9:00:00	4.0
26. 9. 8. 오전 9:30:00	3.9
26. 9. 8. 오전 10:00:00	3.8
26. 9. 8. 오전 10:30:00	3.8
26. 9. 8. 오전 11:00:00	3.9
26. 9. 8. 오전 11:30:00	3.9
26. 9. 8. 오후 12:00:00	3.8
26. 9. 8. 오후 12:30:00	3.7
26. 9. 8. 오후 1:00:00	3.8
26. 9. 8. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 8일 오전 12시 0분 0초 KST - 2026년 9월 9일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 9일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 8. 오후 2:00:00	3.7
26. 9. 8. 오후 2:30:00	3.8
26. 9. 8. 오후 3:00:00	3.9
26. 9. 8. 오후 3:30:00	3.8
26. 9. 8. 오후 4:00:00	3.8
26. 9. 8. 오후 4:30:00	3.9
26. 9. 8. 오후 5:00:00	3.8
26. 9. 8. 오후 5:30:00	3.7
26. 9. 8. 오후 6:00:00	3.9
26. 9. 8. 오후 6:30:00	3.9
26. 9. 8. 오후 7:00:00	3.9
26. 9. 8. 오후 7:30:00	3.9
26. 9. 8. 오후 8:00:00	3.8
26. 9. 8. 오후 8:30:00	3.8
26. 9. 8. 오후 9:00:00	3.9
26. 9. 8. 오후 9:30:00	3.8
26. 9. 8. 오후 10:00:00	3.8
26. 9. 8. 오후 10:30:00	3.9
26. 9. 8. 오후 11:00:00	3.8
26. 9. 8. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 9일 오전 12시 0분 0초 KST - 2026년 9월 10일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 10일 오전 6시 0분 30초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.9
26. 9. 9. 오전 12:00:00	3.9
26. 9. 9. 오전 12:30:00	3.9
26. 9. 9. 오전 1:00:00	3.8
26. 9. 9. 오전 1:30:00	3.9
26. 9. 9. 오전 2:00:00	3.9
26. 9. 9. 오전 2:30:00	3.8
26. 9. 9. 오전 3:00:00	3.8
26. 9. 9. 오전 3:30:00	3.9
26. 9. 9. 오전 4:00:00	3.8
26. 9. 9. 오전 4:30:00	3.9
26. 9. 9. 오전 5:00:00	4.0
26. 9. 9. 오전 5:30:00	4.0
26. 9. 9. 오전 6:00:00	4.0
26. 9. 9. 오전 6:30:00	3.9
26. 9. 9. 오전 7:00:00	3.9
26. 9. 9. 오전 7:30:00	3.9
26. 9. 9. 오전 8:00:00	4.0
26. 9. 9. 오전 8:30:00	4.0
26. 9. 9. 오전 9:00:00	4.0
26. 9. 9. 오전 9:30:00	3.9
26. 9. 9. 오전 10:00:00	4.0
26. 9. 9. 오전 10:30:00	3.9
26. 9. 9. 오전 11:00:00	3.8
26. 9. 9. 오전 11:30:00	4.0
26. 9. 9. 오후 12:00:00	4.0
26. 9. 9. 오후 12:30:00	4.0
26. 9. 9. 오후 1:00:00	3.9
26. 9. 9. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 9일 오전 12시 0분 0초 KST - 2026년 9월 10일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 10일 오전 6시 0분 30초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 9. 오후 2:00:00	3.9
26. 9. 9. 오후 2:30:00	3.9
26. 9. 9. 오후 3:00:00	3.7
26. 9. 9. 오후 3:30:00	3.8
26. 9. 9. 오후 4:00:00	3.8
26. 9. 9. 오후 4:30:00	3.7
26. 9. 9. 오후 5:00:00	3.8
26. 9. 9. 오후 5:30:00	3.8
26. 9. 9. 오후 6:00:00	3.9
26. 9. 9. 오후 6:30:00	3.9
26. 9. 9. 오후 7:00:00	4.0
26. 9. 9. 오후 7:30:00	3.9
26. 9. 9. 오후 8:00:00	3.8
26. 9. 9. 오후 8:30:00	3.8
26. 9. 9. 오후 9:00:00	3.8
26. 9. 9. 오후 9:30:00	3.8
26. 9. 9. 오후 10:00:00	3.8
26. 9. 9. 오후 10:30:00	3.9
26. 9. 9. 오후 11:00:00	3.9
26. 9. 9. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 10일 오전 12시 0분 0초 KST - 2026년 9월 11일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 11일 오전 6시 0분 18초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 10. 오전 12:00:00	3.8
26. 9. 10. 오전 12:30:00	3.9
26. 9. 10. 오전 1:00:00	3.8
26. 9. 10. 오전 1:30:00	3.7
26. 9. 10. 오전 2:00:00	3.8
26. 9. 10. 오전 2:30:00	3.8
26. 9. 10. 오전 3:00:00	3.6
26. 9. 10. 오전 3:30:00	3.8
26. 9. 10. 오전 4:00:00	3.7
26. 9. 10. 오전 4:30:00	3.7
26. 9. 10. 오전 5:00:00	3.8
26. 9. 10. 오전 5:30:00	3.6
26. 9. 10. 오전 6:00:00	3.7
26. 9. 10. 오전 6:30:00	3.8
26. 9. 10. 오전 7:00:00	3.8
26. 9. 10. 오전 7:30:00	3.9
26. 9. 10. 오전 8:00:00	4.0
26. 9. 10. 오전 8:30:00	3.9
26. 9. 10. 오전 9:00:00	3.9
26. 9. 10. 오전 9:30:00	3.9
26. 9. 10. 오전 10:00:00	3.8
26. 9. 10. 오전 10:30:00	3.9
26. 9. 10. 오전 11:00:00	3.9
26. 9. 10. 오전 11:30:00	3.7
26. 9. 10. 오후 12:00:00	3.7
26. 9. 10. 오후 12:30:00	3.7
26. 9. 10. 오후 1:00:00	3.8
26. 9. 10. 오후 1:30:00	3.6



IP Storage Temperature Log_Daily report

2026년 9월 10일 오전 12시 0분 0초 KST - 2026년 9월 11일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 11일 오전 6시 0분 18초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 10. 오후 2:00:00	3.7
26. 9. 10. 오후 2:30:00	3.7
26. 9. 10. 오후 3:00:00	3.7
26. 9. 10. 오후 3:30:00	3.7
26. 9. 10. 오후 4:00:00	3.8
26. 9. 10. 오후 4:30:00	3.7
26. 9. 10. 오후 5:00:00	3.8
26. 9. 10. 오후 5:30:00	3.9
26. 9. 10. 오후 6:00:00	3.8
26. 9. 10. 오후 6:30:00	3.7
26. 9. 10. 오후 7:00:00	3.8
26. 9. 10. 오후 7:30:00	3.9
26. 9. 10. 오후 8:00:00	4.0
26. 9. 10. 오후 8:30:00	3.9
26. 9. 10. 오후 9:00:00	3.9
26. 9. 10. 오후 9:30:00	3.9
26. 9. 10. 오후 10:00:00	3.8
26. 9. 10. 오후 10:30:00	3.9
26. 9. 10. 오후 11:00:00	3.9
26. 9. 10. 오후 11:30:00	3.9



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 11일 오전 12시 0분 0초 KST - 2026년 9월 12일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 12일 오전 6시 0분 17초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.8
26. 9. 11. 오전 12:00:00	3.9
26. 9. 11. 오전 12:30:00	3.7
26. 9. 11. 오전 1:00:00	3.7
26. 9. 11. 오전 1:30:00	3.8
26. 9. 11. 오전 2:00:00	3.7
26. 9. 11. 오전 2:30:00	3.7
26. 9. 11. 오전 3:00:00	3.8
26. 9. 11. 오전 3:30:00	3.7
26. 9. 11. 오전 4:00:00	3.7
26. 9. 11. 오전 4:30:00	3.8
26. 9. 11. 오전 5:00:00	3.9
26. 9. 11. 오전 5:30:00	3.9
26. 9. 11. 오전 6:00:00	3.8
26. 9. 11. 오전 6:30:00	3.8
26. 9. 11. 오전 7:00:00	3.7
26. 9. 11. 오전 7:30:00	3.7
26. 9. 11. 오전 8:00:00	3.7
26. 9. 11. 오전 8:30:00	3.7
26. 9. 11. 오전 9:00:00	3.7
26. 9. 11. 오전 9:30:00	3.7
26. 9. 11. 오전 10:00:00	3.8
26. 9. 11. 오전 10:30:00	3.9
26. 9. 11. 오전 11:00:00	3.9
26. 9. 11. 오전 11:30:00	3.7
26. 9. 11. 오후 12:00:00	3.7
26. 9. 11. 오후 12:30:00	3.8
26. 9. 11. 오후 1:00:00	3.9
26. 9. 11. 오후 1:30:00	4.0



IP Storage Temperature Log_Daily report

2026년 9월 11일 오전 12시 0분 0초 KST - 2026년 9월 12일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 12일 오전 6시 0분 17초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 11. 오후 2:00:00	3.9
26. 9. 11. 오후 2:30:00	3.9
26. 9. 11. 오후 3:00:00	3.9
26. 9. 11. 오후 3:30:00	4.0
26. 9. 11. 오후 4:00:00	4.0
26. 9. 11. 오후 4:30:00	3.9
26. 9. 11. 오후 5:00:00	3.9
26. 9. 11. 오후 5:30:00	3.9
26. 9. 11. 오후 6:00:00	3.9
26. 9. 11. 오후 6:30:00	3.9
26. 9. 11. 오후 7:00:00	3.9
26. 9. 11. 오후 7:30:00	4.0
26. 9. 11. 오후 8:00:00	4.0
26. 9. 11. 오후 8:30:00	3.9
26. 9. 11. 오후 9:00:00	3.8
26. 9. 11. 오후 9:30:00	3.8
26. 9. 11. 오후 10:00:00	3.8
26. 9. 11. 오후 10:30:00	3.7
26. 9. 11. 오후 11:00:00	3.7
26. 9. 11. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 12일 오전 12시 0분 0초 KST - 2026년 9월 13일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 13일 오전 6시 0분 18초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 12. 오전 12:00:00	3.7
26. 9. 12. 오전 12:30:00	3.6
26. 9. 12. 오전 1:00:00	3.8
26. 9. 12. 오전 1:30:00	3.9
26. 9. 12. 오전 2:00:00	3.9
26. 9. 12. 오전 2:30:00	3.9
26. 9. 12. 오전 3:00:00	3.9
26. 9. 12. 오전 3:30:00	3.9
26. 9. 12. 오전 4:00:00	3.9
26. 9. 12. 오전 4:30:00	3.9
26. 9. 12. 오전 5:00:00	4.0
26. 9. 12. 오전 5:30:00	4.0
26. 9. 12. 오전 6:00:00	4.0
26. 9. 12. 오전 6:30:00	4.0
26. 9. 12. 오전 7:00:00	4.0
26. 9. 12. 오전 7:30:00	3.9
26. 9. 12. 오전 8:00:00	3.7
26. 9. 12. 오전 8:30:00	3.8
26. 9. 12. 오전 9:00:00	3.8
26. 9. 12. 오전 9:30:00	3.6
26. 9. 12. 오전 10:00:00	3.7
26. 9. 12. 오전 10:30:00	3.8
26. 9. 12. 오전 11:00:00	3.8
26. 9. 12. 오전 11:30:00	3.8
26. 9. 12. 오후 12:00:00	3.8
26. 9. 12. 오후 12:30:00	3.8
26. 9. 12. 오후 1:00:00	3.9
26. 9. 12. 오후 1:30:00	4.0



IP Storage Temperature Log_Daily report

2026년 9월 12일 오전 12시 0분 0초 KST - 2026년 9월 13일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 13일 오전 6시 0분 18초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 12. 오후 2:00:00	4.0
26. 9. 12. 오후 2:30:00	3.9
26. 9. 12. 오후 3:00:00	3.8
26. 9. 12. 오후 3:30:00	3.9
26. 9. 12. 오후 4:00:00	3.7
26. 9. 12. 오후 4:30:00	3.7
26. 9. 12. 오후 5:00:00	3.7
26. 9. 12. 오후 5:30:00	3.8
26. 9. 12. 오후 6:00:00	3.7
26. 9. 12. 오후 6:30:00	3.7
26. 9. 12. 오후 7:00:00	3.8
26. 9. 12. 오후 7:30:00	3.7
26. 9. 12. 오후 8:00:00	3.7
26. 9. 12. 오후 8:30:00	3.7
26. 9. 12. 오후 9:00:00	3.8
26. 9. 12. 오후 9:30:00	3.7
26. 9. 12. 오후 10:00:00	3.7
26. 9. 12. 오후 10:30:00	3.8
26. 9. 12. 오후 11:00:00	3.9
26. 9. 12. 오후 11:30:00	3.9



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 13일 오전 12시 0분 0초 KST - 2026년 9월 14일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 14일 오전 6시 1분 55초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.8
최대값	4.0
평균	3.9
26. 9. 13. 오전 12:00:00	3.8
26. 9. 13. 오전 12:30:00	3.8
26. 9. 13. 오전 1:00:00	3.8
26. 9. 13. 오전 1:30:00	3.9
26. 9. 13. 오전 2:00:00	3.8
26. 9. 13. 오전 2:30:00	3.8
26. 9. 13. 오전 3:00:00	3.9
26. 9. 13. 오전 3:30:00	4.0
26. 9. 13. 오전 4:00:00	4.0
26. 9. 13. 오전 4:30:00	4.0
26. 9. 13. 오전 5:00:00	3.9
26. 9. 13. 오전 5:30:00	3.8
26. 9. 13. 오전 6:00:00	3.9
26. 9. 13. 오전 6:30:00	3.9
26. 9. 13. 오전 7:00:00	3.9
26. 9. 13. 오전 7:30:00	3.8
26. 9. 13. 오전 8:00:00	3.9
26. 9. 13. 오전 8:30:00	3.9
26. 9. 13. 오전 9:00:00	3.9
26. 9. 13. 오전 9:30:00	3.9
26. 9. 13. 오전 10:00:00	3.9
26. 9. 13. 오전 10:30:00	3.9
26. 9. 13. 오전 11:00:00	3.8
26. 9. 13. 오전 11:30:00	3.8
26. 9. 13. 오후 12:00:00	3.8
26. 9. 13. 오후 12:30:00	3.9
26. 9. 13. 오후 1:00:00	3.9
26. 9. 13. 오후 1:30:00	4.0



IP Storage Temperature Log_Daily report

2026년 9월 13일 오전 12시 0분 0초 KST - 2026년 9월 14일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 14일 오전 6시 1분 55초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 13. 오후 2:00:00	3.9
26. 9. 13. 오후 2:30:00	3.8
26. 9. 13. 오후 3:00:00	3.8
26. 9. 13. 오후 3:30:00	3.9
26. 9. 13. 오후 4:00:00	4.0
26. 9. 13. 오후 4:30:00	4.0
26. 9. 13. 오후 5:00:00	3.9
26. 9. 13. 오후 5:30:00	3.9
26. 9. 13. 오후 6:00:00	3.9
26. 9. 13. 오후 6:30:00	3.9
26. 9. 13. 오후 7:00:00	3.9
26. 9. 13. 오후 7:30:00	4.0
26. 9. 13. 오후 8:00:00	4.0
26. 9. 13. 오후 8:30:00	3.9
26. 9. 13. 오후 9:00:00	3.9
26. 9. 13. 오후 9:30:00	3.9
26. 9. 13. 오후 10:00:00	3.9
26. 9. 13. 오후 10:30:00	4.0
26. 9. 13. 오후 11:00:00	4.0
26. 9. 13. 오후 11:30:00	3.9



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 14일 오전 12시 0분 0초 KST - 2026년 9월 15일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 15일 오전 6시 0분 17초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.9
26. 9. 14. 오전 12:00:00	3.9
26. 9. 14. 오전 12:30:00	3.9
26. 9. 14. 오전 1:00:00	3.9
26. 9. 14. 오전 1:30:00	3.9
26. 9. 14. 오전 2:00:00	3.9
26. 9. 14. 오전 2:30:00	3.7
26. 9. 14. 오전 3:00:00	3.8
26. 9. 14. 오전 3:30:00	3.8
26. 9. 14. 오전 4:00:00	3.7
26. 9. 14. 오전 4:30:00	3.7
26. 9. 14. 오전 5:00:00	3.8
26. 9. 14. 오전 5:30:00	3.6
26. 9. 14. 오전 6:00:00	3.9
26. 9. 14. 오전 6:30:00	3.8
26. 9. 14. 오전 7:00:00	3.8
26. 9. 14. 오전 7:30:00	3.9
26. 9. 14. 오전 8:00:00	4.0
26. 9. 14. 오전 8:30:00	3.9
26. 9. 14. 오전 9:00:00	3.8
26. 9. 14. 오전 9:30:00	3.8
26. 9. 14. 오전 10:00:00	3.9
26. 9. 14. 오전 10:30:00	3.9
26. 9. 14. 오전 11:00:00	3.9
26. 9. 14. 오전 11:30:00	3.9
26. 9. 14. 오후 12:00:00	4.0
26. 9. 14. 오후 12:30:00	3.9
26. 9. 14. 오후 1:00:00	3.8
26. 9. 14. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 14일 오전 12시 0분 0초 KST - 2026년 9월 15일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 15일 오전 6시 0분 17초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 14. 오후 2:00:00	3.9
26. 9. 14. 오후 2:30:00	3.9
26. 9. 14. 오후 3:00:00	3.9
26. 9. 14. 오후 3:30:00	3.8
26. 9. 14. 오후 4:00:00	3.9
26. 9. 14. 오후 4:30:00	4.0
26. 9. 14. 오후 5:00:00	4.0
26. 9. 14. 오후 5:30:00	3.9
26. 9. 14. 오후 6:00:00	4.0
26. 9. 14. 오후 6:30:00	4.0
26. 9. 14. 오후 7:00:00	4.0
26. 9. 14. 오후 7:30:00	4.0
26. 9. 14. 오후 8:00:00	3.9
26. 9. 14. 오후 8:30:00	3.9
26. 9. 14. 오후 9:00:00	4.0
26. 9. 14. 오후 9:30:00	4.0
26. 9. 14. 오후 10:00:00	3.8
26. 9. 14. 오후 10:30:00	3.8
26. 9. 14. 오후 11:00:00	3.9
26. 9. 14. 오후 11:30:00	3.7



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 15일 오전 12시 0분 0초 KST - 2026년 9월 16일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 16일 오전 6시 0분 17초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 15. 오전 12:00:00	3.8
26. 9. 15. 오전 12:30:00	3.8
26. 9. 15. 오전 1:00:00	3.9
26. 9. 15. 오전 1:30:00	3.7
26. 9. 15. 오전 2:00:00	3.8
26. 9. 15. 오전 2:30:00	3.7
26. 9. 15. 오전 3:00:00	3.7
26. 9. 15. 오전 3:30:00	3.8
26. 9. 15. 오전 4:00:00	3.9
26. 9. 15. 오전 4:30:00	3.8
26. 9. 15. 오전 5:00:00	3.7
26. 9. 15. 오전 5:30:00	3.8
26. 9. 15. 오전 6:00:00	3.7
26. 9. 15. 오전 6:30:00	3.7
26. 9. 15. 오전 7:00:00	3.8
26. 9. 15. 오전 7:30:00	3.9
26. 9. 15. 오전 8:00:00	4.0
26. 9. 15. 오전 8:30:00	3.8
26. 9. 15. 오전 9:00:00	3.7
26. 9. 15. 오전 9:30:00	3.7
26. 9. 15. 오전 10:00:00	3.7
26. 9. 15. 오전 10:30:00	3.6
26. 9. 15. 오전 11:00:00	3.7
26. 9. 15. 오전 11:30:00	3.7
26. 9. 15. 오후 12:00:00	3.8
26. 9. 15. 오후 12:30:00	3.6
26. 9. 15. 오후 1:00:00	3.7
26. 9. 15. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 15일 오전 12시 0분 0초 KST - 2026년 9월 16일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 16일 오전 6시 0분 17초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 15. 오후 2:00:00	3.9
26. 9. 15. 오후 2:30:00	3.9
26. 9. 15. 오후 3:00:00	4.0
26. 9. 15. 오후 3:30:00	3.9
26. 9. 15. 오후 4:00:00	3.8
26. 9. 15. 오후 4:30:00	3.9
26. 9. 15. 오후 5:00:00	3.8
26. 9. 15. 오후 5:30:00	3.7
26. 9. 15. 오후 6:00:00	3.9
26. 9. 15. 오후 6:30:00	3.9
26. 9. 15. 오후 7:00:00	3.8
26. 9. 15. 오후 7:30:00	3.8
26. 9. 15. 오후 8:00:00	3.9
26. 9. 15. 오후 8:30:00	3.8
26. 9. 15. 오후 9:00:00	3.7
26. 9. 15. 오후 9:30:00	3.8
26. 9. 15. 오후 10:00:00	3.7
26. 9. 15. 오후 10:30:00	3.7
26. 9. 15. 오후 11:00:00	3.8
26. 9. 15. 오후 11:30:00	3.6



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 16일 오전 12시 0분 0초 KST - 2026년 9월 17일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 17일 오전 6시 0분 17초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 16. 오전 12:00:00	3.7
26. 9. 16. 오전 12:30:00	3.8
26. 9. 16. 오전 1:00:00	3.9
26. 9. 16. 오전 1:30:00	3.9
26. 9. 16. 오전 2:00:00	3.9
26. 9. 16. 오전 2:30:00	3.9
26. 9. 16. 오전 3:00:00	3.9
26. 9. 16. 오전 3:30:00	4.0
26. 9. 16. 오전 4:00:00	4.0
26. 9. 16. 오전 4:30:00	3.9
26. 9. 16. 오전 5:00:00	3.9
26. 9. 16. 오전 5:30:00	3.8
26. 9. 16. 오전 6:00:00	3.8
26. 9. 16. 오전 6:30:00	3.9
26. 9. 16. 오전 7:00:00	3.9
26. 9. 16. 오전 7:30:00	4.0
26. 9. 16. 오전 8:00:00	4.0
26. 9. 16. 오전 8:30:00	3.8
26. 9. 16. 오전 9:00:00	3.7
26. 9. 16. 오전 9:30:00	3.8
26. 9. 16. 오전 10:00:00	3.6
26. 9. 16. 오전 10:30:00	3.7
26. 9. 16. 오전 11:00:00	3.6
26. 9. 16. 오전 11:30:00	3.6
26. 9. 16. 오후 12:00:00	3.8
26. 9. 16. 오후 12:30:00	3.7
26. 9. 16. 오후 1:00:00	3.8
26. 9. 16. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 16일 오전 12시 0분 0초 KST - 2026년 9월 17일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 17일 오전 6시 0분 17초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 16. 오후 2:00:00	3.7
26. 9. 16. 오후 2:30:00	3.7
26. 9. 16. 오후 3:00:00	3.8
26. 9. 16. 오후 3:30:00	3.7
26. 9. 16. 오후 4:00:00	3.8
26. 9. 16. 오후 4:30:00	3.7
26. 9. 16. 오후 5:00:00	3.7
26. 9. 16. 오후 5:30:00	3.8
26. 9. 16. 오후 6:00:00	3.7
26. 9. 16. 오후 6:30:00	3.8
26. 9. 16. 오후 7:00:00	3.8
26. 9. 16. 오후 7:30:00	3.8
26. 9. 16. 오후 8:00:00	3.7
26. 9. 16. 오후 8:30:00	3.8
26. 9. 16. 오후 9:00:00	3.7
26. 9. 16. 오후 9:30:00	3.7
26. 9. 16. 오후 10:00:00	3.8
26. 9. 16. 오후 10:30:00	3.7
26. 9. 16. 오후 11:00:00	3.7
26. 9. 16. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 17일 오전 12시 0분 0초 KST - 2026년 9월 18일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 18일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 17. 오전 12:00:00	3.7
26. 9. 17. 오전 12:30:00	3.7
26. 9. 17. 오전 1:00:00	3.7
26. 9. 17. 오전 1:30:00	3.8
26. 9. 17. 오전 2:00:00	3.8
26. 9. 17. 오전 2:30:00	3.6
26. 9. 17. 오전 3:00:00	3.7
26. 9. 17. 오전 3:30:00	3.8
26. 9. 17. 오전 4:00:00	3.6
26. 9. 17. 오전 4:30:00	3.7
26. 9. 17. 오전 5:00:00	3.7
26. 9. 17. 오전 5:30:00	3.7
26. 9. 17. 오전 6:00:00	3.7
26. 9. 17. 오전 6:30:00	3.8
26. 9. 17. 오전 7:00:00	3.8
26. 9. 17. 오전 7:30:00	3.7
26. 9. 17. 오전 8:00:00	3.8
26. 9. 17. 오전 8:30:00	3.6
26. 9. 17. 오전 9:00:00	3.7
26. 9. 17. 오전 9:30:00	3.7
26. 9. 17. 오전 10:00:00	3.7
26. 9. 17. 오전 10:30:00	3.7
26. 9. 17. 오전 11:00:00	3.6
26. 9. 17. 오전 11:30:00	3.6
26. 9. 17. 오후 12:00:00	3.8
26. 9. 17. 오후 12:30:00	3.6
26. 9. 17. 오후 1:00:00	3.7
26. 9. 17. 오후 1:30:00	3.7



IP Storage Temperature Log_Daily report

2026년 9월 17일 오전 12시 0분 0초 KST - 2026년 9월 18일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 18일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 17. 오후 2:00:00	3.7
26. 9. 17. 오후 2:30:00	3.8
26. 9. 17. 오후 3:00:00	3.7
26. 9. 17. 오후 3:30:00	3.8
26. 9. 17. 오후 4:00:00	3.7
26. 9. 17. 오후 4:30:00	3.7
26. 9. 17. 오후 5:00:00	3.8
26. 9. 17. 오후 5:30:00	3.8
26. 9. 17. 오후 6:00:00	3.8
26. 9. 17. 오후 6:30:00	3.9
26. 9. 17. 오후 7:00:00	3.9
26. 9. 17. 오후 7:30:00	4.0
26. 9. 17. 오후 8:00:00	4.0
26. 9. 17. 오후 8:30:00	3.9
26. 9. 17. 오후 9:00:00	3.9
26. 9. 17. 오후 9:30:00	3.9
26. 9. 17. 오후 10:00:00	4.0
26. 9. 17. 오후 10:30:00	3.8
26. 9. 17. 오후 11:00:00	3.7
26. 9. 17. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 18일 오전 12시 0분 0초 KST - 2026년 9월 19일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 19일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	3.9
평균	3.8
26. 9. 18. 오전 12:00:00	3.7
26. 9. 18. 오전 12:30:00	3.9
26. 9. 18. 오전 1:00:00	3.8
26. 9. 18. 오전 1:30:00	3.9
26. 9. 18. 오전 2:00:00	3.9
26. 9. 18. 오전 2:30:00	3.8
26. 9. 18. 오전 3:00:00	3.7
26. 9. 18. 오전 3:30:00	3.8
26. 9. 18. 오전 4:00:00	3.6
26. 9. 18. 오전 4:30:00	3.6
26. 9. 18. 오전 5:00:00	3.8
26. 9. 18. 오전 5:30:00	3.7
26. 9. 18. 오전 6:00:00	3.7
26. 9. 18. 오전 6:30:00	3.7
26. 9. 18. 오전 7:00:00	3.8
26. 9. 18. 오전 7:30:00	3.8
26. 9. 18. 오전 8:00:00	3.9
26. 9. 18. 오전 8:30:00	3.9
26. 9. 18. 오전 9:00:00	3.7
26. 9. 18. 오전 9:30:00	3.8
26. 9. 18. 오전 10:00:00	3.8
26. 9. 18. 오전 10:30:00	3.9
26. 9. 18. 오전 11:00:00	3.8
26. 9. 18. 오전 11:30:00	3.8
26. 9. 18. 오후 12:00:00	3.9
26. 9. 18. 오후 12:30:00	3.9
26. 9. 18. 오후 1:00:00	3.8
26. 9. 18. 오후 1:30:00	3.7



IP Storage Temperature Log_Daily report

2026년 9월 18일 오전 12시 0분 0초 KST - 2026년 9월 19일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 19일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 18. 오후 2:00:00	3.8
26. 9. 18. 오후 2:30:00	3.7
26. 9. 18. 오후 3:00:00	3.6
26. 9. 18. 오후 3:30:00	3.8
26. 9. 18. 오후 4:00:00	3.7
26. 9. 18. 오후 4:30:00	3.8
26. 9. 18. 오후 5:00:00	3.8
26. 9. 18. 오후 5:30:00	3.7
26. 9. 18. 오후 6:00:00	3.7
26. 9. 18. 오후 6:30:00	3.9
26. 9. 18. 오후 7:00:00	3.9
26. 9. 18. 오후 7:30:00	3.9
26. 9. 18. 오후 8:00:00	3.9
26. 9. 18. 오후 8:30:00	3.9
26. 9. 18. 오후 9:00:00	3.9
26. 9. 18. 오후 9:30:00	3.9
26. 9. 18. 오후 10:00:00	3.8
26. 9. 18. 오후 10:30:00	3.7
26. 9. 18. 오후 11:00:00	3.7
26. 9. 18. 오후 11:30:00	3.7



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 19일 오전 12시 0분 0초 KST - 2026년 9월 20일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 20일 오전 6시 0분 25초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.9
26. 9. 19. 오전 12:00:00	3.7
26. 9. 19. 오전 12:30:00	3.7
26. 9. 19. 오전 1:00:00	3.8
26. 9. 19. 오전 1:30:00	3.9
26. 9. 19. 오전 2:00:00	4.0
26. 9. 19. 오전 2:30:00	4.0
26. 9. 19. 오전 3:00:00	3.9
26. 9. 19. 오전 3:30:00	3.9
26. 9. 19. 오전 4:00:00	3.9
26. 9. 19. 오전 4:30:00	3.9
26. 9. 19. 오전 5:00:00	3.9
26. 9. 19. 오전 5:30:00	3.9
26. 9. 19. 오전 6:00:00	3.9
26. 9. 19. 오전 6:30:00	3.9
26. 9. 19. 오전 7:00:00	3.9
26. 9. 19. 오전 7:30:00	3.9
26. 9. 19. 오전 8:00:00	3.9
26. 9. 19. 오전 8:30:00	3.8
26. 9. 19. 오전 9:00:00	3.7
26. 9. 19. 오전 9:30:00	3.8
26. 9. 19. 오전 10:00:00	3.9
26. 9. 19. 오전 10:30:00	3.8
26. 9. 19. 오전 11:00:00	3.8
26. 9. 19. 오전 11:30:00	3.8
26. 9. 19. 오후 12:00:00	3.9
26. 9. 19. 오후 12:30:00	3.8
26. 9. 19. 오후 1:00:00	3.8
26. 9. 19. 오후 1:30:00	3.7



IP Storage Temperature Log_Daily report

2026년 9월 19일 오전 12시 0분 0초 KST - 2026년 9월 20일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 20일 오전 6시 0분 25초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 19. 오후 2:00:00	3.8
26. 9. 19. 오후 2:30:00	3.9
26. 9. 19. 오후 3:00:00	3.9
26. 9. 19. 오후 3:30:00	3.9
26. 9. 19. 오후 4:00:00	4.0
26. 9. 19. 오후 4:30:00	3.8
26. 9. 19. 오후 5:00:00	3.7
26. 9. 19. 오후 5:30:00	3.8
26. 9. 19. 오후 6:00:00	3.8
26. 9. 19. 오후 6:30:00	3.7
26. 9. 19. 오후 7:00:00	3.8
26. 9. 19. 오후 7:30:00	3.9
26. 9. 19. 오후 8:00:00	3.9
26. 9. 19. 오후 8:30:00	3.9
26. 9. 19. 오후 9:00:00	3.9
26. 9. 19. 오후 9:30:00	4.0
26. 9. 19. 오후 10:00:00	3.9
26. 9. 19. 오후 10:30:00	3.9
26. 9. 19. 오후 11:00:00	3.9
26. 9. 19. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 20일 오전 12시 0분 0초 KST - 2026년 9월 21일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 21일 오전 6시 2분 9초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	4.0
평균	3.8
26. 9. 20. 오전 12:00:00	4.0
26. 9. 20. 오전 12:30:00	3.9
26. 9. 20. 오전 1:00:00	3.9
26. 9. 20. 오전 1:30:00	4.0
26. 9. 20. 오전 2:00:00	3.9
26. 9. 20. 오전 2:30:00	3.7
26. 9. 20. 오전 3:00:00	3.8
26. 9. 20. 오전 3:30:00	3.8
26. 9. 20. 오전 4:00:00	3.8
26. 9. 20. 오전 4:30:00	3.9
26. 9. 20. 오전 5:00:00	4.0
26. 9. 20. 오전 5:30:00	4.0
26. 9. 20. 오전 6:00:00	3.9
26. 9. 20. 오전 6:30:00	3.9
26. 9. 20. 오전 7:00:00	3.8
26. 9. 20. 오전 7:30:00	3.9
26. 9. 20. 오전 8:00:00	3.8
26. 9. 20. 오전 8:30:00	3.8
26. 9. 20. 오전 9:00:00	3.9
26. 9. 20. 오전 9:30:00	3.9
26. 9. 20. 오전 10:00:00	3.9
26. 9. 20. 오전 10:30:00	3.7
26. 9. 20. 오전 11:00:00	3.8
26. 9. 20. 오전 11:30:00	3.6
26. 9. 20. 오후 12:00:00	3.6
26. 9. 20. 오후 12:30:00	3.7
26. 9. 20. 오후 1:00:00	3.6
26. 9. 20. 오후 1:30:00	3.7



IP Storage Temperature Log_Daily report

2026년 9월 20일 오전 12시 0분 0초 KST - 2026년 9월 21일 오전 12시 0분 0초 KST

kyon410@daum.net
생성됨: 2026년 9월 21일 오전 6시 2분 9초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 20. 오후 2:00:00	3.8
26. 9. 20. 오후 2:30:00	3.7
26. 9. 20. 오후 3:00:00	3.7
26. 9. 20. 오후 3:30:00	3.7
26. 9. 20. 오후 4:00:00	3.7
26. 9. 20. 오후 4:30:00	3.8
26. 9. 20. 오후 5:00:00	3.8
26. 9. 20. 오후 5:30:00	3.7
26. 9. 20. 오후 6:00:00	3.8
26. 9. 20. 오후 6:30:00	3.7
26. 9. 20. 오후 7:00:00	3.8
26. 9. 20. 오후 7:30:00	3.7
26. 9. 20. 오후 8:00:00	3.8
26. 9. 20. 오후 8:30:00	3.7
26. 9. 20. 오후 9:00:00	3.8
26. 9. 20. 오후 9:30:00	3.7
26. 9. 20. 오후 10:00:00	3.7
26. 9. 20. 오후 10:30:00	3.8
26. 9. 20. 오후 11:00:00	3.7
26. 9. 20. 오후 11:30:00	3.7



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 21일 오전 12시 0분 0초 KST - 2026년 9월 22일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 22일 오전 6시 0분 21초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.6
최대값	3.9
평균	3.7
26. 9. 21. 오전 12:00:00	3.8
26. 9. 21. 오전 12:30:00	3.7
26. 9. 21. 오전 1:00:00	3.8
26. 9. 21. 오전 1:30:00	3.7
26. 9. 21. 오전 2:00:00	3.7
26. 9. 21. 오전 2:30:00	3.8
26. 9. 21. 오전 3:00:00	3.7
26. 9. 21. 오전 3:30:00	3.7
26. 9. 21. 오전 4:00:00	3.8
26. 9. 21. 오전 4:30:00	3.7
26. 9. 21. 오전 5:00:00	3.8
26. 9. 21. 오전 5:30:00	3.8
26. 9. 21. 오전 6:00:00	3.9
26. 9. 21. 오전 6:30:00	3.8
26. 9. 21. 오전 7:00:00	3.8
26. 9. 21. 오전 7:30:00	3.9
26. 9. 21. 오전 8:00:00	3.7
26. 9. 21. 오전 8:30:00	3.7
26. 9. 21. 오전 9:00:00	3.8
26. 9. 21. 오전 9:30:00	3.7
26. 9. 21. 오전 10:00:00	3.7
26. 9. 21. 오전 10:30:00	3.7
26. 9. 21. 오전 11:00:00	3.6
26. 9. 21. 오전 11:30:00	3.8
26. 9. 21. 오후 12:00:00	3.6
26. 9. 21. 오후 12:30:00	3.7
26. 9. 21. 오후 1:00:00	3.7
26. 9. 21. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 21일 오전 12시 0분 0초 KST - 2026년 9월 22일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 22일 오전 6시 0분 21초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 21. 오후 2:00:00	3.8
26. 9. 21. 오후 2:30:00	3.6
26. 9. 21. 오후 3:00:00	3.8
26. 9. 21. 오후 3:30:00	3.7
26. 9. 21. 오후 4:00:00	3.7
26. 9. 21. 오후 4:30:00	3.8
26. 9. 21. 오후 5:00:00	3.7
26. 9. 21. 오후 5:30:00	3.8
26. 9. 21. 오후 6:00:00	3.8
26. 9. 21. 오후 6:30:00	3.7
26. 9. 21. 오후 7:00:00	3.7
26. 9. 21. 오후 7:30:00	3.7
26. 9. 21. 오후 8:00:00	3.8
26. 9. 21. 오후 8:30:00	3.7
26. 9. 21. 오후 9:00:00	3.7
26. 9. 21. 오후 9:30:00	3.8
26. 9. 21. 오후 10:00:00	3.8
26. 9. 21. 오후 10:30:00	3.7
26. 9. 21. 오후 11:00:00	3.8
26. 9. 21. 오후 11:30:00	3.7



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 22일 오전 12시 0분 0초 KST - 2026년 9월 23일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 23일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.8
26. 9. 22. 오전 12:00:00	3.7
26. 9. 22. 오전 12:30:00	3.8
26. 9. 22. 오전 1:00:00	3.9
26. 9. 22. 오전 1:30:00	3.7
26. 9. 22. 오전 2:00:00	3.8
26. 9. 22. 오전 2:30:00	3.8
26. 9. 22. 오전 3:00:00	3.7
26. 9. 22. 오전 3:30:00	3.7
26. 9. 22. 오전 4:00:00	3.7
26. 9. 22. 오전 4:30:00	3.7
26. 9. 22. 오전 5:00:00	3.8
26. 9. 22. 오전 5:30:00	3.7
26. 9. 22. 오전 6:00:00	3.7
26. 9. 22. 오전 6:30:00	3.8
26. 9. 22. 오전 7:00:00	3.9
26. 9. 22. 오전 7:30:00	3.9
26. 9. 22. 오전 8:00:00	3.8
26. 9. 22. 오전 8:30:00	3.8
26. 9. 22. 오전 9:00:00	3.8
26. 9. 22. 오전 9:30:00	3.7
26. 9. 22. 오전 10:00:00	3.8
26. 9. 22. 오전 10:30:00	3.8
26. 9. 22. 오전 11:00:00	3.8
26. 9. 22. 오전 11:30:00	3.9
26. 9. 22. 오후 12:00:00	3.9
26. 9. 22. 오후 12:30:00	3.9
26. 9. 22. 오후 1:00:00	3.8
26. 9. 22. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 22일 오전 12시 0분 0초 KST - 2026년 9월 23일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 23일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 22. 오후 2:00:00	3.9
26. 9. 22. 오후 2:30:00	4.0
26. 9. 22. 오후 3:00:00	3.8
26. 9. 22. 오후 3:30:00	3.8
26. 9. 22. 오후 4:00:00	3.9
26. 9. 22. 오후 4:30:00	4.0
26. 9. 22. 오후 5:00:00	3.9
26. 9. 22. 오후 5:30:00	3.8
26. 9. 22. 오후 6:00:00	3.9
26. 9. 22. 오후 6:30:00	3.8
26. 9. 22. 오후 7:00:00	3.7
26. 9. 22. 오후 7:30:00	3.8
26. 9. 22. 오후 8:00:00	3.7
26. 9. 22. 오후 8:30:00	3.7
26. 9. 22. 오후 9:00:00	3.8
26. 9. 22. 오후 9:30:00	3.7
26. 9. 22. 오후 10:00:00	3.8
26. 9. 22. 오후 10:30:00	3.8
26. 9. 22. 오후 11:00:00	3.8
26. 9. 22. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 23일 오전 12시 0분 0초 KST - 2026년 9월 24일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 24일 오전 6시 0분 20초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 23. 오전 12:00:00	3.8
26. 9. 23. 오전 12:30:00	3.7
26. 9. 23. 오전 1:00:00	3.8
26. 9. 23. 오전 1:30:00	3.9
26. 9. 23. 오전 2:00:00	4.0
26. 9. 23. 오전 2:30:00	4.0
26. 9. 23. 오전 3:00:00	4.1
26. 9. 23. 오전 3:30:00	4.1
26. 9. 23. 오전 4:00:00	4.0
26. 9. 23. 오전 4:30:00	4.0
26. 9. 23. 오전 5:00:00	4.0
26. 9. 23. 오전 5:30:00	4.0
26. 9. 23. 오전 6:00:00	4.1
26. 9. 23. 오전 6:30:00	4.0
26. 9. 23. 오전 7:00:00	4.0
26. 9. 23. 오전 7:30:00	4.0
26. 9. 23. 오전 8:00:00	4.1
26. 9. 23. 오전 8:30:00	4.1
26. 9. 23. 오전 9:00:00	4.0
26. 9. 23. 오전 9:30:00	4.0
26. 9. 23. 오전 10:00:00	4.0
26. 9. 23. 오전 10:30:00	3.9
26. 9. 23. 오전 11:00:00	3.9
26. 9. 23. 오전 11:30:00	3.9
26. 9. 23. 오후 12:00:00	4.0
26. 9. 23. 오후 12:30:00	3.8
26. 9. 23. 오후 1:00:00	3.8
26. 9. 23. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 23일 오전 12시 0분 0초 KST - 2026년 9월 24일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 24일 오전 6시 0분 20초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 23. 오후 2:00:00	3.7
26. 9. 23. 오후 2:30:00	3.8
26. 9. 23. 오후 3:00:00	3.8
26. 9. 23. 오후 3:30:00	3.7
26. 9. 23. 오후 4:00:00	3.9
26. 9. 23. 오후 4:30:00	3.8
26. 9. 23. 오후 5:00:00	3.8
26. 9. 23. 오후 5:30:00	3.9
26. 9. 23. 오후 6:00:00	3.8
26. 9. 23. 오후 6:30:00	3.7
26. 9. 23. 오후 7:00:00	3.8
26. 9. 23. 오후 7:30:00	3.7
26. 9. 23. 오후 8:00:00	3.8
26. 9. 23. 오후 8:30:00	3.8
26. 9. 23. 오후 9:00:00	3.8
26. 9. 23. 오후 9:30:00	3.7
26. 9. 23. 오후 10:00:00	3.8
26. 9. 23. 오후 10:30:00	3.8
26. 9. 23. 오후 11:00:00	3.9
26. 9. 23. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 24일 오전 12시 0분 0초 KST - 2026년 9월 25일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 25일 오전 6시 0분 21초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 24. 오전 12:00:00	3.9
26. 9. 24. 오전 12:30:00	3.8
26. 9. 24. 오전 1:00:00	3.8
26. 9. 24. 오전 1:30:00	3.9
26. 9. 24. 오전 2:00:00	3.8
26. 9. 24. 오전 2:30:00	3.7
26. 9. 24. 오전 3:00:00	3.9
26. 9. 24. 오전 3:30:00	3.9
26. 9. 24. 오전 4:00:00	3.8
26. 9. 24. 오전 4:30:00	3.8
26. 9. 24. 오전 5:00:00	3.9
26. 9. 24. 오전 5:30:00	3.9
26. 9. 24. 오전 6:00:00	3.9
26. 9. 24. 오전 6:30:00	4.0
26. 9. 24. 오전 7:00:00	4.1
26. 9. 24. 오전 7:30:00	4.0
26. 9. 24. 오전 8:00:00	4.0
26. 9. 24. 오전 8:30:00	4.0
26. 9. 24. 오전 9:00:00	4.0
26. 9. 24. 오전 9:30:00	4.0
26. 9. 24. 오전 10:00:00	4.0
26. 9. 24. 오전 10:30:00	3.9
26. 9. 24. 오전 11:00:00	3.9
26. 9. 24. 오전 11:30:00	3.9
26. 9. 24. 오후 12:00:00	4.0
26. 9. 24. 오후 12:30:00	4.0
26. 9. 24. 오후 1:00:00	4.0
26. 9. 24. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 24일 오전 12시 0분 0초 KST - 2026년 9월 25일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 25일 오전 6시 0분 21초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 24. 오후 2:00:00	3.9
26. 9. 24. 오후 2:30:00	4.1
26. 9. 24. 오후 3:00:00	4.0
26. 9. 24. 오후 3:30:00	4.1
26. 9. 24. 오후 4:00:00	4.1
26. 9. 24. 오후 4:30:00	3.9
26. 9. 24. 오후 5:00:00	3.9
26. 9. 24. 오후 5:30:00	3.9
26. 9. 24. 오후 6:00:00	3.8
26. 9. 24. 오후 6:30:00	3.9
26. 9. 24. 오후 7:00:00	3.8
26. 9. 24. 오후 7:30:00	3.9
26. 9. 24. 오후 8:00:00	4.0
26. 9. 24. 오후 8:30:00	3.8
26. 9. 24. 오후 9:00:00	3.8
26. 9. 24. 오후 9:30:00	3.9
26. 9. 24. 오후 10:00:00	3.9
26. 9. 24. 오후 10:30:00	3.8
26. 9. 24. 오후 11:00:00	3.8
26. 9. 24. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 25일 오전 12시 0분 0초 KST - 2026년 9월 26일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 26일 오전 6시 0분 18초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.8
최대값	4.1
평균	4.0
26. 9. 25. 오전 12:00:00	4.1
26. 9. 25. 오전 12:30:00	4.0
26. 9. 25. 오전 1:00:00	4.0
26. 9. 25. 오전 1:30:00	4.0
26. 9. 25. 오전 2:00:00	4.0
26. 9. 25. 오전 2:30:00	3.8
26. 9. 25. 오전 3:00:00	3.8
26. 9. 25. 오전 3:30:00	3.9
26. 9. 25. 오전 4:00:00	3.9
26. 9. 25. 오전 4:30:00	4.0
26. 9. 25. 오전 5:00:00	4.1
26. 9. 25. 오전 5:30:00	4.0
26. 9. 25. 오전 6:00:00	4.0
26. 9. 25. 오전 6:30:00	3.9
26. 9. 25. 오전 7:00:00	4.1
26. 9. 25. 오전 7:30:00	4.1
26. 9. 25. 오전 8:00:00	4.0
26. 9. 25. 오전 8:30:00	4.0
26. 9. 25. 오전 9:00:00	4.1
26. 9. 25. 오전 9:30:00	4.1
26. 9. 25. 오전 10:00:00	4.1
26. 9. 25. 오전 10:30:00	4.0
26. 9. 25. 오전 11:00:00	4.1
26. 9. 25. 오전 11:30:00	4.1
26. 9. 25. 오후 12:00:00	4.0
26. 9. 25. 오후 12:30:00	4.0
26. 9. 25. 오후 1:00:00	4.0
26. 9. 25. 오후 1:30:00	4.1



IP Storage Temperature Log_Daily report

2026년 9월 25일 오전 12시 0분 0초 KST - 2026년 9월 26일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 26일 오전 6시 0분 18초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 25. 오후 2:00:00	4.1
26. 9. 25. 오후 2:30:00	4.0
26. 9. 25. 오후 3:00:00	4.0
26. 9. 25. 오후 3:30:00	4.0
26. 9. 25. 오후 4:00:00	4.0
26. 9. 25. 오후 4:30:00	4.0
26. 9. 25. 오후 5:00:00	3.9
26. 9. 25. 오후 5:30:00	4.0
26. 9. 25. 오후 6:00:00	4.0
26. 9. 25. 오후 6:30:00	4.1
26. 9. 25. 오후 7:00:00	4.0
26. 9. 25. 오후 7:30:00	4.0
26. 9. 25. 오후 8:00:00	4.0
26. 9. 25. 오후 8:30:00	4.1
26. 9. 25. 오후 9:00:00	4.0
26. 9. 25. 오후 9:30:00	4.0
26. 9. 25. 오후 10:00:00	4.0
26. 9. 25. 오후 10:30:00	3.8
26. 9. 25. 오후 11:00:00	3.9
26. 9. 25. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 26일 오전 12시 0분 0초 KST - 2026년 9월 27일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 27일 오전 6시 0분 29초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 26. 오전 12:00:00	4.0
26. 9. 26. 오전 12:30:00	4.0
26. 9. 26. 오전 1:00:00	4.0
26. 9. 26. 오전 1:30:00	4.0
26. 9. 26. 오전 2:00:00	3.9
26. 9. 26. 오전 2:30:00	3.8
26. 9. 26. 오전 3:00:00	4.0
26. 9. 26. 오전 3:30:00	3.9
26. 9. 26. 오전 4:00:00	4.0
26. 9. 26. 오전 4:30:00	4.1
26. 9. 26. 오전 5:00:00	4.1
26. 9. 26. 오전 5:30:00	4.0
26. 9. 26. 오전 6:00:00	4.0
26. 9. 26. 오전 6:30:00	4.1
26. 9. 26. 오전 7:00:00	4.1
26. 9. 26. 오전 7:30:00	4.0
26. 9. 26. 오전 8:00:00	3.9
26. 9. 26. 오전 8:30:00	4.0
26. 9. 26. 오전 9:00:00	4.0
26. 9. 26. 오전 9:30:00	4.0
26. 9. 26. 오전 10:00:00	3.9
26. 9. 26. 오전 10:30:00	3.8
26. 9. 26. 오전 11:00:00	3.7
26. 9. 26. 오전 11:30:00	3.9
26. 9. 26. 오후 12:00:00	3.8
26. 9. 26. 오후 12:30:00	3.8
26. 9. 26. 오후 1:00:00	3.8
26. 9. 26. 오후 1:30:00	3.7




IP Storage Temperature Log_Daily report

2026년 9월 26일 오전 12시 0분 0초 KST - 2026년 9월 27일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 27일 오전 6시 0분 29초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 26. 오후 2:00:00	3.8
26. 9. 26. 오후 2:30:00	3.8
26. 9. 26. 오후 3:00:00	3.7
26. 9. 26. 오후 3:30:00	3.8
26. 9. 26. 오후 4:00:00	3.9
26. 9. 26. 오후 4:30:00	3.9
26. 9. 26. 오후 5:00:00	3.9
26. 9. 26. 오후 5:30:00	4.0
26. 9. 26. 오후 6:00:00	4.1
26. 9. 26. 오후 6:30:00	4.0
26. 9. 26. 오후 7:00:00	3.9
26. 9. 26. 오후 7:30:00	4.0
26. 9. 26. 오후 8:00:00	3.8
26. 9. 26. 오후 8:30:00	3.9
26. 9. 26. 오후 9:00:00	3.9
26. 9. 26. 오후 9:30:00	3.9
26. 9. 26. 오후 10:00:00	3.9
26. 9. 26. 오후 10:30:00	3.9
26. 9. 26. 오후 11:00:00	3.9
26. 9. 26. 오후 11:30:00	4.0





IP Storage Temperature Log_Daily report

2026년 9월 27일 오전 12시 0분 0초 KST - 2026년 9월 28일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 28일 오전 6시 1분 58초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.0
평균	3.8
26. 9. 27. 오전 12:00:00	4.0
26. 9. 27. 오전 12:30:00	3.9
26. 9. 27. 오전 1:00:00	3.8
26. 9. 27. 오전 1:30:00	3.9
26. 9. 27. 오전 2:00:00	3.8
26. 9. 27. 오전 2:30:00	3.8
26. 9. 27. 오전 3:00:00	3.9
26. 9. 27. 오전 3:30:00	4.0
26. 9. 27. 오전 4:00:00	4.0
26. 9. 27. 오전 4:30:00	3.8
26. 9. 27. 오전 5:00:00	3.9
26. 9. 27. 오전 5:30:00	3.9
26. 9. 27. 오전 6:00:00	3.7
26. 9. 27. 오전 6:30:00	3.9
26. 9. 27. 오전 7:00:00	3.8
26. 9. 27. 오전 7:30:00	3.8
26. 9. 27. 오전 8:00:00	3.8
26. 9. 27. 오전 8:30:00	3.7
26. 9. 27. 오전 9:00:00	3.8
26. 9. 27. 오전 9:30:00	3.8
26. 9. 27. 오전 10:00:00	3.8
26. 9. 27. 오전 10:30:00	3.9
26. 9. 27. 오전 11:00:00	4.0
26. 9. 27. 오전 11:30:00	3.8
26. 9. 27. 오후 12:00:00	3.7
26. 9. 27. 오후 12:30:00	3.9
26. 9. 27. 오후 1:00:00	3.7
26. 9. 27. 오후 1:30:00	3.8



IP Storage Temperature Log_Daily report

2026년 9월 27일 오전 12시 0분 0초 KST - 2026년 9월 28일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 28일 오전 6시 1분 58초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 27. 오후 2:00:00	3.8
26. 9. 27. 오후 2:30:00	3.7
26. 9. 27. 오후 3:00:00	3.8
26. 9. 27. 오후 3:30:00	3.7
26. 9. 27. 오후 4:00:00	3.7
26. 9. 27. 오후 4:30:00	3.8
26. 9. 27. 오후 5:00:00	3.8
26. 9. 27. 오후 5:30:00	3.8
26. 9. 27. 오후 6:00:00	3.9
26. 9. 27. 오후 6:30:00	3.8
26. 9. 27. 오후 7:00:00	3.8
26. 9. 27. 오후 7:30:00	3.8
26. 9. 27. 오후 8:00:00	3.8
26. 9. 27. 오후 8:30:00	3.9
26. 9. 27. 오후 9:00:00	3.8
26. 9. 27. 오후 9:30:00	3.8
26. 9. 27. 오후 10:00:00	4.0
26. 9. 27. 오후 10:30:00	4.0
26. 9. 27. 오후 11:00:00	3.9
26. 9. 27. 오후 11:30:00	4.0



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 28일 오전 12시 0분 0초 KST - 2026년 9월 29일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 29일 오전 6시 0분 19초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.7
최대값	4.1
평균	3.9
26. 9. 28. 오전 12:00:00	4.0
26. 9. 28. 오전 12:30:00	4.0
26. 9. 28. 오전 1:00:00	4.0
26. 9. 28. 오전 1:30:00	4.0
26. 9. 28. 오전 2:00:00	4.0
26. 9. 28. 오전 2:30:00	4.1
26. 9. 28. 오전 3:00:00	4.1
26. 9. 28. 오전 3:30:00	3.9
26. 9. 28. 오전 4:00:00	4.0
26. 9. 28. 오전 4:30:00	3.8
26. 9. 28. 오전 5:00:00	3.9
26. 9. 28. 오전 5:30:00	3.9
26. 9. 28. 오전 6:00:00	3.8
26. 9. 28. 오전 6:30:00	3.9
26. 9. 28. 오전 7:00:00	3.9
26. 9. 28. 오전 7:30:00	3.9
26. 9. 28. 오전 8:00:00	4.0
26. 9. 28. 오전 8:30:00	4.0
26. 9. 28. 오전 9:00:00	3.8
26. 9. 28. 오전 9:30:00	3.9
26. 9. 28. 오전 10:00:00	4.0
26. 9. 28. 오전 10:30:00	3.9
26. 9. 28. 오전 11:00:00	3.9
26. 9. 28. 오전 11:30:00	3.9
26. 9. 28. 오후 12:00:00	4.0
26. 9. 28. 오후 12:30:00	3.9
26. 9. 28. 오후 1:00:00	3.7
26. 9. 28. 오후 1:30:00	3.9



IP Storage Temperature Log_Daily report

2026년 9월 28일 오전 12시 0분 0초 KST - 2026년 9월 29일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 29일 오전 6시 0분 19초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 28. 오후 2:00:00	3.7
26. 9. 28. 오후 2:30:00	3.8
26. 9. 28. 오후 3:00:00	3.8
26. 9. 28. 오후 3:30:00	3.7
26. 9. 28. 오후 4:00:00	3.9
26. 9. 28. 오후 4:30:00	3.9
26. 9. 28. 오후 5:00:00	4.0
26. 9. 28. 오후 5:30:00	4.0
26. 9. 28. 오후 6:00:00	4.0
26. 9. 28. 오후 6:30:00	3.9
26. 9. 28. 오후 7:00:00	3.9
26. 9. 28. 오후 7:30:00	3.8
26. 9. 28. 오후 8:00:00	3.8
26. 9. 28. 오후 8:30:00	3.9
26. 9. 28. 오후 9:00:00	3.8
26. 9. 28. 오후 9:30:00	3.7
26. 9. 28. 오후 10:00:00	3.8
26. 9. 28. 오후 10:30:00	3.9
26. 9. 28. 오후 11:00:00	3.9
26. 9. 28. 오후 11:30:00	3.8



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 29일 오전 12시 0분 0초 KST - 2026년 9월 30일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 30일 오전 6시 0분 21초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	3.8
최대값	4.3
평균	4.0
26. 9. 29. 오전 12:00:00	3.8
26. 9. 29. 오전 12:30:00	3.9
26. 9. 29. 오전 1:00:00	3.8
26. 9. 29. 오전 1:30:00	3.8
26. 9. 29. 오전 2:00:00	3.8
26. 9. 29. 오전 2:30:00	3.8
26. 9. 29. 오전 3:00:00	3.8
26. 9. 29. 오전 3:30:00	3.8
26. 9. 29. 오전 4:00:00	3.9
26. 9. 29. 오전 4:30:00	3.9
26. 9. 29. 오전 5:00:00	4.0
26. 9. 29. 오전 5:30:00	4.0
26. 9. 29. 오전 6:00:00	3.9
26. 9. 29. 오전 6:30:00	4.0
26. 9. 29. 오전 7:00:00	4.0
26. 9. 29. 오전 7:30:00	4.1
26. 9. 29. 오전 8:00:00	4.0
26. 9. 29. 오전 8:30:00	3.9
26. 9. 29. 오전 9:00:00	3.9
26. 9. 29. 오전 9:30:00	3.9
26. 9. 29. 오전 10:00:00	4.0
26. 9. 29. 오전 10:30:00	3.9
26. 9. 29. 오전 11:00:00	3.9
26. 9. 29. 오전 11:30:00	4.0
26. 9. 29. 오후 12:00:00	4.1
26. 9. 29. 오후 12:30:00	4.0
26. 9. 29. 오후 1:00:00	4.0
26. 9. 29. 오후 1:30:00	4.1



IP Storage Temperature Log_Daily report

2026년 9월 29일 오전 12시 0분 0초 KST - 2026년 9월 30일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 9월 30일 오전 6시 0분 21초 KST


측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 29. 오후 2:00:00	4.2
26. 9. 29. 오후 2:30:00	4.1
26. 9. 29. 오후 3:00:00	4.1
26. 9. 29. 오후 3:30:00	4.2
26. 9. 29. 오후 4:00:00	4.2
26. 9. 29. 오후 4:30:00	4.3
26. 9. 29. 오후 5:00:00	4.2
26. 9. 29. 오후 5:30:00	4.2
26. 9. 29. 오후 6:00:00	4.1
26. 9. 29. 오후 6:30:00	4.2
26. 9. 29. 오후 7:00:00	4.3
26. 9. 29. 오후 7:30:00	4.2
26. 9. 29. 오후 8:00:00	4.1
26. 9. 29. 오후 8:30:00	4.2
26. 9. 29. 오후 9:00:00	4.2
26. 9. 29. 오후 9:30:00	4.1
26. 9. 29. 오후 10:00:00	4.0
26. 9. 29. 오후 10:30:00	4.1
26. 9. 29. 오후 11:00:00	4.2
26. 9. 29. 오후 11:30:00	4.2



측정 포인트 요약

동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1

임상시험용 의약품(냉장고B 2~8°C)_0054704396_CH1





IP Storage Temperature Log_Daily report

2026년 9월 30일 오전 12시 0분 0초 KST - 2026년 10월 1일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 10월 1일 오전 6시 7분 40초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃) 0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
최소값	4.0
최대값	4.3
평균	4.1
26. 9. 30. 오전 12:00:00	4.1
26. 9. 30. 오전 12:30:00	4.1
26. 9. 30. 오전 1:00:00	4.2
26. 9. 30. 오전 1:30:00	4.2
26. 9. 30. 오전 2:00:00	4.1
26. 9. 30. 오전 2:30:00	4.2
26. 9. 30. 오전 3:00:00	4.2
26. 9. 30. 오전 3:30:00	4.2
26. 9. 30. 오전 4:00:00	4.1
26. 9. 30. 오전 4:30:00	4.1
26. 9. 30. 오전 5:00:00	4.2
26. 9. 30. 오전 5:30:00	4.1
26. 9. 30. 오전 6:00:00	4.0
26. 9. 30. 오전 6:30:00	4.0
26. 9. 30. 오전 7:00:00	4.1
26. 9. 30. 오전 7:30:00	4.0
26. 9. 30. 오전 8:00:00	4.0
26. 9. 30. 오전 8:30:00	4.0
26. 9. 30. 오전 9:00:00	4.1
26. 9. 30. 오전 9:30:00	4.0
26. 9. 30. 오전 10:00:00	4.0
26. 9. 30. 오전 10:30:00	4.0
26. 9. 30. 오전 11:00:00	4.0
26. 9. 30. 오전 11:30:00	4.1
26. 9. 30. 오후 12:00:00	4.1
26. 9. 30. 오후 12:30:00	4.0
26. 9. 30. 오후 1:00:00	4.0
26. 9. 30. 오후 1:30:00	4.1



IP Storage Temperature Log_Daily report

2026년 9월 30일 오전 12시 0분 0초 KST - 2026년 10월 1일 오전 12시 0분 0초 KST

kyon410@daum.net

생성됨: 2026년 10월 1일 오전 6시 7분 40초 KST

측정 포인트	동국대학교 일산병원 임상시험센터 > 임상시험용의약품 보관실 > 임상시험용 의약품(냉장고B 2~8℃)_0054704396_CH1
타임 스탬프 [Asia/Seoul]	온도 [°C]
26. 9. 30. 오후 2:00:00	4.0
26. 9. 30. 오후 2:30:00	4.1
26. 9. 30. 오후 3:00:00	4.1
26. 9. 30. 오후 3:30:00	4.2
26. 9. 30. 오후 4:00:00	4.3
26. 9. 30. 오후 4:30:00	4.3
26. 9. 30. 오후 5:00:00	4.2
26. 9. 30. 오후 5:30:00	4.1
26. 9. 30. 오후 6:00:00	4.1
26. 9. 30. 오후 6:30:00	4.2
26. 9. 30. 오후 7:00:00	4.2
26. 9. 30. 오후 7:30:00	4.1
26. 9. 30. 오후 8:00:00	4.0
26. 9. 30. 오후 8:30:00	4.1
26. 9. 30. 오후 9:00:00	4.2
26. 9. 30. 오후 9:30:00	4.1
26. 9. 30. 오후 10:00:00	4.1
26. 9. 30. 오후 10:30:00	4.2
26. 9. 30. 오후 11:00:00	4.1
26. 9. 30. 오후 11:30:00	4.1